# **PRICE LIST / Listino prezzi**



QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version





# PRICE LIST Listino Prezzi

**CURRENCY / Valuta: EUR** 

### **INDEX**

Page 2 § T LABORATORY TESTS / Prove di laboratorio
Page 224 § P TESTS PACKAGES / Pacchetti di prove

NOTE: THIS IS A TECHNICAL DOCUMENT, CONTAINING THE PRICE OF ALL THE OFFERED SERVICES.

IT MAY BE DIFFICULT TO SELECT THE APPROPRIATE ONE: TO GET A QUOTATION PLEASE SEND A REQUEST TO: info@albarubens.it

Nota: questo è un documento tecnico, contenente il prezzo di tutti i servizi offerti.

Può essere difficile selezionare quello appropriato: per ottenere un preventivo inviare una richiesta a: info@albarubens.it

SSL CERTIFICATE Certificato SSL



page 1/234



# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



# § T LABORATORY TESTS / Prove di laboratorio

UNIT PRICES AND DETAILS FOR EACH SINGLE PERFORMED TEST, WITH DIFFERENT LEVELS IF APPLICABLE Prezzi unitari e dettagli per ogni singola prova eseguita, con differenti livelli ove applicabile

# 2021.05.05 RFQ TR Testing FRONT PANELS

(id 1814) TR testing to be performed for OEM Volvo Truck

#### 1 - Heat resistance, short term

(id 2441) A method for assessing shrinkage, deformation and changes in appearance of components made of plastics material, when exposed to head (in a ventilated oven).

Test according to TR84199760 - 3.1.1.3.3

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

#### 2 - Resistance to ageing in heat

(id 2442) A method for assessing shrinkage, deformation, and changes in appearance of components made of plastics material, when exposed to heat (in a ventilated oven).

Test according to TR 84199760 - 3.1.1.3.4

Accredited [IEC17025]: NO Execution time [day]: 42
Unit price #1 [EUR]: 4,500.00 Limit #1: Applicable to any sample

#### **Unref. - Cold conditioning**

(id 2443) A method for assessing shrinkage, deformation, and changes in appearance of components made of plastics material, when exposed to cold.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

## 49 CFR 393.67:2023

(id 1976) Transportation - Other Regulations Relating to Transportation

#### 393.67 - e1 - Drop Test

(id 2706) Test for side-mounted liquid fuel tank.

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1 [EUR]: 400.00 Limit #1: Applicable to any sample

#### 393.67 - e2 - Fill Pipe Test

(id 2707) Test for side-mounted liquid fuel tank.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 400.00 Limit #1: Applicable to any sample

## ABYC H-33:2005

(id 2041) These standards are guides for the design, choice of materials, construction, installation, repair, and maintenance of permanently installed diesel fuel systems.

#### 33.17.5 - Overpressure

(id 2783) ---

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 2,200.00 Limit #1: Applicable to any sample

#### 33.21.5 - Pressure Impulse Test

(id 2784) ---

Accredited [IEC17025]: NO Execution time [day]: 4
Unit price #1 [EUR]: 4,000.00 Limit #1: Applicable to any sample



# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX

#### 33.21.6 - Slosh Test

(id 2785) ---

Accredited [IEC17025]: NO Execution time [day]: 23
Unit price #1 [EUR]: 8,200.00 Limit #1: Applicable to any sample

#### 33.20.7.4 - Leak Test

(id 2787) ---

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

# API SPEC 6A 21ST ED (E1):2019

(id 1719) Specification for Wellhead and Tree Equipment, Twenty-First Edition; Errata, April 2019

#### Ann. F/F.1.11 - Pressure and Temperature Cycles

(id 2178) This test involves pressure/temperature cycles in order to check for any damages/leaks at the end of the test, caused by the deterioration of the material.

# **API STD 607 (2016)**

(id 1038) Fire Test for Quarter-turn Valves and Valves Equipped with Nonmetallic Seats, Seventh Edition

#### 5 - Fire test

(id 1547) The purpose of this test is confirming the pressure-containing capability of a valve under pressure during and after the fire test.

Accredited [IEC17025]	: NO	Execution	time [day]	1: 0
Unit price #1 [EUR]:	2,000.00		Limit #1:	Up to 3"
Unit price #2 [EUR]:	2,800.00		Limit #2:	from 4" to 10"
Unit price #3 [EUR]:	4,500.00		Limit #3:	from 12" to 20"

#### 6.2 - Through-seat Leakage During Burn Period

(id 1548) ---

Accredited [IEC17025]: NO	Execution time [day]: 0
Unit price #1 [EUR]: 150.00	Limit #1: Up to 3"
Unit price #2 [EUR]: 250.00	Limit #2: from 4" to 10"
Unit price #3 [EUR]: 500.00	Limit #3: from 12" to 20"

#### 6.3 - External Leakage During Burn and Cool-down Periods

(id 1549) ---

Accredited [IEC17025]: NO	Execution time [day]: 0
Unit price #1 [EUR]: 200.00	Limit #1: Up to 3"
Unit price #2 [EUR]: 400.00	Limit #2: from 4" to 10"
Unit price #3 [EUR]: 800.00	Limit #3: from 12" to 20"

#### 6.4 - Low Pressure Test Through-seat Leakage after Cool-down

(id 1550) ---

Accredited [IEC17025]: NO	Execution time [day]: 0
Unit price #1 [EUR]: 150.00	Limit #1: Up to 3"
Unit price #2 [EUR]: 250.00	Limit #2: from 4" to 10"
Unit price #3 [EUR]: 500.00	Limit #3: from 12" to 20"

#### 6.5 - Operability

(id 1551) ---

Accredited [IEC17025]: NO	Execution time [day]: 0
Unit price #1 [EUR]: 400.00	Limit #1: Up to 3"
Unit price #2 [EUR]: 800.00	Limit #2: from 4" to 10"
Unit price #3 [EUR]: 1,200.00	Limit #3: from 12" to 20"



# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



#### 6.6 - External Leakage Following Operational Test

(id 1552) ---

Accredited [IEC17025]	]: NO	Execution	time [	day:	1: 0			
Unit price #1 [EUR]:	200.00		Limit	#1:	Up to	3 "		
Unit price #2 [EUR]:	400.00		Limit	#2:	from	4" 1	to 1	L0"
Unit price #3 [EUR]:	800.00		Limit	#3:	from	12"	to	20"

## **API STD 622:2018**

(id 1185) Type Testing of Process Valve Packing for Fugitive Emissions, Third Edition STANDARD by American Petroleum Institute, 10/01/2018

# **4.1.2/4.4.2 - Fugitive Emissions Test: Mechanical and thermal cycling** (id 1617) ---

Accredited [IEC17025]: YES Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

#### 4.1.4 - Fugitive emissions test: Temperature Monitoring

(id 1618) ---

```
Accredited [IEC17025]: YES Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant
```

#### 4.4.3 - Fugitive emissions test: Leak Measurement

(id 1619) ---

### API STD 641:2016

(id 1186) Type Testing of Quarter-turn Valves for Fugitive Emissions, First Edition STANDARD by American Petroleum Institute, 10/01/2016

#### - Type testing of Quarter-turn valves for fugitive emissions

(id 1621) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

## **ASME B16:33**

(id 1979) Manually Operated Metallic Gas Valves for Use in Gas Piping Systems up to 175 psi (Sizes NPS 1/2 through NPS 2)

#### 4.3.2 - Temperature Resistance - cold

(id 2717) A valve should be operable at tenperature of -29°C without affecting the capability of the valve to control the flow of gas.

#### 4.3.3 - Temperature Resistance - heat

(id 2718) A valve should be operable at tenperature of 66°C without affecting the capability of the valve to control the flow of gas.

## **ASTM B117 ? 19**

(id 2123) Standard Practice for Operating Salt Spray (Fog) Apparatus

#### Unref. - Salt spray (fog) exposure

(id 2935) This practice covers the apparatus, procedure, and conditions required to create and maintain the salt spray (fog) test environment.

Accredited [IEC17025]: NO Execution time [day]: 4
Unit price #1 [EUR]: 1,400.00 Limit #1: Applicable to any sample





QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



## **ASTM D5868 - 01**

(id 1782) Standard Test Method for Lap Shear Adhesion for Fiber Reinforced Plastic (FRP) Bonding

#### 8.4 - Lap Shear Adhesion for Fiber Reinforced Plastic (FRP) Bonding

(id 2356) This test method describes a lap shear test for use in measuring the bonding characteristics of adhesives for joining fiber reinforced plastics to themselves and to metals. The method is applicable to random and fiber oriented FRP. The specimen loading rate is 13 mm/min (0.0002166 m/s).

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 600.00 Limit #1: Not relevant

## **ASTM D5894?10**

(id 1756) Cyclic Salt Fog/UV Exposure of Painted Metal, (Alternating Exposures in a Fog/Dry Cabinet and a UV/Condensation Cabinet)

#### 8.1 - Fluorescent UV-Condensation Exposure

(id 2292) This practice covers basic principles and operating practice for cyclic corrosion/UV exposure of paints on metal, using alternating periods of exposure in two different cabinets: a cycling salt fog/dry cabinet, and a fluorescent UV/ condensation cabinet.

Accredited [IEC17025]: NO Execution time [day]: 28
Unit price #1 [EUR]: 4,800.00 Limit #1: Applicable to any sample

#### 8.2 - Cyclic Salt Fog/Dry Exposure

(id 2293) This practice covers basic principles and operating practice for cyclic corrosion/UV exposure of paints on metal, using alternating periods of exposure in two different cabinets: a cycling salt fog/dry cabinet, and a fluorescent UV/ condensation cabinet.

Accredited [IEC17025]: NO Execution time [day]: 7
Unit price #1 [EUR]: 2,200.00 Limit #1: Applicable to any sample

# **ASTM G85?11**

(id 836) Standard Practice for Modified Salt Spray (Fog) Testing

#### A.2 - Cyclic acidified salt fog testing

(id 1372) ---

Accredited [IEC17025]: NO	Execution time [day]: 0
Unit price #1 [EUR]: 1,500.00	Limit #1: 1 weeks
Unit price #2 [EUR]: 3,000.00	Limit #2: 4 weeks
Unit price #3 [EUR]: 5,000.00	Limit #3: 8 weeks

#### 8.1 - Intermediate evaluation: visual inspection and leakage test

(id 1373) The salt spray testing requires a careful and immediate examination for the extent of corrosion and for any other failure as required by the specifications covering the material or product being tested or by agreement between the purchaser and the seller. In this case, a leakage test has been requested and a visual inspection is performed in order to verify the presence of bubbles when the sample is immersed in water and subject to the pressure required.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 100.00 Limit #1: Not relevant

#### A.3 - Acidified synthetic sea water (fog) testing

(id 2350) This practice covers and sets forth conditions for five modifications in salt spray (fog) testing for specification purposes. A.3 refers to seawater acidified test, cyclic (SWAAT).

Accredited [IEC17025]: NO Execution time [day]: 42
Unit price #1 [EUR]: 1,400.00 Limit #1: 100 cycles
Unit price #2 [EUR]: 5,500.00 Limit #2: 500 cycles

#### A.5 - Dilute electrolyte cyclic fog-dry test

(id 2780) This test consists of cycles of 1-h dry-off and 1-h fog.

The electrolyte is a solution of sodium chloride and ammonium sulfate, and is much more dilute than traditional salt fog. The fog is performed at room temperature, while the dry-off is at elevated temperature. In addition, the spray atomizing air is not saturated with water.

This test is particularly applicable to paints on steel.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant



# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



## Aston Martin AM370

(id 1935) Powertrain Cooling Systems Validation Requirements

#### 1.1 - Leak test

(id 2604) ---

Accredited [IEC17025]: NO Execution time [day]: 2

Unit price #1 [EUR]: 300.00 Limit #1: Applicable to any sample

#### 1.2 - Burst Pressure Test

(id 2605) ---

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

## Automotive internal test

(id 404) Automotive internal test methods

#### Al 02 - Drop test on fuel and urea tanks

(id 1204) This test is performed in order to verify if tanks are capable to withstand a drop without leakages.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### Al 01 - Vacuum test on urea tanks

(id 1205) Test to be performed with the method described in IVECO Std. 18-3310.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 300.00 Limit #1: Not relevant

#### Al 04 - Fuel Resistance Test

(id 1206) Ageing of the materials used in a gasoline fuel / vapor assembly due to exposure to liquid fuel or fuel vapors, internal and/or external.

### Al 05 - Breakaway valve nozzle functional test

(id 1266) Functional verification

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 800.00 Limit #1: Not relevant

#### Al 06 - Pipes disconnection test

(id 1305) The aim of this test is to verify the pipe disconnection when applying a pull force.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 400.00 Limit #1: Applicable to any sample

#### Al 17 - RENAULT Standard D17-1058--L - Neutral salt spray test

(id 1382) This test is to be applied to compare the resistance to deterioration from salt mist of specimens of similar construction. It is useful for evaluating the quality and uniformity of protective coatings.

Accredited [IEC17025]: NO Execution time [day]: 25
Unit price #1 [EUR]: 4,800.00 Limit #1: 600 hours
Unit price #2 [EUR]: 2,200.00 Limit #2: 288 hours

#### Al 16 - Vacuum test

(id 1388) Additional test to Spec 9.02137/05 Automotive blowby piping endurance tests

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 300.00 Limit #1: Not relevant

#### Al 09 - Wet leakage current test

(id 1615) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 150.00 Limit #1: Not relevant



# PRICE LIST / Listino prezzi

# § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX

# -

#### Al 10 - Temperature cycles and immersion in fluorescent saline solution

(id 1633) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 4,000.00 Limit #1: Applicable to any sample

#### Al 11 - Special mixture preparation

(id 1677) --

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant

## Al 12 - OLD---CONVERTED IN Hutchinson FMS MSO CDC cba 05-2020 Date : 21/04/2020 5.2.2.8 Aging ---BMW Group - Aging time lap

(id 1855) OLD---CONVERTED IN Hutchinson FMS MSO CDC cba 05-2020 Date: 21/04/2020 5.2.2.8 Aging ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 4,000.00 Limit #1: 1000h
Unit price #2 [EUR]: 5,500.00 Limit #2: 2000h

#### Al 13 - Temperature tests on containers for batteries transportation

(id 1889) ---

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 2,800.00 Limit #1: Not relevant

#### Al 14 - Damper performance test

(id 2084) Evaluation of the efficency of a damper in damping the pressure of pulsations in fuel hoses.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 600.00 Limit #1: Not relevant

#### Al 19 - Resistance to fuel combined to temperature cycles (Renault 34-04-831)

(id 2205) The samples are subjected to the temperature cycles described in the specification Renault 34-04-831/--J annex 1, in number defined by the customer, when filled with requested fuel under uncontrolled static conditions. Every week the fuel inside the pieces is renewed.

#### Al 07 - Sumiriko internal test: Fuel conditioning test

(id 2231) Moved to Sumiriko proprietary test methods on 31-Aug-2021

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### Al 08 - Sumiriko internal test: resistance to thermal cycles

(id 2232) Moved to Sumiriko proprietary test methods on 31-Aug-2021

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### Al 18 - Sumiriko internal specification named 80-80

(id 2233) Moved to Sumiriko proprietary test methods on 31-Aug-2021

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### Al 20 - Sumiriko internal test: Fuel exposure preconditioning (SAEJ2260)

(id 2234) Moved to Sumiriko proprietary test methods on 31-Aug-2021

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

## Al 15 - Hutchinson internal method: BMW Ncar long aging study

(id 2235) Moved to Hutchinson proprietary test methods on 31-Aug-2021

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### Al 21 - Resistance to molten aluminum

(id 2355) --

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample



# PRICE LIST / Listino prezzi

# § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX

#### Al 22 - High temperature shock with circulating media

(id 2373) High temperature thermal cycles with rapid thermal excursion on piping circuits with circulating medium at controlled temperature.

Accredited [IEC17025]: NO Execution time [day]: 4

Unit price #1 [EUR]: 1,800.00 Limit #1: Applicable to any sample

#### Al 23 - Fuel soaking preconditioning

(id 2450) --

Accredited [IEC17025]: NO Execution time [day]: 2

Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

#### Al 24 - Rejection to flame on carbon canister

(id 2451) The sample is subjected to a flame obtained with an LPG burner to verify the possibility of explosive phenomena.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 2,000.00 Limit #1: Not relevant

#### Al 25 - Pressure pulsation test

(id 2479) This test is intended to verify the resistance of automotive samples under different pressures and temperatures.

Accredited [IEC17025]: NO Execution time [day]: 64
Unit price #1 [EUR]: 7,000.00 Limit #1: Air as medium

#### Al 26 - Pressure pulsation test up to 35 bar

(id 2501) This test is intended to verify the resistance of automotive samples under a fixed temperature (120°C) and different pressures (up to 35 bar).

Accredited [IEC17025]: NO Execution time [day]: 3

Unit price #1 [EUR]: 3,800.00 Limit #1: Azoto as medium

#### Al 27 - Speed control at no load [Ferrari specific]

(id 2531) This test is intended to verify the rotation ability of a motor by alternately power the two motors for a short time.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 300.00 Limit #1: Applicable to any sample

#### Al 28 - Nissan Cooling Test

(id 2588) ---

Accredited [IEC17025]: NO Execution time [day]: 21

Unit price #1 [EUR]: 5,500.00 Limit #1: Applicable to any sample

#### Al 29 - SoLiSi - Sun simulation

(id 2596) Application of superficial heat to the bumper by infrared lamps (IR), simulating the solar radiation.

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 1,000.00 Limit #1: Applicable to any sample

#### Al 30 - Breaking point test at hot temperature

(id 2637) Evaluation of temperature and pressure at breaking point of pipes filled with the extinguishing fluid FK 5 1 12.

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 800.00 Limit #1: Up to 4 samples at the same time

#### Al 31 - DC-DC Burn-In Test [Ferrari specific]

(id 2644) Functional verification of the integreted DC/DC (3kW flow from HV to MV).

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 300.00 Limit #1: Applicable to any sample

#### Al 32 - Pressure Drop Test

(id 2650) Pressure drop test on pipes with a specified substance.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 300.00 Limit #1: single flow rate
Unit price #2 [EUR]: 600.00 Limit #2: up to 5 flow rates

#### Al 33 - Leak test [Ferrari specific]

(id 2665) Hydraulic seal test, which shall be done on two distinct volumes and with different parameters.

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 300.00 Limit #1: Laboratory temperature, Pmax=2bar

Unit price #2 [EUR]: 800.00 Limit #2: Non standard temperature (controlled in climatic chamber: -40 +100°C range)

Unit price #3 [EUR]: 1,200.00 Limit #3: Laboratory temperature, high pressure (>2 bar)



# PRICE LIST / Listino prezzi § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



#### Al 34 - Thermal Dunk Test - Modified IPX8 as per Tesla Specification

(id 2684) This dunk test is similar to IPX8 with the exception that the sample shall be submerged in 1m depth salted water, using two tanks at two different temperatures (0°C and 85°C). The test is comprised of 20 cycles with each dunk lasting 10 minutes. The internal parts of the sample shall be sprayed with a water sensitive substance (e.g. WDP-217) in order to detect any penetration.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 3,200.00 Limit #1: Applicable to any sample

#### Al 35 - Parameter Tests E-axle Ferrari

(id 2867) This test comprises the following parameter tests:

- MI 33 Visual inspection
- §18.3 of IEC 60204-1:2016+AMD1:2021 Insulation resistance test
- Al 33 Leak Test
- Al 31 Speed control at no load
- Al 27 DC-DC Burn-In Test

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

#### Al 36 - MAN Truck SCR Lines

(id 2902) Thermal cycles on SCR lines with urea as fluid.

Accredited [IEC17025]: NO Execution time [day]: 250
Unit price #1 [EUR]: 12,000.00 Limit #1: Applicable to any sample

#### Al 37 - Water Spray Test 12GX India

(id 2915) This test is intended to verify the potential ingress of water inside a DUT during a water spray test.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,500.00 Limit #1: Applicable to any sample up to 50x50x50 cm

#### Al 38 - Thermal Propagation Test

(id 2989) This test is intended to verify the thermal propagation of cells and batteries

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 1,500.00 Limit #1: Test performed by customer

#### Al 39 - Cooling and Housing Tightness Test

(id 3013) ---

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

## **AVL DVP Plan**

(id 1969) Design and validation plan for e-AXLE testing

#### **Unref. - Lubrication test**

(id 2678) ---

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### **Unref. - Leakage Test Verification**

(id 2679) The unit must be checked for different combinations of speed, oil temperature, longitudinal and lateral slope.

Accredited [IEC17025]: NO Execution time [day]: 6
Unit price #1 [EUR]: 5,200.00 Limit #1: Applicable to any sample

#### Unref. - Leakage Check Verification

(id 2680) The breather, static seals and boxes, and running or dynamic seals must be checked during every other step of the procedure.

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### Unref. - Oil temperature rise and stabilization procedure

(id 2681) ---

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 2,800.00 Limit #1: Applicable to any sample



# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

TNDEX

## **Unref. - Drag Torque Test**

(id 2682) ---

Accredited [IEC17025]: NO

Unit price #1 [EUR]: 7,500.00

**Unref. - Vibration Analysis** (id 2683) ---

Accredited [IEC17025]: NO

Unit price #1 [EUR]: 2,000.00

Execution time [day]: 2

Limit #1: Applicable to any sample

Execution time [day]: 1

Limit #1: Applicable to any sample

## AW101-SM-fQTP-2867 Iss.1 – 07-01-2019

(id 1244) Secondo Mona S.p.A. Formal qualification test procedure: Deploy/stow actuator

#### 4.2 - Waterproofness

(id 1714) ---

Accredited [IEC17025]: NO

Execution time [day]: 0 Unit price #1: ask to office

Limit #1: Not relevant

## BMW 11942577 000 03:2025

(id 2231) LH Fluidfuehrung EES GEN5 WE SE61

#### 5.3.1.4.4 - Aging test

(id 3137) Long time aging is for testing tightness and function after aging.

Execution time [day]: 125 Accredited [IEC17025]: NO

Unit price #1 [EUR]: 7,000.00 Limit #1: Only aging, 3000h duration.

# BMW Group specification 17.10.2018

(id 1638) Requirements Specification in progress provided by the customer - SAP Titel des Lastenheftes - SAP status/ in progress 17.10.2018

#### 5.2.2.8 - Aging time lapse test - Transmission oil cooling pipes

(id 2139) ---

Accredited [IEC17025]: NO Execution time [day]: 130 Unit price #1 [EUR]: 4,000.00 Limit #1: 1000h Unit price #2 [EUR]: 5,500.00 Limit #2: 2000h Unit price #3 [EUR]: 7,000.00 Limit #3: 3000h

#### 5.2.2.3 - Pressure threshold test - Transmission oil lines - Test step 1

(id 2140) ---

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 3,000.00 Limit #1: Applicable to any sample

#### 5.2.2.3 - Pressure threshold test - Transmission oil lines - Test step 2

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 4,500.00 Limit #1: Applicable to any sample

#### 5.1.2.1 - Leak test after the function check - Transmission oil lines

(id 2142) ---

Accredited [IEC17025]: NO Execution time [day]: 5

Unit price #1 [EUR]: 200.00 Limit #1: Applicable to any sample

#### 5.2.2.6 - Pull-out strength test for the connections

(id 2143) The objective of this test is checking the connection quality of the plastic pipe with its coupling or connections of different material pairings.

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 300.00 Limit #1: Applicable to any sample



# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



#### 5.2.2.2 - Burst pressure test

(id 2144) The objective of this test is to ensure sufficient compressive strength of components and assemblies, particularly at high temperature (safeguarding transmission failure/emergency engine oil program)

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 500.00 Limit #1: Applicable to any sample

# BMW Group 10957962-000-02 01.08.2018

(id 1633) Requirements specification Kuelmittelleitungen FAAR WE

#### 5.2.2.13 - 3000 hour endurance test

(id 2061) Objective: check leak-tightness and functionality of component under actual vehicle-specific loads; the vehicle service life is simulated by increased temperature requirements. No leaking may occur during the test.

Accredited [IEC17025]: NO Execution time [day]: 125 Unit price #1 [EUR]: 7,000.00 Limit #1: 3000h

# BMW Group 11224613 000 02

(id 1880) Coolant pipe CLAR WE G7x G6x

#### 5.1.2.1 - Leak test

(id 2510) This test is intended to check the leak-tightness of components and connection points.

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 300.00 Limit #1: Applicable to any sample

#### 5.2.2.3 - Burst pressure test

(id 2511) This test is intended to evaluate the compressive load withstand capability of components and assemblies.

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 500.00 Limit #1: Applicable to any sample

#### 5.2.2.10 - Media resistance test

(id 2512) Resistance to operating media as laid down in section 3.4 shall be assured throughout the entire product service life.

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

# BMW Group 11229184 – 000 – 03 08\_04\_2020

(id 1646) Medienleitung Oelkuehlung eFTC Getr.Gen4

#### 5.2.2.8 - Aging - Transmission oil cooling pipes

(id 2095) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 4,000.00 Limit #1: 1000h (only visual inspection afer ageing)
Unit price #2 [EUR]: 5,500.00 Limit #2: 2000h (only visual inspection afer ageing)
Unit price #3 [EUR]: 7,000.00 Limit #3: 3000h (only visual inspection afer ageing)

# BMW Group 11701376 - 000 - 01:2021

(id 1767) Requirements specification for Kuehlmittelleitungen (coolant pipes) NCAR

#### 5.2.2.13 - 3000 hour endurance test

(id 2339) Check leak-tightness and functionality of component under actual vehicle-specific loads; the vehicle service life is simulated by increased temperature requirements.

Accredited [IEC17025]: NO Execution time [day]: 125
Unit price #1 [EUR]: 7,000.00 Limit #1: Applicable to any sample



# PRICE LIST / Listino prezzi your value certified § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



# BMW Group 11759048-000-01:2022

(id 1904) BMW Group specification

#### 5.2.2.13 - 3000 hour endurance test

(id 2557) 3000 hour endurance test for automotive pipes.

Accredited [IEC17025]: NO Execution time [day]: 125

Unit price #1 [EUR]: 7,000.00 Limit #1: Applicable to any sample

## BMW GS 95024-3-1:2019

(id 1985) Electrical and Electronic Components in Motor Vehicles, Environmental Requirements Test Standard

#### M-02 - Stone Impact Test

(id 2722) ---

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,400.00 Limit #1: Up to 3 locations (7x7cm)

#### K-12 - Thermal shock test with splash water

(id 2723) ---

#### K-18 - Corrosion test with flowing mixed gas

(id 2724) ---

Accredited	[IEC17025]: NO	Execution time [day]: 21
Unit price	#1 [EUR]: 3,200.00	Limit #1: 14 days
Unit price	#2 [EUR]: 3,800.00	Limit #2: 21 days
Unit price	#3 [EUR]: 6,000.00	Limit #3: huge size sample

#### K-06 - Salt spray test

(id 2725) ---

Accredited [IEC17025]: NO Execution time [day]: 6
Unit price #1 [EUR]: 3,600.00 Limit #1: Not connected and not powered sample (always II.b)

#### M-03 - Dust intrusion - IP 6XK

(id 2726) ---

```
Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 600.00 Limit #1: Standard IP chamber (small sample not powered)
Unit price #2 [EUR]: 1,200.00 Limit #2: Medium IP chamber (small sample not powered)
Unit price #3 [EUR]: 2,200.00 Limit #3: Powered on sample
```

### K-11 - High pressure steam-jet cleaning - IP X9K

(id 2727) ---

Accredited [IEC17025]: NO	Execution time [day]: 1
Unit price #1 [EUR]: 1,200.00	Limit #1: small sample (automatic spraying)
Unit price #2 [EUR]: 3,000.00	Limit #2: big sample (manual spraying)

#### K-10 - Water protection - IP X6K

(id 2868) ---

```
Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 400.00 Limit #1: Applicable to any sample
Unit price #2 [EUR]: 1,000.00 Limit #2: up to 5 pipes contemporary
```



# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



## BMW LH 11230536 - 000 - 02

(id 1713) Capitolato BMW LH 11230536 - 000 - 02

#### 5.3.1.3.3 - Aging test

(id 2085) Long time aging is for testing tightness and function after aging.

## BMW LH11717564 - 000 - 02

(id 2192) Capitolato BMW LH11717564 - 000 - 02

#### 5.2.2.14 - Longterm test 1500h for quick connectors

(id 3080) Assessment of the long-term use characteristics and fatigue stability under vehicle-specific loads. Increased temperature requirements simulate the vehicle lifetime.

Accredited [IEC17025]: NO Execution time [day]: 63
Unit price #1 [EUR]: 5,000.00 Limit #1: Applicable to any sample

# BMW QV 16024 September 2019

(id 1683) Fuel lines with joining technology

#### 4.2.8 - Fuel resistance including change in length

(id 2133) ---

Accredited [IEC17025]: NO Execution time [day]: 8
Unit price #1 [EUR]: 2,200.00 Limit #1: Not relevant

# BMW\_PR\_303.5\_EN\_2010-01-21

(id 1752) Alternating climate test for trim parts

#### - Alternating climate test for trim parts

(id 2244) the test is intended to verify the resistance of vehicle trim parts to alternating climatic conditions.

Explanation of cycles code letters: A, B, C, D:

- -Cycle part A applies to parts in footwell, and below door strip, rocker panel cover;
- -Cycle part B applies to parts above door strip, rocker panel cover and below top shoulder area;
- -Cycle part C applies to parts not exposed to sun, above top shoulder;
- -Cycle part D applies to parts in exterior fittings area;

Accredited [IEC17025]: NO Execution time [day]: 10 Unit price #1 [EUR]: 2,400.00 Limit #1: Applicable to any sample

# **Bombardier Transportation VS BAE 02014**

(id 2204) Lifespan testing of EPDM water hoses for MPC and HVIM power converter

#### **Unref. - Durability Test**

(id 3097) The goal of this durability test is to test the new hoses using internal pressure load cycles and various temperatures to guarantee a minimum service life of 5 years. However, a service life of over 10 years is still the goal. For this reason, all four hoses will continue to be used until at least two hoses fail, provided all four hoses have survived the 700,000 load cycles without damage. During the same test, two different hose clamps are to be tested. The hose clamps must be tightened to a specific torque, which must be recorded in the test report, to ensure reproducibility of the test.

Accredited [IEC17025]: NO Execution time [day]: 50
Unit price #1 [EUR]: 5,000.00 Limit #1: Applicable to any sample



# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



# cap. 9.02138/10 ed. February 2019

(id 1452) Untitled

#### 2.6.4.2 - Cicli termici - Thermal cycles

(id 1920) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 2,500.00 Limit #1: Not relevant

## CEI EN 60068-2-14:2011

(id 174) Environmental testing - Part 2-14: Tests - Test N: Change of temperature

#### 7 - Test Na: Rapid change of temperature with prescribed time of transfer

(id 1015) This test determines the ability of components, equipment or other articles to withstand rapid changes of ambient temperature. The exposure times adequate to accomplish this will depend upon the nature of the specimen. The specimen shall be either in the unpacked, switched-off, ready for use state, or as otherwise specified in the relevant specification. The specimen is exposed to rapid changes of temperature in air, or in a suitable inert gas, by alternate exposure to low temperature and to high temperature.

```
Accredited [IEC17025]: YES
Unit price #1 [EUR]: 2,000.00
Unit price #2 [EUR]: 8,000.00
Unit price #3 [EUR]: 13,000.00
Limit #3: Multiple change of temperature (up to 7 days)
Limit #2: Multiple change of temperature (up to 30days)
Limit #3: Multiple change of temperature (up to 50days)
```

#### 8 - Test Nb: Change of temperature with specified rate of change

(id 1066) ---

## 9 - Test Nc: Rapid change of temperature, two-fluid-bath method

(id 1067) ---

#### CEI EN 60068-2-30:2006

(id 178) Environmental testing - Part 2-30: Tests - Test Db: Damp heat, cyclic (12 h + 12 h cycle)

#### **Db - Damp heat (12h+12h)**

(id 1033) ---

## CEI EN 60079-31:2015

(id 97) Explosive atmospheres - Part 31: Equipment dust ignition protection by enclosure "t"

#### 6.1.1.1 - Type tests for dust exclusion by enclosures

(id 117) Test sequence composed of thermal endurance to heat, thermal endurance to cold, impact test and and drop test if applicable, as specified in IEC 60079-0, followed by:

- IEC 60079-31 § 6.1.1.2 Impact test for supplementary enclosures, if present
- IEC 60079-31 § 6.1.1.3 Pressure test
- IEC 60079-31 § 6.1.1.4 IP test.

Accredited [IEC17025]: YES Execution time [day]: 30
Unit price #1: ask to office Limit #1: not relevant



# PRICE LIST / Listino prezzi § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX

# .

#### 6.1.2 - Thermal tests

(id 118) These tests are carried out as described in IEC 60079-0 with specific requirements for 'ta'.

The maximum surface temperature shall be measured on the external surfaces of the enclosure and the surfaces of the internal components.

Accredited [IEC17025]: YES Execution time [day]: 1
Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

#### 6.1.1.3 - Pressure test

(id 119) ---

Accredited [IEC17025]: YES
Unit price #1 [EUR]: 300.00
Limit #1: Applicable to any sample
Unit price #2 [EUR]: 300.00
Limit #2: Applicable to any sample
Unit price #3 [EUR]: 300.00
Limit #3: Applicable to any sample

#### 5 - Construction

(id 524) Evaluation of construction parameters and materials - Check of documentation

Accredited [IEC17025]: NO Execution time [day]: 7
Unit price #1 [EUR]: 500.00 Limit #1: Applicable to any sample

#### 6.1.1.4 - IP test

(id 576) Degree of protection (IP) - IEC 60529

Accredited [IEC17025]: YES Execution time [day]: 30 Unit price #1 [EUR]: 300.00 Limit #1: IP5X or IP6X

#### 6.1.1.2 - Impact test for supplementary enclosures

(id 888) The impact test shall be performed in accordance with §26.4.2. of IEC60079-0.

## CEI EN 60950-1:2007

(id 2053) Apparecchiature per la tecnologia dell'informazione - Sicurezza

Parte 1: Requisiti generali

#### 5.2 - Dielectric Strength Test

(id 2806) The dielectric strength of solid insulation used in equipment must be adequate. Compliance is verified according to 5.2.2 while the equipment is still hot, immediately after the temperature test specified in 4.5.2.

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 300.00 Limit #1: Up to 5 kV dc

#### 4.5.2 - Temperature tests

(id 2807) Materials used in the components and construction of the equipment shall be selected so that, under normal load conditions, temperatures do not exceed the safe values ??of this standard.

Components operating at high temperatures must be effectively shielded or separated to prevent overheating of surrounding materials and components.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

## CEI EN 62208:2012

(id 2001) Involucri vuoti per apparecchiature assiemate di protezione e di manovra per bassa tensione - Prescrizioni generali

#### 9.10 - Tenuta dielettrica

(id 2768) Questa prova si applica agli involucri costruiti in materiale isolante, anche in combinazione con materiali non isolanti.

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 300.00 Limit #1: Up to 5kV DC without preconditioning
Unit price #2 [EUR]: 600.00 Limit #2: Up to 5kV DC with preconditioning



# PRICE LIST / Listino prezzi

# § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

TNDEX



# **CEMP specification:2022 Ed.1**

(id 1815) Descrizione tecnica esecuzione test per certificazione RMRS

#### 03.03 - Test for exposure to temperature changes

(id 2444) Test performed under RMRS 2-020101-156E - 10.5.4.3

Accredited [IEC17025]: NO Execution time [dav]: 6

Unit price #1 [EUR]: 1,600.00 Limit #1: Applicable to any sample

#### 03.05 - Tests for exposure to hoarfrost and dew

(id 2445) Test performed under RMRS 2-020101-156E - 10.5.4.5

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

#### 03.06 - Test for exposure to salt mist

(id 2446) Test performed under RMRS 2-020101-156E - 10.5.4.6

Accredited [IEC17025]: NO Execution time [day]: 7

Unit price #1 [EUR]: 2,500.00 Limit #1: Applicable to any sample

#### 03.07 - Tests for exposure to solar radiation

(id 2447) Test performed under RMRS 2-020101-156E - 10.5.4.8

Accredited [IEC17025]: NO Execution time [day]: 5

Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 03.08 - Test of enclosure protection

(id 2448) Protection against penetration of hard objects. Test performed under RMRS 2-020101-156E - 10.5.5

Accredited [IEC17025]: NO Execution time [day]: 2

Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

# Chrysler Group PF-8950:2010 Rev. B (id 1888) FUEL TANK ASSEMBLIES – HIGH DENSITY POLYETHYLENE

#### 5.14 - Contain Correct Fuel Volume

(id 2524) The fuel system shall be able to consistently meet the rated fuel capacity under all commercial refueling conditions and vehicle duty cycles for the intended useful life of the vehicle. The fuel tank assembly shall meet the Chrysler Global Refueling Specification when installed in the vehicle system.

Accredited [IEC17025]: NO Execution time [dav]: 2

Unit price #1 [EUR]: 600.00 Limit #1: Applicable to any sample

#### 5.15 - Measure Fuel in Tank

(id 2525) The fuel gauge shall accurately represent the amount of fuel in the fuel tank under all vehicle duty cycles including all driving and parking conditions.

Accredited [IEC17025]: NO Execution time [day]: 8

Unit price #1 [EUR]: 850.00 Limit #1: Applicable to any sample

### Clivet

(id 2052) Internal Clivet specification

#### 40 - Impulse test

(id 2805) ---

Accredited [IEC17025]: NO Execution time [day]: 5

Unit price #1 [EUR]: 5,500.00 Limit #1: Applicable to any sample

## CNH 82032356:2013 Rev. C

(id 2060) Fuel Tank Assembly - Rotomoulded and Fabricated Steel

#### 5.9 - Drop Test

(id 2826) This test is intended to verify the resistance of a tank dropped from a height of 2 m after a conditioning at -40°C.

Accredited [IEC17025]: NO Execution time [dav]: 1

Unit price #1 [EUR]: 1,000.00 Limit #1: Applicable to any sample

page 16/234



albarubens

# PRICE LIST / Listino prezzi § T - LABORATORY TESTS / Prove di laboratorio

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



QM ref: ALL 7.1.4

## CNH ENS0310 82036374 Rev. R

(id 2145) Engineering Specification - Environmental Test Specification - Electronic Components

#### 10.4.7 - Pre-Paint Wash Test

(id 2973) Pre-paint wash is an integral part of the overall Ag/CE paint process. This test assesses resistance to general corrosion. It is useful for evaluating the quality and uniformity of protective coatings. It also ensures that supplied parts meet CNH Material Specification MAT0103.

This Pre-paint test checks for the ability to withstand Moisture ingress and 60°C heat. Moisture and heat can cause physical changes in materials and corrosion.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

## CNH ENS0310 82036374 Rev.J

(id 1409) Fiat Industrial CNH Engineering Specifications Environmental Test Specification - Electronic Components

#### 9.1.4 - Temperature Shock test under CNH ENS0310 82036374 Rev.J

(id 1869) The Temperature Shock test confirms the ability of the unit under test to withstand sudden changes in temperature of the surrounding atmosphere that could occur during storage or transportation.

Exposure to rapid temperature changes can permanently affect operation of the unit. Examples of symptoms are:

- Shattering of glass or plastics.
- Permanent changes in electronic component performance.
- Failure of seals.
- Slackening of fixings.

Accredited [IEC17025]: NO Execution time [day]: 30
Unit price #1 [EUR]: 5,600.00 Limit #1: Not relevant

## CNH ENS0310 82036374 Rev.M

(id 1408) CNH Industrial Engineering Specifications - Environmental Test Specification - Electronic Components

#### 10.1.3 - Temperature Cycle Test (Slow Changes, Operational)

(id 1867) The Temperature Cycle test determines the ability of the unit under test to start-up and to operate throughout the operating temperature range encountered in service

Temperature cycling can cause temporary or permanent deterioration in performance of the unit, or failure of structural parts. Examples of symptoms are:

- Failure of unit to start or operate due to changes in electronic components (resistors, capacitors, etc.).
- Binding of moving parts due to differential expansion.
- Cracking, crazing or embrittlement of materials.
- Failure of seals and adhesive bonds.

Accredited [IEC17025]: NO Execution time [day]: 10
Unit price #1 [EUR]: 2,400.00 Limit #1: Not relevant

### 10.1.4.4 - Temperature Shock Test (Sudden Temperature Changes) - Procedure 1 (Sudden Changes in Temperature)

(id 1868) The Temperature Shock test confirms the ability of the unit under test to withstand sudden changes in temperature of the surrounding atmosphere.

The shipping and storage temperature test is mandatory for all electrical/electronic components subjected to following conditions:

- If the component is designed to operate normally over 95°C then Method Cold-Only shall apply to the components. For all other applications, the component shall comply with the Method Hot-and-Cold.
- The component shall be unpowered during the test, simulating the shipping/storage condition.

Exposure to rapid temperature changes can permanently affect operation of the unit. Examples of symptoms are:

- Shattering of glass or plastics.
- Permanent changes in electronic component performance.
- Failure of seals.
- Slackening of fixings.

Accredited [IEC17025]: NO Execution time [day]: 25
Unit price #1 [EUR]: 8,000.00 Limit #1: Not relevant

#### 10.3.2 - Dust Ingress Test IP5K-6K

(id 1870) The dust ingress test checks for the effects of dust laden environments generated by soil working operations in dry conditions.

Dust can penetrate joints or bearing surfaces of moving parts. Examples of symptoms are:

- Penetration of seals.
- Abrasion or seizing, or disturbance in motion of bearings, axles, shafts and other moving parts,
- Change in performance of electrical circuits due to buildup of dust,
- Overheating caused by buildup of dust on heat sink surfaces or ducts.

page 17/234



# § T - LABORATORY TESTS / Prove di laboratorio

PRICE LIST / Listino prezzi

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

#### INDEX

2		

Accredited [IEC17025]: NO	Execution	time [day]: 3	
Unit price #1 [EUR]: 800.00		Limit #1: sample movable by one operator	
Unit price #2 [EUR]: 1,100.00		Limit #2: sample movable by two operator	
Unit price #3 [EUR]: 1,800.00		Limit #3: huge size samples	

#### 10.4.2.6 - Fast Icing Cycle Sealant (Procedure 2) - 1) COLD

(id 2018) This test is part of the verifications of compatibility and performance of the silicon sealant versus housing material and temperature/ice cycles. This test applies to all modules using a form-in-place sealant technology (i.e. IP67 and above) and with a vent valve.

```
Accredited [IEC17025]: NO Execution time [day]: 4
Unit price #1: ask to office Limit #1: Not relevant
```

#### 10.4.2.6 - Fast Icing Cycle Sealant (Procedure 2) - 2) WATER/ICE CYCLES

(id 2019) This test is part of the verifications of compatibility and performance of the silicon sealant versus housing material and temperature/ice cycles. This test applies to all modules using a form-in-place sealant technology (i.e. IP67 and above) and with a vent valve.

```
Accredited [IEC17025]: NO Execution time [day]: 5
Unit price #1: ask to office Limit #1: Not relevant
```

#### 10.4.2.6 - Fast Icing Cycle Sealant (Procedure 2) - 3) HIGH TEMPERATURE SOAK

(id 2020) This test is part of the verifications of compatibility and performance of the silicon sealant versus housing material and temperature/ice cycles. This test applies to all modules using a form-in-place sealant technology (i.e. IP67 and above) and with a vent valve.

```
Accredited [IEC17025]: NO Execution time [day]: 5
Unit price #1: ask to office Limit #1: Not relevant
```

#### 10.4.2.6 - Fast Icing Cycle Sealant (Procedure 2) - 4) CHEMICAL RESISTANCE

(id 2021) This test is part of the verifications of compatibility and performance of the silicon sealant versus housing material and temperature/ice cycles. This test applies to all modules using a form-in-place sealant technology (i.e. IP67 and above) and with a vent valve.

```
Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant
```

#### 10.3.3.5 - Water Ingress Test - High pressure jet washing IPX9K

(id 2223) Water during high-pressure/steam-jet cleaning

```
Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 800.00 Limit #1: sample movable by one operator
Unit price #2 [EUR]: 1,200.00 Limit #2: sample movable by two operator
Unit price #3 [EUR]: 1,800.00 Limit #3: huge size sample
```

### 10.3.1.5 - Altitude test - Procedure 2 (Operational)

(id 2364) Altitude tests are low pressure tests to ensure that the unit can withstand the low pressures involved in air transport, and can operate in high altitudes.

```
Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 2,000.00 Limit #1: Not relevant
```

#### 10.3.3.4 - Water ingress test - Procedure 1 (Hose Washing)

(id 2365) Water Ingress Test checks for resistance to water from the normal working environment.

#### 10.3.3.6 - Water ingress test - Procedure 3 (Heavy rain)

(id 2366) Water Ingress Test checks for resistance to water from the normal working environment.

```
Accredited [IEC17025]: NO Execution time [day]: 0 Unit price \#1 [EUR]: 600.00 Limit \#1: Up to 2 hours, no rotation.
```

#### 10.3.3.7 - Water ingress test - Procedure 4 (Rain and Shine)

(id 2367) Water Ingress Test checks for resistance to water from the normal working environment.

```
Accredited [IEC17025]: NO Execution time [day]: 9
Unit price #1 [EUR]: 1,800.00 Limit #1: Applicable to any sample
```

#### 10.4.4.4 - Salt Spray Test - Procedure 1 (Corrosion)

(id 2368) Salt spray is an established empirical test to assess resistance to general corrosion. It is useful for evaluating the quality and uniformity of protective coatings.

Accredited [IEC17025]: NO	Execution time [day]: 19
Unit price #1 [EUR]: 1,500.00	Limit #1: 1 cycle
Unit price #2 [EUR]: 3 500 00	Limit #2: 4 cycles

page 18/234



# PRICE LIST / Listino prezzi

# § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



#### 10.4.5 - Kesternich - Resistance to industrial atmosphere (T and B only)

(id 2369) This test verifies, within a sufficiently short time, the component/system capacity to stand or operate in corrosive environments, submitting device, equipped with its connectors and guards, to test in industrial atmosphere. This test applies to metallic materials with inorganic and electrolytic finishes.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 2,500.00 Limit #1: Applicable to any sample

#### 10.4.6.6 - Chemical Resistance Tests - Procedure 3 (Resistance to fluids)

(id 2370) Chemical resistance testing checks for the ability to withstand chemicals encountered in the normal working agricultural and construction environments.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 500.00 Limit #1: Up to 1 hour and 1 fluid

## 10.4.6.7 - Chemical Resistance Tests - Procedure 4 (Resistance to flushing hydrocarbon mixtures)

(id 2371) Chemical resistance testing checks for the ability to withstand chemicals encountered in the normal working agricultural and construction environments.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 800.00 Limit #1: Not relevant

#### 10.5.1 - Material Flammability test

(id 2372) This test applies to plastic material exposed to fire. It also applies to parts mounted inside and outside of cabs. If the compliance for the used material is fully documented then this test may not be necessary.

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 10.3.1.4 - Altitude Test Procedure 1 (Not operational, Transport)

(id 2377) Altitude tests are low pressure tests to ensure that the unit can withstand the low pressures involved in air transport, and can operate in high altitudes.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,600.00 Limit #1: Not relevant

#### 10.4.3.3 - Moisture resistance - Procedure 2 (All locations except 3A and 6)

(id 2378) The aim of this test is to verify the capability of components, equipped with connectors and any guards, to withstand in moisture environments.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

#### 10.5.2 - Explosive environments

(id 2379) This section applies to those components to be installed on vehicles carrying dangerous goods.

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 800.00 Limit #1: Not relevant

#### 10.4.7 - Pre-Paint Wash Test

(id 2381) Pre-paint wash is an integral part of the overall Ag/CE paint process. This test assesses resistance to general corrosion. It is useful for evaluating the quality and uniformity of protective coatings. It also ensures that supplied parts meet CNH Material Specification MAT0103.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 500.00 Limit #1: Not relevant

#### 10.4.6.4 - Chemical Resistance Tests Procedure 1 (Resistance to Chemicals)

(id 2627) ---

Accredited [IEC17025]: NO Execution time [day]: 5
Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

# CNH ENS0310 REV.K 12-Dec-2013

(id 1158) FIAT INDUSTRIAL Engineering Specifications CNH ENS0310

#### 10.4.4 - Salt Spray Test according to CNH ENS0310 specification

(id 1582) Salt spray is an established empirical test to assess resistance to general corrosion. It is useful for evaluating the quality and uniformity of protective

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 4,500.00 Limit #1: Not relevant



# § T - LABORATORY TESTS / Prove di laboratorio

PRICE LIST / Listino prezzi

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



# Commission Delegated Regulation EU No 44/2014

(id 2057) Commission Delegated Regulation (EU) No 44/2014 of 21 November 2013 supplementing Regulation (EU) No 168/2013 of the European Parliament and of the Council with regard to the vehicle construction and general requirements for the approval of two- or three-wheel vehicles and quadricycles.

#### Annex IX - 2.1 - Overturn Test

(id 2817) ---

#### Annex IX - 2.2 - Hydraulic test

(id 2818) The tank shall be subjected to a hydraulic internal pressure test which shall be carried out on an isolated unit complete with all its accessories.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,000.00 Limit #1: Applicable to any sample

#### Annex IX - 3.2 - Permeability Test

(id 2819) This test is intended to determine the maximum permissible average loss of fuel of a plastic tank per 24 hours testing time.

Accredited [IEC17025]: NO Execution time [day]: 84
Unit price #1 [EUR]: 4,000.00 Limit #1: Applicable to any sample

#### Annex IX - 3.3 - Shock Test

(id 2820) This is intended to verify the resistance in cold conditions of a plastic tank to an impact energy of 30 J.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,000.00 Limit #1: Up to 5 points

#### Annex IX - 3.4 - Mechanical Strength Test

(id 2821) This test is intended to verify the resistance of a water-filled plastic tank under 30 kPa pressure for 5 hours.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

#### Annex IX - 3.5 - Fuel Resistance Test

(id 2822) This test is intended to determine the tensile strength and elastic limits of six test pieces obtained from new tanks and aged tanks.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 600.00 Limit #1: Applicable to any sample

#### Annex IX - 3.6 - High Temperature Test

(id 2823) This test is intended to verify the resistance of a water-filled plastic tank at high temperatures.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 600.00 Limit #1: Applicable to any sample

#### Annex IX - 3.7 - Fire Resistance Test

(id 2824) This test is intended to determine the medium burning speed rate.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,400.00 Limit #1: Applicable to any sample

# CSA 3.16.21

(id 1980) New external document

#### 5.4.1 - Temperature Resistance - heat

(id 2719) ---

#### 5.4.2 - Temperature Resistance - cold

(id 2720) ---



# **PRICE LIST / Listino prezzi**

# § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

TNDEX

#### 3.6.2 - Bending test

(id 2908) A valve, in the least favourable orientation, shall be subjected to a bending moment as specified in Table 4.

```
Accredited [IEC17025]: NO
                                  Execution time [day]: 1
Unit price #1 [EUR]: 300.00
                                            Limit #1: up to 1"
Unit price #2 [EUR]: 400.00
                                            Limit #2: greater than 1" up to 3"
```

#### 5.2.1 - External leakage (shell)

```
Accredited [IEC17025]: NO
                                  Execution time [day]: 1
Unit price #1 [EUR]: 200.00
                                            Limit #1: up to 1"
Unit price #2 [EUR]: 250.00
                                            Limit #2: greater than 1" up to 3"
```

#### 5.2.2 - Seat leakage

(id 2910) ---

```
Accredited [IEC17025]: NO
                                  Execution time [day]: 1
                                            Limit #1: up to 1"
Unit price #1 [EUR]: 200.00
Unit price #2 [EUR]: 250.00
                                            Limit #2: greater than 1" up to 3"
```

#### 5.6.3 - Operating torque

```
Accredited [IEC17025]: NO
                                  Execution time [day]: 0
Unit price #1 [EUR]: 200.00
                                            Limit #1: up to 1"
Unit price #2 [EUR]: 300.00
                                            Limit #2: greater than 1" up to 3"
```

#### 5.6.4 - Valve stop mechanism strength test

```
Accredited [IEC17025]: NO
                                  Execution time [day]: 0
                                            Limit #1: up to 1"
Unit price #1 [EUR]: 200.00
Unit price #2 [EUR]: 300.00
                                            Limit #2: greater than 1" up to 3"
```

## CSA 9.4:21

(id 1978) Standard for manually operated metallic gas valves for use on piping systems up to 5 psig

#### 5.6.2 - Minimum temperature test

(id 2715) A valve shall be operable at ambient temperatures of 0°C or -40°C as applicable, whichever is lower without affecting the capability of valve to control the flow of gas when tested in accordance to 5.6.2.

```
Accredited [IEC17025]: NO
                                  Execution time [day]: 2
Unit price #1 [EUR]: 300.00
                                            Limit #1: up to 1"
Unit price #2 [EUR]: 500.00
                                            Limit #2: greater than 1" up to 3"
```

#### 5.6.3 - Maximum temperature test

(id 2716) A valve shall be operable at ambient temperatures of 52°C or 82°C as applicable, whichever is lower without affecting the capability of valve to control the flow of gas when tested in accordance to 5.6.2.

```
Accredited [IEC17025]: NO
                                 Execution time [dav]: 2
Unit price #1 [EUR]: 300.00
                                            Limit #1: up to 1"
Unit price #2 [EUR]: 500.00
                                            Limit #2: greater than 1" up to 3"
```

## 5.4.3 - Bending moment

(id 2904) A valve shall be capable of operation without cracking, breaking, or leaking while being subject to a bending moment using the weight specified in Table 5 in accordance with Clause 5.4.3.2.

```
Accredited [IEC17025]: NO
                                  Execution time [day]: 1
Unit price #1 [EUR]: 300.00
                                            Limit #1: up to 1"
Unit price #2 [EUR]: 400.00
                                            Limit \#2: greater than 1" up to 3"
```

#### 5.2.1 - External leakage (shell)

(id 2905) A manual valve shall not leak in excess of 1.22in^3/h (20cm^3/h) of air corrected to 30 in Hg column (101.3 kPa) pressure at 15.2°C when tested in accordance with Clause 5.2.1.2.

Accredited [IEC17025]: NO	Execution time [day]: 1
Unit price #1 [EUR]: 200.00	Limit #1: up to 1"
Unit price #2 [EUR]: 250.00	Limit #2: greater than 1" up to 3"



# § T - LABORATORY TESTS / Prove di laboratorio

PRICE LIST / Listino prezzi

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



#### 5.2.2 - Seat leakage

(id 2906) The leakage rate shall not exceed 50cm^3/h of air corrected to 30.0 in Hg column (101.3 kPa) pressure and 15.2°C.

## **CSA/ANSI Z21.15:2021**

(id 1891) Manually operated gas valves for appliances, appliance connector valves and hose-end valves.

#### 5.2 - Leakage

(id 2548) A manual valve shall not leak under any test condition in excess of 20cm^3/h of air corrected to 30 in Hg column (101.3 kPa) pressure at 15.5°C.

#### **5.4.1 - Continued Operation Test**

(id 2549) The valve, as received and at the end of this test, shall comply with Clause 5.2 and shall completely open and close on application of a torque or force not to exceed that specified in Table 11.

#### 5.5 - Low-temperature Operation

(id 2550) The valve shall comply with clause 5.5.2 at either 0°C, or lower temperature, as specified by the manufacturer.

#### 4.8.1 - Bending moment

(id 2907) A valve shall be capable of operation without cracking, breaking, or leaking while being subjected to a bending moment using the weight specified in Table 4, in accordance with Clause 4.8.1.2.

# D.M. 22 gennaio 2008 n. 37

(id 1023) Regolamento concernente l'attuazione dell'articolo 11-quaterdecies, comma 13, lettera a) della legge n. 248 del 2 dicembre 2005, recante riordino delle disposizioni in materia di attività di installazione degli impianti all'interno degli edifici

#### 5-6 - Impianti elettrici bassa tensione (> 6 kw)

(id 1535) Progettazione, realizzazione, installazione impianti per tutte le utenze condominiali e per utenze domestiche di singole unità abitative aventi potenza impegnata superiore a 6 kw o per utenze domestiche di singole unità abitative di superficie superiore a 400 mg

```
Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant
```

#### 5-6 - Impianti elettrici media tensione (> 6 kw)

(id 1536) Progettazione, realizzazione, installazione impianti relativi agli immobili adibiti ad attività produttive, al commercio, al terziario e ad altri usi, quando le utenze sono alimentate a tensione superiore a 1000 V, inclusa la parte in bassa tensione, o quando le utenze sono alimentate in bassa tensione aventi potenza impegnata superiore a 6 kw o qualora la superficie superi i 200 mq

```
Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1: ask to office Limit #1: Not relevant
```

#### 7 - Verifiche di legge e rilascio della dichiarazione di conformità degli impianti

(id 1537) Al termine dei lavori, previa effettuazione delle verifiche previste dalla normativa vigente, comprese quelle di funzionalità dell'impianto, l'impresa installatrice rilascia al committente la dichiarazione di conformità degli impianti realizzati nel rispetto delle norme di cui all'articolo 6. Di tale dichiarazione, resa sulla base del modello di cui all'allegato I, fanno parte integrante la relazione contenente la tipologia dei materiali impiegati, nonché il progetto di cui all'articolo 5.

Accredited [IEC17025]: NO	Execution time [day]: 2
Unit price #1 [EUR]: 1,000.00	Limit #1: Not relevant



albarubens

# PRICE LIST / Listino prezzi § T - LABORATORY TESTS / Prove di laboratorio

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



# Delegated Regulation EU No 134/2014

(id 1975) This supplements Regulation (EU) No 168/2013 of the European Parliament and of the Council with regard to environmental and propulsion unit performance requirements and amending Annex V thereof.

## Appendix 2 - 3 - Preconditioning fuel soak for the fuel tank permeation test

(id 2701) ---

#### Appendix 2 - 4 - Fuel tank permeation test procedure

(id 2702) Measure permeation emissions by weighing a sealed fuel tank before and after a temperature-controlled soak.

Accredited [IEC17025]: NO Execution time [day]: 14
Unit price #1 [EUR]: 1,000.00 Limit #1: Applicable to any sample

#### Appendix 2 - 6.1.1 - Pressure cycling

(id 2703) A separate durability demonstration for each substantially different combination of treatment approaches and non-metallic tank materials shall be performed by taking a series of steps.

Accredited [IEC17025]: NO Execution time [day]: 7
Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

#### Appendix 2 - 6.1.2 - UV exposure

(id 2704) A separate durability demonstration for each substantially different combination of treatment approaches and non-metallic tank materials shall be performed by taking a series of steps.

Accredited [IEC17025]: NO Execution time [day]: 19
Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

#### Appendix 2 - 6.1.3 - Slosh testing

(id 2705) A separate durability demonstration for each substantially different combination of treatment approaches and non-metallic tank materials shall be performed by taking a series of steps.

Accredited [IEC17025]: NO Execution time [day]: 46
Unit price #1 [EUR]: 2,500.00 Limit #1: Applicable to any sample

## DIN 53530:1981

(id 1939) Separation test on fabric plies bonded together

#### **Unref. - Initial Separation Resistance Test**

(id 2607) The separation test is used to determine the adhesion strength of fabric plies bonded together with elastomers or other layers of material.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 600.00 Limit #1: Applicable to any sample

#### **Unref. - Final Separation Resistance Test**

(id 2608) The separation test is used to determine the adhesion strength of fabric plies bonded together with elastomers or other layers of material.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 600.00 Limit #1: Applicable to any sample

#### DIN 73411-1:1996

(id 1934) Cooling pipes in motor vehicles - Hoses and compounds - Part 1: Dimensions, materials, types

#### 3.12.2 - Long Term Cycle Test

(id 2601) The test specimens shall be subjected to pressure cycles and mechanical oscillation.

Accredited [IEC17025]: NO Execution time [day]: 37
Unit price #1 [EUR]: 6,000.00 Limit #1: Applicable to any sample



# PRICE LIST / Listino prezzi

# § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

TNDEX

# 3.12.3 - Short Term Cycle Test

(id 2602) The test specimens shall be subjected to pressure cycles.

Accredited [IEC17025]: NO Execution time [day]: 4

Unit price #1 [EUR]: 2,000.00 Limit #1: Applicable to any sample

3.13.2 - Build-up Test

(id 2603) ---

Accredited [IEC17025]: NO Execution time [day]: 2

Unit price #1 [EUR]: 500.00 Limit #1: Applicable to any sample

# DIN EN 62093 (VDE 0126-20):2018-02

(id 1271) Photovoltaic system power conversion equipment - Design qualification testing

6.1 - Visual inspection

(id 1763) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant

6.2 - Functionality test

Execution time [day]: 0 Accredited [IEC17025]: NO

Unit price #1: ask to office Limit #1: Not relevant

6.3 - Insulation test

(id 1765) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant

6.4 - Thermal cycling test

(id 1766) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 1,600.00 Limit #1: Not relevant

6.5 - Dry heat test

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 2,400.00 Limit #1: Not relevant

6.6 - Damp heat test

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 1,200.00

Limit #1: Not relevant

6.7 - Humidity freeze test

(id 1769) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 1,500.00 Limit #1: Not relevant

6.8 - Rain intrusion test

(id 1770) ---

Execution time [day]: 0 Accredited [IEC17025]: NO

Unit price #1: ask to office Limit #1: Not relevant

6.9 - Wind driven rain test

(id 1771) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant



# PRICE LIST / Listino prezzi § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

TNDEX

#### 7.1 - Shipping vibration test

(id 1772) --

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1: ask to office

Limit #1: Not relevant

7.2 - Shock test

(id 1773) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant

7.3 - Salt mist test

(id 1774) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant

7.4 - Dust and sand test

(id 1775) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant

7.5 - Mixed gas corrosion test

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant

7.6 - Ammonia corrosion test

(id 1777) ---

Execution time [day]: 0 Accredited [IEC17025]: NO

Limit #1: Not relevant Unit price #1: ask to office

## DT.M.42:2009

(id 1813) TUBI FLESSIBILI PER CONDUZIONE ARIA IN MATERIALE PLASTICO E/O ELASTOMERI TERMOPLASTICI

### 8.9 - Pressione pulsata

(id 2437) ---

Accredited [IEC17025]: NO Execution time [day]: 10

Unit price #1 [EUR]: 5,500.00 Limit #1: Applicable to any sample

8.7 - Dilatazione diametrale

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

8.8 - Pressione di scoppio

(id 2439) ---

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 500.00 Limit #1: One temperature only (destructive testing)

## Ducati Standard DMH00-04.011

(id 1610) Prove di Resistenza agli Agenti Chimici PROVA della GOCCIA

#### unref - Spotting test with Brake Fluid (DOT 4)

(id 2042) The aim of this test is to investigate on the resistance to chemical agents of materials or surface finishing treatments.

Accredited [IEC17025]: NO Execution time [day]: 2

Unit price #1 [EUR]: 300.00 Limit #1: Not relevant

page 25/234



# PRICE LIST / Listino prezzi § T - LABORATORY TESTS / Prove di laboratorio

# alb**ar**ubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX

# unref - Spotting test: variation of the test fluid

(id 2043) The aim of this test is to investigate on the resistance to chemical agents of materials or surface finishing treatments.

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 300.00 Limit #1: Not relevant

# **Ducati Standard DMH00-04.013**

(id 1608) Prove di Resistenza agli Agenti Chimici PROVA di STROFINAMENTO

#### unref - Rubbing test with denatured Ethyl Alcohol

(id 2039) The aim of this test is to investigate on the resistance to chemical agents of materials or surface finishing treatments.

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1: ask to office Limit #1: Not relevant

#### unref - Rubbing test with Unleaded Petrol 95 RON

(id 2040) The aim of this test is to investigate on the resistance to chemical agents of materials or surface finishing treatments.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 200.00 Limit #1: Not relevant

#### unref - Rubbing test with Brake Fluid (DOT 4)

(id 2041) The aim of this test is to investigate on the resistance to chemical agents of materials or surface finishing treatments.

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1: ask to office Limit #1: Not relevant

## DVP 02 06 2015

(id 1247) DVP 02 06 2015

## 1.16 - Line system lifespan tightness

(id 1718) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

### 5.8 - Overheating protection of pipe heating

(id 1719) ---

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1: ask to office Limit #1: Not relevant

#### 5.9 - Heating duration switch at cold storage, Part C

(id 1720) ---

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1: ask to office Limit #1: Not relevant

#### 5.11 - Defreezing test

(id 1721) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

#### 11 - Fordability

(id 1722) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

#### 13 - Relative movements

(id 1723) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant



# PRICE LIST / Listino prezzi

# § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



## **DVP DAIMLER**

(id 1224) Hutchinson S.r.l. validation plan (24/01/2019)

#### 3.1.4 - Temperature changes (Life test)

(id 1637) ---

Accredited [IEC17025]: NO Execution time [day]: 63
Unit price #1 [EUR]: 8,400.00 Limit #1: Applicable to any sample up to 15 days
Unit price #2 [EUR]: 33,600.00 Limit #2: Applicable to any sample up to 63 days

#### 3.1.18 - Heating continuos operation

(id 1638) ---

Accredited [IEC17025]: NO Execution time [day]: 52
Unit price #1 [EUR]: 3,000.00 Limit #1: Up to 240h
Unit price #2 [EUR]: 7,500.00 Limit #2: 1250h

#### unref - Temperature resistance

(id 1737) Resistance of DUT at high temperatures and internal pressurized fluid

Accredited [IEC17025]: NO Execution time [day]: 42
Unit price #1 [EUR]: 5,500.00 Limit #1: duration 1000h
Unit price #2 [EUR]: 8,000.00 Limit #2: duration 1500h

## ECE R-100:2015/505

(id 1987) Regulation No 100 of the Economic Commission for Europe of the United Nations (UNECE) — Uniform provisions concerning the approval of vehicles with regard to specific requirements for the electric power train [2015/505]

#### 8A - Vibration test

(id 2731) The purpose of this test is to verify the safety performance of the REESS under a vibration environment which the REESS will likely experience during the normal operation of the vehicle.

Accredited [IEC17025]: NO Execution time [day]: 1 Subcontract: BPS srl
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 8B - Thermal shock and cycling test

(id 2732) The purpose of this test is to verify the resistance of the REESS to sudden changes in temperature. The REESS shall undergo a specified number of temperature cycles, which start at ambient temperature followed by high and low temperature cycling. It simulates a rapid environmental temperature change which a REESS would likely experience during its life.

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### Annex 8 - Appendix - Procedure for conducting a standard cycle

(id 2733) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 8C - Mechanical shock

(id 2734) The purpose of this test is to verify the safety performance of the REESS under inertial loads which may occur during a vehicle crash.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 8D - Mechanical integrity

(id 2735) The purpose of this test is to verify the safety performance of the REESS under contact loads which may occur during vehicle crash situation.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 8E - Fire resistance

(id 2736) The purpose of this test is to verify the resistance of the REESS, against exposure to fire from outside of the vehicle due to e.g. a fuel spill from a vehicle (either the vehicle itself or a nearby vehicle). This situation should leave the driver and passengers with enough time to evacuate.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant



# § T - LABORATORY TESTS / Prove di laboratorio

PRICE LIST / Listino prezzi

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



#### 8F - External short circuit protection

(id 2737) The purpose of this test is to verify the performance of the short circuit protection. This functionality, if implemented, shall interrupt or limit the short circuit current to prevent the REESS from any further related severe events caused by short circuit current.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 8G - Overcharge protection

(id 2738) The purpose of this test is to verify the performance of the overcharge protection.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 8H - Over-discharge protection

(id 2739) The purpose of this test is to verify the performance of the over-discharge protection. This functionality, if implemented, shall interrupt or limit the discharge current to prevent the REESS from any severe events caused by a too low SOC as specified by the manufacturer.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 8I - Over-temperature protection

(id 2740) The purpose of this test is to verify the performance of the protection measures of the REESS against internal overheating during the operation, even under the failure of the cooling function if applicable. In the case that no specific protection measures are necessary to prevent the REESS from reaching an unsafe state due to internal

over-temperature, this safe operation must be demonstrated.

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

## ECE R-100:2022 Rev. 3

(id 2113) Regulation No 100 of the Economic Commission for Europe of the United Nations (UNECE) — Uniform provisions concerning the approval of vehicles with regard to specific requirements for the electric power train

#### Annex 9 - Appendix 1 - Procedure for conducting a standard cycle

(id 2940) A standard cycle shall start with a standard discharge followed by a standard charge. The standard cycle shall be conducted at an ambient temperature of 20 ± 10 °C.

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### Annex 9 - Appendix 2 - Procedure for SOC adjustment

(id 2941) ---

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 9A - Vibration test

(id 2942) The purpose of this test is to verify the safety performance of the REESS under a vibration environment which the REESS will likely experience during the normal operation of the vehicle.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,800.00 Limit #1: Up to 8 hours

#### 9B - Thermal shock and cycling test

(id 2943) The purpose of this test is to verify the resistance of the REESS to sudden changes in temperature. The REESS shall undergo a specified number of temperature cycles, which start at ambient temperature followed by high and low temperature cycling. It simulates a rapid environmental temperature change which a REESS would likely experience during its life.

Accredited [IEC17025]: NO Execution time [day]: 3
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 9C - Mechanical shock

(id 2944) The purpose of this test is to verify the safety performance of the REESS under inertial loads which may occur during a vehicle crash.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant



# PRICE LIST / Listino prezzi § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



#### 9D - Mechanical integrity

(id 2945) The purpose of this test is to verify the safety performance of the REESS under contact loads which may occur during vehicle crash situation.

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 9E - Fire resistance

(id 2946) The purpose of this test is to verify the resistance of the REESS, against exposure to fire from outside of the vehicle due to e.g. a fuel spill from a vehicle (either the vehicle itself or a nearby vehicle). This situation should leave the driver and passengers with enough time to evacuate.

Accredited [IEC17025]: NO Execution time [day]: 2 Subcontract: RT Cert Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 9F - External short circuit protection

(id 2947) The purpose of this test is to verify the performance of the short circuit protection to prevent the REESS from any further related severe events caused by short circuit current.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 9G - Overcharge protection

(id 2948) The purpose of this test is to verify the performance of the overcharge protection to prevent the REESS from any further related severe events caused by a too high SOC.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 9H - Over-discharge protection

(id 2949) The purpose of this test is to verify the performance of the over-discharge protection to prevent the REESS from any severe events caused by a too low SOC.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 9I - Over-temperature protection

(id 2950) The purpose of this test is to verify the performance of the protection

measures of the REESS against internal overheating during the operation. In the case that no specific protection measures are necessary to prevent the REESS from reaching an unsafe state due to internal overtemperature, this safe operation must be demonstrated.

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 9J - Over-current protection

(id 2951) The purpose of this test is to verify the performance of the overcurrent

protection during DC external charging to prevent the REESS from any

severe events caused by excessive levels of charge current as specified by the manufacturer.

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

## ECE R-34:2016/1428

(id 2054) Regulation No 34 of the Economic Commission for Europe of the United Nations (UNECE) — Uniform provisions concerning the approval of vehicles with regard to the prevention of fire risks [2016/1428]

#### 6.1 - Hydraulic test

(id 2808) The tank shall be subjected to a hydraulic internal pressure test which shall be carried out on an isolated unit complete with all its accessories.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

#### 6.2 - Overturn test

(id 2809) ---

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 800.00 Limit #1: Up to 401
Unit price #2 [EUR]: 1,200.00 Limit #2: From 401 a 701



# § T - LABORATORY TESTS / Prove di laboratorio

PRICE LIST / Listino prezzi

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



#### Annex 5 - 1 - Collision Resistance

(id 2811) This is intended to verify the resistance in cold conditions of a plastic tank to an impact energy of 30 J.

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 1,200.00 Limit #1: Applicable to any sample

#### Annex 5 - 2 - Mechanical strength

(id 2812) This test is intended to verify the resistance of a water-filled plastic tank under 30 kPa pressure for 5 hours.

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 1,000.00 Limit #1: Applicable to any sample

#### Annex 5 - 3 - Fuel Permeability

(id 2813) This test is intended to determine the maximum permissible average loss of fuel of a plastic tank per 24 hours testing time.

Accredited [IEC17025]: NO Execution time [day]: 84

Unit price #1 [EUR]: 5,500.00 Limit #1: Applicable to any sample

#### Annex 5 - 4 - Resistance to Fuel

(id 2814) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant

#### Annex 5 - 5 - Resistance to Fire

(id 2815) For two minutes, the plastic tank, fixed as on the vehicle, shall be exposed to flame. There shall be no leakage of liquid fuel from the tank.

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 4,200.00 Limit #1: Applicable to any sample

#### Annex 5 - 6 - Resistance to High Temperature

(id 2816) This test is intended to verify the resistance of a water-filled plastic tank at high temperatures.

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

## EN 12101-3:2015

(id 209) Smoke and heat control systems - Part 3: Specification for powered smoke and heat exhaust ventilators

Annex C - Electrical cabinets - Test method for the determination of fire resistance of powered smoke and heat control ventilators (fa (id 2022) This test method is an adaptation to electrical cabinets of the test method for the determination of fire resistance of powered smoke and heat control ventilators (fans) described in Annex C of EN 12101-3:2015.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 3,600.00 Limit #1: not relevant

#### Annex C - Test method for the determination of fire resistance of powered smoke and heat control ventilators (fans)

(id 2023) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 4,200.00 Limit #1: Diameter more equal 1000mm

#### Annex D - 2.4.3 - High temperature test

(id 2204) Test method for resistance to temperature of electric motors

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 700.00 Limit #1: Small size

## EN 12259-5:2002

(id 1630) Fixed firefighting systems - Components for sprinkler and water spray systems - Part 5: Water flow detectors

#### F.2.1 - Contact resistance test

(id 2055) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant





QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



## EN 12806:2003

(id 664) Automotive liquified petroleum gas components other than containers

#### unref - Vehicle connection

(id 1265) ---

Accredited [IEC17025]: NO
Unit price #1 [EUR]: 300.00

Execution time [day]: 1

Limit #1: Not relevant

## EN 12941:1998/A2:2008

(id 1789) Respiratory protective devices - Powered filtering devices incorporating a helmet or a hood - Requirements, testing, marking

#### 7.1.2 - Complete Device

(id 2376) Humidity and temperature cycle on respiratory protective devices.

Accredited [IEC17025]: NO Execution time [day]: 7

Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

## EN 12975-1:2006+A1:2010

(id 54) Thermal solar system and components. Solar collectors. General requirements.

#### - General requirements checklist

(id 851) Verification of general requirements satisfation

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 1,000.00 Limit #1: --

# EN 13012:2012

(id 117) Petrol filling stations - Construction and performance of automatic nozzles for use on fuel dispensers

#### **B.1 - Electrical resistance test**

(id 952) The aim of this test is to determine the electrical resistance through the body of the nozzle. The electrical resistance is measured between the nozzle input and the spout and on four other points on the spout.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 300.00 Limit #1: Applicable to any sample

#### **B.2** - Pressure test

(id 953) The aim of this test is to confirm that the pressurized compartments of the nozzle body withstand a pressure of 14 bar, applied to the nozzle body, up to the valve seat.

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 300.00 Limit #1: Applicable to any sample

#### **B.3** - Drop test

(id 954) The aim of this test is to confirm that the nozzle withstands the impact of being dropped by a user

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 200.00 Limit #1: Applicable to any sample

#### **B.4** - Tightness test

(id 955) The aim of this test is to confirm that the nozzle does not leak

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 400.00 Limit #1: Applicable to any sample

#### A.2 - Bending moment preconditioning

(id 956) Preconditioning of the nozzle applying a bending moment representative of those expected to be applied during normal use

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Applicable to any sample



# § T - LABORATORY TESTS / Prove di laboratorio

PRICE LIST / Listino prezzi

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



#### A.3 - Fuel compatibility preconditioning

(id 957) The test is performed with the test liquid for grade 1 tanks according to EN 976-1: 1997 §6.9.2, with the following composition:

41.5% Toluene

41.5% Isooctane

15% Methanol

2% Isobutanol

Accredited [IEC17025]: NO Execution time [day]: 8

Unit price #1: ask to office Limit #1: Applicable to any sample

#### B.5 - Automatic shut-off device test 1

(id 958) The aim of this test is to confirm that the automatic shut-off device stops fluid flow when spout is immersed in fluid

Accredited [IEC17025]: NO Execution time [day]: -1

Unit price #1: ask to office Limit #1: Applicable to any sample

#### **B.6** - Drain test

(id 959) This test aims to confirm that draining of the hose through the nozzle is restricted

Accredited [IEC17025]: NO Execution time [day]: -1

Unit price #1: ask to office Limit #1: Applicable to any sample

#### B.7 - Attitude device test 1

(id 960) This test aims to confirm that the attitude device prevents flow of fluid when the spout axis is pointing at or above the spout axis angle

Accredited [IEC17025]: NO Execution time [day]: -1

Unit price #1: ask to office Limit #1: Applicable to any sample

#### B.8 - Attitude device test 2

(id 961) The aim of this test is to measure the fluid emission when the attitude device operates under defined conditions

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Applicable to any sample

#### B.9 - Automatic de-activating mechanism test

(id 962) This test aims to confirm that nozzles with an automatic de-activating mechanism entirely within the nozzle can not deliver fuel after flow has been stopped until the nozzle has been manually re-opened

Accredited [IEC17025]: NO Execution time [day]: -1

Unit price #1: ask to office Limit #1: Applicable to any sample

#### B.10 - Line shock generated test

(id 963) This test aims to confirm that the nozzle does not generate potentially destructive pressure during operation

Accredited [IEC17025]: NO Execution time [day]: -1

Unit price #1: ask to office Limit #1: Applicable to any sample

## EN 13274-7:2019

(id 1536) Respiratory protective devices - Methods of test - Part 7: Determination of particle filter penetration

#### 6 - Particle filter penetration: sodium chloride test method

(id 2008) An aerosol of sodium chloride particles is generated by atomizing an aqueous solution of the salt and evaporating the water. The concentration of this aerosol is measured before and after the filter under test by means of flame photometry.

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 600.00 Limit #1: Medical type mask
Unit price #2 [EUR]: 900.00 Limit #2: Pre-formed type mask

## EN 136:1998

(id 811) Respiratory protective devices - Full face masks - Requirements, testing, marking

### 8.2 - Endurance

(id 1347) ---

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 1,200.00 Limit #1: Not relevant



# **PRICE LIST / Listino prezzi** § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

TNDEX



## EN 13760:2003

(id 26) Automotive LPG filling system for light and heavy vehicles - Nozzle, test requirements and dimensions

#### 5.2 - Overpressure

(id 510) --

Accredited [IEC17025]: NO Unit price #1 [EUR]: 300.00 Execution time [day]: 1

Limit #1: Applicable to any sample

#### 5.3 - External leak tests at -20°C or -40°C

(id 511) ---

Accredited [IEC17025]: NO Unit price #1 [EUR]: 800.00 Execution time [day]: 3

Limit #1: Applicable to any sample

#### 5.4 - Endurance

(id 512) ---

Accredited [IEC17025]: NO

Execution time [day]: 10

Unit price #1 [EUR]: 1,200.00 Limit #1: Applicable to any sample

#### 5.5 - LPG compatibility (for rubber materials)

Accredited [IEC17025]: NO Unit price #1 [EUR]: 500.00 Execution time [day]: 15

Limit #1: Applicable to any sample

#### 5.6 - Corrosion resistance

(id 514) ---

Accredited [IEC17025]: NO

Execution time [day]: 15 Unit price #1 [EUR]: 2,200.00

Limit #1: Applicable to any sample

#### 5.7 - Resistance to dry heat

(id 515) ---

Accredited [IEC17025]: NO Unit price #1 [EUR]: 500.00 Execution time [day]: 15 Limit #1: Applicable to any sample

#### 5.8 - Ozone ageing

(id 516) ---

Accredited [IEC17025]: NO

Unit price #1 [EUR]: 600.00 Limit #1: Applicable to any sample

Execution time [day]: 15

### 5.9 - Temperature cycle

Accredited [IEC17025]: NO

Execution time [day]: 15 Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

5.10 - Drop test

(id 518) ---

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 500.00 Limit #1: Applicable to any sample

#### 5.11 - Electrical continuity test of the filling nozzle

(id 519) ---

Accredited [IEC17025]: NO Execution time [day]: 5

Unit price #1 [EUR]: 400.00 Limit #1: Applicable to any sample

5.12 - Freezing

(id 520) ---

Accredited [IEC17025]: NO Execution time [day]: 15

Unit price #1 [EUR]: 500.00 Limit #1: Applicable to any sample

page 33/234



# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



#### 5.3 - External leak tests at 65°C

(id 1564) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 400.00 Limit #1: Applicable to any sample

#### 5.3 - External leak tests at 20°C

(id 1565) --

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 400.00 Limit #1: Applicable to any sample

## EN 140:1998

(id 1864) Respiratory protective devices - Half masks and quarter masks - Requirements, testing, marking

#### **Unref. - Conditioning**

(id 2480) This test is inteded to condition a PPE sample with a thermal cycle of 48 hours.

Accredited [IEC17025]: NO Execution time [day]: 2

Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

## EN 144-1:2018

(id 1739) Respiratory protective devices - Gas cylinder valves - Part 1: Inlet connections

#### 5 - Impact resistance

(id 2202) This test method is based on a verification of the impact resistance of inlet connections for connecting cylinder valves to gas cylinders for respiratory protective devices (RPD).

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant

## EN 14460:2018

(id 1066) Explosion resistant equipment

#### 6.3.2 - Pressure test

(id 2930) Pressure testing of explosion-pressure-shock resistant equipment.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 6.3.3 - Explosion test

(id 2931) ---

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### EN 14678-1:2013

(id 51) LPG equipment and accessories. Construction and performance of LPG equipment for automotive filling stations. Dispensers

# - LPG equipment and accessories – construction and performance of LPG equipment for auto motive filling stations (id 845) ---

Accredited [IEC17025]: NO Execution time [day]: -1

Unit price #1: ask to office Limit #1: Applicable to any sample

### EN 14683:2019

(id 1516) Medical face masks - Requirements and test methods

#### 5.1 - General: materials and construction, design

(id 1983) Evaluation of the construction, design, and face fit of a face mask.



# **PRICE LIST / Listino prezzi** your value certified § T - LABORATORY TESTS / Prove di laboratorio

# QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

#### TNDEX

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 300.00 Limit #1: Not relevant

#### 5.2.2 - Filtration efficiency - internally modified test

(id 1985) Test not performed in accordance with § 5.2.2 of EN14683, but following some details of EN149, and internal variations: the test is not performed with the Staphylococcus aureus required by the standard but with NaCl. Instead of control plates of bacterial culture, a Flame Photometer is used, as per 8.5.2.2.3 of EN 149. The mechanical setup is similar to Figure B.3 of EN14683.

When a mask consists of two or more areas with different characteristics or different layer-composition, each area shall be tested individually. The lowest performing panel or area shall determine the Filtration Efficiency value of the complete mask.

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1 [EUR]: 600.00 Limit #1: Oualification in package Unit price #2 [EUR]: 600.00 Limit #2: Single measurement

#### 5.2.3 - Breathability (differential pressure)

albarubens

(id 1986) The differential pressure generated by a constant air flow (8 I / min) through a known section (4.9 cm^2) of filtering material is used to determine the breathability of the sample under test.

At least 5 areas are tested for each sample, or more if the sample is made up of different areas of materials (at least 2 measurements for each material). For each tested sample an average ?P (mbar) is calculated.

The maximum ?P (mbar) obtained is divided by the known sampling area (4.9 cm^2) obtaining the breathability value in mbar / cm^2.

This value must be suitable for the class chosen for the mask under test according to the data acceptance criteria in Table 1 of EN 14683: 2019

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 600.00 Limit #1: Qualification in package Unit price #2 [EUR]: 200.00 Limit #2: Single measurement

#### 5.2.5 - Microbial cleanliness (Bioburden)

(id 1988) Test subcontracted to other laboratory

Accredited [IEC17025]: NO Execution time [day]: 7 Subcontract: MICRONA Laboratorio Unit price #1 [EUR]: 1,100.00 Limit #1: Not relevant

## EN 14797:2006

(id 696) Explosion venting devices

#### 7.3.2 - Explosion testing for function and mechanical integrity

(id 1370) Mechanical strength and venting efficiency of explosion venting devices shall be determined.

Accredited [IEC17025]: NO Execution time [day]: 2 Unit price #1 [EUR]: 300.00 Limit #1: Only documentation analysis Unit price #2 [EUR]: 3,000.00 Limit #2: Surface < = 0.5 m2 Unit price #3 [EUR]: 6,000.00 Limit #3: Surface > = 0.5 m2

#### 7.2.3 - Static activation pressure - Mechanical test method

(id 1371) All explosion venting devices shall be tested for static activation pressure. The static activation pressure is observed:

- when discharge begins;
- when the venting element releases from the retaining element.

Accredited [IEC17025]: NO Execution time [day]: 2 Unit price #1 [EUR]: 1,000.00 Limit #1: Surface < = 0.5 m2 Unit price #2 [EUR]: 2,000.00 Limit #2: Surface > 0.5 m2

#### 7.2.2 - Static activation pressure - Pressure test method

(id 2464) All explosion venting devices shall be tested for static activation pressure. The static activation pressure is observed:

- when discharge begins;
- when the venting element releases from the retaining element.

Accredited [IEC17025]: NO Execution time [day]: 2 Unit price #1 [EUR]: 1,000.00 Limit #1: Surface < = 0.5 m2 Unit price #2 [EUR]: 2,000.00 Limit #2: Surface > 0.5 m2

# EN 149:2003+A1:2009

(id 1518) Respiratory protective devices - Filtering half masks to protect against particles - Requirements, testing, marking

#### 8.3.1 - Trattamento di indossamento simulato

(id 1990) ---

Accredited [TEC17025]: NO Execution time [day]: 0 Unit price #1: ask to office Limit #1: Not relevant

page 35/234



# PRICE LIST / Listino prezzi § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



#### 8.3.2 - Condizionamento a temperatura

(id 1991) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

#### 7.9.2 - Penetration of filter material

(id 2009) The test is intended to evaluate the maximum percentage of penetration through the face mask material.

It's performed as described in EN13274-7, using a water aerosol with NaCl as a tracking substance, monitored using a flame photometer.

The mechanical setup depends on the mask shape.

When the mask consists of two or more areas with different characteristics or different layer-composition, each area is tested individually. The lowest-performing panel or area shall determine the Penetration value of the complete mask.

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 600.00 Limit #1: Single measurement

# EN 15089:2009

(id 700) Explosion isolation systems

#### 7.2.2 - Module A: Explosion resistance testing

(id 1977) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

#### 7.2.3 a) - Module B: Flame transmission test (high pressure conditions)

(id 1978) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

#### 7.2.4.2 - Module C: Functional testing: passive explosion protection valves

(id 1979) The main objective of the functional testing is the determination/verification of the minimum/maximum installation distance of the isolation device dependent on the method of the detection and to assess the efficacy of the device/system as an explosion isolation device/system according to the intended use as specified by the manufacturer. The test will assess the efficacy of the device and the minimum and maximum installation distance. This shall be determined by varying the location of the ignition source in the test enclosure and other influencing parameters e.g., enclosure volume, maximum explosion overpressure in the enclosure according to the intended use.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

#### 7.2.4.3 - Module C: Functional testing: active isolation valves

(id 1980) The main objective of the functional testing is the determination/verification of the minimum/maximum installation distance of the isolation device dependent on the method of the detection and to assess the efficacy of the device/system as an explosion isolation device/system according to the intended use as specified by the manufacturer. The test will assess the efficacy of the device and the minimum and maximum installation distance. This shall be determined by varying the location of the ignition source in the test enclosure and other influencing parameters e.g., enclosure volume, maximum reduced explosion overpressure in the enclosure, type of detector or sensor used or combinations of detectors and sensors, activation pressure, sensitivity of flame detector, all according to the intended use.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

## 7.2.3 b) - Module B: Flame transmission test (low pressure conditions)

(id 1981) ---

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1: ask to office Limit #1: Not relevant

## EN 15276-1:2019

(id 1788) Fixed firefighting systems - Condensed aerosol extinguishing systems - Part 1: Requirements and test methods for components

### 7.7 - Accelerated ageing test

(id 2374) The objective of the test is to demonstrate the ability of the condensed aerosol generators to function correctly during the service life period of 15 years when stored in a dry, non-corrosive environment.

Accredited [IEC17025]: NO Execution time [day]: 17

Unit price #1 [EUR]: 4,000.00 Limit #1: Applicable to any sample



# PRICE LIST / Listino prezzi § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX

#### Unref. - Characterization

(id 2375) This test represents all the phases needed as acceptance criteria for Accelerated ageing test (§7.7).

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 500.00 Limit #1: Applicable to any sample

# EN 15714-3:2022

(id 2111) Industrial valves - Actuators - Part 3: Pneumatic part-turn actuators for industrial valves - Basic requirements

#### 7.4 - Endurance of part?turn actuators

(id 2925) The actuator shall be designed to guarantee a minimum endurance, without maintenance, in accordance with values given in Table 1; minimum endurance shall be demonstrated by a suitable test.

Accredited [IEC17025]: NO Execution time [day]: 21

Unit price #1 [EUR]: 2,500.00 Limit #1: Applicable to any sample

#### A.4.1/3 - Initial and final tests

(id 2937) Initial and final angular stroke based on 0°-90° shall be set and recorded.

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 1,200.00 Limit #1: Applicable to any sample

#### A.4.2 - Endurance test

(id 2938) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 2,500.00 Limit #1: Applicable to any sample

# EN 16314:2013

(id 1456) Gas meters - Additional functionalities

#### 7.13.4.5 - Valve closing

(id 1924) ---

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 350.00 Limit #1: Not relevant

# EN 175301-803:2006

(id 1180) Detail Specification: Rectangular connectors - Flat contacts, 0,8 mm thickness, locking screw not detachable

# CP2 - Retreatment Cleaning of specimen by washing briefly in light petrol

(id 1778) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

# EN 1775:2007

(id 42) Gas supply - Gas pipework for buildings - Maximum operating pressure less than or equal to 5 bar - Functional recommendations

# Annex A - (informative) Resistance to high temperatures

(id 720) Test method for resistance to temperature for Gas supply - Gas pipework for buildings

Accredited [IEC17025]: NO Execution time [day]: 5
Unit price #1 [EUR]: 400.00 Limit #1: cad for type

# EN 1809:2014+A1:2016

(id 1159) Diving equipment - Buoyancy compensators - Functional and safety requirements, test methods

# 5.3.3 - Wearing and environmental resistance tests - low temperatures

(id 1583) ---

page 37/234



# PRICE LIST / Listino prezzi your value certified § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

#### TNDEX

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1 [EUR]: 300.00 Limit #1: Applicable to any sample

# 5.3.2 - Wearing and environmental resistance tests - high temperatures

(id 1584) ---

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1 [EUR]: 200.00 Limit #1: Applicable to any sample

# EN 1834-1:2000

(id 50) Reciprocating Internal Combustion Engines - Safety Requirements for Design and Construction of Engines for Use in Potentially Explosives Atmospheres - Part 1: Group II Engines for Use in Flammable Gas and Vapour Atmospheres

# 6.4.2 - Spark arrester visual test

(id 846) The test is performed in dark room conditions.

For the test fresh grounded wood charcoal is used.

Before starting the test, the charcoal is sifted, in order to use only a fraction between 0.5 and 1 mm, and then weighed.

• Charcoal rate [g/s]: L/4, where L= engine swept volume in litres

The test shall be conducted at the following engine conditions or simulating them:

- 1) at maximum power;
- 2) at no load high idle speed;
- 3) accelerating from low idle to high idle over the period of 30 s but with no load.

A LPG burner and air system is usually employed for the test.

Each test is performed applying different LPG pressure and flow:

- 1: with high flow and maximum pressure for the burner employed;
- 2: with low flow and minimum pressure for the burner employed;
- 3: gradually increasing pressure and flow in order to simulate acceleration.

A different number of burners and different LPG pressures may be employed in order to simulate different engines and engine conditions/exhoust flow rate:

- Exhaust gas /LPG ratio: 10/1
- Exhaust gas flow rate = LPG flow rate \* Nr of Burners \*10
- For each burner used:

0.4 bar corresponds with a LPG flow rate of 40 Kg/h,

1.5 bar corresponds with a LPG flow rate of 200 Kg/h

```
Accredited [IEC17025]: NO
                                  Execution time [day]: 5
Unit price #1 [EUR]: 2,200.00
                                            Limit #1: small size up to 2 inches
Unit price #2 [EUR]: 2,800.00
                                            Limit #2: medium size up to 4 inches
Unit price #3 [EUR]: 3,500.00
                                            Limit #3: big size over 4 inches
```

# 6.2.1 - Determination of the maximum explosion pressure

(id 896) The purpose of these tests is to produce and measure the most severe explosion that is expected to occur in a flameproof enclosure (e.g. in an inlet or exhaust system). The explosion test is conducted with the maximum gap of the flameproof enxclosure defined by the manufacturer

```
Accredited [IEC17025]: NO
                                  Execution time [day]: 4
Unit price #1 [EUR]: 1,800.00
                                            Limit #1: No precompression
Unit price #2 [EUR]: 2,500.00
                                            Limit #2: With precompression
```

## 6.2.3 - Non transmission test

(id 897) The purpose of these tests is to produce the most severe explosion that is expected to occur in an inlet or exhaust system, and to verify that, under these conditions the explosion is contained within the flameproof enclosure and not transmitted to the surrounding atmosphere by assemblies and/or flame arresters. The non transmission test shall be conducted with the maximum gap of the flame proof enclosure defined by the manufacturer.

The non transmission test shall be performed with the gas mixture at atmospheric pressure for atmospheric systems and non pressurised parts of the air inlet

boosting device and at atmospheric pressure times pressure ratio for pressurised parts of the air inlet boosting device.

```
Accredited [IEC17025]: NO
                                  Execution time [day]: 4
Unit price #1 [EUR]: 2,000.00
                                            Limit #1: No precompression
Unit price #2 [EUR]: 2,800.00
                                            Limit #2: With precompression (turbo charged)
Unit price #3 [EUR]: 3,500.00
                                            Limit #3: Precompression with high temperature
```

## 6.2.2 - Overpressure test

(id 898) The purpose of this test is to apply a controlled pressure in excess of the greatest pressure that will be experienced under the most severe explosion and to show that under these circumstances the flameproof enclosure retains its integrity

Accredited	[IEC17025]: NO	Execution time [day]: 2
Unit price	#1 [EUR]: 800.00	Limit #1: Any engine



# PRICE LIST / Listino prezzi your value certified § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

TNDEX



# unref - Not ignition of hot insulated parts into explosive atmosfere

(id 904) ---

Accredited [IEC17025]: NO Execution time [day]: 2 Unit price #1 [EUR]: 2,000.00 Limit #1: Applicable to any sample

#### unref - Vibration test

(id 946) ---

Accredited [IEC17025]: NO Execution time [day]: 10

Unit price #1 [EUR]: 2,500.00 Limit #1: Applicable to any sample

# unref - Impact test

(id 947) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 200.00 Limit #1: Applicable to any sample

# EN 1834-2:2000

(id 226) Reciprocating internal combustion engines - Safety requirements for design and construction of engines for use in potentially explosive atmospheres - Part 2: Group I engines for use in underground workings susceptible to firedamp and/or combustible dust

# 6.3.2 - Determination of the maximum explosion pressure

(id 1045) The purpose of these tests is to produce and measure the most severe explosion that is expected to occur in a flameproof enclosure (e.g. in an inlet or exhaust system). The explosion test is conducted with the maximum gap of the flameproof enxclosure defined by the manufacturer

Accredited [IEC17025]: NO Execution time [day]: 2 Unit price #1 [EUR]: 1,500.00 Limit #1: No precompression Unit price #2 [EUR]: 2,200.00 Limit #2: With precompression (turbo charged or low temperature)

#### 6.3.3 - Overpressure test

(id 1046) The purpose of this test is to apply a controlled pressure in excess of the greatest pressure that will be experienced under the most severe explosion and to show that under these circumstances the flameproof enclosure retains its integrity

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1 [EUR]: 800.00 Limit #1: Any engine

#### 6.3.4 - Non transmission tests

(id 1047) The purpose of these tests is to produce the most severe explosion that is expected to occur in an inlet or exhaust system, and to verify that, under these conditions the explosion is contained within the flameproof enclosure and not transmitted to the surrounding atmosphere by assemblies and/or flame arresters. The non transmission test shall be conducted with the maximum gap of the flame proof enclosure defined by the manufacturer. The non transmission test shall be performed with the gas mixture at atmospheric pressure for atmospheric systems and non pressurised parts of the air inlet boosting device and at atmospheric pressure times pressure ratio for pressurised parts of the air inlet boosting device.

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1 [EUR]: 2,000.00 Limit #1: No precompression Unit price #2 [EUR]: 2,800.00 Limit #2: With precompression (turbo charged or low temperature)

# 6.4.2 - Spark arrester visual test

(id 1048) The test is performed in dark room conditions.

For the test fresh grounded wood charcoal is used.

Before starting the test, the charcoal is sifted, in order to use only a fraction between 0.5 and 1 mm, and then weighed.

• Charcoal rate [g/s]: L/4, where L= engine swept volume in litres

The test shall be conducted at the following engine conditions or simulating them:

- 1) at maximum power:
- 2) at no load high idle speed:
- 3) accelerating from low idle to high idle over the period of 30 s but with no load.

A LPG burner and air system is usually employed for the test.

Each test is performed applying different LPG pressure and flow.

- 1: with high flow and maximum pressure for the burner employed;
- 2: with low flow and minimum pressure for the burner employed:
- 3: gradually increasing pressure and flow in order to simulate acceleration.

A different number of burners and different LPG pressures may be employed in order to simulate different engines and engine conditions/exhoust flow rate:

- Exhaust gas /LPG ratio: 10/1
- Exhaust gas flow rate = LPG flow rate \* Nr of Burners \*10

page 39/234



# PRICE LIST / Listino prezzi § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

#### INDEX

• For each burner used:

0.4 bar corresponds with a LPG flow rate of 40 Kg/h, 1.5 bar corresponds with a LPG flow rate of 200 Kg/h

# EN 1834-3:2000

(id 227) Reciprocating internal combustion engines - Safety requirements for design and construction of engines for use in potentially explosive atmospheres - Part 3: Group II engines for use in flammable dust atmospheres

#### 6.3.2 - Spark arrester visual test

(id 1049) The test is performed in dark room conditions. Before starting the test, the charcoal is sifted, in order to use only a fraction between 0.5 and 1 mm, and weighed. The test shall be conducted at the following engine conditions: 1) at maximum power; 2) at no load high idle speed; 3) accelerating from low idle to high idle over the period of 30 s but with no load. So, each test is performed by applying different LPG pressure: 1: 1.5 bar, the maximum pressure for the burner employed; 2: 0.4 bar, the minimum pressure for the burner employed; 3: a gradually increasing pressure in order to simulate acceleration. A different number of burners may be employed in order to simulate different engine conditions. • Exhaust gas /LPG ratio: 10/1 • Exhaust gas flow rate = LPG flow rate \* Nr of Burners \*10 • For each burner: 0.4 bar corresponds with a LPG flow rate of 40 Kg/h, 1.5 bar corresponds with a LPG flow rate of 200 Kg/h • Charcoal rate [g/s]: L/4, where L= engine swept volume in litres

Accredited [IEC17025]: NO Execution time [day]: 5
Unit price #1 [EUR]: 3,800.00 Limit #1: over 4 inches
Unit price #2 [EUR]: 2,200.00 Limit #2: up to 2 inches

# EN 250:2014

(id 1737) Respiratory equipment - Open-circuit self-contained compressed air diving apparatus - Requirements, testing and marking

# 6.12.1 - Resistance to specific temperature: testing after storage at 70C

(id 2198) The temperature resistance tests are performed prior to conducting any other test on components or sub-assemblies of the resipratory equipment apparatus.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 400.00 Limit #1: Conditioning only
Unit price #2 [EUR]: 550.00 Limit #2: Conditioning and pressure test

## 6.12.2 - Resistance to specific temperature: testing after storage at -30C

(id 2199) The temperature resistance tests are performed prior to conducting any other test on components or sub-assemblies of the resipratory equipment apparatus.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 400.00 Limit #1: Conditioning only
Unit price #2 [EUR]: 550.00 Limit #2: Conditioning and pressure test

### 6.12.3 - Resistance to specific temperature: testing at 55C

(id 2200) The temperature resistance tests are performed prior to conducting any other test on components or sub-assemblies of the resipratory equipment apparatus.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 4,000.00 Limit #1: Conditioning only
Unit price #2 [EUR]: 550.00 Limit #2: Conditioning and pressure test

# 6.12.4 - Resistance to specific temperature: testing at -20C

(id 2201) The temperature resistance tests are performed prior to conducting any other test on components or sub-assemblies of the resipratory equipment apparatus.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 550.00 Limit #1: Applicable to any sample

# EN 2591-315:2015

(id 1286) Aerospace series - Elements of electrical and optical connection - Test methods - Part 315: Fluid resistance

# - Fluid resistance

(id 1798) ---

page 40/234



# § T - LABORATORY TESTS / Prove di laboratorio

PRICE LIST / Listino prezzi

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1: ask to office Limit #1: Not relevant

# EN 331:2015

(id 1974) Manually operated ball valves and closed bottom taper plug valves for gas installations for buildings

## 5.2 - Internal pressure and leak-tightness

(id 2689) A valve is leak-tight when the measured leakage rate does not exceed the values of Table 5, when tested as indicated in 5.2.

## 5.7 - Angular seal

(id 2690) This test is intended to measure the angle of rotation of the actuator.

#### 5.3 - Rated flow rate

(id 2691) ---

Accredited [IEC17025]: NO	Execution time [day]: 0
Unit price #1 [EUR]: 200.00	Limit #1: up to 1"
Unit price #2 [EUR]: 250.00	Limit #2: greater than 1" up to 3"
Unit price #3 [EUR]: 300.00	Limit #3: grater than 3"

## 5.5.2.1.1 - Torque MT1

(id 2692) The stresses MT1 (torque) and MF1 (bending moment) represent the installation stresses.

# 5.4 - Operating torque

(id 2693) ---

Accredited	[IEC17025]: NO	Execution time [day]: 2
Unit price	#1 [EUR]: 200.00	Limit #1: up to 1"
Unit price	#2 [EUR]: 250.00	Limit #2: greater than 1" up to 3"
Unit price	#3 [EIR]: 300 00	Limit #3: greater than 3"

# 5.8 - Safeguard against overloading of the handle - Stop resistance

(id 2694) ---

## 5.6.1 - Endurance test

(id 2695) The manual actuator of the valve shall withstand, at ambient temperature, a series of operating cycles (see Table 9).

After the endurance test the valve shall conform to the requirements for external and internal leaktightness at ambient temperature and at  $(60 \pm 5)$  °C in accordance with 4.4. It shall comply with the requirements for operating torque in accordance with 4.7.1.

# 5.6.2 - Resistance to low temperature

(id 2696) ---

Accredited [IEC17025]: NO	Execution time [day]: 2
Unit price #1 [EUR]: 400.00	Limit #1: up to 1"
Unit price #2 [EUR]: 600.00	Limit #2: greater than 1" up to 3"
Unit price #3 [EUR]: 800.00	Limit #3: greater than 3"

page 41/234



# PRICE LIST / Listino prezzi § T - LABORATORY TESTS / Prove di laboratorio

# alb**ar**ubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



### 5.6.3 - Salt spray resistance

(id 2697) This test shall be carried out on device N° 3 according to Table 10 following the requirements of EN ISO 9227, for a neutral salt spray (NSS).

```
Accredited [IEC17025]: NO Execution time [day]: 4
Unit price #1 [EUR]: 1,200.00 Limit #1: Applicable to any sample
Unit price #2 [EUR]: 1,200.00 Limit #2: Applicable to any sample
Unit price #3 [EUR]: 1,200.00 Limit #3: Applicable to any sample
```

#### 5.5.2.1.2 - Torque MT2

(id 2698) The stresses MT2 and MF2 represent the stresses to which the valve may be submitted during service.

# 5.5.2.2.1 - Bending moment MF1 at the axis of the obturator

(id 2699) The stresses MT1 (torque) and MF1 (bending moment) represent the installation stresses.

# 5.5.2.2.2 - Bending moment MF2 at axis of the obturator

(id 2700) The stresses MT2 and MF2 represent the stresses to which the valve may be submitted during service.

# EN 378-2:2016

(id 2172) Refrigerating systems and heat pumps - Safety and environmental requirements - Part 2: Design, construction, testing, marking and documentation

# Annex I - Leak simulation test for A2L, A2, A3, B2L, B2, B3 refrigerants

(id 3020) ---

```
Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 2,400.00 Limit #1: Applicable to any sample
```

# unref - Leak test inspired to Annex I

(id 3021) ---

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,600.00 Limit #1: Applicable to any sample

# EN 397:2025

(id 2156) Industrial protective helmets

#### 5.13.5 - Transferred charge test for insulating and isolated dissipative helmet components

(id 2983) The purpose of this test is to determine if insulating or isolated dissipative helmet components can be charged to the extent that brush discharges capable of igniting a potentially explosive atmosphere are likely to occur. The probability of a brush discharge igniting a potentially explosive atmosphere is determined by

measuring transferred charge in electrostatic discharges from insulating or isolated dissipative helmet components.

Rubbing with different cloths, such as cotton, polyamide, etc. simulates normal charge generating processes, while whipping with a leather glove and charging with a corona electrode is equivalent to strong charge generating processes.

```
Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 800.00 Limit #1: Applicable to one material
```

# 5.13.1 - Atmosphere for conditioning and testing

(id 2984) Prior to any measurement the specimens and any rubbing cloths (see 5.13.5) shall be conditioned in accordance with EN 60079-32-2:2015, 4.1.

```
Accredited [IEC17025]: YES Execution time [day]: 1
Unit price #1 [EUR]: 300.00 Limit #1: Applicable to any sample
```



# PRICE LIST / Listino prezzi

# § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



# 5.13.2 - Surface resistance test for helmets comprising dissipative or conductive components

(id 2985) ---

Accredited [IEC17025]: YES Execution time [day]: 1

Unit price #1 [EUR]: 300.00 Limit #1: Applicable to one material

## 5.13.3 - Leakage resistance test for helmets comprising dissipative or conductive components

(id 2986) The purpose of this test is to determine if any dissipative or conductive component of the helmet can be properly earthed via contact to the human body of the wearer, either directly or via other components of the helmet. Resistance measurements shall made from each dissipative or conductive component, via any intermediate components, including neck strap, cradle, sweat band, etc., to an electrode representing a human head.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 800.00 Limit #1: Applicable to one material

#### 5.13.4 - Capacitance test for isolated conductive helmet components

(id 2987) ---

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 800.00 Limit #1: Applicable to one material

# EN 403:2004

(id 797) Respiratory protective devices for self-rescue - Filtering devices with hood for escape from fire - Requirements, testing, marking

#### 7.4.5 - Temperature

(id 1333) ---

Accredited [IEC17025]: NO Execution time [day]: 7
Unit price #1 [EUR]: 1,200.00 Limit #1: Not relevant

#### 7.4.6 - Pressure changes

(id 1334) ---

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 1,800.00 Limit #1: Applicable to any sample

# EN 417:2012

(id 2010) Non-refillable metallic gas cartridges for liquefied petroleum gases, with or without a valve, for use with portable appliance — Construction, inspection, testing and marking

# 6.5 - Gas tightness of cartridges

(id 2771) The conformity of a gas cartridge design type with the applicable provisions and this standard shall be assessed in accordance with the conformity assessments procedures of ADR and RID for non-UN pressure receptacles, or for gas cartridges.

The test specified in 6.5 shall be performed on samples taken from 100 cartridges selected at random from a batch of filled cartridges produced in 1 h. The test is intended to verify the gas tightness of the cartridges after being subjacted to different temperatures.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

#### 6.6 - Gas tightness of valves

(id 2772) The conformity of a gas cartridge design type with the applicable provisions and this standard shall be assessed in accordance with the conformity assessments procedures of ADR and RID for non-UN pressure receptacles, or for gas cartridges. The test specified in 6.6 shall be performed on samples taken from 100 cartridges selected at random from a batch of filled cartridges produced in 1 h. The test is intended to verify the gas tightness of the cartridges after 50 manual operations of the valve.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 500.00 Limit #1: Applicable to any sample

# EN 50050-2:2013

(id 368) Electrostatic hand-held spraying equipment - Safety requirements Part 2: Hand-held spraying equipment for ignitable coating powder

### 5.2.1 - Test of earth connections

(id 1236) This test verifies that hand-held spraying equipment has an earth connection complying with § 4.1.7, 4.1.9 and 4.3.2 of EN 50050-2:2013.

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 200.00 Limit #1: Applicable to any sample



# **PRICE LIST / Listino prezzi**

# your value certified § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

TNDEX

# 5.2.2 - Resistance to earth test

(id 1237) See § 4.1.2, 4.1.5, 4.1.7, 4.3.2 and 4.3.5 of of EN 50050-2:2013.

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 300.00 Limit #1: Apply to any sample

## 5.2.3 - Test of contact surface area

(id 1238) The size of the contact surface area shall be proven by measurement, see § 4.1.5

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 100.00 Limit #1: Not relevant

#### 5.2.4 - Test of OFF-position

(id 1239) The function of the OFF-position shall be proven by measurement, see § 4.1.6.

Accredited [IEC17025]: NO Execution time [dav]: 0

Unit price #1 [EUR]: 200.00 Limit #1: Not relevant

#### 5.3.1 - Cable pull test

(id 1240) Each cable of the applicator, except for cables connecting intrinsically safe circuits among each other, shall be submitted to a cable pull test.

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 200.00 Limit #1: Not relevant

#### 5.3.2 - Impact test

(id 1241) This test is not applicable to triboelectrical applicators

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 200.00 Limit #1: Not relevant

#### 5.3.3 - Drop test

(id 1242) This test is not applicable to triboelectrical applicators

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 100.00 Limit #1: Not relevant

# 5.4.1.1 - Measurement of maximum output high voltage

(id 1243) The measurement of maximum output high voltage is carried out without load at the end of the high voltage electrode at the spraying device of the applicator. This test is not applicable to triboelectrical applicators.

Accredited [IEC17025]: NO Execution time [day]: 2

Unit price #1 [EUR]: 300.00 Limit #1: up to1000 V

# 5.4.1.2 - Measurement of maximum spraying current

(id 1244) This test is not applicable to triboelectrical applicators.

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 300.00 Limit #1: Not relevant

# 5.4.1.3 - Measurement of maximum short-circuit current

(id 1245) Measurement of the maximum short-circuit current, see § 4.2.1. This test is not applicable to triboelectrical applicators.

Execution time [day]: 0 Accredited [IEC17025]: NO

Unit price #1 [EUR]: 500.00 Limit #1: Not relevant

#### 5.4.2.1 - Test of parts at high voltage

(id 1246) The parts at high voltage of the applicator, which produce hazard of incendive discharges inside the applicator, are tested at Umax. All parts of the applicator which are at earth potential for operational reasons are earthed during the test. This test is not applicable to triboelectrical applicators.

Accredited [TEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 1,800.00 Limit #1: Not relevant

# 5.4.2.2 - Test of the high voltage cable

(id 1247) This test is not applicable to triboelectrical applicators.

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 1,800.00 Limit #1: Not relevant

## 5.4.3 - Test of the degree of protection of applicators

(id 1248) Compliance with degree of protection IP 64 to be tested according to EN 60529. This test is not applicable to triboelectrical applicators.

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample



# § T - LABORATORY TESTS / Prove di laboratorio

PRICE LIST / Listino prezzi

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



#### 5.5.1.2 - Ignition test of applicators with high voltage supply: test within the gas mixture

(id 1250) Test for observance of the limit value for maximum discharge energy of single sparks (2 mJ) according to Annex B.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 1,200.00 Limit #1: Applicable to any sample

#### 5.5.2 - Ignition test for triboelectrical applicators

(id 1251) Test for observance of the limit value for maximum transferred charge of single sparks (200 nC) by measurement according to Annex A.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

# 5.6.1 - Test of the accessories: the degree of protection

(id 1252) Compliance with degree of protection IP54 according to EN 60529 shall be tested on samples of equipment category 3D.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 600.00 Limit #1: Not relevant

# EN 50155:2017

(id 1424) Railway applications - Rolling stock - Electronic equipment

#### 13.4.4 - Low temperature start-up test

(id 1901) This test is carried out in accordance with EN 60068-2-1 (test Ad), using natural ventilation.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 500.00 Limit #1: up to 24h small sample
Unit price #2 [EUR]: 800.00 Limit #2: Big size sample

## 13.4.5.3 - Dry heat thermal test: cycle B

(id 1902) This test is carried out in accordance with EN 60068-2-2 (test Be), using natural ventilation unless a different type of cooling is normally used or provided to the equipment.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 800.00 Limit #1: up to 24 hrs

# 13.4.6 - Low temperature storage test

(id 1903) This test is carried out in accordance with EN 60068-2-1 (test Ab)

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 500.00 Limit #1: up to 24 hrs

#### 13.4.7 - Cyclic damp heat test

(id 1904) This test is carried out in accordance with EN 60068-2-30, test Db, Variant 2.

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 800.00 Limit #1: 2 cycles small sample
Unit price #2 [EUR]: 1,000.00 Limit #2: 2 cycles big sample

#### 13.4.9 - Insulation test

(id 1919) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

### 13.4.5.2 - Dry heat thermal test cycle A EN 50155

(id 1976) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

# EN 50177:2009+A1:2012

(id 2185) Stationary electrostatic application equipment for ignitable coating powders - Safety requirements

## 6.1 - Tests of the high voltage cables

(id 3066) ---

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant



# PRICE LIST / Listino prezzi

# § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX

# DEX

### 6.2.1 - Occupationally safe state

(id 3067) The stationary equipment for electrostatic coating with ignitable coating powders shall be tested for occupationally safe state.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

# 6.2.2 - Effectivity of forced ventilation (exhaust air systems)

(id 3068) --

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

# 6.2.3 - Voltage-controlled and constant voltage operation

(id 3069) Safe disconnection of high voltage shall be tested for voltage-controlled and constant voltage operation.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

## 6.2.4 - Constant current operation

(id 3070) Safe disconnection of high voltage supply shall be tested for constant current operation.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

### 6.2.5 - Protection against too high discharge energy

(id 3071) ---

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

# 6.2.6 - Protection against ignition of cleaning agents

(id 3072) ---

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

# 6.2.7 - Effectivity of measures for protection against direct contact

(id 3073) ---

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

# 6.2.8 - Effectivity of grounding measures

(id 3074) ---

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

## 6.2.9 - Interlocks - protection against entrance

(id 3075) ---

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

# 6.2.10 - Effectivity of locally acting fire extinguishing system

(id 3076) ---

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

## 6.2.11 - Further tests

(id 3077) In compliance with EN 12981:2005.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

# 6.3 - Specific test requirements for spraying systems of type B-P, type C-P category 2D

(id 3078) The test according to this subclause is made for proving that the requirements for equipment of category 2D are observed.

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant



# PRICE LIST / Listino prezzi

# § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



# EN 50365:2023

(id 2131) Live Working - Electrically insulating helmets for use on low and medium voltage installations

# 5.3.4 - AC Proof test voltage

(id 2956) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 300.00 Limit #1: Applicable to any sample

#### 5.3.6 - DC Proof voltage test

(id 2958) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 300.00 Limit #1: Applicable to any sample

## 5.3.3 - Preconditioning

(id 2970) ---

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 300.00 Limit #1: Applicable to any sample

#### 5.3.5 - AC Withstand test voltage

(id 3101) ---

Accredited [IEC17025]: NO Execution time [day]: 2

Unit price #1 [EUR]: 300.00 Limit #1: Applicable to any sample

# EN 50381:2004+AC:2005

(id 717) Transportable ventilated rooms with or without an internal source of release

# 14.3 - Overpressure test

(id 1539) ---

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 600.00 Limit #1: Typical sample on field
Unit price #2 [EUR]: 1,000.00 Limit #2: Performed at Laboratory
Unit price #3 [EUR]: 600.00 Limit #3: Huge sample on field

# 14.4/14.4.1.1 - Purging test for Type of protection vM2

(id 1540) ---

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 1,800.00 Limit #1: Typical sample on field
Unit price #2 [EUR]: 3,500.00 Limit #2: Performed at Laboratory
Unit price #3 [EUR]: 4,200.00 Limit #3: Huge sample on field

# 14.4/14.4.1.2 - Purging test for type of protection v2, v3 and v4

(id 1541) ---

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 1,800.00 Limit #1: Typical sample on field
Unit price #2 [EUR]: 3,500.00 Limit #2: Performed at Laboratory
Unit price #3 [EUR]: 5,000.00 Limit #3: Huge sample on field

#### 14.5 - Verification of minimum pressure differential

(id 1542) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 600.00 Limit #1: Typical sample on field
Unit price #2 [EUR]: 1,000.00 Limit #2: Performed at Laboratory
Unit price #3 [EUR]: 600.00 Limit #3: Huge sample on field

# 14.6 - Ventilation flow rate tests for TVR s with an internal source of release $(id\ 1543)$ ---

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 800.00 Limit #1: Typical sample on field
Unit price #2 [EUR]: 1,200.00 Limit #2: Performed at Laboratory
Unit price #3 [EUR]: 800.00 Limit #3: Huge sample on field



# PRICE LIST / Listino prezzi § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



# 14.7 - Overpressure test for containment systems with limited release

(id 1544) ---

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 600.00 Limit #1: Typical sample on field
Unit price #2 [EUR]: 1,000.00 Limit #2: Performed at Laboratory
Unit price #3 [EUR]: 600.00 Limit #3: Huge sample on field

# 14.8 - Verification of sequence of operation of the safety devices

(id 1545) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 300.00 Limit #1: Typical sample on field
Unit price #2 [EUR]: 300.00 Limit #2: Performed at Laboratory
Unit price #3 [EUR]: 1,000.00 Limit #3: Huge sample on field

#### 14.2 - Thermal tests

(id 1546) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 800.00 Limit #1: Typical sample on field
Unit price #2 [EUR]: 1,600.00 Limit #2: Performed at Laboratory
Unit price #3 [EUR]: 2,200.00 Limit #3: Huge sample on field

# EN 50521:2008

(id 768) Connectors for photovoltaic systems - Safety requirements and tests

# 6.2 - Preparation of specimens

(id 1318) ---

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1: ask to office Limit #1: Not relevant

### 6.3.2 - Durability of marking

(id 1319) ---

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1: ask to office Limit #1: Not relevant

# 6.3.3 - Protection against electric shock

(id 1320) ---

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1: ask to office Limit #1: Not relevant

# 6.3.4 - Temperature rise

(id 1321) ---

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1: ask to office Limit #1: Not relevant

# 6.3.5 - Mechanical operation

(id 1322) ---

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1: ask to office Limit #1: Not relevant

# 6.3.6 - Bending (flexing) test

(id 1323) ---

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1: ask to office Limit #1: Not relevant

# 6.3.7 - Measurement of clearances and creepage distances

(id 1324) ---

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1: ask to office Limit #1: Not relevant



# **PRICE LIST / Listino prezzi**

# § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

TNDEX

# 6.3.8 - Dielectric strength

(id 1325) ---

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1: ask to office

Limit #1: Not relevant

## 6.3.9 - Corrosion test

(id 1326) ---

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1: ask to office Limit #1: Not relevant

# 6.3.10 - Mechanical strength at lower temperatures

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1: ask to office Limit #1: Not relevant

## 6.3.11 - Thermal cycle test (EN 60068-2-14 test Nb)

(id 1328) ---

Accredited [IEC17025]: NO Execution time [day]: 10

Unit price #1: ask to office Limit #1: Not relevant

## 6.3.12 - Damp heat test

(id 1329) ---

Accredited [IEC17025]: NO Execution time [day]: 42

Unit price #1: ask to office Limit #1: Not relevant

#### 6.3.13 - Insertion and withdrawal force

(id 1330) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant

# 6.3.14 - Effectiveness of connector coupling device

(id 1331) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant

# 6.4 - Test schedule (routine test) for non-rewirable free connectors

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant

# EN 50545-1:2011+A1:2016

(id 2090) Electrical apparatus for the detection and measurement of toxic and combustible gases in car parks and tunnels - Part 1: General performance requirements and test methods for the detection and measurement of carbon monoxide and nitrogen oxides

### 6.2 - Unpowered storage

(id 1799) ---

Accredited [IEC17025]: NO Execution time [day]: 4

Unit price #1 [EUR]: 1,200.00 Limit #1: Applicable to any sample

# 6.3 - Linearity

(id 1800) ---

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

## 6.4 - Alarm set points and outputs (car parks only)

Accredited [IEC17025]: NO Execution time [day]: 2

Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample



# **PRICE LIST / Listino prezzi** § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

TNDEX

# 6.5 - Repeatability

(id 1802) ---

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 500.00 Limit #1: Applicable to any sample

6.6 - Temperature

(id 1803) --

Accredited [IEC17025]: NO Execution time [day]: 3

Unit price #1 [EUR]: 2,800.00 Limit #1: Applicable to any sample

6.7 - Humidity

(id 1804) ---

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 2.400.00 Limit #1: Applicable to any sample

6.9 - Interfering gases

(id 1805) ---

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

6.10 - Recovery from high gas concentrations

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 500.00 Limit #1: Applicable to any sample

6.11 - Mechanical strength

(id 1807) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 300.00 Limit #1: Applicable to any sample

6.12 - Warm-up time

(id 1808) ---

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 500.00 Limit #1: Applicable to any sample

6.13 - Response time

(id 1809) ---

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 500.00 Limit #1: Applicable to any sample

6.14 - Power supply variations

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 300.00 Limit #1: Applicable to any sample

6.15 - Long-term stability

(id 1811) ---

Accredited [IEC17025]: NO Execution time [day]: 60

Unit price #1 [EUR]: 2,400.00 Limit #1: Applicable to any sample

6.8 - OBSOLETE - Air velocity (for diffusion apparatus only)

(id 2880) ---

Accredited [IEC17025]: NO Execution time [day]: 2

Unit price #1: ask to office Limit #1: Not relevant

6.8 - Flow rate (for aspirated apparatus only)

(id 2881) ---

Accredited [IEC17025]: NO Execution time [day]: 2

Unit price #1: ask to office Limit #1: Not relevant



# PRICE LIST / Listino prezzi

# § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

TNDEX

# 6.16 - Field calibration, or field verification kit

(id 2882) ---

Accredited [IEC17025]: NO Execution time [day]: 2 Unit price #1: ask to office Limit #1: Not relevant

## 6.17 - Calibration mask

(id 2883) ---

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

#### 6.18 - Testing of multiple point selector

(id 2884) ---

Accredited [IEC17025]: NO Execution time [day]: 2

Limit #1: Applicable to any sample Unit price #1 [EUR]: 500.00

### 6.19 - Vibration

(id 2885) ---

Accredited [IEC17025]: NO Execution time [day]: 2

Limit #1: Not relevant Unit price #1: ask to office

#### 6.20 - Electromagnetic compatibility

(id 2886) ---

Accredited [IEC17025]: NO Execution time [day]: 2

Unit price #1: ask to office Limit #1: Not relevant

#### 6.1 - General

(id 2887) ---

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 200.00 Limit #1: Applicable to any sample

# EN 60068-2-1:2007

(id 2197) Environmental testing - Part 2-1: Tests - Test A: Cold

# 5.2 - Test Ab: Cold for non heat-dissipating specimens with gradual change of temperature

(id 936) Cold tests for non heat-dissipating specimens

Execution time [day]: 4 Accredited [IEC17025]: YES

Unit price #1 [EUR]: 800.00 Limit #1: up to 24 hours (medium/big size sample)

Unit price #2 [EUR]: 1,500.00 Limit #2: 96h (medium/big size sample)

# 5.3 - Test Ad: Cold for heat-dissipating specimens with gradual change of temperature that are powered after initial temperature sta (id 1062) Cold tests for heat-dissipating specimens

Accredited [IEC17025]: YES Execution time [day]: 1

Limit #1: up to 24 h (medium/big size sample) Unit price #1 [EUR]: 800.00 Unit price #2 [EUR]: 1,500.00 Limit #2: from 1 day to 3 days (medium/big size sample)

## 5.4 - Test Ae: Cold for heat-dissipating specimens with gradual change of temperature that are required to be powered throughout t (id 1063) Cold test for heat-dissipating specimens

Accredited [IEC17025]: YES Execution time [day]: 4 Unit price #1 [EUR]: 800.00

Limit #1: Up to 24 h (medium/big size sample) Limit #2: from 1 day to 5 days (medium/big size sample) Unit price #2 [EUR]: 1,500.00

Unit price #3 [EUR]: 4,000.00 Limit #3: long duration

# EN 60068-2-2:2007

(id 2200) Environmental testing - Part 2-2: Tests - Test B: Dry heat

# 5.2 - Test Bb: Dry heat for non heat-dissipating specimens with gradual change of temperature

(id 1020) Dry heat test for non heat-dissipating specimens

Accredited [IEC17025]: YES Execution time [day]: 7 Unit price #1 [EUR]: 800.00 Limit #1: up to 24 h Unit price #2 [EUR]: 1,000.00 Limit #2: 96h Unit price #3 [EUR]: 2,800.00 Limit #3: up to 1000h

page 51/234



# § T - LABORATORY TESTS / Prove di laboratorio

PRICE LIST / Listino prezzi

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

TNDEX



5.3 - Test Bd: Dry heat for heat-dissipating specimens with gradual change of temperature that are not powered during the condition (id 1064) Dry heat for heat-dissipating specimens

Accredited [IEC17025]: YES Execution time [day]: 42 Unit price #1 [EUR]: 800.00 Limit #1: up to 24 h Unit price #2 [EUR]: 1,200.00 Limit #2: from 1 day to 3 days Unit price #3 [EUR]: 2,800.00 Limit #3: Up to 1000h

#### 5.4 - Test Be: Dry heat for heat-dissipating specimens with gradual change of temperature that are required to be powered throughout (id 1065) Dry heat for heat-dissipating specimens

Accredited [IEC17025]: YES

Execution time [day]: 4 Unit price #1 [EUR]: 800.00 Limit #1: Up to 24 h Unit price #2 [EUR]: 1,200.00 Limit #2: From 1 day to 3 days Unit price #3 [EUR]: 1,800.00 Limit #3: Big sample

# EN 60068-2-30:2005

(id 251) Environmental testing - Part 2-30: Tests - Test Db: Damp heat, cyclic (12 h + 12 h cycle)

# Db - Damp heat cyclic (12h +12h cycle)

(id 1037) This test comprises one or more temperature cycles in which the relative humidity is maintained at a high level, in order to determine the suitability of components, equipment or other articles for use, transportation and storage under conditions of high humidity – combined with cyclic temperature changes and, in general, producing condensation on the surface of the specimen.

The severity shall be chosen from the following:

- a) upper temperature: 40 °C, number of cycles: 2, 6, 12, 21, 56;
- b) upper temperature: 55 °C, number of cycles: 1, 2, 6.

After stabilization, the temperature of the chamber is raised to the appropriate upper temperature prescribed by the relevant specification. The upper temperature is achieved in a period of 3 h  $\pm$  30 min.

During this period, the relative humidity is less than 95 % RH. During the last 15 min it is less than 90 % RH.

The temperature is then be maintained within the prescribed limits for the upper temperature (±2 K) until 12 h ± 30 min from the start of the cycle.

During this period, the relative humidity is 93 % RH ± 3 %RH. During the first and last 15 min it is between 90 % RH and 100 % RH. The temperature is then lowered in accordance with one of the two variants given below:

Variant 1: the temperature is lowered to 25 °C ± 3 K within 3 h to 6 h. The rate of fall for the first one and one half hours is such that it would result in a temperature of 25 °C ±3 K being attained in 3 h ± 15 min. The relative humidity is not less than 95 % RH. During the first 15 min it is not less than 90 % RH.

Variant 2: the temperature shall be lowered to 25 °C ± 3 K within 3 h to 6 h, but without the additional requirement for the first hour and one half as in variant 1. The relative humidity is not less than 80 % RH.

The temperature is then be maintained at 25 °C ± 3 K with a relative humidity of not less than 95 % RH until the 24 h cycle is completed.

Accredited [IEC17025]: YES Execution time [day]: 4 Unit price #1 [EUR]: 500.00 Limit #1: 1 cycle Unit price #2 [EUR]: 800.00 Limit #2: 2 cycles Unit price #3 [EUR]: 1,000.00 Limit #3: 6 cycles

# EN 60068-2-78:2013

(id 2187) Environmental testing - Part 2-78: Tests - Test Cab: Damp heat, steady state

# 4.4 - Damp heat, steady state

(id 1024) The object of this standard is to investigate the effect of high humidity at constant temperature without condensation on the specimen over a prescribed period.

Accredited [IEC17025]: YES Execution time [day]: 10 Unit price #1 [EUR]: 600.00 Limit #1: Up to 2 days Unit price #2 [EUR]: 1,000.00 Limit #2: Up to 10 days Unit price #3 [EUR]: 1,500.00 Limit #3: Up to 20 days

# 4.4 - Damp heat, steady state (big samples)

(id 1574) The object of this standard is to investigate the effect of high humidity at constant temperature without condensation on the specimen over a prescribed period.

Accredited [IEC17025]: YES Execution time [day]: 2 Unit price #1 [EUR]: 1,500.00 Limit #1: Up to 48 hours Unit price #2 [EUR]: 2,200.00 Limit #2: Up to 96 hours



# PRICE LIST / Listino prezzi § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



# EN 60079-11:2012

(id 81) Explosive atmospheres - Part 11: Equipment protection by intrinsic safety "i"

#### 10.1 - Spark ignition test

(id 122) Investigation on effect of opening and closing circuit in explosive atmosphere.

As prescribed by the strandard at §10.1.2, in order to perform this test, the instrument 4.162 Spark Ignition Testing Apparatus is used and it is calibrated before use.

Accredited [IEC17025]: YES

Unit price #1 [EUR]: 800.00

Limit #1: Up to n. 10 test points with same sample configurations and same gas mixture

Unit price #2 [EUR]: 1,000.00

Limit #2: Up to n. 10 test points with different sample configurations and same gas mixture

Unit price #3 [EUR]: 1,500.00

Limit #3: Up to n. 20 test points with different sample configurations and same gas mixture

#### 10.2 - Temperature tests

(id 123) ---

Accredited [IEC17025]: YES Execution time [day]: 1
Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

#### 10.3 - Dielectric strength tests

(id 124) ---

#### 10.6.1 - Casting compound

(id 125) Mechanical tests

Accredited [IEC17025]: YES Execution time [day]: 0
Unit price #1 [EUR]: 300.00 Limit #1: Applicable to any sample

# 10.9 - Cable pull test

(id 126) ---

Accredited [IEC17025]: YES Execution time [day]: 0
Unit price #1 [EUR]: 200.00 Limit #1: Applicable to any sample

#### 10.5.3 - Spark ignition and surface temperature of cells and batteries

(id 275) Determination of sparking short circuit current and temperature rise for cells and batteries.

# 10.4 - Determination of parameters of loosely specified components

(id 423) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 200.00 Limit #1: Single measurement
Unit price #2 [EUR]: 500.00 Limit #2: Up to 5 measurements on the same sample
Unit price #3 [EUR]: 800.00 Limit #3: Up to 10 measurements on the same sample

# 10.5.1 - General (tests for cells and batteries)

(id 424) The rechargeable battery is fully charged and then discharged twice in order to verify if its capacity is within the manufacturer's specification.

Accredited [IEC17025]: YES
Unit price #1 [EUR]: 150.00
Unit price #2 [EUR]: 300.00

Execution time [day]: 0
Limit #1: Only visual inspection and voltage measurements (usually primary batteries)
Limit #2: Visual inspection, voltage measurements and charge-dicharge cycles (secondary batteries)

# 10.7 - Tests for apparatus containing piezoelectric devices

(id 425) ---

Accredited [IEC17025]: YES Execution time [day]: 1
Unit price #1 [EUR]: 500.00 Limit #1: Applicable to any sample



# PRICE LIST / Listino prezzi

# § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

TNDEX

# 10.8 - Type tests for diode safety barriers and safety shunts

(id 426) These tests are used to demonstrate that the safety barrier or safety shunt can withstand the effects of transients.

Accredited [IEC17025]: YES Execution time [day]: 0 Unit price #1: ask to office Limit #1: Attenzione test non più esequibile

## 10.10 - Transformer tests

(id 427) Verification of the insulation and measurement of the temperature of the object with particular attention to the temperature developed by the windings of

Accredited [IEC17025]: YES Execution time [day]: 0 Unit price #1 [EUR]: 800.00 Limit #1: Small size Unit price #2 [EUR]: 1,200.00 Limit #2: Medium size Unit price #3 [EUR]: 1,600.00 Limit #3: Big size

## Annex H - Ignition testing of semi-conductors limiting power supply circuits

Accredited [IEC17025]: NO Execution time [day]: 30 Unit price #1: ask to office Limit #1: Applicable to any sample

# 10.12 - Current carrying capacity of infallible printed circuit board connections

Accredited [IEC17025]: YES Execution time [day]: 0 Unit price #1: ask to office Limit #1: Applicable to any sample

### 10.6.2 - Determination of the acceptability of fuses requiring encapsulation

Accredited [IEC17025]: YES Execution time [day]: 0 Limit #1: Applicable to any sample Unit price #1 [EUR]: 300.00

#### 10.6.3 - Partitions

(id 619) Mechanical tests

Accredited [IEC17025]: YES Execution time [day]: 0 Unit price #1 [EUR]: 300.00 Limit #1: Applicable to any sample

# 10.11.2.4.1 - Optical isolators - Carbonisation test on receiver side

(id 620) A carbonisation test shall be conducted to check the formation of an internal creepage path caused by the heated plastic material (carbonisation).

Accredited [IEC17025]: YES Execution time [day]: 0 Unit price #1 [EUR]: 200.00 Limit #1: Applicable to any sample

# 10.11.3.2 - Optical isolators - Pre-test dielectric

(id 628) Prior to the short-circuit current tests, the samples of the optical isolator shall be capable of withstanding a dielectric strength test without breakdown.

Accredited [IEC17025]: YES Execution time [day]: 0 Unit price #1 [EUR]: 500.00 Limit #1: Applicable to any sample

#### 10.5.4 - Battery container pressure test

(id 891) Determination of venting pressure of battery container

Accredited [IEC17025]: YES Execution time [day]: 0 Unit price #1 [EUR]: 300.00 Limit #1: up to 4 Ah Unit price #2 [EUR]: 500.00 Limit #2: from 4 to 25 Ah Unit price #3 [EUR]: 800.00 Limit #3: more than 25 Ah

# 7.4 - Battery construction

(id 915) Verification of constructional requirements and basic specification

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 200.00 Limit #1: Applicable to any sample Unit price #2 [EUR]: 200.00 Limit #2: Applicable to any sample

# 10.5.2 - Electrolyte leakage test for cells and batteries

(id 916) Ten test samples shall be subjected to the most onerous test condition but with no use of an external charging circuit which exceeds the charging rates recommended by the manufacturer of the cell or battery.

The samples shall be placed with any case discontinuities, for example seals, facing downward or in the orientation specified by the manufacturer of the device, over a piece of blotting paper for a period of at least 12 h after the test then subjected to a visual inspection.

Accredited [IEC17025]: YES Execution time [day]: 1 Unit price #1 [EUR]: 200.00 Limit #1: up to 4 Ah Unit price #2 [EUR]: 400.00 Limit #2: from 4 to 25 Ah Unit price #3 [EUR]: 800.00 Limit #3: more than 25 Ah



# **PRICE LIST / Listino prezzi**

# § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

TNDEX

#### 10.11.2.1 - Overload test at the receiver side

(id 1725) This test is intended to define the maximum temperature at the receiver side when the device experiences an overload.

Accredited [IEC17025]: YES Execution time [day]: 0 Unit price #1 [EUR]: 100.00 Limit #1: Not relevant

#### 10.11.2.2 - Overload test at the transmitter side

(id 1726) This test is intended to define the maximum temperature at the transmitter side when the device experiences an overload.

Accredited [IEC17025]: YES Execution time [day]: 0 Unit price #1 [EUR]: 100.00 Limit #1: Not relevant

#### 10.11.2.3 - Thermal conditioning and dielectric strength test

(id 1727) After overloading, the devices shall be subjected to thermal conditioning and dielectric strength tests.

Accredited [IEC17025]: YES Execution time [day]: 0 Unit price #1 [EUR]: 300.00 Limit #1: Applicable to any sample

## 10.11.2.4.2 - Optical isolators - Carbonisation test on transmitter side

(id 2343) A carbonisation test shall be conducted to check the formation of an internal creepage path caused by the heated plastic material (carbonisation).

Accredited [IEC17025]: YES Execution time [dav]: 0 Unit price #1 [EUR]: 200.00 Limit #1: Applicable to any sample

#### 10.11.3.3 - Optical isolators - Short-circuit current test

(id 2344) A test circuit shall be connected to the optical isolator so that the test current flows through the non-intrinsically safe side of the optical isolator. Protective components or assemblies that form part of the circuit are permitted to remain connected for the test.

Accredited [IEC17025]: YES Execution time [day]: 0 Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

## 10.11.3.4 - Optical isolators - Current limited short-circuit current test

(id 2345) This test applies when the optical isolators have protective series fuses or current-limiting resistors.

Accredited [IEC17025]: YES Execution time [day]: 2 Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

# 10.11.3.5 - Optical isolators - Dielectric strength test

(id 2346) The optical isolators shall withstand a dielectric strength test without breakdown.

Accredited [IEC17025]: YES Execution time [day]: 2 Unit price #1 [EUR]: 500.00 Limit #1: Applicable to any sample

# 7.3 - Cold resistance test for fuses

(id 2465) The cold resistance of the fuse at the minimum specified ambient temperature may be taken as an infallible resistance complying with 8.5 for current limiting purposes.

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1 [EUR]: 300.00 Limit #1: Only one parameter

# EN 60079-15:2010

(id 84) Explosive atmospheres - Part 15: Equipment protection by type of protection "n"

#### 22.3.1.1 - Thermal endurance to heat

(id 152) Tests for enclosures on which the type of protection depends. Thermal endurance tests

Accredited [IEC17025]: NO Execution time [day]: 30 Limit #1: Applicable to any sample Unit price #1 [EUR]: 1,600.00

# 6.4 - Clearances, creepage distances and separations

(id 153) ---

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 500.00 Limit #1: Applicable to any sample

## 22.4.3.2 - Tests for enclosed break devices

(id 429) ---

Accredited [IEC17025]: NO Execution time [day]: 30 Unit price #1 [EUR]: 800.00 Limit #1: Applicable to standard relays Unit price #2 [EUR]: 1,200.00 Limit #2: Applicable to standard equipment



# PRICE LIST / Listino prezzi

# § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX

# 22.5.2 - Voltage test for sealed devices

(id 430) Tests for sealed devices

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 300.00 Limit #1: Applicable to any sample

# 22.5.3 - Tests on devices with free space

(id 431) Tests for sealed devices

Accredited [IEC17025]: NO Execution time [day]: 30

Unit price #1 [EUR]: 300.00 Limit #1: Applicable to any sample

#### 22.5.4 - Test for sealed devices for luminaires

(id 432) ---

Accredited [IEC17025]: NO Execution time [day]: 30

Unit price #1: ask to office Limit #1: Applicable to any sample

## 22.6.2.1/22.6.2.2/22.6.2.3 - Type test requirements for restricted breathing enclosures

(id 433) Equipment where the nominal volume does not change due to pressure

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

#### 22.7 - Test for screw lampholders

(id 434) ---

Accredited [IEC17025]: NO Execution time [day]: 30

Unit price #1: ask to office Limit #1: Applicable to any sample

#### 22.8 - Test for starter holders for luminaires

(id 435) ---

Accredited [IEC17025]: NO Execution time [day]: 30

Unit price #1: ask to office Limit #1: Applicable to any sample

# 22.9.2 - Moisture resistance, insulation and electric strength

(id 436) ---

Accredited [IEC17025]: NO Execution time [day]: 30

Unit price #1: ask to office Limit #1: Applicable to any sample

# 22.9.3 - Cut-out device test

(id 437) ---

Accredited [IEC17025]: NO Execution time [day]: 30

Unit price #1: ask to office Limit #1: Applicable to any sample

#### 22.9.4 - Life test

(id 438) ---

Accredited [IEC17025]: NO Execution time [day]: 30

Unit price #1: ask to office Limit #1: Applicable to any sample

# 22.10 - Test for wiring of luminaires subject to high-voltage impulses from ignitors

(id 439) ---

Accredited [IEC17025]: NO Execution time [day]: 30

Unit price #1: ask to office Limit #1: Applicable to any sample

## 22.11 - Mechanical shock test for batteries

(id 440) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 300.00 Limit #1: Applicable to any sample

# 22.12 - Insulation resistance test for batteries

(id 441) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 400.00 Limit #1: Test voltage: 1000 V



# PRICE LIST / Listino prezzi § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



# 22.13 - Additional ignition tests for large or high-voltage machines

(id 442) ---

Accredited [IEC17025]: NO Execution time [day]: 30
Unit price #1 [EUR]: 2,000.00 Limit #1: Applicable to any sample

#### 6.5 - Electric strength

(id 614) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 300.00 Limit #1: Applicable to any sample

#### 22.5.1 - Conditioning for tests for sealed devices

(id 615) Tests for sealed devices

Accredited [IEC17025]: NO Execution time [day]: 10

Unit price #1 [EUR]: 500.00 Limit #1: Applicable to any sample

#### 22.3.1.2 - Drop test for hand-held equipment

(id 617) Clause 26.4.3 of IEC 60079-0 applies.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 100.00 Limit #1: Applicable to any sample

# 22.6.3 - Alternative type test for equipment where the nominal volume of the enclosure changes due to pressure

(id 919) Type tests for nR restricted breathing enclosures.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 1,000.00 Limit #1: Applicable to any sample

## 22.4.3.3 - Tests for enclosed non-incendive components

(id 1256) ---

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 1,200.00 Limit #1: Up to 2kW resistive load
Unit price #2 [EUR]: 1,500.00 Limit #2: Up to 2kW inductive load (motor)

# 11.2 - Construction requirements verification

(id 1377) --

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 300.00 Limit #1: Not relevant

# EN 60079-18:2015+A1:2017

(id 89) Explosive atmospheres - Part 18: Equipment protection by encapsulation "m"

# 8.2.2 - Maximum temperature

(id 110) ---

Accredited [IEC17025]: YES	Execution time [day]: 0
Unit price #1 [EUR]: 600.00	Limit #1: < 500 [dm3]
Unit price #2 [EUR]: 1,000.00	Limit #2: 500 < 3000 [dm3]
Unit price #3 [EUR]: 1,200.00	Limit #3: > 3000 [dm3]

### 8.2.3.1 - Thermal endurance to heat

(id 111) Evaluation of the thermal endurance to heat of non-metallic equipment or parts of equipment.

# 8.2.3.2 - Thermal endurance to cold

(id 112) Evaluation of the thermal endurance to cold of non-metallic parts of equipment.



# **PRICE LIST / Listino prezzi**

# § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

TNDEX

#### 8.2.4 - Dielectric strength test

(id 114) Test on the apparatus.

The test is carried out on the following arrangements of circuits as applicable:

- A. between galvanically isolated circuits;
- B. between each circuit and all earthed parts;
- C. between each circuit and the surface of the compound or the non-metallic enclosure that, if necessary, can be clad with a conductive foil.

Accredited [IEC17025]: YES Execution time [day]: 0 Unit price #1 [EUR]: 300.00 Limit #1: Up to 5kV dc Unit price #2 [EUR]: 1,200.00 Limit #2: From 5 to 10kV ac/dc Unit price #3 [EUR]: 2,400.00 Limit #3: From 10 to 60kV dc only

### 8.2.5 - Cable pull test

(id 115) This test has to be performed on two samples:

1) on a previously unstressed sample at 21 °C ± 2 °C.

2) on a further test sample previously conditioned according to 8.2.3.1 (thermal endurance to heat) at the maximum temperature at the cable entry point. The cable under test shall be tested with a force 20 times the value in millimetres of the diameter of the cable or 50 times the mass (in kilograms) of the "m" equipment, whichever is the lower value, for a minimum tensile force of 1N. This value can be reduced to 25 % of the required value in the case of permanent installations. The force shall be applied in the least favourable direction.

Accredited [IEC17025]: YES Execution time [day]: 0 Unit price #1 [EUR]: 200.00 Limit #1: Applicable to any sample

# 8.2.6 - Pressure test for Group I and Group II electrical equipment

(id 116) This test is required for level of protection "ma" equipment with any individual free spaces between 1 cm3 and 10 cm3 and level of protection "mb" equipment with any individual free spaces between 10 cm3 and 100 cm3.

Accredited [IEC17025]: YES Execution time [day]: 0 Unit price #1 [EUR]: 300.00 Limit #1: Applicable to any sample

### 5 - Requirements for compounds

(id 537) Evaluation of the properties of the the compound(s) on which the encapsulation "m" depends - Check of documentation

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 50.00 Limit #1: Applicable to any sample

# 7.3 - Free space in the encapsulation

(id 538) The sum of the free spaces and the thickness of the compound surrounding such free spaces shall meet the requirements of §7.3

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1: ask to office Limit #1: Applicable to any sample

## 8.2.7 - Test for resettable thermal protective device

(id 543) This test shall be performed after the thermal endurance test.

Accredited [IEC17025]: YES Execution time [day]: 0 Unit price #1 [EUR]: 400.00 Limit #1: Applicable to any sample

# 8.1.1 - Water absorption test

(id 544) Test on the compound

Accredited [IEC17025]: YES Execution time [day]: 2 Unit price #1 [EUR]: 200.00 Limit #1: Applicable to any sample

# 8.1.2 - Dielectric strength test

(id 545) Test on the compound

Accredited [IEC17025]: YES Execution time [day]: 0 Unit price #1 [EUR]: 300.00 Limit #1: Up to 5kV dc Unit price #2 [EUR]: 1,200.00 Limit #2: From 5 to 10kV ac/dc Unit price #3 [EUR]: 2,200.00 Limit #3: From 10 to 50kV ac/dc

#### 8.2.8 - Sealing test for build-in protective devices

(id 618) ---

Accredited [IEC17025]: YES Execution time [day]: 0 Unit price #1 [EUR]: 400.00 Limit #1: Applicable to any sample

# 7.4.1 - Thickness of the compound of "m" equipment

(id 1222) ---

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1: ask to office Limit #1: Not relevant



# § T - LABORATORY TESTS / Prove di laboratorio

**PRICE LIST / Listino prezzi** 

# alb**ar**ubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



# 7.4.2 - Thickness of the compound. Windings for electrical machines

(id 1223) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant

# 7.4.3 - Thickness of the compound. Rigid, multi-layer printed wiring boards with through connections

(id 1224) --

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant

# 7.5 - Switching contacts

(id 1225) Verification of the requirements specified at §7.5 about switching contacts and their additional enclosure.

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 100.00 Limit #1: Applicable to any sample

# 7.9.2 - Electrical protective devices

(id 1226) Verification of the requirements specified at § 7.9.2 about electrical protective devices

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant

#### 7.2.1 - Fault examination

(id 1227) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant

## 7.2.2 - Components considered as not subject to fail

(id 1228) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant

# 7.2.3 - Isolating components

(id 1229) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant

# 7.2.4.1 - Infallible separation distances. General

(id 1230) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant

#### 7.2.4.2 - Distances through the compound

(id 1231) ---

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1: ask to office Limit #1: Not relevant

# 7.2.4.3 - Distances through solid insulation

(id 1232) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant

## 7.9.3 - Thermal protective devices

(id 1233) Verification of the requirements specified at § 7.9.3 about thermal protective devices

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

# 7.9.4 - Built-in protective devices

(id 1234) Verification of the requirements specified at § 7.9.4 about built-in protective devices

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant



# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



#### 7.9.1 - Protective devices. General

(id 1235) Verification of general requirements of §7.9.1

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 100.00 Limit #1: Applicable to any sample

# 7.8.5 - Current limitation: maximum surface temperature with the highest discharge current

(id 1738) ---

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 300.00 Limit #1: Not relevant

# EN 60079-26:2015

(id 93) Explosive atmospheres - Part 26: Equipment with equipment protection level (EPL) Ga

#### 5.2 - Separation elements

(id 1080) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant

### 5.3 - Temperature evaluation

(id 1081) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant

# EN 60512-11-9:2002

(id 1284) Connectors for electronic equipment - Tests and measurements - Part 11-9: Climatic tests - Test 11i: Dry heat

# unref - Test 11i: Dry heat

(id 1796) This test is carried out in accordance with IEC 60068-2-2,

- Test Ba: Dry heat for non-heat-dissipating specimens with sudden change of temperature when no measurement is required during the test;

- Test Bb: Dry heat for non-heat-dissipating specimens with gradual change of temperature when measurements are required during the test; using the degree of severity specified in the detail specification.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 250.00 Limit #1: Up to 96 h (small oven)

# EN 60512-12-2:2006

(id 1279) Connectors for electronic equipment - Tests and measurements Part 12-2: Soldering tests - Test 12b: Solderability, wetting, soldering iron method

# unref - Test 12b: Solderability, wetting, soldering iron method

(id 1789) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

# EN 60512-13-2:2006

(id 1282) Connectors for electronic equipment - Tests and measurements Part 13-2: Mechanical operation tests - Test 13b: Insertion and withdrawal forces

# unref - Test 13b: Insertion and withdrawal forces

(id 1793) The object of this test is to measure the insertion and withdrawal forces of mating connectors, or of a connector with the mating insertion and/or withdrawal gauge(s) specified in the connector detail specification, without the effect of any locking, latching, sealing, engaging, separating or similar device.





QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



# EN 60512-13-5:2006

(id 1274) Connectors for electronic equipment - Tests and measurements Part 13-5: Mechanical operation tests - Test 13e: Polarizing and keying method

# unref - Test 13e: Polarizing and keying method

(id 1784) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

# EN 60512-2-2:2003

(id 1275) Connectors for electronic equipment - Tests and measurements - Part 2-2: Electrical continuity and contact resistance tests - Test 2b: Contact resistance - Specified test current method

# unref - Test 2b: Contact resistance - Specified test current method

(id 1785) The object of this test is to measure the electrical resistance across a pair of mated contacts or a contact with a measuring gauge.

The contact resistance is calculated as follows:

R=(|Vmf - Vmr|)/(|If| + |Ir|),

where R is the resistance;

Vmf is the measured forward voltage;

Vmr is the measured reverse voltage;

If is the forward current;

Ir is the reverse current.

Accredited [IEC17025]: NO
Unit price #1 [EUR]: 150.00

Execution time [day]: 0

Limit #1: Up to 3 contacts

# EN 60512-3-1:2002

(id 1276) Connectors for electronic equipment Tests and measurements - Part 3-1: Insulation tests Test 3a: Insulation resistance

# unref - Test 3a: Insulation resistance of electromechanical components

(id 1786) The object of this test is to assess the insulation resistance of electromechanical components.

The insulation resistance is measured on specimens using the specified test voltage applied in turn between each termination being tested and all others connected together and to the housing and/or the mounting plate.

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 150.00 Limit #1: Applicable to any sample

# EN 60512-4-1:2003

(id 1277) Connectors for electronic equipment - Tests and measurements - Part 4-1: Voltage stress tests - Test 4a: Voltage proof

## unref - Test 4a: Voltage proof - Method A

(id 1787) The object of this test is to determine the ability of a component to withstand specified test voltages applied in a specified manner.

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 150.00 Limit #1: Applicable to any sample

# EN 60512-7-2:2012

(id 1285) Connectors for electronic equipment - Tests and measurements - Part 7-2: Impact tests (free connectors) - Test 7b: Mechanical strength impact

# unref - Test 7b: Mechanical strength impact

(id 1797) The object of this test is to assess the ability of a free connector on the end of a cable or wire bundle to withstand impacts it could receive when dropped onto a hard surface.

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 150.00 Limit #1: Up to 8 cycles



# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



# EN 60512-9-1:2010

(id 1278) Connectors for electronic equipment - Tests and measurements - Part 9-1: Endurance tests - Test 9a: Mechanical operation

# unref - Test 9a: Mechanical operation

(id 1788) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

# EN 60999-1:2000

(id 1280) Connecting devices - Electrical copper conductors - Safety requirements for screw-type and screwless-type clamping units - Part 1: General requirements and particular requirements for clamping units for conductors from 0, 2 mm2 up to 35 mm2 (included)

#### - Screw terminal

(id 1791) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

# EN 61701:2012

(id 1582) Salt mist corrosion testing of photovoltaic (PV) modules

## unref - Salt Mist (crystalline PV) - Severity 3 to 6

(id 245) Determination of the resistance of the module to cyclic corrosion from salt mist. This test is useful for evaluating the compatibility of materials, and the quality and uniformity of protective coatings.

Severities 3 to 6 are defined as follows:

- Severity (3): one test cycle consisting of: four spray periods, each of 2h, with a humidity storage period between 20h and 22h after each; afterwards one storage period of three days under a standard atmosphere for testing at 23°C ± 2°C and 45% to 55% humidity:
- Severity (4): two test cycles as specified in Severity (3);
- Severity (5): four test cycles as specified in Severity (3);
- Severity (6): eight test cycles as specified in Severity (3)

Accredited [IEC17025]: NO	Execution time [day]: 32
Unit price #1 [EUR]: 4,500.00	Limit #1: Severity 4
Unit price #2 [EUR]: 6,000.00	Limit #2: Severity 5
Unit price #3 [EUR]: 8,000.00	Limit #3: Severity 6

# unref - Salt Mist (thin-film PV)

(id 402) Determination of the resistance of the module to cyclic corrosion from salt mist. This test is useful for evaluating the compatibility of materials, and the quality and uniformity of protective coatings

Accredited [IEC17025]: NO Execution time [day]: 70
Unit price #1 [EUR]: 8,000.00 Limit #1: Severity 1

### unref - Salt Mist (crystalline PV) - Severity 1

(id 1096) Determination of the resistance of the module to cyclic corrosion from salt mist. This test is useful for evaluating the compatibility of materials, and the quality and uniformity of protective coatings

# 4.2 - Bypass diode functionality test

(id 1099) The purpose of this test is to verify that the bypass diode(s) of the test

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 200.00 Limit #1: Not relevant



# PRICE LIST / Listino prezzi § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



# EN 81-71:2018+AC:2019

(id 1710) Safety rules for the construction and installation of lifts - Particular applications to passenger lifts and goods passenger lifts - Part 71: Vandal resistant lifts

#### Annex B - Impact test

(id 2153) Impact test with a pointed impactor with a mass of 1 Kg and 10 mm radius point.

```
Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 300.00 Limit #1: Applicable to any sample
```

#### Annex F - Fire tests

(id 2154) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 500.00 Limit #1: Applicable to any sample

# EN ISO 10497:2010

(id 1040) Testing of valves - Fire type-testing requirements

## 5 - Fire test

(id 1908) The purpose of this test is confirming the pressure-containing capability of a valve under pressure during and after the fire test.

Accredited [IEC17025]: NO	Execution time [day]: 0
Unit price #1 [EUR]: 2,500.00	Limit #1: Up to 3"
Unit price #2 [EUR]: 4,000.00	Limit #2: From 4" to 10"
Unit price #3 [EUR]: 6,500.00	Limit #3: From 12" to 20"

## 6.2 - Through seat leakage during burn period

(id 1909) ---

Accredited [IEC17025]: NO	Execution time [day]: 0
Unit price #1 [EUR]: 150.00	Limit #1: Up to 3"
Unit price #2 [EUR]: 250.00	Limit #2: From 4" to 10"
Unit price #3 [EUR]: 500.00	Limit #3: From 12" to 20"

# 6.3 - External leakage during burn and cool-down periods

(id 1910) ---

Accredited [IEC17025]	: NO Execution	time [day	]: 0
Unit price #1 [EUR]:	200.00	Limit #1:	Up to 3"
Unit price #2 [EUR]:	400.00	Limit #2:	From 4" to 10"
Unit price #3 [FUR]:	800 00	T.imit #3:	From 12" to 20"

# 6.4 - Low Pressure Test Through-seat Leakage after Cool-down

(id 1911) ---

Accredited [IEC17025]	: NO	Execution	time [	day	]: 0				
Unit price #1 [EUR]:	150.00		Limit	#1:	Up to	3 "			
Unit price #2 [EUR]:	250.00		Limit	#2:	From	4"	to	10	"
Unit price #3 [EUR]:	500.00		Limit	#3:	From	12	to	20	"

# 6.5 - Operability

(id 1912) ---

Accredited [IEC17025]: NO	Execution time [day]: 0
Unit price #1 [EUR]: 400.00	Limit #1: Up to 3"
Unit price #2 [EUR]: 800.00	Limit #2: From 4" to 10"
Unit price #3 [EUR]: 1.200 00	Limit #3: From 12" to 20"

# 6.6 - External Leakage Following Operational Test

(id 1913) ---

Accredited [IEC17025	]: NO	Execution	time [	day	]: 0			
Unit price #1 [EUR]:	200.00		Limit	#1:	Up to	3 "		
Unit price #2 [EUR]:	400.00		Limit	#2:	From	4 "	to :	10"
Unit price #3 [EUR]:	800.00		Limit	#3:	From	12"	to	20"



# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



# **EN ISO 11117:2019**

(id 1839) Gas cylinders - Valve protection caps and guards - Design, construction and tests

# 7.7 - Drop test

(id 2467) The valve protection device shall be tested to prove that it functions in such a way that it protects the valve. The test valve shall maintain its operability.

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 1,800.00 Limit #1: Applicable to any sample

# EN ISO 9806:2013

(id 55) Solar energy - Solar thermal collectors - Test methods

# 7 - Leakage test

(id 853) Closed loop air heating collectors only

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 1,000.00 Limit #1: --

#### 8 - Rupture or collapse test

(id 854) Air heating collectors only

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 800.00 Limit #1: --

#### 9 - High-temperature resistance test

(id 855) Capability of the collector to withstand high irradiance levels without failures.

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 1,000.00 Limit #1: Applicable to any sample

# 10 - Standard stagnation temperature of liquid heating collectors

(id 856) ---

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 500.00 Limit #1: --

# 11 - Exposure and pre-exposure test

(id 857) Low-cost reliability test sequence

Accredited [IEC17025]: NO Execution time [day]: 30 Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

# 12 - External thermal shock test

(id 858) Collectors may from time to time be exposed to sudden rainstorms on hot sunny days, causing a severe external thermal shock. This test is intended to assess the capability of a collector to withstand such thermal shocks without a failure.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

## 13 - Internal thermal shock test

(id 859) Collectors may be exposed to a sudden intake of cold fluid on hot sunny days, causing an internal thermal shock after a period of shutdown, when the installation is brought back into operation while the collector is at its stagnation temperature. This test is to assess the capability of a collector to withstand such shocks without failure.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

## 14 - Rain penetration test

(id 860) This test is applicable only for glazed collectors and is intended to assess the extent to which glazed collectors are substantially resistant to rain penetration.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 1,000.00 Limit #1: Applicable to any sample

## 15 - Freeze resistance test

(id 861) This test is intended to assess the extent to which water heating collectors which are claimed to be freeze resistant can withstand freezing, and freeze/thaw cycling. This test is not intended for use with collectors for which it is clearly stated in the installation manual that they may only be used with an antiferent fluid.

Accredited [IEC17025]: NO Execution time [day]: 4
Unit price #1 [EUR]: 1,200.00 Limit #1: Applicable to any sample

page 64/234



# § T - LABORATORY TESTS / Prove di laboratorio

PRICE LIST / Listino prezzi

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX

# 16 - Mechanical load test with positive or negative pressure

(id 862) --
Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 800.00 Limit #1: --

#### 17 - Impact resistance test

(id 863) Verification that the sample is capable of withstanding the impact of hailstones

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 600.00 Limit #1: Applicable to any sample

#### 18 - Solar collectors final visual inspection

(id 864) Final checklist for collectors examination after all the tests.

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 600.00 Limit #1: --

# 20.28 - Performance testing of fluid heating collectors

(id 865) ---

Accredited [IEC17025]: NO Execution time [day]: 5 Unit price #1 [EUR]: 2,500.00 Limit #1: --

#### 6 - Internal pressure tests for fluid channels

(id 870) Capability of the collector to be pressurized to nominal value.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 500.00 Limit #1: Applicable to any sample

# EPA 40 CFR 1051.501:2025

(id 2222) This section describes test procedures that you use to determine whether vehicles meet the emission standards of this part.

# c - Soaking

(id 3107) ---

Accredited [IEC17025]: YES Execution time [day]: 28
Unit price #1 [EUR]: 2,800.00 Limit #1: Applicable to any sample

# c - Permeation testing

(id 3108) ---

Accredited [IEC17025]: NO Execution time [day]: 14
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

# EPA 40 CFR 1051.515:2023

(id 1993) Title 40 —Protection of Environment

Chapter I — Environmental Protection Agency

Subchapter U —Air Pollution Controls

Part 1051 —Control of Emissions from Recreational Engines and Vehicles

Subpart F —Test Procedures

# a - Preconditioning fuel soak

(id 2753) ---

# b - Permeation test run

(id 2754) --

Accredited [IEC17025]: NO Execution time [day]: 14
Unit price #1 [EUR]: 1,000.00 Limit #1: Applicable to any sample



# PRICE LIST / Listino prezzi § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX

# d1 - Pressure cycling

(id 2755) ---

Accredited [IEC17025]: NO Execution time [day]: 7

Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

# d2 - UV exposure

(id 2756) ---

Accredited [IEC17025]: NO Execution time [day]: 19

Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

# d3 - Slosh testing

(id 2757) ---

Accredited [IEC17025]: NO Execution time [day]: 46

Unit price #1 [EUR]: 2,500.00 Limit #1: Applicable to any sample

# EPA 40 CFR 1060.515:2023

(id 2171) How do I test EPA Nonroad Fuel Lines and EPA Cold-Weather Fuel Lines for permeation emissions?

### Unref. - Soaking

(id 3011) ---

Accredited [IEC17025]: NO Execution time [day]: 28

Unit price #1 [EUR]: 2,000.00 Limit #1: Applicable to any sample

# ETAG 004:2013

(id 2130) EXTERNAL THERMAL INSULATION COMPOSITE SYSTEMS (ETICS) WITH RENDERING

# 5.1.3.1 - Water absorption (capillarity test)

(id 2952) These tests have 3 purposes, to determine:

- the water absorption, in order to assess, in Chapter 6, whether it is acceptable,
- which finishing coats should be applied on the rig to be subjected to hygrothermal testing (5.1.3.2.1),
- whether the freeze-thaw testing described in 5.1.3.2.2 is necessary.

Accredited [IEC17025]: NO Execution time [day]: 15

Unit price #1 [EUR]: 2,400.00 Limit #1: Without pre-conditioning
Unit price #2 [EUR]: 3,200.00 Limit #2: With pre-conditioning

# 5.1.3.2.1 - Hygrothermal Behaviour

(id 2953) ---

Accredited [IEC17025]: NO Execution time [day]: 55

Unit price #1 [EUR]: 3,500.00 Limit #1: Without pre-conditioning
Unit price #2 [EUR]: 6,500.00 Limit #2: With pre-conditioning

#### 5.1.3.2.2 - Freeze-thaw behaviour

(id 2954) The freeze-thaw test shall be carried out as determined by the analysis of the capillarity test (§ 5.1.3.1), i.e. shall be conducted except if the water absorption after 24 hours of both the reinforced base coat and the rendering system determined with each type of finishing coat is less than 0,5 kg/m².

Accredited [IEC17025]: NO Execution time [day]: 58
Unit price #1 [EUR]: 4,500.00 Limit #1: Without pre-conditioning
Unit price #2 [EUR]: 7,500.00 Limit #2: With pre-contioning

# 5.1.7.1.2 - Finishing coat not tested on the rig

(id 2955) The test is performed on an insulation panel faced with the rendering system applied in accordance with the manufacturer's instructions

Accredited [IEC17025]: NO Execution time [day]: 42

Unit price #1 [EUR]: 1,800.00 Limit #1: Without pre-conditioning
Unit price #2 [EUR]: 4,800.00 Limit #2: With pre-conditioning



# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



# **EU Regulation 2018/1832**

(id 1977) Commission Regulation (EU) 2018/1832 of 5 November 2018 amending Directive 2007/46/EC of the European Parliament and of the Council, Commission Regulation (EC) No 692/2008 and Commission Regulation (EU) 2017/1151 for the purpose of improving the emission type approval tests and procedures for light passenger and commercial vehicles, including those for in-service conformity and real-driving emissions and introducing devices for monitoring the consumption of fuel and electric energy.

# Annex VI - 5.1.1 - Ageing through exposure to temperature cycling

(id 2710) ---

Accredited [IEC17025]: NO Execution time [day]: 8
Unit price #1 [EUR]: 2,200.00 Limit #1: Applicable to any sample

# Annex VI - 5.1.2 - Ageing through exposure to vibration

(id 2711) ---

Accredited [IEC17025]: NO Execution time [day]: 1 Subcontract: BPS srl Unit price #1 [EUR]: 1,500.00 Limit #1: Applicable to any sample

#### Annex VI - 5.1.3 - Ageing through exposure to fuel vapour and determining BWC300

(id 2712) Ageing shall consist of repeatedly loading with fuel vapour and purging with laboratory air.

Accredited [IEC17025]: NO Execution time [day]: 30
Unit price #1 [EUR]: 3,500.00 Limit #1: Applicable to any sample

# EVS EN 15267-3:2008

(id 1753) Air quality - Certification of automated

measuring systems - Part 3: Performance specifications and test procedures for automated measuring systems for monitoring emissions from stationary sources

### 10.3 - Security

(id 2265) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

# 10.4 - Output ranges and zero point

(id 2266) The test laboratory shall check whether the output ranges on the AMS can be adjusted and whether such ranges are appropriate for the intended applications.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 10.5 - Additional data outputs

(id 2267) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 10.6 - Display of operational status signals

(id 2268) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

# 10.7 - Prevention or compensation for optical contamination

(id 2269) For optical techniques, the test laboratory shall assess whether contamination of the optical boundary surfaces interferes with the measuring technique.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

# 10.8 - Degrees of protection provided by enclosures

(id 2270) The effect of liquid water on the AMS shall be assessed by inspection in relation to EN 60529.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant



# PRICE LIST / Listino prezzi

# your value certified § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

TNDEX

#### 10.9 - Response time

(id 2271) The test laboratory shall determine the AMS response time using zero and span reference materials.

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

# 10.10 - Repeatability standard deviation at zero point

(id 2272) The repeatability standard deviation at zero point shall be determined by application of a reference material at the zero point.

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

# 10.11 - Repeatability standard deviation at span point

(id 2273) The repeatability standard deviation at span point shall be determined by application of a reference material at the span point.

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1 [EUR]: 1.000.00 Limit #1: Not relevant

#### 10.12 - Lack of fit

(id 2274) The test laboratory shall perform the lack of fit test according to Annex C. The linearity of the response of the AMS shall be checked using at least seven different reference materials, including a zero concentration.

Accredited [IEC17025]: NO Execution time [day]: 2 Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 10.14 - Influence of ambient temperature

(id 2275) The test laboratory shall determine how the zero and span values of the AMS are influenced by changes in ambient temperature by using a climatic chamber, which can control ambient temperature from -20 °C to +50 °C, within limits of ±1,0 K.

Accredited [IEC17025]: NO Execution time [day]: 2 Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

# 10.15 - Influence of sample gas pressure

(id 2276) The test laboratory shall determine the influence of variations in sample gas pressure on the response of the AMS.

Accredited [IEC17025]: NO Execution time [day]: 0 Limit #1: Not relevant Unit price #1 [EUR]: 1,000.00

# 10.16 - Influence of the sample gas flow for extractive AMS

(id 2277) The AMS shall initially be operated with the flow rate prescribed by the manufacturer. This flow rate shall then be changed to the lowest flow rate specified by the manufacturer.

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

# 10.17 - Influence of voltage variations

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 10.18 - Influence of vibration

(id 2279) ---

Accredited [IEC17025]: NO Execution time [day]: 1 Subcontract: GESTLABS S.R.L. Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

### 10.19 - Cross-sensitivity

(id 2280) The influence of potentially interfering substances also present in the waste gas shall be determined by admitting test gas mixtures to the input of the complete AMS.

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

### 10.20 - Excursion of measurement beam of cross-stack in-situ AMS

(id 2281) The test laboratory shall gradually and precisely deflect the transmitter and receiver assemblies of the AMS in the horizontal and vertical planes, and then record the measured signals using reference materials.

Accredited [TEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant



# PRICE LIST / Listino prezzi § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



### 10.22 - Response factors

(id 2282) The response factors (weighting factors) for total organic carbon measuring AMS shall be evaluated using defined test gas concentrations admitted to the AMS from test gas containers or by evaporating produced mixtures.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 12.1 - Calibration function (Field test)

(id 2283) The test laboratory shall determine the calibration function for the individual measured components of the AMS during the field test, by performing parallel measurements with the SRM.

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 12.2 - Response time (Field tests)

(id 2284) The response time shall be evaluated by admitting reference material at the zero and span points to the input of the complete AMS.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 12.3 - Lack of fit (Field test)

(id 2285) ---

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

## 12.4 - Maintenance interval (Field test)

(id 2286) The test laboratory shall determine the maintenance work that is necessary for the AMS to work properly as well as the intervals at which such maintenance work shall be performed.

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

## 12.5 - Zero and span drift (Field test)

(id 2287) The test shall be performed using two AMS of identical design as part of the field test in the form of paired measurements in the smallest measuring range tested.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 12.6 - Availability (Field test)

(id 2288) The test laboratory shall determine the availability of the AMS by recording the duration of the field test and all interruptions to the normal monitoring functionality of the AMS.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 12.7 - Reproducibility (Field tests)

(id 2289) Reproducibility shall be determined during the three month field test from simultaneous, continuous measurements by means of two identical AMS at the same measurement point (paired measurements) and an electronic data recording system with a memory capacity of at least four weeks and a sampling rate of at least four times during the AMS averaging period.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

# 12.8 - Contamination check of in-situ systems

(id 2290) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

# EVS EN 15267-4:2017

(id 1754) Air quality - Certification of automated measuring systems - Part 4: Performance criteria and test procedures for automated measuring systems for periodic measurements of emissions from stationary sources

#### 10.3 - Output ranges and zero point

(id 2246) The test laboratory shall check whether the output ranges on the P-AMS can be adjusted and whether such ranges are appropriate for the intended applications.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

page 69/234



# PRICE LIST / Listino prezzi

# § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

TNDEX



#### 10.4 - Display of operational status signals

(id 2247) The test laboratory shall assess whether the P-AMS has a means of displaying and provide data for recording the relevant operational status (e.g. standby, service, malfunction).

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

## 10.5 - Degrees of protection provided by enclosures

(id 2248) The effect of liquid water on the P-AMS shall be assessed by inspection in relation to EN 60529.

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 10.6 - Response time

(id 2249) The test laboratory shall determine the P-AMS response time using zero and span reference materials.

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

# 10.7 - Repeatability standard deviation at zero point

(id 2250) The repeatability standard deviation at zero point shall be determined by application of a reference material at the zero point.

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

# 10.8 - Repeatability standard deviation at span point

(id 2251) The repeatability standard deviation at span point shall be determined by application of a reference material at the span point.

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 10.9 - Lack of fit

(id 2252) The test laboratory shall perform the lack of fit test according to Annex C at the beginning of the laboratory test. The linearity of the response of the P-AMS shall be checked using at least seven different reference materials, including a zero concentration.

Accredited [IEC17025]: NO Execution time [day]: 2 Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

# 10.10 - Short-term zero and span drift

(id 2253) ---

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 10.11 - Set-up time after transportation and influence of ambient temperature

(id 2254) The test of setting the P-AMS into operation and the test of the influence of a slow change in ambient temperature shall be carried out with the P-AMS placed in a climatic chamber. Only one test cycle is necessary.

Accredited [IEC17025]: NO Execution time [day]: 2 Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 10.12 - Influence of voltage variations

(id 2255) The supply voltage to the P-AMS shall be varied, using an isolating transformer, in steps of 5 % from the nominal supply voltage to at least the upper and the lower limits specified in Clause 8.

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

## 10.13 - Influence of vibration

(id 2256) The P-AMS shall be examined in the laboratory and in the field in respect to whether normal vibrations and transportation affect the performance of the P-AMS.

Accredited [IEC17025]: NO Execution time [day]: 1 Subcontract: GESTLABS S.R.L. Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

## 10.14 - Influence of sample gas pressure

(id 2257) The test laboratory shall determine the influence of variations in sample gas pressure on the response of the P-AMS.

Accredited [TEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant



# **PRICE LIST / Listino prezzi**

# § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

TNDEX



# 10.15 - Influence of the sample gas flow for extractive P-AMS

(id 2258) The P-AMS shall initially be operated with the flow rate prescribed by the manufacturer. This flow rate shall then be changed to the lowest flow rate specified by the manufacturer.

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 10.16 - Cross-sensitivity

(id 2259) The influence of potentially interfering substances also present in the stack gas shall be determined by admitting test gas mixtures to the input of the complete P-AMS.

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 10.18 - Response factors

(id 2260) The response factors (weighting factors) for total organic carbon measuring P-AMS shall be evaluated using defined test gas concentrations admitted to the P-AMS from test gas containers or by evaporating produced mixtures.

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

# 12.1 - Response time (Field tests)

(id 2261) The response time shall be evaluated by admitting reference material at the zero and span points to the input of the complete P-AMS.

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

### 12.2 - Short-term zero and span drift (Field tests)

Execution time [day]: 1 Accredited [IEC17025]: NO Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

## 12.3 - Reproducibility (Field tests)

(id 2263) Reproducibility shall be determined during field tests from simultaneous measurements by means of two identical P-AMS at the same measurement point (paired measurements).

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

# FCA 9.02135/03 10-MAR-2006

(id 995) Vacuum brake piping: Carryover/Frozen

# 2.4.10 - Prolonged thermal ageing

(id 1526) ---

Accredited [IEC17025]: NO Execution time [day]: 8 Unit price #1 [EUR]: 700.00 Limit #1: Not relevant

# 2.4.14.1 - Resistance to hot vacuum

(id 1527) ---

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 500.00 Limit #1: Not relevant

## 2.4.6 - Cold impact strength

(id 1528) ---

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1 [EUR]: 200.00 Limit #1: Standard temperature (-40°C)

# FCA 9.02136/03 12-APR-2016

(id 944) Plastic pipes for engine coolant lines

# 2.5.1 - Extended thermal ageing in coolant

(id 1385) ---

Accredited [IEC17025]: NO Execution time [day]: 2 Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

page 71/234



# PRICE LIST / Listino prezzi § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX

# - 🛮 🖠

### 2.5.2 - Thermal cycle resistance

(id 1386) ---

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 2,500.00 Limit #1: Not relevant

#### 2.5.8 - Pulsating pressure

(id 1387) --

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 1,500.00 Limit #1: Not relevant

# FCA 9.02159/01:2014 Rev. B

(id 1889) Plastic fuel tanks (1D) - FROZEN

#### 2.8.5 - Verifica dei volumi e della capacità

(id 2526) ---

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 400.00 Limit #1: Applicable to any sample

#### 2.8.7 - Verifica curva di livello

(id 2527) Check that the measured level of fluid inside the tank corresponds to that indicated by technical drawings when the tank is positioned at 0° on level ground.

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

## 2.7 - Determinazione della quantità e delle dimensioni degli sfridi interni

(id 2536) ---

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 2.8.1 - Helium test

(id 2537) Il complessivo serbatoio carburante deve superare la prova di tenuta in Elio utilizzando un sistema hard-vacuum. Il valore di perdita minima ammessa per il complessivo serbatoio carburante deve essere correlato con un valore di minima perdita ammessa che ne assicuri la conformità per i valori di emissione e non comprometta la prestazione di permeazione.

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 200.00 Limit #1: Applicable to any sample

# 2.8.3 - Tenuta a depressione

(id 2538) Suction test at -20mbar on tanks.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 500.00 Limit #1: Applicable to any sample

#### 2.10.1 - Resistenza alla sovrappressione

(id 2558) This test is intended to verify the overpressure resistance of a tank.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 300.00 Limit #1: Applicable to any sample

### 2.8.2 - Prova di tenuta pneumatica in vasca d'acqua

(id 2595) ---

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 200.00 Limit #1: Applicable to any sample

# 2.10.5 - Impact strength on green parts

(id 2878) ---

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,200.00 Limit #1: Applicable to any sample



## PRICE LIST / Listino prezzi

### § T - LABORATORY TESTS / Prove di laboratorio

## albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



#### FCA CS.00056 10-DEC-2018

(id 1489) FCA Environmental specification for electrical/Electronic (E/E) components

#### 5.5.1 - Dust intrusion - IP6K

(id 1941) The dust intrusion shall apply to electronic components when the dust intrusion during vehicle operation can degrade the component or its function

#### 5.5.3 - Water or steam intrusion - IPX9K

(id 1942) The purpose is to determine if the componenet housing provides sufficient protection against water intrusion. A substance like UV dye or fluorescine shall be added to the water for assessing water ingress.

#### 5.5.5 - Salt fog

(id 1943) This test simulates costal conditions where fine droplets of salt water are suspended in air.

Accredited	[IEC17025]: NO	Execution time [day]: 7
Unit price	#1 [EUR]: 600.00	Limit #1: 24h
Unit price	#2 [EUR]: 1,400.00	Limit #2: 96h
Unit price	#3 [EUR]: 2,200.00	Limit #3: 168h

#### 5.6.2 - Chemical exposure

(id 1944) The fluids selected for the test shall be fluids that the component may encounter during the vehicle service life.

```
Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample
```

#### 5.3.8 - Solar radiation soak according to DIN 75220:2012

(id 1945) ---

```
Accredited [IEC17025]: NO Execution time [day]: 50 Unit price #1 [EUR]: 1,800.00 Limit #1: 7 days
```

#### 5.3.4 - Power thermal cycle endurance (PTCE)

(id 1946) ---

```
Accredited [IEC17025]: NO Execution time [day]: 0 Unit price \#1 [EUR]: 8,000.00 Limit \#1: Applicable to any sample
```

#### 5.1.3.7 - Functional checks - Before and after test

(id 1948) ---

```
Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1: ask to office Limit #1: Not relevant
```

#### 5.6.1 - Mixed Flowing Gas

(id 2516) This is an accelerated test to simulate the effects of atmospheric pollutants, corrosive gases on electronics. To accelerate the test, the gases are much more concentrated than seen in field conditions.

```
Accredited [IEC17025]: NO Execution time [day]: 14
Unit price #1 [EUR]: 3,200.00 Limit #1: Applicable to any sample
```

#### 5.5.2 - Mud Resistance

(id 2647) This test is to verify that the component ability to dissipate heat is not affected so much that it impairs the functionality when covered by mud. This test will verify that the component can still function and dissipate heat while covered by mud.

```
Accredited [IEC17025]: NO Execution time [day]: 3
Unit price #1 [EUR]: 3,300.00 Limit #1: Not relevant
```

#### 5.5.4 - Salt Water Immersion

(id 2810) This test simulates the immersion of components when a vehicle passes through or stops in water puddles. Leaks can occur across seals or through connector pins.

A substance like UV dye or fluorescine shall be added to the water for assessing water ingress.

```
Accredited [IEC17025]: NO Execution time [day]: 3
Unit price #1 [EUR]: 2,400.00 Limit #1: Applicable to any sample
```



### PRICE LIST / Listino prezzi § T - LABORATORY TESTS / Prove di laboratorio



QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

TNDEX



#### FCA LP.7A009:2024

(id 2203) FUEL SYSTEM - FUEL GAUGE ACCURACY TEST BENCH

#### 12.1.1 - Tank Volume vs. Resistance Test on level ground

(id 3084) ---

Accredited [IEC17025]: NO Unit price #1 [EUR]: 200.00 Execution time [day]: 1

Limit #1: Applicable to any sample

#### 12.2.1 - Resistance signal of tank unit with tank inclined on uphill-downhill and inclined right-left

(id 3085) ---

Accredited [IEC17025]: NO Unit price #1 [EUR]: 400.00 Execution time [day]: 1

Limit #1: Applicable to any sample

#### FCA LP.7A010:2014

(id 1862) Fuel System Refueling - Bench or Vehicle Level Test

#### 12.1.2 - Overfilling evaluation

(id 2554) This is a fuel fill performance test. It includes the following three procedures: Top Off, Trickle Fill and Round Up.

Accredited [IEC17025]: NO Execution time [day]: 8

Unit price #1 [EUR]: 850.00 Limit #1: Applicable to any sample

#### FCA LP.7A010:2022

(id 1999) Fuel System Refueling - Bench or Vehicle Level Test

#### 11.1.2 - Rated Capacity

(id 2765) ---

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 400.00 Limit #1: Applicable to any sample

#### FCA LP.7A010:2024

(id 2050) FUEL SYSTEM REFUELING BENCH OR VEHICLE LEVEL TEST - (STELLANTIS HARMONIZED)

#### 12.1.2 - Rated capacity

(id 2803) ---

Accredited [IEC17025]: NO

Execution time [day]: 1 Unit price #1 [EUR]: 400.00

Limit #1: Applicable to any sample, without 9.4 pre-conditioning, with one type of nozzle

#### 12.1.4 - Filling on grade

(id 2920) ---

Accredited [IEC17025]: NO Execution time [day]: 2

Unit price #1 [EUR]: 600.00 Limit #1: Without preconditioning, one type of nozzle

#### 12.1.3.3 - Overfilling evaluation - Trickle fill

(id 2921) ---

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 200.00 Limit #1: Applicable to any sample, without 9.4 pre-conditioning, with one type of nozzle

#### 12.2.5 - Interrupted fill

(id 2922) ---

Accredited [IEC17025]: NO Execution time [day]: 1

Limit #1: Applicable to any sample, without 9.4 pre-conditioning, with one type of nozzle Unit price #1 [EUR]: 400.00



# PRICE LIST / Listino prezzi § T - LABORATORY TESTS / Prove di laboratorio

## alb**ar**ubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX

### .

#### 12.2.6 - Nozzle orientation

Unit price #2 [EUR]: 800.00

(id 2923) ---

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 400.00 Limit #1: Applicable to any sample, without 9.4 pre-conditioning, with one type of nozzle

one degree position

Limit #2: Applicable to any sample, standard execution

#### 12.2.10 - Flow Rate Robustness Test

(id 2924) ---

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 400.00 Limit #1: Applicable to any sample, without 9.4 pre-conditioning, with one type of nozzle

one flow rate

#### 12.1.3.1 - Overfilling evaluation - Round Up

(id 3089) ---

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 200.00 Limit #1: Applicable to any sample, without 9.4 pre-conditioning, with one type of nozzle

#### 12.1.3.2 - Overfilling evaluation - Top Off

(id 3090) ---

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 200.00 Limit #1: Applicable to any sample, without 9.4 pre-conditioning, with one type of nozzle

#### FCA MS.90089 08-Apr-2015

(id 1415) Requisiti prestazionali rivestimenti per componenti sotto-cofano e telaio verniciati

#### Tabella 6 - Spessore ED - Pannello DV/PV - Particolare

(id 1876) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 100.00 Limit #1: Not relevant

#### Tabella 6 - Adesione iniziale Quadrettatore (Cross Hatch)

(id 1877) ---

Accredited [IEC17025]: NO

Execution time [day]: 1

Unit price #1 [EUR]: 300.00 Limit #1: Not relevant

#### Tabella 6 - DI Immersione in acqua

(id 1878) ---

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 200.00 Limit #1: Not relevant

#### Tabella 6 - Adesione all'umido

(id 1879) ---

Accredited [IEC17025]: NO Execution time [day]: 25
Unit price #1 [EUR]: 1,300.00 Limit #1: Not relevant

#### Tabella 6 - Resistenza alla ghiaia dopo esposizione all'umidità

(id 1880) ---

Accredited [IEC17025]: NO Execution time [day]: 5
Unit price #1 [EUR]: 5,500.00 Limit #1: Not relevant

#### FCA PF.90064 09-MAY-2018

(id 1492) Liquid vapor separator drain valve-fuel

#### 6.2.4 - Port strength

(id 1950) The LVSDV (Liquid Vapor Separator Drain Valve) port, along with the ports located on the valve to tank internal attachment, shall be capable of withstandning all assembly forces, durability, and FMVSS impact forces, while remaining leak free through the useful life of the vehicle. Adequate port strength is required to ensure the port/connection integrity during manufacturing, shipping, assembly, and in use durability performance of the design. A port strength load test confirms the strength of the port when subjected to external forces.

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant

page 75/234



### **PRICE LIST / Listino prezzi** your value certified § T - LABORATORY TESTS / Prove di laboratorio

### albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

TNDEX



#### 5.2 - Resistance to fluids/Vapors

(id 1951) This test demonstrates the performance of the LVSDV (Liquid Vapor Separator Drain Valve) when exposed to expected fuels. The LVSDV must withstand exposure to fluids present in the operating environment of the fuel tank for the useful life of the vehicle, 15 years/150.000 miles.

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1: ask to office Limit #1: Not relevant

#### 6.2.5 - Drop impact test

(id 1952) The LVSDV (Liquid Vapor Separator Drain Valve) must be capable of funcitioning properly after a drop impact from a specified height to insure the robustness of the design to normal handling conditions. Prior to the start of the test, all orientations that will be subject to the dropimpact test must be approved by Fuel Systems Engineering. This test confirms if the LVSDV is boviously damaged during handling, or if not obviously damaged, will still function when handled manually during assembly facilities or during fuel system servicing situations at dealership locations.

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1: ask to office Limit #1: Not relevant

#### 6.2.7 - Low speed impact test

(id 1953) The LVSDV (Liquid vapor separator drain valve-fuel) assembly internal components to the fuel tank assembly shall meet functional requirements following low speed internal load conditions without damage, disassembly, or loss of function. A low speed impact test confirms the internal components of the LVSDV maintain the ability to function after a low speed internal load condition.

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1: ask to office Limit #1: Not relevant

#### 6.2.10 - Valve to Upper housing Pull Out Force Test

(id 1954) The retention between the upper housing and lower housing shall withstand a high force load condition tested to failure.

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1: ask to office Limit #1: Not relevant

(id 1955) The LVSDV (Liquid Vapor Separator Drain Valve) must meet flow requirements to prevent vapor pressure build-up in the fuel tank. The LVSDV must provide sufficient refueling vapor flow such that the fuel system tank pressure is minimized to avoid premature shut off, well back, and/or spit back during a reueling scenario.

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 700.00 Limit #1: Applicable to any sample

#### FCA PF.90068:2020

(id 1768) DIESEL EXHAUST FLUID - HEATED SUPPLY LINES

#### 5.5.2 - Calcium Chloride Resistance

(id 2340) ---

Accredited [IEC17025]: NO Execution time [day]: 9 Unit price #1 [EUR]: 1,800.00 Limit #1: Applicable to any sample

#### 5.5.3 - Zinc Chloride Resistance

(id 2341) ---

Accredited [IEC17025]: NO Execution time [day]: 7 Unit price #1 [EUR]: 1,500.00 Limit #1: Applicable to any sample

#### 5.7.1 - Thermal cycle

(id 2779) ---

Accredited [IEC17025]: NO Execution time [day]: 65 Unit price #1 [EUR]: 18,500.00 Limit #1: Applicable to any sample

#### FCA PF.90068:2025

(id 2196) DIESEL EXHAUST FLUID - HEATED SUPPLY LINES (STELLANTIS HARMONIZED)

#### 9.3 - Heater Durability Cycle

(id 3082) The heated DEF supply line assembly shall be able to remain functional and withstand thermal stresses while in vehicle position, during all vehicle conditions, thaw cycles and environments for the intended life of the vehicle. The DEF supply line should also be robust against reduced and intermittent functionality of the purge function of the DEF supply module.



# PRICE LIST / Listino prezzi § T - LABORATORY TESTS / Prove di laboratorio

### albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX

Accredited [IEC17025]: NO Execution time [day]: 31

Unit price #1 [EUR]: 4,500.00 Limit #1: Applicable to any sample

### FCA PF.90083:2020 Change level B

(id 1703) Fuel Tank assembly - Plastic

#### 8.2.1.1 - Preconditioning for Approved Tank Constructions

(id 2145) ---

Accredited [IEC17025]: NO Execution time [day]: 140

Unit price #1 [EUR]: 3,200.00 Limit #1: Applicable to any sample

#### 8.2.1.2 - Preconditioning New Tank Constructions

(id 2146) ---

Accredited [IEC17025]: NO Execution time [day]: 140

Unit price #1 [EUR]: 8,000.00 Limit #1: Applicable to any sample

#### 9.5 - Slosh Test

(id 2462) The objective of the Slosh Durability test is to prove the robustness of the fuel tank assembly and components when subjected to sloshing events that can occur within the life of a vehicle. This test intended to test all internal components including slosh baffle attachments and unique internal design features.

Accredited [IEC17025]: NO Execution time [day]: 70
Unit price #1 [EUR]: 15,000.00 Limit #1: Applicable to any sample

#### 6.3.8 - Vacuum Strength

(id 2520) The purpose of this test is to determine the ability of the Fuel Tank Assembly to withstand excessive vacuum.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,200.00 Limit #1: Applicable to any sample

#### 6.3.6 - Tank Shell Burst

(id 2521) The fuel tank shell must resist burst during dynamic pressure loading, for the intended useful life of the vehicle.

The purpose of this test is to verify the structural integrity of the fuel tank shell when exposed to high pressure. The burst test is to be conducted on the fuel tank as molded.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 400.00 Limit #1: Applicable to any sample

#### 6.3.7 - Tank Assembly Burst

(id 2522) The purpose of this test is to verify the structural integrity of the Fuel Tank Assembly when exposed to high internal pressure.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,200.00 Limit #1: Applicable to any sample

#### 7.5.1 - Fuel level Sensor Accuracy

(id 2551) The Fuel Delivery Module, mounted in the fuel tank, must provide the correct level sensor output at key points (i.e. LFW, ¼, ½, ¾, and full) in static conditions that occur in vehicle use.

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 500.00 Limit #1: Applicable to any sample

#### 7.12.1 - Fuel Measurement Test

(id 2552) The fuel gauge shall accurately represent the amount of fuel in the fuel tank under all vehicle duty cycles including all driving and parking conditions.

Accredited [IEC17025]: NO Execution time [day]: 7
Unit price #1 [EUR]: 850.00 Limit #1: Applicable to any sample (without preconditioning)

#### 7.11 - Fuel Tank Volumes and Capacity

(id 2553) The fuel system, including the fuel rank assembly, shall meet all the volume characteristics of the fuel system including total volume, active and passive vapor domes, Rated Capacity, and partial refueling.

The Fuel Tank Assembly shall meet the vehicle requirements for Rated Capacity, vapor venting, and dynamic unusable fuel capacity when the tank is installed in the vehicle environment.

Accredited [IEC17025]: NO Execution time [day]: 2 Unit price #1 [EUR]: 1,600.00 Limit #1: Applicable to any sample



### PRICE LIST / Listino prezzi § T - LABORATORY TESTS / Prove di laboratorio

## albarubens QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



#### 6.3.10 - Fuel Tank Hose and Line Pull-Off

(id 2788) Determine the loads to separate the hoses and lines from the Fuel Tank Assembly.

Accredited [IEC17025]: NO Execution time [day]: 21
Unit price #1 [EUR]: 800.00 Limit #1: Applicable to gasoline components

#### FCA PF.90098:2018

(id 1766) This standard covers the environmental, physical, electrical, mechanical, functional, regulatory, and durability requirements for the Inlet Check Valve (ICV).

#### 5.2 - Resistance to Fluids - Vapors and Preconditioning Test for Subsequent Testing

(id 2338) This test demonstrates the performance of the ICV when exposed to expected fuels. The ICV must withstand exposure to fluids present in the operating environment of the fuel tank for the useful life of the vehicle, 15 years / 150,000 miles.

Accredited [IEC17025]: NO Execution time [day]: 42
Unit price #1 [EUR]: 2,500.00 Limit #1: Applicable to any sample

#### FCA PF.90141 20/02/2017

(id 1647) Return hose assembly - hydraulic steering

#### 7.3 - Burst test

(id 2096) ---

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 300.00 Limit #1: One temperature only (destructive testing)

#### FCA PF.90197 15-JUN-2018

(id 1330) Fuel bundle and fuel system Jumpers - Plastic, rubber and metallic

#### 6.3.5 - Internal fuel resistance test

(id 1960) ---

Accredited [IEC17025]: NO Execution time [day]: 44
Unit price #1 [EUR]: 3,500.00 Limit #1: Single pipe
Unit price #2 [EUR]: 4,500.00 Limit #2: Assembly of multiple pipes

#### 9.5 - Pulsating pressure

(id 1961) The fuel/vapor line assembly must withstand internal pressure and pressure pulsations for the useful life of the vehicle.

The assembled components are subjected to 300000 pressure cycles for traditional Stop & Start systems, 600000 cycles for Stop & start systems with Freewheels.

Pressure cycle is described in Tab. 10 of FCA PF.90197 ed.2018.

Test pressures depends on the class of the components that determines the working pressures of the line assembly, in any case the minimum pressure applied during pulsating pressure is P min indicated in Tab. 9 of FCA PF.90197 ed.2018. and the maximum one is 2 times Ps.

Chamber temperature cycle is described in Fig.4 - Time/Temperature of FCA PF.90197 ed.2018.

#### 9.3 - Life cycle test

(id 1962) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

#### 7.4 - Burst test

(id 1963) ---

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 300.00 Limit #1: One temperature only (destructive testing)



albarubens

### your value certified

### PRICE LIST / Listino prezzi § T - LABORATORY TESTS / Prove di laboratorio

#### QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO

Copy uncontrolled when printed: check the updated version

TNDEX



#### 8.2 - Assembly Hydrocarbon loss (MiniSHED) Test - NON NAFTA

(id 1971) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 6,000.00 Limit #1: Applicable to any sample

#### 9.4 - Vibration cycle test

(id 1972) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant

#### 7.2 - Leak Resistance

(id 1973) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant

#### 6.3.4 - Fitting Pull-Off test

(id 1974) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant

#### 6.1 - Appearance requirements

(id 1975) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant

#### 7.3 - Pressure drop/Flow resistance test

(id 1982) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Limit #1: Applicable to any sample Unit price #1: ask to office

#### FCA PF.90197:2021

(id 1743) Fuel bundle and fuel system jumpers - plastic, rubber and metallic

#### 6.3.5 - Internal fuel resistance test

(id 2222) The materials used in a gasoline fuel / vapor line assembly must function with continuous exposure to liquid fuel and fuel vapors, including alcohol

Accredited [IEC17025]: NO Execution time [day]: 44 Unit price #1 [EUR]: 3,800.00 Limit #1: Single pipe

Unit price #2 [EUR]: 4,800.00 Limit #2: Assembly of multiple pipes

#### 7.2 - Leak resistance

(id 2225) Checking the tightness without leaks of fhe fuel / vapor line assembly

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 300.00 Limit #1: Applicable to any sample

#### 6.3.4 - Fitting Pull-Off Test

(id 2228) Testing the robustness of the connections between the fuel / vapor line and the fuel / vapor connectors.

Accredited [IEC17025]: NO Execution time [dav]: 0

Limit #1: Applicable to any sample Unit price #1 [EUR]: 300.00

#### 7.4 - Burst Test

(id 2229) Verification of the ability of the fuel / vapor line assembly to withstand internal pressure.

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 300.00 Limit #1: One temperature only (destructive testing)



## **PRICE LIST / Listino prezzi**

### § T - LABORATORY TESTS / Prove di laboratorio

### albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version



#### FCA PF.90254:2015

(id 1959) CAPLESS REFUELING MODULE

#### 5.6.2 - Soak Test for Subsequent Testing

(id 2648) The Capless Refueling Module must withstand exposure to any fluids present in the operating environment of the fuel filler pipe for the useful life of the

Accredited [IEC17025]: NO Execution time [day]: 21 Unit price #1 [EUR]: 1,400.00 Limit #1: Applicable to any sample

#### 5.4 - Dust Ingestion

(id 2649) The Capless Refueling Module must not allow foreign environmental material to enter into the fuel tank assembly.

Accredited [IEC17025]: NO Execution time [day]: 5 Limit #1: Applicable to any sample Unit price #1 [EUR]: 1,800.00

### FCA PF.90258 (S,E,D) 8-JAN-2019

(id 1197) FCA technical specification - Fuel tank assembly - Metallic

#### 5.7.1 - Soak Test - Fuel Compatibility

(id 1623) ---

Accredited [IEC17025]: NO Execution time [day]: 70 Unit price #1 [EUR]: 5,000.00 Limit #1: Up to 20 litres - 1000h Unit price #2 [EUR]: 7,500.00 Limit #2: Up to 20 litres - 1680h Unit price #3 [EUR]: 13,500.00 Limit #3: More than 20 litres - up to 1680h

#### 5.4.1 - Corrosion: Functional requirement

(id 1819) ---

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 9,000.00 Limit #1: Not relevant

#### 5.4.2 - Corrosion: Cosmetic requirement

(id 1820) ---

Accredited [IEC17025]: NO Execution time [day]: 40 Unit price #1 [EUR]: 6,500.00 Limit #1: Not relevant

#### 5.6.1 - Chemical Exposure Test

(id 1821) ---

Execution time [day]: 0 Accredited [IEC17025]: NO Unit price #1: ask to office Limit #1: Not relevant

#### 6.2.1 - Physical Requirements: Fuel Tank Shell Mass

(id 1822) ---

Accredited [IEC17025]: NO Execution time [dav]: 0 Unit price #1: ask to office Limit #1: Not relevant

#### 6.2.3 - Physical Requirements: Internal Cleanliness

(id 1823) ---

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1: ask to office Limit #1: Not relevant

#### 6.2.5 - Physical Requirements: Material Treatment Requirements

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1: ask to office Limit #1: Not relevant

#### 6.2.6 - Physical Requirements: Dimensional Requirements

(id 1825) ---

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1: ask to office Limit #1: Not relevant



### your value certified § T - LABORATORY TESTS / Prove di laboratorio

**PRICE LIST / Listino prezzi** 

## albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

TNDEX



#### 6.2.8.2.1 - Welded Components - Roller Welding - Micrographic check test

(id 1826) ---

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1: ask to office Limit #1: Not relevant

#### 6.2.8.2.2 - Welded Components - Roller Welding - Peel check

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1: ask to office Limit #1: Not relevant

#### 6.2.8.3.1 - Brazing Welding - Visual test

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1: ask to office Limit #1: Not relevant

#### 6.2.8.3.3 - Brazing Welding - Peel check

(id 1829) ---

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1: ask to office Limit #1: Not relevant

#### 6.2.8.4 - Welded spots

(id 1830) ---

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1: ask to office Limit #1: Not relevant

#### 6.2.9 - Non-hermetic Internal Component Retention

(id 1831) Non-hermetic internal components, such Grade Vent Valve, Control Valve, brackets, etc. that are attached to the internal surface of fuel tank but not welded, must meet a pull off force that exceeds the supported component loading seen in the vehicle environment.

Accredited [IEC17025]: NO Execution time [day]: 2 Unit price #1 [EUR]: 2,500.00 Limit #1: No soaking - up to 10 points Unit price #2 [EUR]: 3,500.00 Limit #2: With soaking - up to 10 points

#### 6.2.11 - Cold Roll Test

(id 1832) ---

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1: ask to office Limit #1: Acceptance criteria after fuel compatibility test Unit price #2 [EUR]: 3,000.00 Limit #2: Single measurement Unit price #3 [EUR]: 1,500.00 Limit #3: Repetition with set up ready

#### 6.3.5 - Mechanical Requirements: Stability Slow Speed and High Speed Sled Impact Test

(id 1833) ---

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1: ask to office Limit #1: Not relevant

#### 6.3.6 - Mechanical Requirements: Tank Assembly Burst

(id 1834) ---

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 6.3.7 - Mechanical Requirements: Vacuum Strength

(id 1835) ---

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 570.00 Limit #1: Not relevant

#### 6.3.8 - Mechanical Requirements: Tank Assembly Handling Impact

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 500.00 Limit #1: Not relevant

#### 6.3.9 - Mechanical Requirements: Fuel Tank Hose and Line Pull-Off

(id 1837) ---

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 2,500.00 Limit #1: Fuel assembly Unit price #2 [EUR]: 4,500.00 Limit #2: Diesel assembly



### **PRICE LIST / Listino prezzi** § T - LABORATORY TESTS / Prove di laboratorio

## albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

TNDEX



#### 6.4.6.1 - Electrical Requirements: Sensor Function and Continuity Verification

(id 1838) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant

#### 7.2 - Customer Refueling Fill Quality

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant

#### 7.6.2 - Tip Angle Venting - Static and Dynamic Liquid Carryover

(id 1840) --

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant

#### 7.7.1 - Fuel Level Sensor Accuracy

(id 1841) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant

#### 7.7.2 - Fuel System Priming/Unpriming Checks on Test Bench

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant

#### 7.9 - Vacuum Leakage

(id 1843) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 240.00 Limit #1: Not relevant

#### 7.10 .1 - External Leakage - Static Leak Integrity S,E - Fuel Tank Leak Test S,E

(id 1844) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 4,000.00 Limit #1: Not relevant

#### 7.12 - Fuel Tank Volumes and Capacity <D>

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant

#### 7.13.1 - Fuel Measurement Test

(id 1846) The fuel gauge shall accurately represent the amount of fuel in the fuel tank under all vehicle duty cycles including all driving and parking conditions.

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 800.00 Limit #1: With one fuel

#### 8.1 - ORVR SHED Refueling <E>

(id 1847) ---

Execution time [day]: 0 Accredited [IEC17025]: NO

Unit price #1: ask to office Limit #1: Not relevant

#### 8.2.1.1 - Preconditioning for NAFTA permeation test

(id 1848) ---

Accredited [IEC17025]: NO Execution time [day]: 70

Unit price #1 [EUR]: 4,000.00 Limit #1: Applicable to any sample

#### 8.3.1.x - Preconditioning for Non-NAFTA permeation test

(id 1849) ---

Accredited [IEC17025]: NO Execution time [day]: 140

Unit price #1 [EUR]: 4,000.00 Limit #1: Up to 10 weeks Unit price #2 [EUR]: 8,000.00 Limit #2: Up to 20 weeks



## PRICE LIST / Listino prezzi

### § T - LABORATORY TESTS / Prove di laboratorio

### albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

TNDEX

#### 8.6.1 - Hydraulic Test

(id 1850) ---

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1: ask to office Limit #1: Not relevant

#### 8.6.2 - Ambient Roll Test (Overturning Test)

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 800.00 Limit #1: Up to 401 Unit price #2 [EUR]: 1,200.00 Limit #2: From 401 a 701

#### 9.6 - Assembly Shake/Vibration

Accredited [IEC17025]: NO Execution time [dayl: 0 Unit price #1: ask to office Limit #1: Not relevant

#### 6.4.7 - Electrostatic Discharge (ESD)

(id 1917) See SAE J1645 REV. AUG2006

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1: ask to office Limit #1: Not relevant

#### 5.4.3 - Internal corrosion

(id 1940) ---

Accredited [IEC17025]: NO Execution time [day]: 2 Unit price #1 [EUR]: 1,000.00 Limit #1: Up to 48 h

#### 8.2.1 - Permeation test for NAFTA

(id 2037) ---

Accredited [IEC17025]: NO Execution time [day]: 4 Unit price #1 [EUR]: 2,000.00 Limit #1: Not relevant

#### 8.3.1 - Permeation test for Non-NAFTA

Execution time [day]: 4 Accredited [IEC17025]: NO Unit price #1 [EUR]: 2,000.00 Limit #1: Not relevant

#### Amendment - Soak Test - Fuel Compatibility (2016h)

(id 2215) ---

Accredited [IEC17025]: NO Execution time [day]: 84 Unit price #1 [EUR]: 12,500.00 Limit #1: With cold roll as acceptance criteria Unit price #2 [EUR]: 9,500.00 Limit #2: Only aging without cold roll test

#### FCA PF.90258:2022

(id 2228) FUEL TANK ASSEMBLY - METALLIC

#### 5.7.2 - Soak Test - Fuel Compatibility - Aluminum Fuel Tank

(id 3110) Verify the compatibility of the Aluminum fuel tank and component materials with applicable fuels.

Accredited [IEC17025]: NO Execution time [day]: 70 Unit price #1 [EUR]: 7,500.00 Limit #1: Up to 20 litres - 1680h Unit price #2 [EUR]: 9,000.00 Limit #2: More than 20 litres - dimension up to 60x160x60cm - 1680h Unit price #3 [EUR]: 13,500.00 Limit #3: More than 20 litres - big dimensions - 1680h

### FCA PF.90303:2022 Change Level: A

(id 2091) FCA automotive specification

#### 5.6.2 - Water Spray Test for Grommets

(id 2888) To be carried out on wiring harness or a section of a wiring harness after the three phase temperature/humidity/accelerated aging test. This test is performed on grommets in their design intent position on a wire assembly or a portion of a wire assembly.

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1 [EUR]: 300.00 Limit #1: Not relevant

page 83/234



# PRICE LIST / Listino prezzi § T - LABORATORY TESTS / Prove di laboratorio

## albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



#### FCA PF90177:2018

(id 1190) FCA Performance Specification. CANISTER - EVAPORATIVE EMISSIONS 28-MAR-2018

#### 8.7.2 - Wall Permeation: canister pre-conditioning

(id 1616) ---

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 2,600.00 Limit #1: Not relevant

#### 9.3 - Pressure Cycling

(id 1884) This test will verify structural durability of the canister housing, cover and welded interface. The volume compensation ability of the canister under pressure will also be verified.

Accredited [IEC17025]: NO Execution time [day]: 7
Unit price #1 [EUR]: 2,000.00 Limit #1: Not relevant

#### 8.7.3 - Permeation test

(id 2056) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 2,200.00 Limit #1: Applicable to any sample

#### Fiat 9.02136-01

(id 1812) Tubi e manicotti in gomma per conduzione liquido di raffreddamento

#### 2.5.5 - Dilatazione diametrale

(id 2434) ---

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 2.5.6 - Pressione di scoppio

(id 2435) ---

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 2.5.11 - Pressione pulsante

(id 2436) ---

Accredited [IEC17025]: NO Execution time [day]: 9
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

### Fiat Auto 50471/01:2000

(id 2129) Prova di resistenza alla luce ultravioletta alternata alla umidità

#### Unref. - Prova di resistenza alla luce ultravioletta alternata alla umidità

(id 2939) Lo scopo di questo test è quello di simulare in tempi sufficientemente brevi i decadimenti delle caratteristiche estetiche/funzionali delle superfici verniciate, causati dalla componente UV della luce solare in presenza e non di umidità ambientale.

Questo metodo è comunemente impiegato per:

- prodotti verniciati adatti a parti esterne della vettura, sia in plastica che in metallo.
- particolari esterno vettura metallici e non metallici, rivestimenti organici.

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 5,000.00 Limit #1: Up to 35 cycles

#### Fiat Auto 50473:2000

(id 2043) Determinazione della resistenza alla benzina delle vernici, degli smalti, ecc.

#### 2.3 - Metodo B (strofinamento)

(id 2789) ---

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

page 84/234



albarubens

## PRICE LIST / Listino prezzi

### § T - LABORATORY TESTS / Prove di laboratorio

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



#### Fiat Auto 9.02144/01:2002 Ed. 6

(id 1890) Bocchettoni introduzione combustibile in materiale plastico

#### 2.9.3 - Low temperature impact resistance

(id 2523) Resistenza all'urto a bassa temperatura su bocchettoni.

Accredited [IEC17025]: NO Execution time [day]: 8

Unit price #1 [EUR]: 400.00 Limit #1: Applicable to any sample

#### 2.6 - Determination of the quantity and size of internal scraps

(id 2534) ---

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 2.7.3 - Check of volumes and capacites

(id 2535) Verification of the filling neck volume compared to the rated volume.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 400.00 Limit #1: Applicable to any sample

#### 2.9.1 - Burst resistance

(id 2581) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 250.00 Limit #1: Applicable to any sample

2.7.1 - High and low pressure tightness

(id 2582) ---

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 500.00 Limit #1: Applicable to any sample

#### 2.7.2 - Suction tightness

(id 2583) ---

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 500.00 Limit #1: Applicable to any sample

#### Fiat Auto 9.93607/01:1995

(id 2112) Serbatoio rigidi per liquido lavacristalli

#### 2.10 - Thermal Cycle Resistance

(id 2926) This test is intended to verify the resistance of tanks complete with electric pumps and windshield washer fluid to 10 thermal cycles.

Accredited [IEC17025]: NO Execution time [day]: 10 Unit price #1 [EUR]: 2,600.00 Limit #1: Applicable to any sample

### Fiat Group Automobiles 50180:2003

(id 2122) Prove di corrosione

#### A3 - Prova di corrosione in nebbia cupro-salino-acetica

(id 2933) Da adottarsi preferibilmente per rivestimenti inorganici, chimici ed elettrolitici poco reattivi (es. cromatura su acciaio o lega di zinco, ossidazione anodica su alluminio. ecc.)

Accredited [IEC17025]: NO Execution time [day]: 10
Unit price #1 [EUR]: 2,600.00 Limit #1: Applicable to any sample



# PRICE LIST / Listino prezzi § T - LABORATORY TESTS / Prove di laboratorio

albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



### Fiat Group Automobiles 50180:2007

(id 1436) PROVE DI CORROSIONE

#### Unref. - Metodo A1 - Prova di corrosione in nebbia salina

(id 3028) Simulare con prove normalizzate, ed in tempi sufficientemente brevi, effetti della stessa natura di quelli riscontrabili su materiali, rivestimenti, particolari e componenti durante l'esercizio su autoveicolo in ambienti corrosivi.

Unit price #1 [EUR]: 3,200.00 Limit #1: Up to 360h with intermediate checks

### Fiat Group Automobiles 55250/03:2013

(id 1885) MULTISTRATO HDPE – EVOH PER SERBATOI COMBUSTIBILE PRODOTTI CON TECNOLOGIA DI ESTRUSIONE / SOFFIAGGIO.

#### 5 - Resilienza IZOD

(id 2528) ---

Accredited [IEC17025]: NO Execution time [day]: 2

Unit price #1 [EUR]: 200.00 Limit #1: Applicable to any sample

#### 5 - Impact resistance from falling dart

(id 2529) ---

Accredited [IEC17025]: NO Execution time [day]: 2

Unit price #1 [EUR]: 200.00 Limit #1: Applicable to any sample

#### 5 - Layer of Regrind HDPE Material SEM Analysis

(id 2530) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 500.00 Limit #1: Applicable to any sample

### Fiat Group Automobiles 9.02137/01:2012

(id 1991) TUBAZIONI PLASTICHE CON TERMINALI PER CONDUZIONE COMBUSTIBILE

#### 2.7.17 - Funzionamento simulato

(id 2751) ---

Accredited [IEC17025]: NO Execution time [day]: 42

Unit price #1 [EUR]: 4,000.00 Limit #1: Applicable to any sample

### Fiat Group Automobiles 9.02137/05:2010

(id 386) Automotive blowby piping endurance tests

#### 2.7.3 - Thermal aging for blow-by circulation systems

(id 1195) ---

Accredited [IEC17025]: NO Execution time [day]: 9
Unit price #1 [EUR]: 800.00 Limit #1: Not relevant

#### 2.7.5 - Long thermal aging in exhausted oil for blow-by circulation systems

(id 1196) The pipe is filled at 50% with exhaust oil supplied by the customer

Accredited [IEC17025]: NO Execution time [day]: 22
Unit price #1 [EUR]: 1,800.00 Limit #1: Not relevant

#### 2.7.7 - Pulsing pressure for blow-by circulation systems

(id 1197) Pulsing pressure obtained with compressed air mixed with 1 drop-per-cycle exhaust motor oil, supplied by the customer

Accredited [IEC17025]: NO Execution time [day]: 2

Unit price #1 [EUR]: 1,500.00 Limit #1: Not relevant



# PRICE LIST / Listino prezzi § T - LABORATORY TESTS / Prove di laboratorio

## albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



#### 2.7.4 - Resistance to thermal cycles for blow-by circulation systems

(id 1202) Thermal cycle repeated as requested

#### 2.7.6 - Burst resistance for blow-by circulation systems

(id 1383) ---

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 300.00 Limit #1: Not relevant

#### Flexible Accredited Methods

(id 2140) Here can be found the flexible accredited methods whose parameters are covered by A2LA Parameter Based Scope

#### **Unref. - Temperatures**

(id 2964) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### **Unref. - Relative humidity**

(id 2965) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### **Unref. - Thermal cycles**

(id 2966) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### Unrerf. - Salt mist

(id 2967) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### Unref. - Insulation resistance

(id 2968) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### Unref. - Dielectric strenght

(id 2969) ---

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### **Unref. - Impacts**

(id 2972) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

### FM Approvals 5511:2021

(id 2208) Firefighting Nozzles for Use with Hose, Monitor Assemblies and Other Firefighting Equipment

#### 4.3 - Spray Character, Reach, and Discharge Angle

(id 3100) Spray nozzles shall be capable of developing discharge patterns varying from straight stream to at least 100 degrees spray angle. Spray pattern settings shall provide a full and uniform spray pattern. The straight stream reach and discharge angle of each distinct pattern selection observed in testing shall be used to verify the manufacturer's published reach and discharge angle for each rating and flow condition.

Spray nozzles designed for Class C fire applications shall be exempt from the straight stream pattern requirement. Instead, the minimum discharge angle shall not be less than 30 degrees.

Accredited [IEC17025]: NO Execution time [day]: 2

Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

page 87/234



### PRICE LIST / Listino prezzi § T - LABORATORY TESTS / Prove di laboratorio



QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

TNDEX



#### FORD ES-FU-5A-9J279-AA Rev.2013-09-25

(id 1493) Ford engineering specification- Tube Assy. - Fuel Supply, Return and Vapor

#### 4.3 - point 2) - Key life tests for generic features of assemblies (Preconditioning only)

(id 1949) The key life test is required to qualify generic component designs, materials, manufacturing processes, and manufacturers of pressurized fuel tube or vapor assemblies.

However, for this specific case, only the preconditioning part indicated at point 2) of §4.3 is performed with test fluid CM15, and only point 1) of the acceptance criteria is considered, as requested by the customer.

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 2,800.00 Limit #1: Applicable to any sample

### FORD ESDS7H-19B591-AA

(id 1749) Engineering Specification - External corrosion resistance - Test Method ASTM B 117

#### 3A.9 - External Corrosion Resistance

(id 2237) This test method is intended for assessing the corrosion resistance of the external parts of components exposed to a continuous salt mist environment.

Accredited [IEC17025]: NO Execution time [day]: 20

Unit price #1 [EUR]: 2,600.00 Limit #1: Applicable to any sample

#### FORD ESDS7H-19E777-AA/K

(id 945) Engineering Specification - APGE - Ford Laboratory Test Method BI 123-01

#### 2 - APGE (Ford Laboratory Test Method BI 123-01)

(id 1451) Objective of this test: To simulate Arizona Proving Ground Accelerated Corrosion Test in a Laboratory

Accredited [IEC17025]: NO Execution time [day]: 56 Unit price #1 [EUR]: 2,700.00 Limit #1: Not relevant

#### 3A.9 - External Corrosion Resistance

(id 2216) ---

Accredited [IEC17025]: NO Execution time [day]: 20

Unit price #1 [EUR]: 2,600.00 Limit #1: Applicable to any sample

### FORD WSK-M98P4-A (2 July 1990)

(id 1495) FUEL LINE, POLYAMIDE (PA) 11/12

#### 3.4.6 - Cold Impact Resistance

(id 1956) ---

Accredited [IEC17025]: NO Execution time [dav]: 3

Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

#### 3.4.7 - Resistance to Fuels

(id 1957) ---

Accredited [IEC17025]: NO Execution time [day]: 31

Unit price #1 [EUR]: 1,500.00 Limit #1: Applicable to any sample

#### 3.4.4 - Flexibility of the sample

(id 1958) ---

Accredited [IEC17025]: NO Execution time [day]: 1

Limit #1: Applicable to any sample Unit price #1 [EUR]: 200.00



## PRICE LIST / Listino prezzi

#### your value certified § T - LABORATORY TESTS / Prove di laboratorio

### albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



### FPI9.EMS011 (Issue 1, 05/11/2015)

(id 1640) FPT Normalized Specification for Engine Control System Components Validation

#### 8.3.2 - Dust Ingress Test

(id 2086) The dust ingress test checks for the effects of dust laden environments generated by soil working operations in dry conditions. The Dust Ingress procedure results can be affected by prior Heat (see section 8.1), Mechanical (see section 8.2) and Altitude (see section 8.3.1) tests, and so is normally conducted subsequent to these. This test is to be performed with dust environment with a suspended Arizona dust (ISO 12103) using the method outlined in EN60529 (BS).

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

#### 8.3.3.4 - Hose Washing - Test severity level IP X7

(id 2087) Water Ingress Test checks for resistance to water from the normal working environment (including leakage).

The Water Ingress procedure results can be affected by prior Heat (see section 8.1), Mechanical (see section 8.2) and Altitude (see section 8.3.1) tests, and so is normally conducted subsequent to these. The test is performed using the method outlined in EN60529 (BS).

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

#### 8.3.3.5 - Pressure Wash/Steam Procedure (High Pressure Jet Washing)

(id 2088) Water Ingress Test checks for resistance to water from the normal working environment (including leakage). The Water Ingress procedure results can be affected by prior Heat (see section 8.1), Mechanical (see section 8.2) and Altitude (see section 8.3.1) tests, and so is normally conducted subsequent to these.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

#### 8.4.2.4 - Corrosion

(id 2089) The test shall be performed in according to IEC 60068-2-52

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

#### 8.4.2.5 - Leakage and Function

(id 2090) Salt spray (mist) test as outlined in IEC 60068-2-11 Test 2 Procedure Ka, method of salt spray (fog) testing.

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1: ask to office Limit #1: Not relevant

#### GB 18296 -2001

(id 1326) NATIONAL STANDARD OF PEOPLE'S REPUBLIC OF CHINA - Safety property requirements and test methods for automobile fuel tank

#### 4.2 - Test for the opening pressure of safety venting system

(id 1816) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 500.00 Limit #1: Not relevant

#### 4.3 - Test for the vibration enduring performance

(id 1817) ---

Accredited [IEC17025]: NO Execution time [day]: 2 Subcontract: GESTLABS S.R.L.

Unit price #1 [EUR]: 2,000.00 Limit #1: Sample up to 50 cm length

Unit price #2 [EUR]: 3,800.00 Limit #2: Sample up to 200 cm length

#### 4.4 - Test for pressure resistance of metal fuel tank

(id 1818) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 500.00 Limit #1: Not relevant

#### GB 18296:2019

(id 1905) Safety property requirements and test methods for automobile fuel tank and its installation

#### 5.2 - Overturn test

(id 2559) ---

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 800.00 Limit #1: Up to 401 with set up ready
Unit price #2 [EUR]: 1,500.00 Limit #2: From 401 a 701 with set up ready

page 89/234



albarubens

# PRICE LIST / Listino prezzi § T - LABORATORY TESTS / Prove di laboratorio

### QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



#### 5.3 - Low temperature shock resistance test of plastic fuel tank

(id 2560) ---

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 1,200.00 Limit #1: Applicable to any sample

#### 5.4 - Mechanical strength test of plastic fuel tank

(id 2561) ---

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 1,000.00 Limit #1: Applicable to any sample

#### 5.5 - Fuel permeability test of plastic fuel tank

(id 2562) ---

Accredited [IEC17025]: NO Execution time [day]: 84

Unit price #1 [EUR]: 5,500.00 Limit #1: Applicable to any sample

#### 5.6 - Fuel resistance test of plastic fuel tank

(id 2563) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant

#### 5.7 - Fire resistance test of plastic fuel tank

(id 2564) ---

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 4,200.00 Limit #1: Applicable to any sample

#### 5.8 - High temperature resistance test of plastic fuel tank

(id 2565) ---

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

#### 5.1 - Pressure resistance test of metal fuel tank

(id 2782) ---

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 1,000.00 Limit #1: Applicable to any sample

#### GB 20800.1:2006

(id 1047) General rules of explosion-protect techniques of reciprocating internal combustion engines for explosive atmospheres - Part 1: Group II engines for use in flammable gas and vapor atmospheres

#### 5.2.2 - Determination of maximum explosion pressures

(id 1561) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: not relevant

#### 5.2.4 - Non transmission tests

(id 1562) ---

Accredited [IEC17025]: NO Execution time [day]: 2

Unit price #1: ask to office Limit #1: Not relevant

#### 5.2.3 - Overpressure test

(id 1563) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant



### PRICE LIST / Listino prezzi

#### § T - LABORATORY TESTS / Prove di laboratorio

### albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



#### GMW14638:2019

(id 1708) GM material specification: Thermoplastic Fuel Tubing

#### 3.15 - Resistance to stone impact

(id 2150) The Stone Impact Test Procedure is as per GMW14700, Method B for ten (10) cycles at -40 °C test temperature.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

#### 3.6 - Burst Test

(id 2151) Evaluation of bursting pressure, using air at room temperature.

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 500.00 Limit #1: Applicable to any sample

#### 3.22 - Fuel Resistance 5040h Recirculation Test

(id 2152) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 15,000.00 Limit #1: Applicable to any sample

#### 3.23 - Sour Gasoline 1008h Recirculation Test

(id 2155) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 5,000.00 Limit #1: Applicable to any sample

#### 3.22.3.1 - Length change after Fuel Resistence

(id 2209) Simple measurement of the length of the tube, performed before and after the aging

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 100.00 Limit #1: Applicable to any sample

#### 3.22.3.2 - Leak test after Fuel Resistance

(id 2210) Simple bubble test

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 300.00 Limit #1: Applicable to any sample

#### 3.9 - Cold Temperature Flexibility Test

(id 2211) Bending on a mandrel at -40°C

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 600.00 Limit #1: Applicable to any sample

#### 3.10 - Surface Resistivity

(id 2213) Simple resistivity measurement, using a megohmmeter and copper pins inserted in the tube

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 300.00 Limit #1: Applicable to any sample

#### 3.16 - Layer Adhesion Test

(id 2214) The tube is cut in helical coil, then it's measured the force to separate the layers at 50mm/min

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

#### GMW14872:2013

(id 1545) Cycling Corrosion Laboratory test

#### unref - Corrosion testing

(id 2005) ---

Accredited [IEC17025]: NO Execution time [day]: 8
Unit price #1 [EUR]: 800.00 Limit #1: 24 hours
Unit price #2 [EUR]: 1,000.00 Limit #2: 48 hours
Unit price #3 [EUR]: 1,400.00 Limit #3: 96 hours



### PRICE LIST / Listino prezzi § T - LABORATORY TESTS / Prove di laboratorio

albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX

### Honda specification

(id 1651) Fuel lines components

#### 6.1.b - Sealing property (filled tube, static+dynamic)

(id 2102) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 300.00 Limit #1: Liquid line (positive pressure)
Unit price #2 [EUR]: 400.00 Limit #2: Vapour line (positive+negative pressure)

#### 6.2.a - Bursting pressure (untreated state)

(id 2103) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 300.00 Limit #1: Not relevant

#### 6.3 - Tube deformation of negative pressure

(id 2104) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

#### 6.5 - Thermo-cycle test

(id 2105) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 3,600.00 Limit #1: Not relevant

#### 6.6 - Vibration durability of fuel circulation

(id 2106) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 4,200.00 Limit #1: Applicable to any sample

#### 6.8 - Low temperature impact property

(id 2107) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 2,800.00 Limit #1: Not relevant

#### 6.9 - Bending test at low temperature

(id 2108) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

#### 6.10 - Calcium Chloride resistance

(id 2109) ---

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 1,000.00 Limit #1: Applicable to any sample

#### 6.11 - Pipe insertion property

(id 2110) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

#### 6.12 - Pipe insertion durability

(id 2111) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 1,200.00 Limit #1: Applicable to any sample

#### 6.13 - Mating pipe, tube, joint pulling-out load

(id 2112) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

page 92/234



### PRICE LIST / Listino prezzi § T - LABORATORY TESTS / Prove di laboratorio

### LABORATORT TEOTO / FTOVO di laboratori

### albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX

#### 6.14 - Tensile resistance

(id 2113) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 1,200.00 Limit #1: Applicable to any sample

#### 6.15 - Sealing performance after rotating joint of pipe and tube

(id 2114) --

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 1,500.00 Limit #1: Applicable to any sample

#### 6.16 - Weatherability

(id 2115) --

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 4,500.00 Limit #1: Applicable to any sample

#### 6.17 - Adhesiveness

(id 2116) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 500.00 Limit #1: Applicable to any sample

#### 6.18/P - Fuel permeability: preconditioning

(id 2117) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 8,500.00 Limit #1: Applicable to any sample

#### 6.19 - Conductivity

(id 2118) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant

#### 6.4 - Dimensional stability

(id 2119) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant

#### 6.21 - Positive and negative pressure durability test

(id 2120) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 1,200.00 Limit #1: Applicable to any sample

#### 6.7 - Vibration durability

(id 2121) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 3,000.00 Limit #1: Applicable to any sample

#### 6.20 - Chipping resistance

(id 2122) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 1,200.00 Limit #1: Applicable to any sample

#### 6.1.a - Sealing property (untreated state, static+dynamic)

(id 2123) ---

Accredited [IEC17025]: NO Execution time [day]: 2

Unit price #1 [EUR]: 300.00 Limit #1: Liquid line (positive pressure)

Unit price #2 [EUR]: 400.00 Limit #2: Vapour line (positive+negative pressure)

#### 6.18/S - Fuel permeability: SHED test

(id 2124) ---

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 2,200.00 Limit #1: Not relevant

page 93/234



## **PRICE LIST / Listino prezzi**

### § T - LABORATORY TESTS / Prove di laboratorio

### albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

TNDEX

#### 6.1.pre - Pretreatment before sealing

(id 2126) ---

Accredited [IEC17025]: NO Execution time [day]: 2

Unit price #1 [EUR]: 300.00 Limit #1: Not relevant

### **Hutchinson proprietary test methods**

(id 1691) Circolazione BMW Cooling Progetto i20/G26/48V - Validazione QC - Metodo Interno HUT

#### unref - 5.2.2.8 Aging according CDC 11229184 - 000 - 02 (07/2019)

(id 2057) ---

Accredited [IEC17025]: NO Execution time [day]: 125 Unit price #1 [EUR]: 4,000.00 Limit #1: 1000h (only visual inspection afer ageing) Unit price #2 [EUR]: 5,500.00 Limit #2: 2000h (only visual inspection afer ageing) Limit #3: 3000h (only visual inspection afer ageing) Unit price #3 [EUR]: 7,000.00

#### unref - Circulation test

(id 2135) ---

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 2,300.00 Limit #1: Not relevant

#### 01759\_21\_00390 §4.7.3.2.2 - Slow ageing test

(id 2173) ---

Accredited [IEC17025]: NO Execution time [day]: 42 Unit price #1 [EUR]: 1,000.00 Limit #1: Lot of small sample (up to 200mm)

#### unref - Hutchinson internal method: BMW Ncar long aging study

Accredited [IEC17025]: NO Execution time [day]: 210 Unit price #1 [EUR]: 3,800.00 Limit #1: Large group of scalar samples

#### IEC 60034-5:2020+COR1:2024

(id 2201) Rotating electrical machines - Part 5: Degrees of protection provided by the integral design of rotating electrical machines (IP code) -Classification

#### 9 - Test for first characteristic numeral 1

(id 2062) Machine protected against solid objects greater than 50 mm

Accredited [IEC17025]: YES Execution time [day]: 0 Unit price #1: ask to office Limit #1: Not relevant

#### 9 - Test for first characteristic numeral 2

(id 2063) Machine protected against solid objects greater than 12 mm

Accredited [IEC17025]: YES Execution time [day]: 0 Unit price #1 [EUR]: 400.00 Limit #1: Not relevant

#### 9 - Test for first characteristic numeral 3

(id 2064) Machine protected against solid objects greater than 2,5 mm

Accredited [IEC17025]: YES Execution time [day]: 0 Unit price #1 [EUR]: 400.00 Limit #1: Not relevant

#### 9 - Test for first characteristic numeral 4

(id 2065) Machine protected against solid objects greater than 1 mm

Accredited [IEC17025]: YES Execution time [day]: 0 Unit price #1 [EUR]: 400.00 Limit #1: Not relevant

#### 9 - Test for first characteristic numeral 5 and 6

(id 2066) IP 5X: Contact with or approach to live or moving parts inside the enclosureIngress of dust is not totally prevented but dust does not enter in sufficient quantity to interfere with satisfactory operation of the machine;

IP 6X: Ingress of dust totally prevented

Accredited [IEC17025]: YES Execution time [day]: 0 Unit price #1 [EUR]: 500.00 Limit #1: up to 30 Kg Limit #2: from 31 to 100 Kg Unit price #2 [EUR]: 1,100.00 Unit price #3 [EUR]: 1,300.00 Limit #3: over 100 Kg

page 94/234



## PRICE LIST / Listino prezzi

### § T - LABORATORY TESTS / Prove di laboratorio

## albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

TNDEX

#### 10 - Test for second characteristic numeral 1

(id 2067) Machine protected against dripping water

Accredited [IEC17025]: YES Execution time [day]: 0 Unit price #1: ask to office Limit #1: Not relevant

#### 10 - Test for second characteristic numeral 2

(id 2068) Machine protected against dripping water when tilted up to 15°

Accredited [IEC17025]: YES Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant

#### 10 - Test for second characteristic numeral 3

(id 2069) Water falling as a spray at an angle up to 60° from the vertical shall have no harmful effect

Accredited [IEC17025]: YES Execution time [day]: 0 Unit price #1 [EUR]: 400.00 Limit #1: Not relevant

#### 10 - Test for second characteristic numeral 4 (hand-held spray nozzle)

(id 2070) Machine protected against splashing water.

Test performed to evaluate the protection of the equipment against the access of water in the presence of a spray nozzle with a water delivery rate of 10 l/min + or

Accredited [IEC17025]: YES Execution time [day]: 0 Limit #1: Up to 30kg Unit price #1 [EUR]: 500.00 Unit price #2 [EUR]: 1,100.00 Limit #2: From 31kg to 100kg Unit price #3 [EUR]: 1,300.00 Limit #3: Over 100kg

#### 10 - Test for second characteristic numeral 5

(id 2071) Machine protected against water jets

Accredited [IEC17025]: YES Execution time [day]: 0 Unit price #1 [EUR]: 500.00 Limit #1: up to 30 kg Unit price #2 [EUR]: 1,100.00 Limit #2: over 30 kg Unit price #3 [EUR]: 1,300.00 Limit #3: over 100 kg

#### 10 - Test for second characteristic numeral 6

(id 2072) Machine protected against heavy seas

Accredited [IEC17025]: YES Execution time [day]: 0 Unit price #1 [EUR]: 500.00 Limit #1: up to 30 Kg Unit price #2 [EUR]: 1,000.00 Limit #2: from 31 to 100 Kg Unit price #3 [EUR]: 1,200.00 Limit #3: over 100 Kg

#### 10 - Test for second characteristic numeral 7

(id 2073) Machine protected against the effects of immersion

Accredited [IEC17025]: YES Execution time [day]: 0 Unit price #1 [EUR]: 500.00 Limit #1: up to 10 Kg Unit price #2 [EUR]: 100.00 Limit #2: over 10 Kg

#### 10 - Test for second characteristic numeral 8

(id 2074) Machine protected against the effects of continuous submersion

Accredited [IEC17025]: YES Execution time [day]: 0 Unit price #1 [EUR]: 800.00 Limit #1: Up to 10 Kg Unit price #2 [EUR]: 1,500.00 Limit #2: Over 10 Kg

#### 10 - Test for second characteristic numeral 9

(id 2075) Machine protected against high pressure and high temperature water jets

Execution time [day]: 0 Accredited [IEC17025]: YES Unit price #1: ask to office Limit #1: Not relevant

#### 10 - OBSOLETE - Test for second characteristic numeral 4 (oscillating tube)

(id 2134) OBSOLETE - Machine protected against splashing water.

Execution time [day]: 1 Accredited [IEC17025]: NO Unit price #1: ask to office Limit #1: Not relevant



# PRICE LIST / Listino prezzi your value certified § T - LABORATORY TESTS / Prove di laboratorio



QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



#### IEC 60068-2-11:2021

(id 1736) Basic environmental testing procedures - Part 2-11: Tests - Test Ka: Salt mist

#### - Test Ka: salt mist

(id 894) This test method is intended for assessing the corrosion resistance of electrotechnical products components, equipment and materials in a salt mist environment. Its objective is to verify that the comparative quality of a metallic material, with or without corrosion protection, is maintained when exposed to salt mist. This test method is useful for evaluating the quality and the uniformity of coatings applied to protect metals against corrosion. It is particularly useful for detecting discontinuities, such as pores and other defects, in certain metallic, organic, anodic oxide and conversion coatings.

#### IEC 60068-2-21:2021

(id 1133) Environmental testing - Part 2-21: Tests - Test U: Robustness of terminations and integral mounting devices

#### 4 - Tests Ua: Robustness of terminals against axial stresses

(id 2333) The purpose of these tests is to verify that the terminations and their attachments to the body of the component will withstand such axial stresses as are likely to be applied during normal assembly or handling operations.

Accredited [IEC17025]: YES Execution time [day]: 0
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 5 - Tests Ub: Robustness of terminals against bending stresses

(id 2334) The purpose of this test is to verify that pliable and rigid terminations and attachment of such terminations to the body of the component shall withstand such bending loads as are likely to be applied during normal assembly or handling operations.

#### 6 - Test Uc: Torsion

(id 2335) The purpose of this test is to verify that the terminations and attachment of the terminations to the body of the component will withstand torsional forces that are likely to be applied during normal assembly or dismantling operations.

Accredited [IEC17025]: YES Execution time [day]: 0
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 7 - Test Ud: Torque

(id 2336) The purpose of this test is to verify that the terminations, the attachment of the terminations to the body of the component and integral mounting means shall withstand torque forces that are likely to be applied during normal assembly or handling operations.

Accredited [IEC17025]: YES Execution time [day]: 0
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### IEC 60068-2-27:2008

(id 310) Environmental testing -

Part 2-27: Tests - Test Ea and guidance: Shock document

#### **Unref. - Shock Vibration**

(id 3105) This part of IEC 60068 provides a standard procedure for determining the ability of a specimen to withstand specified severities of non-repetitive or repetitive shocks.



### PRICE LIST / Listino prezzi

#### § T - LABORATORY TESTS / Prove di laboratorio

## albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



#### IEC 60068-2-38:2021

(id 1995) Environmental testing - Part 2-38: Tests - Test Z/AD: Composite temperature/humidity cyclic test

#### 7.4 - Composite temperature, Humidity cyclic test

(id 2760) ---

Accredited [IEC17025]: NO Execution time [day]: 9
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### IEC 60068-2-52:2017

(id 819) Environmental testing - Part 2-52: Tests - Test Kb: Salt mist, cyclic (sodium chloride solution)

#### 7 - Initial measurement in addition to visual

(id 1354) The test specimen(s) is visually inspected and, if necessary, electrically and mechanically checked as required by the relevant specification.

```
Accredited [IEC17025]: YES Execution time [day]: 1
Unit price #1: ask to office Limit #1: Not relevant
```

#### 9.4.2 - Test method 1

(id 1356) Test method 1 should be used to test products which are exposed to a marine environment (or which are used in close proximity to the sea) for much of their operational life (e.g. ship radar, deck equipment).

```
Accredited [IEC17025]: YES Execution time [day]: 28
Unit price #1 [EUR]: 6,000.00 Limit #1: Applicable to any sample
```

#### 9.4.3 - Test method 2

(id 1357) Test method 2 should be used to test products which may be exposed to the marine environment (or which are used in close proximity to the sea) from time to time but will normally be protected by an enclosure (e.g. navigational equipment which will normally be used on the bridge or in a control room).

```
Accredited [IEC17025]: YES Execution time [day]: 3
Unit price #1 [EUR]: 2,000.00 Limit #1: Small salt chamber
Unit price #2 [EUR]: 2,600.00 Limit #2: Medium salt chamber
Unit price #3 [EUR]: 3,600.00 Limit #3: Big salt chamber
```

#### 9.4.4 - Test method 3

(id 1358) Test method 3 can be used for products where, under normal use, there is a frequent change between salt-laden and dry atmosphere, for example automobiles and their parts.

```
Accredited [IEC17025]: YES Execution time [day]: 7
Unit price #1 [EUR]: 1,600.00 Limit #1: Small salt chamber
Unit price #2 [EUR]: 2,000.00 Limit #2: Medium salt chamber
Unit price #3 [EUR]: 3,200.00 Limit #3: Big salt chamber
```

#### 9.4.5 - Test method 4

(id 1359) Test method 4 can be used for products where, under normal use, there is a frequent change between salt-laden and dry atmosphere, for example automobiles and their parts.

```
Accredited [IEC17025]: YES Execution time [day]: 14
Unit price #1 [EUR]: 3,200.00 Limit #1: Small salt chamber
Unit price #2 [EUR]: 4,000.00 Limit #2: Medium salt chamber
Unit price #3 [EUR]: 5,500.00 Limit #3: Big salt chamber
```

#### 9.4.6 - Test method 5

(id 1360) Test method 5 can be used for products where, under normal use, there is a frequent change between salt-laden and dry atmosphere, for example automobiles and their parts.

```
Accredited [IEC17025]: YES Execution time [day]: 28
Unit price #1 [EUR]: 6,000.00 Limit #1: Not relevant
```

#### 9.4.7 - Test method 6

(id 1361) Test method 6 can be used for products where, under normal use, there is a frequent change between salt-laden and dry atmosphere, for example automobiles and their parts.

```
Accredited [IEC17025]: YES Execution time [day]: 56
Unit price #1 [EUR]: 8,000.00 Limit #1: Not relevant
```



# PRICE LIST / Listino prezzi § T - LABORATORY TESTS / Prove di laboratorio

### albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



#### 9.4.8 - Test method 7

(id 1362) Test method 7 defines a specific number of test cycles that include the spraying of a neutral salt solution, followed by dry conditions and humid conditions. The process can be used as a general corrosion test for many materials including automobiles and their parts.

Accredited [IEC17025]: YES	Execution time [day]: 20
Unit price #1 [EUR]: 3,000.00	Limit #1: from 1 to 10 days
Unit price #2 [EUR]: 4,000.00	Limit #2: from 15 to 30 days
Unit price #3 [EUR]: 5.000.00	Limit #3: from 50 to 60 days

#### 9.4.9 - Test method 8

(id 1363) Test method 8 defines a specific number of test cycles that include the spraying of an acidified salt solution, followed by dry conditions and humid conditions. The process can be used to induce a corrosion that occurs in acidified salt environment.

#### 10 - Final recovery (washing and drying)

(id 1364) As specified by the relevant specification

#### 11 - Final measurements in addition to visual

(id 1365) The test specimen(s) is visually inspected and, if necessary, electrically and mechanically checked as required by the relevant specification.

```
Accredited [IEC17025]: YES Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant
```

#### IEC 60068-2-60:2015

(id 198) Environmental testing - Part 2-60: Tests - Test Ke: Flowing mixed gas corrosion test

#### Ke - Flowing mixed gas corrosion test - Method 4

(id 1025) Flowing mixed gas corrosion test - Method 4

This test determines the corrosive influence of operating and storage indoor environments on electrotechnical products components, equipment and materials, particularly contacts and connections, considered separately, integrated into a subassembly or assembled as a complete equipment.

#### IEC 60068-2-64:2008

(id 2220) Environmental testing - Part 2-64: Tests - Test Fh: Vibration, broadband random and guidance

#### Unref. - Random Vibration

(id 3104) This part of IEC 60068 demonstrates the adequacy of specimens to resist dynamic loads without unacceptable degradation of its functional and/or structural integrity when subjected to the specified random vibration test requirements.

```
Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 2,000.00 Limit #1: Up to 24h
```

#### IEC 60068-2-68:1994

(id 22) Environmental testing - Part 2: Tests - Test L: Dust and sand

#### 6.1 - Test L: Dust and sand - Method Lc1: recirculating chamber

(id 279) This test is performed in order to determine the resistance of equipment to corrosion from dust and sand. It is performed by reproducing simulated environmental conditions of a 'desert storm'.

Where the air pressure within the specimen differs from the atmospheric air pressure, the specimen is Category 1. In this case, a pressure reduction as stated in the relevant standard is applied during the test. Otherwise, if the air pressure in the specimen is that of the ambient air pressure, the sample category is 2 and no pressure reduction is applied.

The particle size distribution of the sand used is:

Smaller than 150 micron: 9.5% by weight Smaller than 212 micron: 17.0% by weight Smaller than 300 micron: 30.5% by weight Smaller than 425 micron: 51.5% by weight

page 98/234



### PRICE LIST / Listino prezzi your value certified § T - LABORATORY TESTS / Prove di laboratorio

### albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

#### TNDEX

Smaller than 600 micron: 77.5% by weight Smaller than 850 micron: 93.5% by weight Bigger than 850 micron: 6.5% by weight

Accredited [IEC17025]: YES Execution time [day]: 5 Unit price #1 [EUR]: 3,000.00 Limit #1: 2hours - Relative humidity <25% - Wind speed  $20\sim25m/s$  - Sand Concentration 5 g / - Sand type: quartz Unit price #2 [EUR]: 4,500.00 Limit #2: 4hours - Relative humidity <25% - Wind speed 20~25m/s - Sand Concentration 5 g / - Sand type: quartz Limit #3: 24hours - Relative humidity <25% - Wind speed 20~25m/s - Sand Concentration 5 g Unit price #3 [EUR]: 8,000.00 m3 - Sand type: quartz

#### unref - Test L: Dust and sand - Method Lc1 (reduced)

(id 1346) (\*) This test is a variation of the test method Lc1 performed in order to prepare the sample to temperature tests. The sand is not recirculating into the chamber but it is just deposited onto the sample in order to form a sand layer.

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 600.00 Limit #1: 3h - Sand type: quartz

#### IEC 60068-2-6:2007

(id 934) Environmental testing - Part 2-6: Tests - Test Fc: Vibration (sinusoidal)

#### unref - Test Fc: Vibration (sinusoidal)

(id 1568) ---

Accredited [IEC17025]: NO Execution time [day]: 0 Subcontract: GESTLABS S.R.L. Unit price #1 [EUR]: 1,500.00 Limit #1: Not relevant

#### IEC 60068-2-74:1999

(id 2138) Environmental testing - Part 2: Tests - Test Xc: Fluid contamination

#### 10 - EXTENDED CONTAMINATION (CLASS C)

(id 2971) This part of IEC 60068 gives a method of test which provides a standard procedure to determine the ability of components, equipments or their constituent materials, hereinafter referred to as specimen, to withstand accidental contact with fluids, without being unacceptably affected. This procedure is not intended to demonstrate the operation of an equipment which is normally immersed in a fluid.

Accredited [IEC17025]: NO Execution time [day]: 5 Unit price #1 [EUR]: 1,200.00 Limit #1: Applicable to any sample

#### IEC 60068-2-75:2014+AMD1:2025

(id 2215) Environmental testing - Part 2: Tests - Test Eh: Hammer tests

#### 5 - Test Eha - Pendulum hammer

(id 668) ---

Accredited [IEC17025]: YES Execution time [dav]: 0 Unit price #1 [EUR]: 300.00 Limit #1: Up to 5 points Unit price #2 [EUR]: 600.00 Limit #2: from 6 to 10 points Unit price #3 [EUR]: 800.00 Limit #3: special condition

#### 7 - Test Ehc - Vertical hammer

Accredited [IEC17025]: YES Execution time [day]: 0 Limit #1: Up to 5 points Unit price #1 [EUR]: 300.00 Unit price #2 [EUR]: 600.00 Limit #2: from 6 to 10 points Unit price #3 [EUR]: 800.00 Limit #3: special condition



### **PRICE LIST / Listino prezzi** your value certified § T - LABORATORY TESTS / Prove di laboratorio

## albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version



#### IEC 60077-2:2017

(id 1746) Railway applications - Electric equipment for rolling stock - Part 2: Electrotechnical components - General rules

#### 9.3.3.2 - Operating Limits

(id 2236) ---

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 120.00 Limit #1: Applicable to any sample

#### IEC 60079-0:2017+COR1:2020

(id 1605) Explosive atmospheres - Part 0: Equipment - General requirements

#### 26.4.2 - Resistance to impact

(id 1421) Evaluation of the resistance to impact of equipment intended for use in explosive gas atmosphere.

```
Accredited [IEC17025]: YES
                                  Execution time [day]: 0
Unit price #1 [EUR]: 200.00
                                            Limit #1: Applicable to any sample
Unit price #2 [EUR]: 200.00
                                            Limit #2: Applicable to any sample
Unit price #3 [EUR]: 200.00
                                            Limit #3: Applicable to any sample
```

#### 26.7.2 - Test temperatures for non metallic enclosures or non metallic parts of enclosures

(id 1422) This is not a test. This clause expresses conditions for upper and lower service temperature during test temperatures.

```
Accredited [IEC17025]: YES
                                  Execution time [day]: 0
Unit price #1: ask to office
                                            Limit #1: Applicable to any sample
```

#### 26.11 - Resistance to chemical agents for Group I electrical equipment

(id 1423) Prolonged immersion of samples in 2 types of fluid (an oil and a hydraulic liquid for mining application) followed by final measurements (resistance to impact, drop test if applicable, test for degrees of protection)

```
Accredited [IEC17025]: NO
                                 Execution time [day]: 3
Unit price #1 [EUR]: 500.00
                                           Limit #1: up to 21
Unit price #2 [EUR]: 1,500.00
                                           Limit #2: from 21 up to 201
Unit price #3 [EUR]: 2,200.00
                                           Limit #3: from 211 up to 501
```

#### 26.12 - Earth continuity

(id 1424) This test applies to non-metallic walled enclosures provided with an internal earth continuity plate.

NOTE An internal earth continuity plate is often fitted, for example, to allow for use of metallic cable glands without the use of separate individual earthing tags. This test includes a conditioning with the same test conditions of 26.8 of IEC 60079-0 (Thermal endurance to heat) plus a 336 h period in an air oven at at least

```
Accredited [IEC17025]: YES
                                  Execution time [day]: 0
                                           Limit #1: Applicable to any sample
Unit price #1 [EUR]: 300.00
```

#### 26.13 - Surface resistance test

(id 1425) Evaluation of the surface resistance of non-metallic materials of equipment intended for use in explosive gas atmosphere. The electrodes are painted with a conducting paint with a solvent which has no significant effect on the surface resistance.

```
Accredited [IEC17025]: YES
                                 Execution time [dav]: 1
                                            Limit #1: Conditioning at 50% Rh
Unit price #1 [EUR]: 300.00
Unit price #2 [EUR]: 500.00
                                            Limit #2: Conditioning at 30% Rh
```

#### 26.14 - Measurement of capacitance

(id 1426) Evaluation of the capacitance of equipment intended for use in explosive gas atmosphere

```
Accredited [IEC17025]: YES
                                 Execution time [day]: 0
                                            Limit #1: Applicable to any sample
Unit price #1 [EUR]: 300.00
```

#### 26.15 - Verification of ratings of ventilating fans

```
Accredited [IEC17025]: NO
                                  Execution time [day]: 0
Unit price #1 [EUR]: 500.00
                                            Limit #1: Applicable to any sample
```

#### 26.16 - Alternative qualification of elastomeric sealing O-rings

(id 1428) Qualification of elastomeric sealing O-ring measuring the compression before and after a thermal endurance to heat and thermal endurance to cold test. The compression set value 'c' describes the ability for the sealing to return to its initial dimension after being subject to compression.

It is calculated as follows:

 $c = (t0 \times t1)/(t0 \times tS)$ , where

t0 is the initial thickness of the sealing ring measured at (20  $\pm$  5) °C temperature,

page 100/234



## **PRICE LIST / Listino prezzi**

### your value certified § T - LABORATORY TESTS / Prove di laboratorio

### albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

#### TNDEX

tS is the thickness when compressed as intened in the equipment,

t1 is the thickness of the sealing ring measured at (20  $\pm$  5) °C after thermal endurance.

Accredited [TEC17025]: NO Execution time [day]: 30 Unit price #1 [EUR]: 1,800.00 Limit #1: Applicable to any sample

#### 26.3 - Tests in explosive test mixtures

(id 1429) This clause of the standard describes how to carry out tests in explosive test mixtures, according to the relevant standards.

In general, the sample under test is operated according to the parameters provided by the customer immersed in an explosive test mixture.

The explosive mixture is dictated by the relevant standard or agreed with the customer / inspector among those for the various gas groups in accordance with the ATEX directive and standards.

```
Accredited [IEC17025]: YES
                                 Execution time [day]: 1
Unit price #1 [EUR]: 1,500.00
                                            Limit #1: Test inside small vessel
Unit price #2 [EUR]: 2,200.00
                                            Limit #2: Test in explosion vented chamber
Unit price #3 [EUR]: 4,000.00
                                           Limit #3: Outdoor test
```

#### 26.4.3 - Drop test

(id 1430) Evaluation of the resistance to drop of hand-held electrical equipment or electrical equipment carried on the person, ready for use.

```
Accredited [IEC17025]: YES
                                 Execution time [day]: 0
Unit price #1 [EUR]: 100.00
                                           Limit #1: Applicable to any sample
```

#### 26.4.5 - Degree of protection (IP) by enclosures

(id 1431) Degree of protection against access to hazardous parts, solid foreign objects or water to be verified in accordance to IEC 60529 or IEC 60034-5 in the case of rotating electrical machines.

```
Accredited [IEC17025]: YES
                                  Execution time [day]: 2
Unit price #1 [EUR]: 600.00
                                            Limit #1: Applicable to any sample
```

#### 26.5.1.2 - Service temperature

(id 1432) Evaluation of the service temperature of an enclosure intended for use in an explosive gas atmosphere

```
Accredited [IEC17025]: YES
                                  Execution time [day]: 2
Unit price #1 [EUR]: 600.00
                                            Limit #1: Applicable to any sample
Unit price #2 [EUR]: 1,000.00
                                            Limit #2: With any additional dust layer
Unit price #3 [EUR]: 1,400.00
                                            Limit #3: Windings/bushings terminals temperature
```

#### 26.5.1.3 - Maximum surface temperature

(id 1435) Evaluation of the maximum surface temperature of equipment intended for use in explosive gas atmosphere. The test is performed by measuring the temperature in the hottest parts of the sample, detected after a termographic inspection with an infrared thermal camera.

```
Accredited [IEC17025]: YES
                                  Execution time [day]: 5
Unit price #1 [EUR]: 800.00
                                            Limit #1: Applicable to any sample
Unit price #2 [EUR]: 1,000.00
                                            Limit #2: With any additional dust layer
Unit price #3 [EUR]: 1,400.00
                                            Limit #3: Windings/bushings terminals temperature
```

#### 26.5.2 - Thermal shock test

(id 1436) Evaluation of the resistance to thermal shock of equipment intended for use in an explosive gas atmosphere. Glass parts of luminaires and windows of electrical equipment have to be subject to a thermal shock caused by a jet of water of about 1 mm diameter at a temperature of 5-15°C sprayed on them when they are at not less than the maximum service temperature.

```
Accredited [IEC17025]: YES
                                 Execution time [day]: 0
                                           Limit #1: Applicable to any sample
Unit price #1 [EUR]: 200.00
Unit price #2 [EUR]: 200.00
                                           Limit #2: Applicable to any sample
Unit price #3 [EUR]: 200.00
                                           Limit #3: Applicable to any sample
```

#### 26.5.3 - Small component ignition test

(id 1437) Test performed on small components in order to demonstrate that they do not cause temperature ignition of a flamable mixture

```
Accredited [IEC17025]: YES
                                 Execution time [day]: 0
Unit price #1 [EUR]: 1,200.00
                                            Limit #1: Samples with maximum ambient temperature <30°C or easily increaseable power during
                                            operation
Unit price #2 [EUR]: 1,600.00
                                            Limit #2: Samples with maximum ambient temperature between 30°C and 60°C or test under
                                            malfunction
```

#### 26.6 - Torque test for bushings

(id 1438) Test for the verification of resistance to torque on bushings subjected to torque during connection or disconnection of conductors

Accredited [IEC17025]: YES	Execution time [day]: 0
Unit price #1 [EUR]: 200.00	Limit #1: Only for bushings from M4 to M20
Unit price #2 [EUR]: 600.00	Limit #2: For bushings more than M20



### PRICE LIST / Listino prezzi § T - LABORATORY TESTS / Prove di laboratorio

### LABORATORT TESTS / FTOVE di laboratorio

### albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX

#### 26.8 - Thermal endurance to heat

(id 1439) Evaluation of the thermal endurance to heat of equipment intended for use in explosive gas atmosphere

#### 26.9 - Thermal endurance to cold

(id 1440) Evaluation of the thermal endurance to cold of equipment intended for use in explosive gas atmosphere

#### A.3.1.4 - Clamping test (non armoured and braided cables)

(id 1443) Tests of clamping of non armoured and braided cables: Tensile strenght

```
Accredited [IEC17025]: YES Execution time [day]: 1
Unit price #1 [EUR]: 300.00 Limit #1: Applicable to any sample
```

#### A.3.1.5 - Mechanical strength (non armoured and braided cables)

(id 1444) Test of clamping of non-armoured and braided cables: Mechanical strength

#### A.3.2.1.2 - Clamping test (armoured cables)

(id 1445) Test of clamping of armoured cables: Tensile test

#### A.3.2.1.3 - Mechanical strength (armoured cables)

(id 1446) Test of clamping of armoured cables: Mechanical strength

#### A.3.3 - Type test for resistance to impact (cable glands)

(id 1447) ---

```
Accredited [IEC17025]: YES Execution time [day]: 2
Unit price #1 [EUR]: 200.00 Limit #1: Not relevant
```

#### A.3.4 - Test for degree of protection (IP) of cable glands

(id 1448) The test shall be carried out in accordance with IEC 60529 with the specifications of A.3.4

```
Accredited [IEC17025]: YES Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant
```

#### 26.10 - Resistance to UV light

(id 1449) This test is an evaluation of the resistance to ultraviolet and visible light on equipment intended for use in explosive gas atmosphere.

Where the preparation of test samples in accordance with ISO 179 is not practical, the laboratory performs the test in the following alternative way (\*):

- The test is performed directly on the sample.
- Since the relevant irradiation to which the sample may be exposed during its lifetime is essentially made of a UV component and a visible component, the sample is submitted to both in two different times:
  - 60 kWh/m2 for the visible component (that corresponds to about 60h)
  - 15 kWh/m2 for the UV component (that corresponds to about 10 days).
- The values above are drawn from PV modules testing (Outdoor exposure test), because the PV samples are exposed to natural sunlight for their entire lifetime. For this reason, this values are deemed as suitable also for an Atex sample that may be exposed to the same natural sunlight.
- In that case, the evaluation criterion (the final impact bending test, see acceptance criteria below) will be replaced by a simple resistance to impact according to 26.9 of IEC 60079-0, with an impact energy of 2/4 J with a test mass of 1 kg falling vertically from a height 20/40 cm. The mass is fitted with an impact head made of hardened steel in the form of a hemisphere of 25 mm diameter.

Accredited	[I]	EC17025	]: NO	Execution	time	[day]	: 15							
Jnit price	#1	[EUR]:	600.00		Limit	#1:	Simplified	test	directly	on	the	sample,	with	impact
Jnit price	#2	[EUR]:	600.00		Limit	#2:	Simplified	test	directly	on	the	sample,	with	impact
Init price	#3	[EIIE]:	3 000 00		Limit	#3:	Test on nla	astic	hare as	T S O 1	179	with he	ndina	



## PRICE LIST / Listino prezzi

### § T - LABORATORY TESTS / Prove di laboratorio

### albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

TNDEX

### 26.17 - Transferred charge test

(id 1450) Possible subcontract (?)

Accredited [IEC17025]: NO Execution time [day]: 2 Unit price #1 [EUR]: 1,100.00 Limit #1: Not relevant

#### 25 - Compliance of prototype or sample with documents

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1 [EUR]: 200.00 Limit #1: Not relevant

#### A.3.1.1 - Tests of clamping of non armoured and braided cables: cable glands with clamping by the sealing ring

(id 1893) Sample preparation and prliminary test

Accredited [IEC17025]: YES Execution time [day]: 0 Limit #1: Applicable to any sample Unit price #1 [EUR]: 200.00

#### A.3.1.3 - Tests of clamping of non armoured and braided cables: cable glands with clamping by means of a clamping device

(id 1894) Sample preparation and preliminary test

Accredited [IEC17025]: YES Execution time [day]: 0 Unit price #1: ask to office Limit #1: Not relevant

#### A.3.1.2 - Tests of clamping of non armoured and braided cables: cable glands with clamping by filling compound

(id 1895) Sample preparation and preliminary test

Accredited [IEC17025]: YES Execution time [day]: 0 Unit price #1: ask to office Limit #1: Not relevant

#### A.3.2.1.1 - Tests of clamping of armoured cables: armouring clamped by a device integral to the gland

(id 1896) Sample preparation and preliminary test

Accredited [IEC17025]: YES Execution time [day]: 0 Unit price #1: ask to office Limit #1: Not relevant

#### Ann.A/G - Tests of cable glands: resting period

(id 1897) ---

Accredited [IEC17025]: YES Execution time [day]: 0 Unit price #1: ask to office Limit #1: Not relevant

#### Ann.A/G - Tests of cable glands: retightening with reference torque

Accredited [IEC17025]: YES Execution time [day]: 0 Unit price #1: ask to office Limit #1: Not relevant

#### 26.5.1.2 - Service temperature on heating resistor

(id 2017) Evaluation of the service temperature of an heating resistor on a customer's simulacrum

Accredited [IEC17025]: YES Execution time [day]: 5 Unit price #1 [EUR]: 1,000.00 Limit #1: up to 6 points Unit price #2 [EUR]: 2,500.00 Limit #2: from 7 to 30 points Unit price #3 [EUR]: 4,500.00 Limit #3: from 31 to 60 points

#### 26.3 - Tests in explosive test mixtures - on-site

(id 2916) This clause of the standard describes how to carry out tests in explosive test mixtures, according to the relevant standards.

In general, the sample under test is operated according to the parameters provided by the customer immersed in an explosive test mixture.

The explosive mixture is dictated by the relevant standard or agreed with the customer / inspector among those for the various gas groups in accordance with the ATEX directive and standards.

This test is performed on-site, by the customer location.

Accredited [IEC17025]: YES Execution time [day]: 1 Limit #1: Small sample (about 30x30x30 cm max) Unit price #1 [EUR]: 2,000.00 Unit price #2 [EUR]: 2,600.00 Limit #2: Medium sample (about 60x60x60 cm max) Unit price #3 [EUR]: 4,000.00 Limit #3: Big sample (100x100x100 cm max)

#### 26.5.1.2 - Service temperature on special sample

(id 2992) Evaluation of the service temperature of an enclosure intended for use in an explosive gas atmosphere

Accredited [IEC17025]: YES Execution time [day]: 2 Unit price #1 [EUR]: 2,400.00 Limit #1: Lead acid battery



# PRICE LIST / Listino prezzi § T - LABORATORY TESTS / Prove di laboratorio

## albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



#### 26.5.1.3 - Maximum surface temperature on special sample

(id 2993) Evaluation of the maximum surface temperature of equipment intended for use in explosive gas atmosphere. The test is performed by measuring the temperature in the hottest parts of the sample, detected after a termographic inspection with an infrared thermal camera.

```
Accredited [IEC17025]: YES Execution time [day]: 5
Unit price #1 [EUR]: 1,000.00 Limit #1: Small heating-dissipating sample
```

#### IEC 60079-11:2023

(id 1883) Explosive atmospheres - Part 11: Equipment protection by intrinsic safety "i"

#### 9.14.1.1 - General (Tests for cells, batteries and supercapacitors)

(id 2540) ---

```
Accredited [IEC17025]: YES

Unit price #1 [EUR]: 150.00

Unit price #2 [EUR]: 300.00

Limit #1: Only visual inspection and voltage measurements (usually primary batteries)

Limit #2: Visual inspection, voltage measurements and charge-dicharge cycles (secondary batteries)
```

#### 9.14.2 - Electrolyte leakage test for cells, batteries and supercapacitors

(id 2541) ---

Accredited [IEC1702!	3]: YES	Execution	time [	day	]: 1			
Unit price #1 [EUR]	300.00		Limit	#1:	up to	4 Ah	1	
Unit price #2 [EUR]	500.00		Limit	#2:	from	4 to	25	Ah
Unit price #3 [EUR]	1,000.00		Limit	#3:	more	than	25	Ah

#### 9.14.3 - Spark ignition and surface temperature of cells, batteries or supercapacitors

(id 2542) ---

#### 9.14.4 - Battery container pressure tests

(id 2543) Determination of venting pressure of battery container.

Accredited [IEC17025]: YES	Execution time [day]: 0
Unit price #1 [EUR]: 300.00	Limit #1: up to 4 Ah
Unit price #2 [EUR]: 500.00	Limit #2: from 4 to 25 Ah
Unit price #3 [EUR]: 800 00	Limit #3: more than 25 Ah

#### 9.14.5 - Battery resistance

(id 2544) ---

#### 9.11 - Tests for intrinsically safe apparatus containing piezoelectric devices

(id 2651) ---

```
Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 500.00 Limit #1: Applicable to any sample
```

#### 9.13 - Determination of parameters of loosely specified components

(id 2652) ---

```
Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 300.00 Limit #1: Single measurement
Unit price #2 [EUR]: 600.00 Limit #2: Up to 5 measurements on the same sample
Unit price #3 [EUR]: 800.00 Limit #3: Up to 10 measurements on the same sample
```

#### 7.11 - Fuses

(id 2653) Where intrinsic safety depends on the opening of a fuse, for the purposes of assessment the fuse shall be considered capable of carrying a current of 1,7 In continuously. A single suitably rated fuse is sufficient to provide protection.

Accredited [IEC17025]: NO	Execution time [day]: 1
Unit price #1 [EUR]: 500.00	Limit #1: Applicable to any sample



### § T - LABORATORY TESTS / Prove di laboratorio

**PRICE LIST / Listino prezzi** 

### albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX

9.2.6 - Circuits with both inductance and capacitance

(id 2654) For circuits containing both inductance and capacitance where the capacitive stored energy can reinforce the power source feeding an inductor or vice versa, the circuit shall be assessed for compliance.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1: ask to office Limit #1: Not relevant

#### 9.1 - Spark ignition test

(id 2655) Investigation on effect of opening and closing circuit in explosive atmosphere.

As prescribed by the strandard at §9.1.2, in order to perform this test, the instrument 4.162 Spark Ignition Testing Apparatus is used and it is calibrated before use.

Accredited [IEC17025]: YES

Unit price #1 [EUR]: 800.00

Limit #1: Up to n. 10 test points with same sample configurations and same gas mixture

Unit price #2 [EUR]: 1,000.00

Limit #2: Up to n. 10 test points with different sample configurations and same gas mixture

Unit price #3 [EUR]: 1,500.00

Limit #3: Up to n. 20 test points with different sample configurations and same gas mixture

#### 9.4.2 - Acceptability of encapsulated or coated fuses

(id 2656) Where fuses are required to be encapsulated or coated, and the material could enter the interior of the fuse and affect safety, the following test shall be performed on five samples of each fuse before encapsulation or coating is applied.

Accredited [IEC17025]: YES Execution time [day]: 0
Unit price #1 [EUR]: 300.00 Limit #1: Applicable to any sample

#### 9.4.4 - Cable pull test

(id 2657) ---

Accredited [IEC17025]: YES Execution time [day]: 0
Unit price #1 [EUR]: 200.00 Limit #1: Applicable to any sample

#### 9.6 - Dielectric strength test

(id 2658) ---

#### 9.5 - Current carrying capacity of infallible printed circuit board connections

(id 2659) ---

Accredited [IEC17025]: YES Execution time [day]: 0
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 9.4.1 - Casting compund

(id 2660) ---

#### 9.4.3 - Partitions

(id 2661) ---

Accredited [IEC17025]: YES Execution time [day]: 0
Unit price #1 [EUR]: 300.00 Limit #1: Applicable to any sample

#### 9.10.2.2 - Overload test at the receiver side

(id 2662) This test is intended to define the maximum temperature at the receiver side when the device experiences an overload.

Accredited [IEC17025]: YES Execution time [day]: 0
Unit price #1 [EUR]: 100.00 Limit #1: Applicable to any sample

#### 9.10.2.3 - Overload test at the transmitter side

(id 2663) This test is intended to define the maximum temperature at the transmitter side when the device experiences an overload.

Accredited [IEC17025]: YES Execution time [day]: 0
Unit price #1 [EUR]: 100.00 Limit #1: Applicable to any sample

#### 9.10.2.4 - Thermal conditioning and dielectric strength test

(id 2664) After overloading, the devices shall be subjected to thermal conditioning and dielectric strength tests.

Accredited [IEC17025]: YES Execution time [day]: 0
Unit price #1 [EUR]: 300.00 Limit #1: Applicable to any sample



# PRICE LIST / Listino prezzi § T - LABORATORY TESTS / Prove di laboratorio

## albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



#### 7.12.2 - Construction of cells and batteries used in intrinsically safe apparatus

(id 2759) Verification of constructional requirements and basic specification

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 200.00 Limit #1: Applicable to any sample

#### IEC 60079-13:2017

(id 397) Explosive atmospheres - Part 13: Equipment protection by pressurized room "p" and artificially ventilated room "v"

#### 7.5.3 - Purging test for for artificially ventilated rooms

(id 1267) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

#### 7.5.4 - Minimum ventilation flow rate test for artificially ventilated rooms

(id 1268) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

#### 7.5.5 - Confirmation of the ratings of the safety devices for artificially ventilated rooms

(id 1269) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

#### 7.5.6 - Verification of sequence of operation of the safety devices for artificially ventilated rooms

(id 1270) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

#### 7.5.7.2 - Ventilation test for artificially ventilated rooms

(id 1271) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

#### 7.5.7.3.1 - Air-flow sensing devices for artificially ventilated rooms

(id 1272) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

#### 7.5.7.3.2 - Gas detectors for artificially ventilated rooms

(id 1273) ---

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1: ask to office Limit #1: Not relevant

#### 6.4.3 - Overpressure test for pressurized rooms

(id 1274) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

#### 6.4.4 - Purging test for pressurized rooms

(id 1275) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

#### 6.4.5 - Minimum pressure differential test for pressurized rooms

(id 1276) ---

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1: ask to office Limit #1: Not relevant



## PRICE LIST / Listino prezzi

### § T - LABORATORY TESTS / Prove di laboratorio

### albarubens QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



#### 6.4.6 - Confirmation of the ratings of the safety devices for pressurized rooms

(id 1277) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

#### 6.4.7 - Verification of sequence of operation of the safety devices for pressurized rooms

(id 1278) --

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

#### IEC 60079-15:2017

(id 790) Explosive atmospheres - Part 15: Equipment protection by type of protection "n"

#### 11.2.2 - Tests for sealed devices: Voltage test

(id 1455) 3 samples

Accredited [IEC17025]: YES Execution time [day]: 0
Unit price #1 [EUR]: 300.00 Limit #1: Applicable to any sample

#### 11.2.3 - Leakage test on sealed devices

(id 1456) 3 samples

Accredited [IEC17025]: YES Execution time [day]: 1
Unit price #1 [EUR]: 300.00 Limit #1: Applicable to any sample

#### 11.3.2 - Type test requirements for restricted breathing enclosures

(id 1458) Equipment where the volume does not change due to pressure

Accredited [IEC17025]: YES Execution time [day]: 1
Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

#### 6.2 - Electric strength insulation from earth or frame

(id 1468) ---

#### 11.2.1 - Tests for sealed devices: Conditioning

(id 1469) Thermal stress as preconditioning of samples.

The conditioning in accordance with IEC 60079-0 may be substituted.

Accredited [IEC17025]: YES Execution time [day]: 10
Unit price #1 [EUR]: 500.00 Limit #1: Applicable to any sample

#### 11.1.2.2 - Non-ignition test for non-incendive components

(id 1472) The device or component, which shall be arranged to have the most adverse dimensions permitted by the construction drawings, shall be filled with and surrounded by an explosive mixture according to the stated group of the equipment and operated at least 50 times without igniting the test mixture.

#### 5 - Maximum surface temperatures

(id 1474) The maximum surface temperature shall be determined in accordance with the determination of maximum temperature classification requirements of IEC 60079-0. The surface to be considered shall be the external surface of the equipment including the surface of internal parts of non-incendive components to which the explosive gas atmosphere has access.

Accredited [IEC17025]: YES Execution time [day]: 5
Unit price #1: ask to office Limit #1: not relevant

#### 11.3.3 - Alternative type test for restricted breathing enclosures (equipment where the volume changes due to pressure)

(id 1476) ---

Accredited [IEC17025]: YES Execution time [day]: 2
Unit price #1 [EUR]: 800.00 Limit #1: Not relevant



### **PRICE LIST / Listino prezzi** your value certified § T - LABORATORY TESTS / Prove di laboratorio

## albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

TNDEX

#### 11.1.1 - Preparation of non-incendive component samples

(id 2316) Preconditioning of samples, consisting in 6000 operations at maximum rating; applicable to switches, relays etc.

```
Accredited [IEC17025]: NO
                                 Execution time [day]: 0
Unit price #1 [EUR]: 300.00
                                           Limit #1: No necessity of automated setup
Unit price #2 [EUR]: 1,000.00
                                           Limit #2: Required automated setup
```

#### IEC 60079-1:2014+COR1: 2018

(id 9) Explosive atmospheres - Part 1: Equipment protection by flameproof enclosures "d"

#### 15.2.2 - Determination of explosion pressure

(id 99) Measurement of enclosure internal overpressure

```
Accredited [IEC17025]: YES
                                 Execution time [dav]: 1
Unit price #1 [EUR]: 800.00
                                           Limit #1: V enclosure < 30 1
Unit price #2 [EUR]: 1,600.00
                                            Limit #2: 30 1 < V enclosure < 150 1
Unit price #3 [EUR]: 2,800.00
                                            Limit \#3: V enclosure > 150 l or special condition (not standard temperatures, powered on,
```

#### 15.2.3.2 - Overpressure test (static)

(id 100) First method (static)

```
Accredited [IEC17025]: YES
                                  Execution time [day]: 1
Unit price #1 [EUR]: 400.00
                                            Limit #1: V enclosure < 30 1
Unit price #2 [EUR]: 800.00
                                            Limit #2: 30 1 < V enclosure < 150 1
Unit price #3 [EUR]: 1,400.00
                                            Limit #3: V enclosure > 150 l
```

#### 15.3 - Test for non-transmission of an internal ignition

(id 101) Capability of flameproof path to stop the flame

```
Accredited [IEC17025]: YES
                                  Execution time [day]: 1
Unit price #1 [EUR]: 1,000.00
                                            Limit #1: V enclosure < 30 l
Unit price #2 [EUR]: 1,800.00
                                            Limit #2: 30 1 < V enclosure < 150 1
Unit price #3 [EUR]: 3,800.00
                                            Limit #3: V enclosure > 150 l
```

#### 15.4.2 - Tests of ability of the enclosure to withstand pressure

(id 397) The overpressure test in accordance with 15.2.3 applies, but with a thin flexible membrane fitted to the inner surfaces of the breathing and draining devices.

```
Accredited [IEC17025]: YES
                                  Execution time [day]: 5
                                            Limit #1: V enclosure < 40 l
Unit price #1 [EUR]: 500.00
Unit price #2 [EUR]: 800.00
                                            Limit #2: 440 1 < V enclosure < 200 1
Unit price #3 [EUR]: 1,000.00
                                            Limit #3: V enclosure > 200 l
```

#### 5 - Flameproof joints evaluation

(id 546) Physical and/or documental evaluation of flameproof joints

```
Accredited [IEC17025]: YES
                                  Execution time [day]: 0
Unit price #1 [EUR]: 600.00
                                           Limit #1: Small enclosure
Unit price #2 [EUR]: 800.00
                                            Limit #2: Medium enclosure
Unit price #3 [EUR]: 1,000.00
                                            Limit #3: Big enclosure
```

#### C.3.1 - Sealing test for flameproof entry devices

(id 745) After verifying the requirements for thermal endurance to heat and thermal endurance to cold and after the required sample preparation, an hydraulic pressure is applied to in order to verify the sealing.

This test may apply to:

- Cable glands and conduit sealing devices with sealing ring
- Cable glands sealed with setting compound
- Conduit sealing devices sealed with setting compound

```
Accredited [IEC17025]: YES
                                  Execution time [day]: 1
Unit price #1 [EUR]: 300.00
                                            Limit #1: Applicable to any sample
```

#### C.3.2 - Test of mechanical strength for flameproof entry devices

```
Accredited [IEC17025]: YES
                                  Execution time [day]: -1
Unit price #1 [EUR]: 200.00
                                            Limit #1: Applicable to any sample
```



# PRICE LIST / Listino prezzi § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



#### C.3.3.1 - Torque test for Ex blanking elements

(id 747) ---

Accredited [IEC17025]: YES Execution time [day]: -1

Unit price #1: ask to office Limit #1: Applicable to any sample

#### C.3.4.1 - Torque test for Exd thread adapters

(id 749) ---

Unit price #1 [EUR]: 200.00 Limit #1: not relevant

#### **B.1 - Sintered metal elements**

(id 836) Possible subcontract

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1: ask to office Limit #1: 1

#### **B.2 - Pressed metal wire elements**

(id 837) Possible subcontract

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Applicable to any sample

#### **B.3 - Metal foam elements**

(id 838) Possible subcontract

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Applicable to any sample

#### G.4.1 - Overpressure test for flameproof enclosures with an internal source of release

(id 879) Additional requirements for flameproof enclosures with an internal source of release

Accredited [IEC17025]: YES Execution time [day]: 0

Unit price #1 [EUR]: 1,000.00 Limit #1: Applicable to any sample

#### G.4.2 - Leakage test for an infallible containment system

(id 880) Additional requirements for flameproof enclosures with an internal source of release

Accredited [IEC17025]: YES Execution time [day]: 0

Unit price #1 [EUR]: 1,000.00 Limit #1: Applicable to any sample

#### G.4.3 - Leakage test for a containment system with a limited release

(id 881) Additional requirements for flameproof enclosures with an internal source of release

Accredited [IEC17025]: YES Execution time [day]: 1

Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

#### E.3.5 - General requirements for cells (or batteries) inside flameproof enclosures

(id 967) Verifications to be performed before and after the tests of enclosures required by IEC 60079-0.

Accredited [IEC17025]: YES Execution time [day]: 0

Unit price #1 [EUR]: 300.00 Limit #1: Applicable to any sample

#### E.4.1.1 - Prevention of excessive temperature and cell damage - Flameproof enclosures

(id 968) Verification of the battery conditions under short circuit discharge or of the presence and compliance of a safety device

Unit price #1 [EUR]: 1,000.00 Limit #1: Applicable to any sample

#### unref - Setup of unsealed equipments with Tamb under -20°C

(id 974) Preparation of heated sensors and valves to operate Pref measurement inside climatic chamber, where precompression is not possible. Including climatic chamber cost.

Accredited [IEC17025]: NO Execution time [day]: 2

Unit price #1 [EUR]: 600.00 Limit #1: Not relevant

#### 19.4 - Test of erosion by flame

(id 1041)

Accredited [IEC17025]: YES Execution time [day]: 4

Unit price #1 [EUR]: 3,000.00 Limit #1: Small equipment



# PRICE LIST / Listino prezzi

## § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX

#### C.3.3.2 - Over-pressure test for Ex blanking elements

(id 1042) ---

Accredited [IEC17025]: YES Execution time [day]: 1
Unit price #1: ask to office Limit #1: Not relevant

#### 15.5 - Tests for "dc" devices

(id 1123) --

#### 15.4.4.3 - Non-transmission test for breathing and draining devices (Groups I, IIA and IIB)

(id 1183) The non-transmission test of 15.3.2 shall be applied

Accredited [IEC17025]: YES Execution time [day]: 2
Unit price #1: ask to office Limit #1: Not relevant

#### C.3.4.2 - Impact test for Ex thread adapters

(id 1192) A sample Ex thread adapter of each size is screwed into a test-block. A solid steel or brass bar is then screwed into the Ex thread adapter with the torque required by the standard. The assembly is then subjected to a test of resistance to impact, according to the requirements given in IEC 60079-0, appliying the impact at right angles to the axis of the bar and as near to the end of the bar as practicable.

Accredited [IEC17025]: YES Execution time [day]: 1
Unit price #1 [EUR]: 200.00 Limit #1: Not relevant

#### C.3.4.3 - Over-pressure test for Ex thread adapters

(id 1193) ---

Accredited [IEC17025]: YES Execution time [day]: 0
Unit price #1 [EUR]: 300.00 Limit #1: Not relevant

#### 10.8 - Mechanical strength

(id 1261) Type tests for breathing and draining devices which form part of a flameproof enclosure

#### 10.9.3.3 - Test for non-transmission of an internal ignition for breathing and draining devices used as Ex components

(id 1262) Type tests for breathing and draining devices used as Ex components of a flameproof enclosure

#### 10.9.3.2 - Thermal tests for breathing and draining devices used as Ex components

(id 1263) Type tests for breathing and draining devices used as Ex components of a flameproof enclosure

#### 10.9.3.4 - Test of the ability of the breathing and draining device to withstand pressure

(id 1264) Type tests for breathing and draining devices used as Ex components of a flameproof enclosure

Accredited [IEC17025]: YES Execution time [day]: 1
Unit price #1 [EUR]: 500.00 Limit #1: Not relevant

#### C.2.2 - Flameproof joints evaluation (cable glands)

(id 1569) ---

Accredited [IEC17025]: YES Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

#### G.3.3 - Containment system with a limited release - Verification of internal pressure increase

(id 1624) --

#### 4.4.2 - Construction of "dc" devices

(id 1900) ---

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 300.00 Limit #1: Not relevant

page 110/234



PRICE LIST / Listino prezzi

## § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

TNDEX

#### 15.2.3.3 - Overpressure test (dynamic)

(id 2208) Second method (dynamic)

Accredited [IEC17025]: NO Execution time [day]: 2 Unit price #1 [EUR]: 500.00 Limit #1: V enclosure < 40 l Unit price #2 [EUR]: 1,000.00 Limit #2: 40 1 < V enclosure < 150 1 Unit price #3 [EUR]: 1,800.00 Limit #3: V enclosure > 150 l

#### - New test

(id 2913) ---

Accredited [IEC17025]: NO Execution time [day]: 2 Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

## IEC 60079-26:2021

(id 1222) Explosive atmospheres - Part 26: Equipment with separation elements or combined Levels of Protection

#### 7.2 - Separation elements

(id 2914) Separation elements in accordance with Clause 5 shall be tested in such a way that the operational parameters (for example, pressure or temperature limits) stated by the manufacturer are verified.

Accredited [IEC17025]: NO Execution time [day]: 2 Unit price #1 [EUR]: 300.00 Limit #1: Applicable to any sample

## IEC 60079-28:2015

(id 109) Explosive atmospheres - Part 28: Protection of equipment and transmission systems using optical radiation

#### 6.3.1 - Ignition tests with continuous wave radiation and pulses above 1 s duration

(id 1213) ---

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1: ask to office Limit #1: Not relevant

#### 6.3.2 - Ignition tests with single pulses less than 1 ms duration

(id 1214) ---

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1: ask to office Limit #1: Not relevant

#### 6.3.3 - Tests for pulse trains and pulses from 1 ms to 1 s duration

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1: ask to office Limit #1: Not relevant

#### 6.2.3 - Reference test for continuous wave radiation and pulses above 1s duration

(id 1216) Verification of the suitability of test set-up for type tests

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1: ask to office Limit #1: Not relevant

#### 6.2.4 - Reference test for pulsed radiation below 1ms pulse duration

(id 1218) Verification of the suitability of test set-up for type tests

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1: ask to office Limit #1: Not relevant

#### 5.2.2laser - Inherently safe optical radiation "op is" - Continuous wave laser radiation

(id 1257) The test is intended to qualify continuous laser sources, ensuring the power is lower than the ignition limits

Accredited [IEC17025]: NO Execution time [day]: 0 Limit #1: Applicable to any sample Unit price #1 [EUR]: 500.00

#### 5.2.3 - Inherently safe optical radiation "op is" - Pulsed radiation

Accredited [IEC17025]: YES Execution time [day]: 2 Unit price #1 [EUR]: 1,100.00 Limit #1: Applicable to any sample



# PRICE LIST / Listino prezzi § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



#### 5.2.2led - Inherently safe optical radiation "op is" - Led lamp radiation

(id 1518) The test is intended to qualify standard led lamps, ensuring that the specific power is lower than the ignition limits

## IEC 60079-29-1:2016+AMD1:2020

(id 1859) Explosive atmospheres - Part 29-1: Gas detectors - Performance requirements of detectors for flammable gases

#### 4 - General requirements verification

(id 1481) ---

Accredited [IEC17025]: YES Execution time [day]: 1
Unit price #1 [EUR]: 500.00 Limit #1: Applicable to any sample

#### 5.4.2 - Unpowered storage

(id 1482) ---

Accredited [IEC17025]: YES Execution time [day]: 4
Unit price #1 [EUR]: 1,200.00 Limit #1: Applicable to any sample

#### 5.4.3.2 - Calibration and adjustment: calibration curve

(id 1483) The equipment shall be calibrated and adjustments shall be carried out to obtain correct indications in accordance with the manufacturer's instruction manual.

#### 5.4.3.3 - Calibration and adjustment: response to different gases (Group II equipment)

(id 1485) ---

#### 5.4.4.2 - Short-term stability

(id 1488) ---

Accredited [IEC17025]: YES Execution time [day]: 1
Unit price #1 [EUR]: 800.00 Limit #1: Not relevant

#### 5.4.4.3 - Long-term stability for fixed and transportable equipment (Group I)

(id 1489) ---

Accredited [IEC17025]: YES Execution time [day]: 30 Unit price #1 [EUR]: 2,200.00 Limit #1: Not relevant

#### 5.4.4.4 - Long-term stability for portable equipment (Group I)

(id 1490) ---

Accredited [IEC17025]: YES Execution time [day]: 25
Unit price #1 [EUR]: 2,400.00 Limit #1: Not relevant

#### 5.4.4.5 - Long-term stability for fixed and transportable equipment (Group II)

(id 1491) ---

Accredited [IEC17025]: YES Execution time [day]: 64
Unit price #1 [EUR]: 4,500.00 Limit #1: Applicable to any sample

#### 5.4.4.6 - Long-term stability for portable equipment (Group II)

(id 1492) ---

Accredited [IEC17025]: YES Execution time [day]: 20
Unit price #1 [EUR]: 2,200.00 Limit #1: Not relevant

#### 5.4.5 - Alarm set points

(id 1493) ---



# PRICE LIST / Listino prezzi § T - LABORATORY TESTS / Prove di laboratorio

# alb**ar**ubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX

# DEX -

5.4.6 - Temperature test

(id 1495) This test shall be performed in a temperature chamber having the capability of holding the remote sensor or equipment at the specified temperature within  $\pm 2$  °C. For battery powered equipment, the test shall be carried out with all batteries specified in the instruction manual.

Accredited [IEC17025]: YES Execution time [day]: 2
Unit price #1 [EUR]: 2,200.00 Limit #1: Applicable to any sample

#### 5.4.7 - Pressure test

(id 1496) ---

#### 5.4.8 - Humidity of test gas

(id 1497) ---

Accredited [IEC17025]: YES Execution time [day]: 1
Unit price #1 [EUR]: 2,200.00 Limit #1: Applicable to any sample

#### 5.4.9 - Air velocity

(id 1498) ---

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1: ask to office Limit #1: Not relevant

#### 5.4.10 - Flow rate for aspirated equipment

(id 1499) ---

Accredited [IEC17025]: YES Execution time [day]: 1
Unit price #1: ask to office Limit #1: Not relevant

#### 5.4.11.1 - Orientation: portable equipment

(id 1500) ---

#### 5.4.11.2 - Orientation: fixed and transportable equipment

(id 1501) ---

#### 5.4.12 - Vibration test

(id 1502) The vibration test shall be performed after unpowered storage testing for pre-conditioning purposes.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 2,000.00 Limit #1: Applicable to any sample

#### 5.4.13 - Drop test for portable and transportable equipment

(id 1503) ---

#### 5.4.14 - Warm-up time for gas detectors

(id 1504) ---

### 5.4.15/Ann B - Time of response

(id 1505) ---

#### 5.4.16 - High gas concentration operation above the measuring range

(id 1506) ---

Accredited [IEC17025]: YES Execution time [day]: 1
Unit price #1 [EUR]: 1,500.00 Limit #1: Applicable to any sample



# § T - LABORATORY TESTS / Prove di laboratorio

**PRICE LIST / Listino prezzi** 

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



#### 5.4.17 - Battery capacity: battery discharge and low battery duration

(id 1507) ---

Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

#### 5.4.18 - Power supply variations

(id 1509) ---

Unit price #1 [EUR]: 500.00 Limit #1: Applicable to any sample

#### 5.4.19 - Addition of sampling probe

(id 1510) ---

Accredited [IEC17025]: YES Execution time [day]: 1

Unit price #1: ask to office Limit #1: Not relevant

#### 5.4.20.1 - Other gases exposure

(id 1511) ---

Unit price #1 [EUR]: 2,000.00 Limit #1: Not relevant

#### 5.4.20.2 - Poisons exposure (Group I equipment with catalytic or semiconductor sensors)

(id 1512) ---

Accredited [IEC17025]: YES Execution time [day]: 1

Unit price #1 [EUR]: 3,000.00 Limit #1: Not relevant

#### 5.4.22 - Field calibration kit

(id 1514) ---

Unit price #1: ask to office Limit #1: Not relevant

#### 5.4.21 - Electromagnetic compatibility

(id 1516) ---

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1: ask to office Limit #1: Not relevant

### 5.4.23 - Software function (EN 60079-29-1:2016)

(id 1517) ---

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 3,500.00 Limit #1: Applicable to any sample

## IEC 60079-2:2014+COR1:2015

(id 30) Explosive atmosphere - Part 2: Equipment protection by pressurized enclosure "p"

#### 16.2 - Maximum overpressure test

(id 602) Verification of the capability of a pressurized enclosure to hold the maximum overpressure rated without permanent deformation.

#### 16.3.2 - Leakage test: static pressurization

(id 603) Monitoring of the internal pressure of a pressurized enclosure at its maximum overpressure in order to verify leakages (static pressurization)

#### 16.6 - Verification of minimum overpressure

(id 604) Capability to mantain the minimum overpressure under normal operating conditions (maintenance conditions)



# **PRICE LIST / Listino prezzi**

## § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

TNDEX



#### 16.7.1 - Overpressure test for infallible containement system

(id 605) Preliminary stress test for infallible containement system

Accredited [IEC17025]: YES Execution time [day]: 0 Unit price #1 [EUR]: 500.00 Limit #1: Applicable to any sample

#### 16.8 - Overpressure test for a containment system with a limited release

(id 606) Overpressure test carried out on a containment system wich has a limited release during normal operation.

```
Accredited [IEC17025]: YES
                                 Execution time [day]: 0
Unit price #1 [EUR]: 500.00
                                           Limit #1: < 500 [dm3] and pressure < 20mbar
Unit price #2 [EUR]: 800.00
                                            Limit #2: 500 < 1000 [dm3] and pressure < 20mbar
Unit price #3 [EUR]: 1,200.00
                                            Limit #3: Every volume and pressure > 20mbar
```

#### 16.7.2 - Infallibility test

(id 629) Check for the presence of leaks when the containment system is flushed and pressurized with pure helium.

```
Accredited [IEC17025]: YES
                                  Execution time [day]: 0
Unit price #1 [EUR]: 1,000.00
                                           Limit #1: Applicable to any sample
```

#### 16.4.2/16.4.4 - Purging test where protective gas is air or inert gas with similar density

(id 639) Purging test for pressurized enclosures with no internal source of release where protective gas is air or an inert gas with a density equal to air ±10 %.

Where the protective gas is inert, the time taken until there is no sample point where there is a test gas concentration in excess of that specified below is measured and noted as the purging time:

- where test(s) were conducted for specific flammable gases, a value equivalent to 25 % of the most onerous LFL;
- where one specific flammable gas is covered, a value equivalent to 25 % of its LFL;
- where all flammable gases are covered, 1 % for the helium test and 0,25 % for the argon or carbon dioxide test.

NOTE These values correspond approximately to 25 % of the LFL for light and heavy flammable gases respectively.

Where the protective gas is inert, The time taken until there is no sample point where there is an oxygen concentration not exceeding 2 % (V/V) is measured and noted as the purging time.

```
Accredited [IEC17025]: YES
                                  Execution time [day]: 0
Unit price #1 [EUR]: 1,200.00
                                            Limit #1: < 500 [dm3] and pressure < 20mbar
Unit price #2 [EUR]: 2,800.00
                                            Limit #2: 500 < 1000 [dm3] and pressure < 20mbar
Unit price #3 [EUR]: 4,000.00
                                            Limit #3: Every volume and pressure > 20mbar
```

#### 16.3.1 - Leakage test: other than static pressurization

(id 877) Maximum leakage flow rate measurement in normal operation conditions (other than static pressurization)

```
Accredited [IEC17025]: YES
                                  Execution time [day]: 0
Unit price #1 [EUR]: 500.00
                                            Limit #1: < 500 [dm3] and pressure < 20mbar
Unit price #2 [EUR]: 800.00
                                            Limit #2: 500 < 1000 [dm3] and pressure < 20mbar
Unit price #3 [EUR]: 1,200.00
                                            Limit #3: Every volume and pressure > 20mbar
```

#### 16.4.5 - Filling procedure test on a pressurized enclosure protected by static pressurization

(id 878) In the case of static pressurization, the enclosure is filled initially with air at normal atmospheric pressure. The equipment is then filled with inert gas in accordance with the manufacturer's specifications, and the oxygen concentration inside the sample is verified.

```
Accredited [IEC17025]: YES
                                  Execution time [day]: 0
Unit price #1 [EUR]: 600.00
                                            Limit #1: Applicable to any sample
```

#### 16.1 - Determination of maximum overpressure rating

(id 895) Measurement of the highest internal operating pressure

```
Accredited [IEC17025]: YES
                                 Execution time [day]: 0
Unit price #1 [EUR]: 500.00
                                            Limit #1: Applicable to any sample
```

#### 16.5.2.1 - Purging test for pressurized enclosures with an internal source of release

(id 901) Purging test for a pressurized enclosure with an internal source of release where the flammable substance has less than 2 % (V/V) oxygen and the protective gas is inert

```
Accredited [IEC17025]: YES
                                  Execution time [day]: 1
Unit price #1 [EUR]: 800.00
                                            Limit \#1: < 500 [dm3] and pressure < 20mbar
Unit price #2 [EUR]: 1,000.00
                                            Limit #2: 500 < 1000 [dm3] and pressure < 20mbar
Unit price #3 [EUR]: 1,500.00
                                            Limit #3: Every volume and pressure > 20mbar
```

#### 16.4.3 - Purging test for pressurized enclosure where the protective gas is inert

(id 1138) Purging test for pressurized enclosures with no internal source of release where the protective gas is inert with a density different from air

Accredited [IEC17025]: YES	Execution time [day]: 1
Unit price #1 [EUR]: 2,000.00	Limit #1: < 500 [dm3] and pressure < 20mbar
Unit price #2 [EUR]: 2,800.00	Limit #2: 500 < 1000 [dm3] and pressure < 20mbar
Unit price #3 [EUR]: 4,000.00	Limit #3: Every volume and pressure > 20mbar



# PRICE LIST / Listino prezzi § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



#### 16.5.3.1 - Purging test for pressurized enclosures with an internal source of release

(id 1143) Purging test for a pressurized enclosure with an internal source of release with pressurization by continuous flow, containment system with less than 21 % (V/V) oxygen and the protective gas is inert

#### 16.5.3.2 - Dilution test on pressurized enclosures with an internal source of release

(id 1144) Dilution test on pressurized enclosures with an internal source of release with pressurization by continuous flow, containment system with less than 21 % (V/V) oxygen and the protective gas is inert.

#### 16.5.4.1 - Dilution test on pressurized enclosures with an internal source of release - Purging test

(id 1145) Purging test on pressurized enclosure with an internal source of release where the flammable substance is not a liquid, with pressurization by continuous flow and where the protective gas is air

#### 16.5.4.2 - Dilution test on pressurized enclosures with an internal source of release - Dilution test

(id 1146) Dilution test on pressurized enclosure with an internal source of release where the flammable substance is not a liquid, with pressurization by continuous flow and where the protective gas is air

#### 5.9 - Spark and particle barriers

(id 1589) ---

```
Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 50.00 Limit #1: Applicable to any sample
```

#### 17.1 - Functional test

(id 2743) The performance of safety devices provided with the pressurized enclosure shall be verified.

```
Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 300.00 Limit #1: Applicable to any sample
```

#### 16.4.2/16.4.4 - Purging test where protective gas is air or inert gas with similar density (on-site)

(id 2932) Purging test for pressurized enclosures with no internal source of release where protective gas is air or an inert gas with a density equal to air ±10 %. This test is intended to be performed at the customer facilty.

```
Accredited [IEC17025]: YES Execution time [day]: 0
Unit price #1 [EUR]: 2,800.00 Limit #1: Test performed at the customer facility
```

#### IEC 60079-31:2022

(id 1763) Explosive atmospheres - Part 31: Equipment dust ignition protection by enclosure "t"

#### 6.1.1.1 - Type tests for dust exclusion by enclosures

(id 2472) Test sequence composed of thermal endurance to heat, thermal endurance to cold, impact test and and drop test if applicable, as specified in IEC 60079-0, followed by:

- IEC 60079-31 § 6.1.1.2 Impact test for supplementary internal enclosures, if present
- IEC 60079-31 § 6.1.1.3 Pressure test
- IEC 60079-31 § 6.1.1.4 IP test.

Accredited [IEC17025]: YES Execution time [day]: 30
Unit price #1: ask to office Limit #1: not relevant



# PRICE LIST / Listino prezzi

## § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

TNDEX

# 6.1.1.2 - Impact test on supplementary internal enclosures

(id 2473) An impact test on supplementary internal enclosures shall be performed in accordance with the resistance to impact test of IEC 60079-0.

Accredited [IEC17025]: YES Execution time [day]: 0 Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 6.1.1.3 - Pressure test

(id 2474) A positive internal pressure shall be applied to the Ex Equipment.

Accredited [IEC17025]: YES Execution time [day]: 0

Limit #1: Applicable to any sample Unit price #1 [EUR]: 300.00

#### 6.1.1.4 - IP test

(id 2475) ---

Accredited [IEC17025]: YES Execution time [day]: 30 Unit price #1 [EUR]: 300.00 Limit #1: TP5X or TP6X

#### 6.1.2 - Tests to determine maximum surface temperature

(id 2476) ---

Accredited [IEC17025]: YES Execution time [day]: 1 Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

## IEC 60079-32-2:2015

(id 116) Explosive atmospheres - Part 32-2: Electrostatics hazards - Tests

#### 4.11 - Transferred charge

(id 950) ---

Accredited [IEC17025]: NO Execution time [day]: 2 Unit price #1 [EUR]: 1,100.00 Limit #1: Applicable to any sample

#### 4.12 - Ignition test

(id 951) Ignition tests are carried out by bringing the ignition probe close to the charged test sample with a flammable gas mixture flowing through the probe. A rubbing sequence with materials from the positive and negative end of the triboelectric series is performed with the following steps: 1 stroke per s with a force of about 40 N for 10 + or - 1 s with a material from the triboelectric series, then discharge according to standard requirements. The rubbing sequence is performed at least 20 times with a first material from the positive end of the triboelectric series, 20 times with a material from the negative end of the triboelectric series, 20 times with a second material from the positive end of the triboelectric series.

Accredited [IEC17025]: YES Execution time [day]: 2 Unit price #1 [EUR]: 1,000.00 Limit #1: Applicable to any sample

#### 4.2 - Surface resistance

(id 1781) ---

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1 [EUR]: 500.00 Limit #1: Not relevant

#### 4.14 - Breakdown voltage

(id 1813) The electrical breakdown voltage shall be measured according to the short-time (rapid rise) test in IEC 60243-1 with the additional requirements of IEC 60243-2 for DC testing.

The breakdown voltage is that in which 1mA of leakage current is exceeded.

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 400.00 Limit #1: Test on plate Unit price #2 [EUR]: 600.00 Limit #2: Test on tube

### IEC 60079-5:2015+AMD1:2022

(id 1846) Explosive atmospheres - Part 5: Equipment protection by powder filling "q"

#### 5.1.1 - Pressure type test of container

(id 1026) Overpressure test on enclosures and containers

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 300.00 Limit #1: Not relevant



# PRICE LIST / Listino prezzi

# § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

TNDEX



#### 5.1.2 - Verification of the degree of protection of the enclosure

(id 1027) The verification of the IP degree of protection is conducted according to IEC 60529, in compliance with the requirements of § 4.1.3

Accredited [IEC17025]: NO Execution time [day]: 2 Unit price #1 [EUR]: 600.00 Limit #1: IP65

#### 5.1.3 - Insulation resistance test of the filling material

(id 1028) Dielectric strength test on the filling material

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 300.00 Limit #1: Apply to any sample

#### 5.1.4 - Maximum temperatures

(id 1029) Measurement of the maximum temperatures under overload conditions or malfunction simulation

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1 [EUR]: 800.00 Limit #1: Not relevant

#### 4.2.1 - Material specification

(id 1043) Requirements for equipment protection by powder filling "q"

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 500.00 Limit #1: Not relevant

#### 4.3.2 - Distances surrounding free space

(id 1044) Requirements for equipment protection by powder filling "q"

Accredited [IEC17025]: NO Execution time [day]: 2 Unit price #1 [EUR]: 200.00 Limit #1: Not relevant

#### 4.3.1 - Distances through filling material

(id 1864) Requirements for equipment protection by powder filling "q"

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 200.00 Limit #1: Not relevant

### IEC 60079-6:2015

(id 78) Explosive atmospheres - Part 6: Equipment protection by oil immersion "o"

#### 6.1.1 - Overpressure test on sealed enclosures

(id 1340) Test to verify the ability of sealed enclosure to withstand an internal pressure

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1 [EUR]: 400.00 Limit #1: Not relevant

#### 6.1.2 - Reduced pressure test on sealed enclosures

(id 1341) Verification of the increase in pressure after 24h during which the internal pressure of the enclosure without protective liquid shall be reduced by an amount equivalent to not less than the pressure that would result from a change in the protective liquid level from the maximum permissible level to the minimum permissible level when appropriately corrected for any ambient temperature variations specified in the documentation

Accredited [IEC17025]: NO Execution time [day]: 2 Unit price #1 [EUR]: 500.00 Limit #1: Not relevant

#### 6.1.3 - Overpressure test on unsealed enclosures

(id 1342) Verification of the ability of unsealed enclosures to withstand an internal pressure equal to 150 kPa, with the breather sealed, for at least 60s.

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1 [EUR]: 500.00 Limit #1: Not relevant

#### 6.1.4 - Maximum temperature

(id 1343) Measurement of maximum surface temperature at maximum load

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1 [EUR]: 600.00 Limit #1: Not relevant

#### 6.1.5 - Switching Tests

(id 1344) This test is performed only if the equipment includes internal switches

Accredited [IEC17025]: NO Execution time [day]: 2 Unit price #1 [EUR]: 1,200.00 Limit #1: Not relevant



# PRICE LIST / Listino prezzi

## § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

TNDEX

#### 4.6 - Immersion depth verification

(id 1345) Measurement of immersion depth of live parts at the minimum oil level possible

Accredited [IEC17025]: NO Execution time [day]: 2 Unit price #1 [EUR]: 300.00 Limit #1: Not relevant

### IEC 60079-7:2015+AMD1:2017

(id 14) Explosive atmospheres - Part 7: Equipment protection by increased safety "e"

#### 6.1 - Dielectric strength

(id 120) ---

Accredited [IEC17025]: YES	Execution time [day]: 0
Unit price #1 [EUR]: 300.00	Limit #1: Up to 5kV dc
Unit price #2 [EUR]: 1,200.00	Limit #2: From 5 to 10kV ac/dc
Unit price #3 [EUR]: 2,200.00	Limit #3: From 10 to 50kV ac/dc

#### 6.8/E.2 - General purpose connection and junction boxes - Maximum dissipated power method

(id 421) ---

```
Accredited [IEC17025]: YES
                                  Execution time [day]: 0
Unit price #1 [EUR]: 300.00
                                            Limit #1: Applicable to any sample
```

#### 6.10 - Terminal insulating material tests

(id 422) A pulling force corresponding to the conductor size shall be applied gradually to each conductor in sequence and maintained at that value for at least 1

```
Accredited [IEC17025]: YES
                                  Execution time [day]: 1
Unit price #1 [EUR]: 300.00
                                            Limit #1: Up to 25mm2
Unit price #2 [EUR]: 500.00
                                            Limit \#2: Up to 250mm2
Unit price #3 [EUR]: 800.00
                                            Limit #3: More than 250mm2
```

#### 6.6.2/6.7.2 - Insulation resistance for Ex eb-ec batteries

(id 608) Insulation resistance test for cells and batteries of Level of Protection 'eb' or 'ec'

```
Accredited [IEC17025]: YES
                                  Execution time [day]: 0
                                            Limit #1: Applicable to any group of samples
Unit price #1 [EUR]: 400.00
```

#### 6.9.2 - Verification of the electrical insulation

(id 610) Tests on resistance heating device

```
Accredited [IEC17025]: YES
                                 Execution time [day]: 0
Unit price #1 [EUR]: 300.00
                                           Limit #1: Applicable to any sample
```

## 6.6.3/6.7.3 - Shock test for Ex eb-ec batteries

(id 657) ---

```
Accredited [IEC17025]: NO
                                  Execution time [day]: 1
Unit price #1 [EUR]: 1,800.00
                                            Limit #1: Applicable to any sample
```

#### 6.6.4/6.7.4 - Test for ventilation of battery container for Ex eb-ec batteries

(id 658) The purpose of the test is the determination of the maximum hydrogen concentration within the battery container and the adequate dimensioning of its ventilation openings

Accredited [IEC17025]: YES	Execution time [day]: 0
Unit price #1 [EUR]: 1,400.00	Limit #1: V container < 20 l with overcharging current
Unit price #2 [EUR]: 1,800.00	Limit #2: 20 1 < V container < 50 1 with overcharging current
Unit price #3 [EUR]: 2,200.00	Limit #3: V container > 50 l or dangerous testing conditions

#### 6.8-E.3 - General purpose connection and junction boxes - Defined arrangement method (id 876) ---

Accredited [IEC17025]: YES	Execution time [day]: 0
Unit price #1 [EUR]: 800.00	Limit #1: up 5 liters
Unit price #2 [EUR]: 1,200.00	Limit #2: for 5 to 100 liters
Unit price #3 [EUR]: 2,000.00	Limit #3: over 100 liters



# PRICE LIST / Listino prezzi

## § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

TNDEX



#### 4 - 5 - Constructional requirements connection

(id 917) Physical and/or documental evaluation of the sample characteristics

Accredited [IEC17025]: NO Execution time [day]: 2 Unit price #1 [EUR]: 300.00 Limit #1: Applicable to any sample

#### B.2 - Resistance heating device intended for immersion insulation test

(id 921) This type test applies to specific forms of resistance heating devices or resistance heating units (other than trace heater).

This test is not intended to verify the suitability of resistance heating devices or units to operate when immersed in liquids other than water or at pressures of more than 500 Pa.

Accredited [IEC17025]: NO Execution time [day]: 15 Unit price #1 [EUR]: 1,000.00 Limit #1: Applicable to any sample

#### 6.8.3 - Dielectric strength test and insulation test for heating resistors

Accredited [IEC17025]: YES Execution time [day]: 0 Unit price #1 [EUR]: 1,000.00 Limit #1: Applicable to any sample

#### 4.2.3 - Factory connection

(id 973) ---

Accredited [IEC17025]: NO Execution time [dav]: 2 Unit price #1 [EUR]: 300.00 Limit #1: Applicable to any sample

#### 4 - 5 - Constructional requirements motor

(id 1022) Filling of ExTR60079-7\_4C

Accredited [IEC17025]: NO Execution time [day]: 2 Limit #1: Applicable to any sample Unit price #1 [EUR]: 500.00

#### A.2 - Determination of maximum service temperatures

(id 1030) Winding temperature - normal operation

Accredited [IEC17025]: NO Execution time [day]: 2 Unit price #1 [EUR]: 1,000.00 Limit #1: up to 2 liters Unit price #2 [EUR]: 2,500.00 Limit #2: over 2 liters

#### A.3 - Determination of maximum surface temperatures

(id 1031) Determination of maximum surface temperature on electrical motors

Accredited [IEC17025]: NO Execution time [day]: 2 Unit price #1 [EUR]: 800.00 Limit #1: up to 2 liters Unit price #2 [EUR]: 1,500.00 Limit #2: over 2 liters

#### 6.9.3 - Thermal stability of the insulating materials of resistance heating devices

(id 1083) Tests on resistance heating device

Accredited [IEC17025]: YES Execution time [day]: 29 Unit price #1 [EUR]: 1,400.00 Limit #1: Applicable to any sample

#### 6.9.4 - Test for resistance to impact

(id 1084) Tests on resistance heating device

Accredited [IEC17025]: YES Execution time [day]: 0 Limit #1: Applicable to any sample Unit price #1 [EUR]: 200.00

#### 6.9.5 - Test for the cold start current

(id 1085) Tests on resistance heating device

Accredited [IEC17025]: YES Execution time [day]: 0 Unit price #1 [EUR]: 300.00 Limit #1: Manufacturer's declared cold start temperature ±2 °C

#### 6.9.6 - Tests for specific forms of resistance heating devices

(id 1086) Tests on resistance heating device

Accredited [IEC17025]: YES Execution time [day]: 2 Unit price #1: ask to office Limit #1: Not relevant

#### 6.9 - Resistance heating equipment

(id 1087) Sequence of tests to be performed on resistance heating equipment

Accredited [IEC17025]: YES Execution time [day]: 0 Unit price #1: ask to office Limit #1: Not relevant

page 120/234



# PRICE LIST / Listino prezzi

## § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

TNDEX



#### B.1 - Resistance heating devices subjected to mechanical stresses

(id 1088) Type tests for specific forms of resistance heating devices or resistance heating units (other than trace heater)

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1: ask to office Limit #1: Not relevant

#### B.2 - Resistance heating devices or units intended for immersion

(id 1089) Type tests for specific forms of resistance heating devices

or resistance heating units (other than trace heater)

Accredited [IEC17025]: NO Execution time [day]: 14 Unit price #1 [EUR]: 500.00 Limit #1: Applicable to any sample

#### B.3 - Resistance heating devices or units having hygroscopic insulating material

(id 1090) Type tests for specific forms of resistance heating devices

or resistance heating units (other than trace heater)

Accredited [IEC17025]: NO Execution time [day]: 28

Unit price #1 [EUR]: 1,000.00 Limit #1: Applicable to any sample

#### B.4.2.2 - Safety device sensing the tempertaure

(id 1091) Verification of limiting temperature of resistance heating devices (other than trace heaters)

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 1,000.00 Limit #1: Up to 6 points

#### B.4.2.3 - Safety device sensing the temperature and at least one other parameter

(id 1092) Verification of limiting temperature of resistance heating devices (other than trace heaters)

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1: ask to office Limit #1: Not relevant

#### B.4.2.4 - Safety device sensing a parameter other than the temperature

(id 1093) Verification of limiting temperature of resistance heating devices (other

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1: ask to office Limit #1: Not relevant

#### B.4.3 - Resistance heating unit of stabilized design

(id 1094) Verification of limiting temperature of resistance heating devices (other than trace heaters)

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 800.00 Limit #1: Not relevant

#### B.4.4 - Heating device with temperature self-limiting characteristic

(id 1095) Verification of limiting temperature of resistance heating devices (other than trace heaters)

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1: ask to office Limit #1: Not relevant

#### 6.3.2 - Impact test

(id 1101) Tests for luminaires

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1: ask to office Limit #1: Not relevant

#### 6.2.1 - Determination of starting current ratio IA-IN and the time tE

(id 1125) ---

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1 [EUR]: 400.00 Limit #1: Not relevant

#### 6.2.3.1.3 - Stator winding insulation

(id 1126) Insulation systems and connection cables shall be tested in an explosive test mixture applying a sinusoidal voltage of at least 1.5 times the rated r.m.s. line voltage for at least 3 min between earth and each phase.

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1 [EUR]: 800.00 Limit #1: Test inside small vessel Unit price #2 [EUR]: 1,300.00 Limit #2: Test in explosion vented chamber Unit price #3 [EUR]: 3,000.00 Limit #3: Outdoor test



# PRICE LIST / Listino prezzi § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX

#### 6.2.3.2 - Cage rotor

(id 1127) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 800.00 Limit #1: Not relevant

#### 6.2.4 - Overspeed test of cemented magnets

(id 1128) An electric motor shall be tested at 1.2 times its maximum rated speed in order to verify the tightness of the adhesives of the magnets as well as the stability of the rotor.

#### 6.3.3 - Mechanical tests for screw lampholders other than E10

(id 1129) ---

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1: ask to office Limit #1: Not relevant

#### 6.3.4 - Abnormal operation of luminaires

(id 1130) ---

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1: ask to office Limit #1: Not relevant

#### 6.3.5 - Sulphur dioxide test for level of protection "eb" for the connection of bi-pin lamp caps to lampholders

(id 1131) ---

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1: ask to office Limit #1: Not relevant

#### 6.3.6 - Vibration test for level of protection "eb" for luminaires with bi-pin lamps

(id 1132) ---

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1: ask to office Limit #1: Not relevant

#### 6.3.7 - Tests for wiring of luminaires subject to high-voltage impulses from ignitors

(id 1133) ---

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1: ask to office Limit #1: Not relevant

# 6.3.8 - Tests for electronic starters for tubular fluorescent lamps and for ignitors in Level of Protection "ec" for discharge lamps

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1: ask to office Limit #1: Not relevant

#### 6.3.2 - Drop test

(id 1137) Tests for luminaires

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

#### 6.6.4/6.7.4 - -ANNULLATO - Determination of minimum ventilation openings of battery container

(id 1338) ---ANNULLATO nella riunione del 05.02.2017, vedere MOM 17/0299---The purpose of the test is the determination of the minimum dimensioning of the ventilation openings so that, with the rated hydrogen flow (calculated for the battery for which the case is prepared), the concentration limits imposed by the standard are not exceeded.

The test is an internal method based on §6.6.4/6.7.4.

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 1,200.00 Limit #1: Not relevant

#### A.3.2.2 - Locked rotor tests - Optional testing at reduced voltage

(id 1350) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant



# PRICE LIST / Listino prezzi § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX

# .

#### A.3.2.3 - Locked rotor tests - Rotor temperature

(id 1351) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

#### A.3.2.4 - Locked rotor tests - Determination of starting current IA

(id 1352) --

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

#### A.3.2.5 - Locked rotor tests - Stator temperature

(id 1353) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

#### B.4 - Verification of limiting temperature of resistance heating devices (other than trace heaters)

(id 1570) The test shall be carried out in accordance with the procedure of B.4.2, B.4.3 or B.4.4

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

#### 4.2.3.4/4.2.3.5/4.2.3.6 - Field wiring connections: cable pull for pluggable connections

(id 1906) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 200.00 Limit #1: Not relevant

### IEC 60086-4:2019

(id 1397) Primary batteries - Part 4: Safety of lithium batteries

#### 6.5.7 - Test K: Thermal abuse

(id 1866) This test simulates the condition when a battery is exposed to an extremely high temperature.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 300.00 Limit #1: up to 2 liters in volume

### IEC 60204-1:2016+AMD1:2021

(id 1849) Amendment 1 - Safety of machinery - Electrical equipment of machines - Part 1: General requirements

#### 18.3 - Insulation resistance tests

(id 1567) This part of IEC 60204 gives general requirements for the electrical equipment of machines. This test is included in the sequence of tests listed in 18.1.

#### 18.4 - Voltage test

(id 2432) The maximum test voltage shall be applied between the power circuit conductors and the protective bonding circuit for a period of approximately 1s

Accredited [IEC17025]: YES Execution time [day]: 1
Unit price #1 [EUR]: 300.00 Limit #1: Up to 5kV dc

### IEC 60335-2-40:2022

(id 1828) Household and similar electrical appliances – Safety – Part 2-40: Particular requirements for electrical heat pumps, air-conditioners and dehumidifiers

#### LL.3 - Response time of the refrigerant detection system

(id 2625) ---

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 500.00 Limit #1: Applicable to any sample



# PRICE LIST / Listino prezzi § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX

# -

#### LL.9 - Temperature test

(id 2626) ---

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 1,200.00 Limit #1: Only one temperature

#### LL.4 - Refrigerant detection system calibration and short-term stability

(id 2892) The refrigerant detection systems shall have an alarm set point and be calibrated from the factory for the refrigerant marked on the appliance.

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 400.00 Limit #1: Applicable to any sample

#### LL.5 - Selectivity test and poisoning test

(id 2893) Refrigerant detection system including refrigerant sensors shall not have false or nuisance trips, and not be subject to poisoning.

Accredited [IEC17025]: NO Execution time [day]: 2

Unit price #1: ask to office Limit #1: Not relevant

#### LL.6 - Refrigerant poisoning and oil spray test

(id 2894) The refrigerant detection system shall not be subject to poisoning or damage by the refrigerant(s) marked on the appliance mixed with oil.

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1: ask to office Limit #1: Applicable to any sample

#### LL.7 - Long term stability

(id 2895) The refrigerant sensor shall be exposed to clean air and energized continuously for a period of 90 days.

Accredited [IEC17025]: NO Execution time [day]: 90

Unit price #1 [EUR]: 3,200.00 Limit #1: Applicable to any sample

#### LL.8 - Humidity test

(id 2896) The refrigerant sensor shall be exposed for 60 min to clean air humidified to (30 ± 5) % RH at (15 ± 2) °C.

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 1,200.00 Limit #1: Applicable to any sample

#### LL.10 - Vibration requirements

(id 2897) Vibration requirements this clause apply for refrigerant sensors. Vibration requirements for refrigerant sensors need not apply to the entire appliance.

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1: ask to office Limit #1: Not relevant

#### LL.11 - Ignition test

(id 2898) Refrigerant sensors shall not be a source of ignition for leaked refrigerant.

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 1,500.00 Limit #1: Applicable to any sample

#### LL.12 - Refrigerant detection system self-test routine

(id 2899) The refrigerant detection system shall include a means for self-testing to determine if a refrigerant sensor electrical malfunction has occurred.

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 100.00 Limit #1: Applicable to any sample

#### LL.13 - Serviceability

(id 2900) Refrigerant sensors shall be accessible for inspection and replacement. Refrigerant sensors for replacement shall be specified by the appliance manufacturer

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 50.00 Limit #1: Applicable to any sample

### LL.14 - Refrigerant sensor identification

(id 2901) Visual inspection.

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 50.00 Limit #1: Applicable to any sample



# PRICE LIST / Listino prezzi § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



## IEC 60512-1-1:2002

(id 1379) Connectors for electronic equipment Tests and measurements - Part 1-1: General examination Test 1a: Visual examination

#### unref - Test 1a: Visual examination

(id 1782) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 50.00 Limit #1: Not relevant

### IEC 60512-1-2:2002

(id 1380) Connectors for electronic equipment Tests and measurements - Part 1-2: General examination Test 1b: Examination of dimension and mass

#### unref - Test 1b: Examination of dimension and mass

(id 1783) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

## IEC 60512-11-10:2002

(id 1381) Connectors for electronic equipment -Tests and measurements - Part 11-10: Climatic tests -Test 11j: Cold

#### unref - Test 11j: Cold

(id 1794) This test shall be carried out in accordance with IEC 60068-2-1:

- Test Aa: Cold for non-heat-dissipating specimens with sudden change of temperature when no measurement during the test is required;
- Test Ab: Cold for non-heat-dissipating specimens with gradual change of temperature when measurements during the test are required; using the degree of severity specified in the detail specification.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 150.00 Limit #1: Up to 2h (chest freezer - small)

### IEC 60512-11-3:2002

(id 1382) Connectors for electronic equipment -Tests and measurements - Part 11-3: Climatic tests -Test 11c: Damp heat, steady state

#### unref - Test 11c: Damp heat, steady state

(id 1792) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 500.00 Limit #1: Applicable to any sample

## IEC 60529:1989+AMD2:2013+COR1:2019

(id 1607) Degrees of protection provided by enclosures (IP Code)

#### 12 - Tests for protection against access to hazardous parts - First characteristic numeral

(id 563) The purpose of this test is to evaluate the protection of persons against access to hazardous parts. It is performed together with the relevant tests for protection of equipment against the penetration of solid foreign objects. According to the first characteristic numeral to be achieved, it can be performed with the following access probes: 1X: sphere 50 mm diam.; 2X: jointed test finger 12 mm diam.; 3X: tool 2.5 mm diam.; 4X, 5X and 6X: wire 1.0 mm diam.

#### 13.2 - Tests for protection against solid foreign objects - First characteristic numeral 2 (IP 2X)

(id 567) This test is performed to evaluate the protection of the equipment against the access of the object probe rigid sphere of 12.5 mm diameter, without handle or guard, with a test force of 30 N.

Accredited [IEC17025]: YES Execution time [day]: 2
Unit price #1 [EUR]: 300.00 Limit #1: Applicable to any sample



# § T - LABORATORY TESTS / Prove di laboratorio

PRICE LIST / Listino prezzi

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



#### 13.2 - Test for protection against solid foreign objects - First characteristic numeral 3 (IP 3X)

(id 568) This test is performed to evaluate the protection of the equipment against the access of the object probe rigid steel rod of 2.5 mm diameter, with edges free from burrs, with a test force of 3 N.

```
Accredited [IEC17025]: YES Execution time [day]: 2
Unit price #1 [EUR]: 300.00 Limit #1: Applicable to any sample
```

#### 13.2 - Tests for protection against solid foreign objects - First characteristic numeral 4 (IP 4X)

(id 569) This test is performed to evaluate the protection of the equipment against the access of the object probe rigid steel rod of 1 mm diameter, with edges free from burrs, with a test force of 1 N.

```
Accredited [IEC17025]: YES Execution time [day]: 1
Unit price #1 [EUR]: 400.00 Limit #1: Applicable to any sample
```

#### 13.4 - Dust test for first characteristic numerals 5 and 6 (IP 5X or 6X)

(id 571) This test is performed to evaluate the protection of the equipment against the access of dust penetration.

For huge size samples the test may be performed into a movable chamber.

Sample tested are of necessity in one of two categories:

Category 1: Enclosures where the normal working cycle of the equipment causes reductions in air pressure within the enclosure below that of the surrounding air, e.g., due to thermal cycling effects.

Category 2: Enclosures where no pressure difference relative to the surrounding air is present.

In the case of IP evaluations under ATEX directive or standards the devices are always considered belonging to Category 1 and tested after an ageing cycle.

```
Accredited [IEC17025]: YES

Unit price #1 [EUR]: 400.00

Unit price #2 [EUR]: 800.00

Unit price #3 [EUR]: 1,200.00

Execution time [day]: 1

Limit #1: Standard IP chamber (very small sample)

Limit #2: Medium IP chamber (sample up to 90cm height,200 lenght)

Limit #3: Huge size samples with movable chamber
```

#### 14.2.5 - Test for second characteristic numeral 5 with the 6\_3 mm nozzle (IP X5)

(id 572) This test is performed to evaluate the protection of the equipment against the access of water in the presence of a water jet hose nozzle of 6.3 mm diameter, and a 12.5 l/min flow rate, spraying water from a 3m distance.

#### 14.2.6 - Test for second characteristic numeral 6 with the 12-5 mm nozzle-IP X6

(id 573) This test is performed to evaluate the protection of the equipment against the access of water in the presence of a water jet hose nozzle of 12.5 mm diameter, and a 100 l/min flow rate, spraying water from a 3m distance.

#### 14.2.7 - Test for second characteristic numeral 7 - temporary immersion (IP X7)

(id 574) This test is performed to evaluate the protection of the equipment against the access of water in case of temporary immersion between 0.15 and 1 m for 30 min.

#### 14.2.8 - Test for second characteristic numeral 8 - continuous immersion (IP X8)

(id 575) This test is performed to evaluate the protection of the equipment against the access of water in case of immersion at a depth and for a duration subject to agreement with the manufacturer. The test parameters selected shall take account of the condition that the enclosure will be continuously immersed in actual use. More severe conditions than for second characteristic numeral 7 shall be applied. Ingress of water in quantities causing harmful effects shall not be possible when the enclosure is continuously immersed in water under the conditions agreed. A pressure difference between the sample and the external ambient pressure may be applied in order to simulate depth > 1m (typically 1bar=10m).

#### 15 - Tests for protection against access to hazardous parts - Additional letter (A,B,C,D)

(id 673) The purpose of this test is to evaluate the protection of persons against access to hazardous parts. It is performed together with the relevant tests for protection of equipment against the penetration of solid foreign objects. It can be performed with the following access probes: A: sphere 50 mm diam.; B: jointed test finger 12 mm diam., 80 mm length; C: tool 2.5 mm diam., 80 mm length; D: wire 1.0 mm diam.

Accredited	[IEC17025]: YES	Execution time [day]: 0
Unit price	#1 [EUR]: 200 00	Limit #1: Applicable to any sample



# **PRICE LIST / Listino prezzi** your value certified § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

TNDEX



#### 14.2.9 - Test for second characteristic numeral 9 by high pressure and temperature water jetting (IP X9)

(id 833) This test is performed to evaluate the protection of the equipment against the access of water in the presence of a test nozzle spraying water at high temperature and pressure. For big enclosures, the spray position during the test is from all practical directions covering the entire surface area and the test duration is 1 min/m2 for a mimimum of 3 min. For small enclosures the spray position is from 0°, 30°, 60°, 90° and the test duration is 30 s per position.

```
Accredited [IEC17025]: YES
                                 Execution time [day]: 1
Unit price #1 [EUR]: 800.00
                                            Limit #1: largest dimension < 250 mm
                                           Limit #2: largest dimension >= 250 mm
Unit price #2 [EUR]: 1,200.00
Unit price #3 [EUR]: 1,800.00
                                           Limit #3: huge size sample
```

#### 14.2.3 - Test for second characteristic numeral 3 with spray nozzle (IP X3)

(id 886) This test is performed to evaluate the protection of the equipment against the access of water in the presence of an oscillating tube spraying water from -60° to + 60° from vertical

```
Accredited [IEC17025]: YES
                                 Execution time [day]: 1
Unit price #1 [EUR]: 400.00
                                            Limit #1: Very small sample (surface < 1m2)
Unit price #2 [EUR]: 800.00
                                            Limit #2: Medium sample (surface up to 3m2)
Unit price #3 [EUR]: 1,200.00
                                            Limit #3: Huge size sample
```

#### 13.2 - Tests for protection against solid foreign objects - First characteristic numeral 1 (IP 1X)

(id 1181) This test is performed to evaluate the protection of the equipment against the access of the object probe rigid sphere of 50 mm diameter without handle or guard, with a test force of 50 N.

```
Accredited [IEC17025]: YES
                                  Execution time [day]: 0
Unit price #1 [EUR]: 300.00
                                            Limit #1: Not relevant
```

#### 14.2.1 - Test for second characteristic numeral 1 with the drip box (IP X1)

(id 1376) This test is performed to evaluate the protection of the equipment against the access of water with a device which produces a uniform flow of water drops over the whole area of the enclosure.

```
Accredited [IEC17025]: YES
                                 Execution time [day]: 1
Unit price #1 [EUR]: 400.00
                                            Limit #1: both dimensions < 300 mm
Unit price #2 [EUR]: 800.00
                                            Limit #2: one dimension <= 300 mm and one dimension >= 300 mm
Unit price #3 [EUR]: 1,200.00
                                            Limit #3: both dimensions > 300 mm
```

#### 14.2.4 - Test for second characteristic numeral 4 with a spray nozzle (IP X4)

(id 1883) This test is performed to evaluate the protection of the equipment against the access of water in the presence of a spray nozzle with a water delivery rate of 10 l/min + or - 5%.

```
Accredited [TEC17025]: YES
                                  Execution time [day]: 2
Unit price #1 [EUR]: 400.00
                                            Limit #1: Very small sample (surface < 1m2)
Unit price #2 [EUR]: 800.00
                                            Limit #2: Medium sample (surface up to 5m2)
Unit price #3 [EUR]: 1,200.00
                                            Limit #3: Huge size sample
```

#### 14.2.2 - Test for second characteristic numeral 2 with the drip box (IP X2)

(id 2044) This test is performed to evaluate the protection of the equipment against the access of water with a device which produces a flow of water drops over the enclosure.

```
Accredited [IEC17025]: YES
                                 Execution time [day]: 1
Unit price #1 [EUR]: 400.00
                                            Limit #1: up to 40x40cm exposed surface
Unit price #2 [EUR]: 800.00
                                            Limit #2: up to 100x100cm exposed surface
Unit price #3 [EUR]: 1,200.00
                                            Limit #3: Huge size sample
```

#### IEC 60855-1:2016

(id 1228) Live working - Insulating foam-filled tubes and solid rods - Part 1: Tubes and rods of a circular cross-section

#### 5.3.2 - Visual checks

(id 1641) ---

```
Accredited [IEC17025]: NO
                                 Execution time [dav]: 0
Unit price #1: ask to office
                                            Limit #1: Not relevant
```

#### 5.3.3 - Dimensional checks

(id 1642) ---

Accredited [IEC17025]: NO Execution time [dav]: 0 Unit price #1: ask to office Limit #1: Not relevant



# PRICE LIST / Listino prezzi

## § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX

#### 5.4.2.1.3 - Tests before exposure to humidity

(id 1643) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant

#### 5.4.2.1.4 - Tests after exposure to humidity

(id 1644) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant

5.4.2.2.1 - Alternative dry test

(id 1645) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant

5.4.2.2.2 - Alternative test after exposure to immersion

(id 1646) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant

5.4.3 - Wet test

(id 1647) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant

5.5.1 - Bending test

(id 1648) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant

5.5.2 - Torsion test

(id 1649) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant

5.5.3 - Crushing test on insulating foam-filled tube

(id 1650) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant

5.5.4.1 - Bending ageing test

(id 1651) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant

5.5.4.2 - Dielectric test after mechanical ageing

(id 1652) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant

5.5.5 - Dye penetration test

(id 1653) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant

5.5.6 - Durability of marking

(id 1654) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant



# PRICE LIST / Listino prezzi § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



## IEC 60947-5-5 rev 2005-04

(id 1750) Low-voltage switchgear and controlgear - Part 5-5: Control circuit devices and switching elements - Electrical emergency stop device with mechanical latching function

#### 7.3.2 - Robustness of a button actuator

(id 2238) A button actuator shall withstand:

- a force applied in three mutually perpendicular axes;
- a torque in both directions of rotation, in each of the latched and unlatched positions, where the resetting action requires rotation of the push-botton.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 7.3.3 - Durability test

(id 2239) The three samples shall be subjected to the following test: the actuator of an emergency stop device shall be moved through its full travel, then it shall be reset in a manner to imitate human operation as closely as possible.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 7.4 - Conditioning procedures

(id 2240) The purpose of the following procedures is to expose the emergency stop devices to various environmental conditions in order to verify their fuctioning after such exposure.

Accredited [IEC17025]: NO Execution time [day]: 16
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 7.5 - Shock test

(id 2241) The three emergency stop devices which have been conditioned in accordance with 7.4 shall be tested each on one of the three mutually perpendicular axes

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

### IEC 61010-1:2010+AMD1:2016

(id 726) Safety requirements for electrical equipment for measurement, control, and laboratory use - Part 1: General requirements

#### 11.7 - Fluid pressure and leakage

(id 2348) ---

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 500.00 Limit #1: Applicable to any sample

#### 13 - Protection against liberated gases and substances, explosion and implosion

(id 2349) Equipment shall not liberate dangerous amounts of poisonous and injurious gases and substances in normal condition and in single fault condition. If potentially-hazardous substances are liberated, the operator shall not be directly exposed to a quantity of the substance that could cause harm.

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1 [EUR]: 1,200.00 Limit #1: Applicable to any sample

#### Unref. - Helium leak test

(id 2903) ---

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

### IEC 61076-2-101:2012

(id 2044) Connectors for electronic equipment - Product requirements - Part 2-101: Circular connectors - Detail specification for M12 connectors with screw-locking

#### 5.2.1 - Rated voltage - Rated impulse voltage - Pollution degree

(id 2790) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 150.00 Limit #1: Applicable to any sample



# PRICE LIST / Listino prezzi § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX

## 5.2.2 - Voltage proof

(id 2791) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 5.2.3 - Current-carrying capacity

(id 2792) --

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 300.00 Limit #1: Applicable to any sample

#### 5.2.4 - Contact resistance

(id 2793) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 150.00 Limit #1: Applicable to any sample

#### 5.2.5 - Insulation resistance

(id 2794) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 5.3.1 - IP degree of protection

(id 2795) ---

Accredited [IEC17025]: NO Execution time [day]: 2

Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 5.3.2 - Mechanical operation

(id 2796) ---

Accredited [IEC17025]: NO Execution time [day]: 2

Unit price #1 [EUR]: 500.00 Limit #1: Applicable to any sample

#### 5.3.3 - Insertion and withdrawal forces

(id 2797) ---

Accredited [IEC17025]: NO Execution time [day]: 2

Unit price #1 [EUR]: 300.00 Limit #1: Applicable to any sample

#### 5.3.5 - Polarizing method

(id 2798) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 5.3.6 - Vibration (sinusoidal)

(id 2799) ---

Accredited [IEC17025]: NO Execution time [day]: 2

Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 5.3.7 - Pressure differential

(id 2800) ---

Accredited [IEC17025]: NO Execution time [day]: 2

Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### IEC 61215-1-1:2016

(id 158) Terrestrial photovoltaic (PV) modules - Design qualification and type approval - Part 1-1: Special requirements for testing of crystalline silicon photovoltaic (PV) modules

#### 11.1 - MQT 01 - Visual inspection

(id 1286) This test of IEC 61215-2:2016 is applicable without modifications.

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant

page 130/234



# PRICE LIST / Listino prezzi

## § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

TNDEX

## 11.2 - MQT 02 - Maximum power determination

(id 1287) This test of IEC 61215-2:2016 is applicable without modifications.

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant

#### 11.3 - MQT 03 - Insulation test

(id 1288) This test of IEC 61215-2:2016 is applicable without modifications.

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant

#### 11.4 - MQT 04 - Measurement of temperature coefficients

(id 1289) This test of IEC 61215-2:2016 is applicable without modifications.

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant

#### 11.5 - MQT 05 - Measurment of nominal operating temperature (NMOT)

(id 1290) This test of IEC 61215-2:2016 is applicable without modifications.

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant

#### 11.6 - MQT 06.1 and MQT 06.2 - Performance at STC and NMOT

(id 1291) This test of IEC 61215-2:2016 is applicable without modifications.

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant

#### 11.7 - MQT 07 - Peformance at low irradiance

(id 1292) This test of IEC 61215-2:2016 is applicable without modifications.

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant

#### 11.8 - MQT 08 - Outdoor exposure test

(id 1293) This test of IEC 61215-2:2016 is applicable without modifications.

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant

#### 11.9 - MQT 09 - Hot-spot endurance test

(id 1294) This test of IEC 61215-2:2016 is applicable without modifications.

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant

#### 11.10 - MQT 10 - UV preconditioning test

(id 1295) This test of IEC 61215-2:2016 is applicable without modifications.

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant

### 11.11 - MQT 11 - Thermal cycling test

(id 1296) This test of IEC 61215-2:2016 is applicable without modifications. The technology specific current which needs to be applied according to test MQT 11 of IEC 61215-2:2016 shall be equal to the STC peak power current.

Execution time [day]: 0 Accredited [IEC17025]: NO

Unit price #1: ask to office Limit #1: Not relevant

#### 11.12 - MQT 12 - Humidity freeze test

(id 1297) This test of IEC 61215-2:2016 is applicable without modifications.

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant

#### 11.13 - MQT 13 - Damp heat test

(id 1298) This test of IEC 61215-2:2016 is applicable without modifications.

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant

page 131/234



# PRICE LIST / Listino prezzi

## § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX

11.14 - MQT 14 - Robustness of terminatons test (id 1299) This test of IEC 61215-2:2016 is applicable without modifications.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

#### 11.15 - MQT 15 - Wet leakage current test

(id 1300) This test of IEC 61215-2:2016 is applicable without modifications.

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant

#### 11.16 - MQT 16 - Static mechanical load test

(id 1301) This test of IEC 61215-2: 2016 is applicable without modifications.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

#### 11.17 - MQT 17 - Hail test

(id 1302) This test of IEC 61215-2: 2016 is applicable without modifications.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

#### 11.18 - MQT 18 - Bypass diode thermal test

(id 1303) This test of IEC 61215-2: 2016 is applicable without modifications.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

#### 11.19 - MQT 19 - Stabilization

(id 1304) This test of IEC 61215-2: 2016 is applicable with the following modifications:

for the definition of stabilization as per test MQT 19 of IEC 61215-2:2016, x = 0.01 shall be used for crystalline silicon PV modules.

Temperature is a critical parameter. For the measurement MQT 6.1 of IEC 61215-2:2016 it shall be ensured that measurement is performed at 25 + or - 2°C module temperature.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

### IEC 61215-1-4:2016

(id 285) TERRESTRIAL PHOTOVOLTAIC (PV) MODULES – DESIGN QUALIFICATION AND TYPE APPROVAL Part 1-4: Special requirements for testing of thin-film Cu(In,Ga)(S,Se)2 based terrestrial photovoltaic (PV) modules

#### 11.1 - MQT 01 - Visual inspection

(id 1739) This test of IEC 61215-2:2016 is applicable without modifications.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

#### 11.2 - MQT 02 - Maximum power determination

(id 1740) This test of IEC 61215-2:2016 is applicable without modifications.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

#### 11.3 - MQT 03 - Insulation test

(id 1741) This test of IEC 61215-2:2016 is applicable without modifications.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

#### 11.4 - MQT 04 - Measurement of temperature coefficients

(id 1742) This test of IEC 61215-2:2016 is applicable without modifications.

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1: ask to office Limit #1: Not relevant

page 132/234



# PRICE LIST / Listino prezzi

## § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX

# .

#### 11.5 - MQT 05 - Measurement of nominal module operating temperature (NMOT)

(id 1743) This test of IEC 61215-2:2016 is applicable without modifications.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

#### 11.6 - MQT 06.1 - Performance at STC

(id 1744) This test of IEC 61215-2:2016 is applicable without modifications.

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant

#### 11.6 - MQT 06.2 - 11.6 Performance at NMOT

(id 1745) This test of IEC 61215-2:2016 is applicable without modifications.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

#### 11.7 - MQT 07 - Performance at low irradiance

(id 1746) This test of IEC 61215-2:2016 is applicable without modifications.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

#### 11.8 - MQT 08 - Outdoor exposure test

(id 1747) This test of IEC 61215-2:2016 is applicable without modifications.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

#### 11.9 - MQT 09 - Hot-spot endurance test

(id 1748) This test of IEC 61215-2:2016 is applicable with the following modifications:

Cu(In,Ga)(S,Se)2 thin-film modules may exhibit performance changes with extended time in storage without light exposure (the "dark soak" effect). In order to minimize the influence of this dark soak effect, limit the time delay between the outdoor exposure or stabilization and the hot spot procedure to within 2 to 3 days; the modules are to be stored in the dark at ? 25 °C.

MQT 09.2 of IEC 61215-2:2016 shall be performed for any MLI module design. If module is constructed by interconnection of cell-like substructures, MQT 09.1 of IEC 61215-2:2016 may be applicable.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

#### 11.10 - MQT10 - UV preconditioning test

(id 1749) This test of IEC 61215-2:2016 is applicable without modifications.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

#### 11.11 - MQT 11 - Thermal cycling test

(id 1750) This test of IEC 61215-2:2016 is applicable with the following modifications:

The technology specific current which needs to be applied according to MQT 11 of IEC 61215-2:2016 shall be equal to 0,1 x STC peak power current.

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1: ask to office Limit #1: Not relevant

#### 11.13 - MQT 13 - Damp heat test

(id 1751) This test of IEC 61215-2:2016 is applicable with the following modifications:

MQT 13 of IEC 61215-2:2016 can be conducted according to the following methods:

Method A) Perform MQT 13 as defined in IEC 61215-2:2016.

Method B) Perform MQT 13 as defined in IEC 61215-2:2016 with applied forward bias as per §11.13.1 of IEC 61215-1-4:2016.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

## 11.12 - MQT 12 - Humidity-freeze test

(id 1752) This test of IEC 61215-2:2016 is applicable without modifications.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

#### 11.14 - MQT 14.1 - Robustness of terminations test - jbox

(id 1753) This test of IEC 61215-2:2016 is applicable without modifications.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

page 133/234



# § T - LABORATORY TESTS / Prove di laboratorio

PRICE LIST / Listino prezzi

# alb**ar**ubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX

#### 11.14 - MQT 14.2 - Robustness of terminations test - cord

(id 1754) This test of IEC 61215-2:2016 is applicable without modifications.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

#### 11.15 - MQT 15 - Wet leakage current test

(id 1755) This test of IEC 61215-2:2016 is applicable without modifications.

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant

#### 11.16 - MQT 16 - Static mechanical load test

(id 1756) This test of IEC 61215-2:2016 is applicable without modifications.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

#### 11.17 - MQT 17 - Hail test

(id 1757) This test of IEC 61215-2:2016 is applicable without modifications.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

#### 11.18 - MQT 18.1 - Bypass diode testing: bypass diode thermal test

(id 1758) This test of IEC 61215-2:2016 is applicable without modifications.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

#### 11.18 - MQT 18.2 - Bypass diode testing: bypass diode functionality test

(id 1759) This test of IEC 61215-2:2016 is applicable without modifications.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

#### 11.19 - MQT 19.1 - Initial stabilization

(id 1760) This test of IEC 61215-2:2016 is applicable with the following modifications:

For the definition of stabilization as per MQT 19 of IEC 61215-2:2016, x = 0.02 shall be used. Any kind of storage shall be done at temperature below 25 °C to avoid thermal activated processes affecting MQT 06.1 of IEC 61215-2:2016 measurement.

Initial stabilization is performed on all modules.

To fulfil MQT 19 requirements using light exposure, a minimum of two intervals each of at least 10 kWh/m2 are required. After this preconditioning all of the test modules shall be measured for STC power (MQT 06.1 of IEC 61215-2:2016).

If stabilization is performed outdoors in general no module temperature limits apply. The outdoor stabilization shall be proven at least with one module using the indoor method following the validation procedure from MQT 19 of IEC 61215-2:2016.

The minimum and maximum module temperatures observed during outdoor light exposure stabilization verification while the irradiance level is above 500 W/m2 shall be the minimum and maximum allowable module temperatures for all modules. If the module temperature falls outside of these limits the new module temperature range has to be re-verified.

Output power determination shall be performed after a minimum cooling time of 30 min and maximum 60 min.

A validated alternative procedure can be used in accordance to MQT 19 of IEC 61215-2:2016.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

#### 11.19 - MQT 19.2 - Final stabilization

(id 1761) For the definition of stabilization as per MQT 19 of IEC 61215-2:2016, x = 0.02 shall be used.

Any kind of storage shall be done at temperature below 25 °C to avoid thermal activated processes affecting MQT 06.1 of IEC 61215-2:2016 measurement. Final stabilization is performed on all modules after the test sequences to prove fulfilment of gate No. 2 requirement of IEC 61215-1:2016.

To fulfil MQT 19 requirements a minimum of two intervals of at least 10 kWh/m2 each are required.

If stabilization is performed outdoors in general no module temperature limits apply. The outdoor stabilization shall be proven at least with one module using the indoor method following MQT 19 of IEC 61215-2:2016.

The minimum and maximum module temperatures observed during outdoor light exposure stabilization verification while the irradiance level is above 500 W/m2

shall be the minimum
and maximum allowable module temperatures for all modules. If the module temperature falls outside of these limits the new module temperature range has to be

and maximum allowable module temperatures for all modules. If the module temperature falls outside of these limits the new module temperature range has to be re-verified.

Output power determination shall be performed after a minimum cooling time of 30 min and maximum 60 min.

A validated alternative procedure can be used in accordance to MQT 19 of IEC 61215-2:2016.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant



# PRICE LIST / Listino prezzi § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

TNDEX



## IEC 61215-1:2016

(id 157) Terrestrial photovoltaic (PV) modules - Design qualification and type approval - Part 1: Test requirements

#### 7.2.1 - Verification of rated label values (Gate 1)

(id 1207) Pmax Verification:

Each individual module shall meet the following criterion:

 $PmaxLab \times (1 + (|m1|[%]/100)) > PmaxNP \times (1-(|t1|[%]/100))$  or

 $PmaxLab \times (1 + (|m1|[%]/100)) = PmaxNP \times (1-(|t1|[%]/100)),$ 

PmaxLab is the measured maximum STC power of each module in the stabilized state;

PmaxNP is the maximum rated nameplate power of each module without tolerances;

m1 is the measurement uncertainty in % of laboratory for Pmax (expanded combined uncertainty (k=2), ISO/IEC Guide 98-3);

t1 is the manufacturer's rated lower production tolerance in % for Pmax.

For AveragePmaxLab, the following criterion shall apply:

AveragePmaxLab x (1 + (|m1|[%]/100)) > PmaxNP or

 $AveragePmaxLab \ x \ (1 + (|m1|[\%]/100)) = PmaxNP,$ 

AveragePmaxLab is the arithmetic average of the measured maximum STC power of the modules in stabilized condition. For multiple bins of power classes this formula has to be applied to each power class under investigation.

#### Voc Verification:

Each individual module shall meet the following criterion:

VocLab x (1 + (|m2|[%]/100)) < VocNP x <math>(1 - (|t2|[%]/100)) or

 $VocLab \times (1 + (|m2|[\%]/100)) = VocNP \times (1 - (|t2|[\%]/100)),$ 

where

VocLab is the measured maximum Voc of each module in the stabilized state:

VocNP is the maximum rated nameplate Voc of each module without tolerances;

m2 is the measurement uncertainty in % of laboratory for Voc;

t2 is the manufacturer's rated upper production tolerance in % for Voc.

#### Isc Verification:

Each individual module shall meet the following criterion:

 $||scLab|| \times (1 + (|m3|[\%]/100)) < ||scNP|| \times (1 - (|t3|[\%]/100))$  or

 $IscLab \times (1 + (|m3|[\%]/100)) = IscNP \times (1-(|t3|[\%]/100),$ 

where

IscLab is the measured maximum Isc of each module in the stabilized state;

IscNP is the maximum rated nameplate Isc of each module without tolerances;

m3 is the measurement uncertainty in % of laboratory for lsc;

t3 is the manufacturer's rated upper production tolerance in % for lsc.

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1: ask to office Limit #1: Not relevant

#### 7.2.2 - Maximum power degradation during type approval testing (Gate 2)

(id 1208) At the end of each test sequence or for sequence B after bypass diode test, the maximum power output drop of each module Pmax (Lab\_Gate No. 2) shall be less than 5 %, referenced to the module's initial measured output power Pmax (Lab\_Gate No. 1).

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1: ask to office Limit #1: Not relevant

#### 7.2.3 - Electrical circuitry

(id 1209) Verification of the presence of an open-circuit during the tests

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1: ask to office Limit #1: Not relevant

#### 7.3 - Visual defects

(id 1210) Verification of the presence of any visual defects that may cause a risk of reliability loss, including power output.

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1: ask to office Limit #1: Not relevant



# PRICE LIST / Listino prezzi

## § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX

#### 7.4 - Electrical safety

(id 1211) Verification of insulation and wet leakage requirements

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1: ask to office Limit #1: Not relevant

#### 5.1 - Name plate and markings

(id 1348) This test is a visual inspection performed in order to verify the compliance of markings and nameplate

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 50.00 Limit #1: Not relevant

### IEC 61215-2:2016+COR1:2018

(id 159) Terrestrial photovoltaic (PV) modules - Design qualification and type approval - Part 2: Test procedures

#### 4.1 - MQT 01 - Visual Inspection

(id 985) The purpose of the visual inspection is to detect any visual defects that may cause a risk of reliability loss, including power output. In some instances further testing may be required to finally decide if major visual defects exist or not. The inspection is performed in Illumination conditions of at least 1000 lx.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 50.00 Limit #1: Applicable to any sample

#### 4.2 - MQT 02 - Maximum Power Determination (Flash test)

(id 986) The purpose of this test is to determine the variation in power of the module after the initial stabilization and after the various environmental stress tests. The values are the result of a differential measurement performed making a comparison between a control module and any other measured module.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 150.00 Limit #1: Applicable to any sample

#### 4.3 - MQT 03 - Insulation Test

(id 987) The purpose of this test is to determine whether or not the module is sufficiently well insulated between live parts and accessible parts.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 100.00 Limit #1: Applicable to any sample

#### 4.4 - MQT 04 - Measurement of temperature coefficients

(id 988) Determination of the temperature coefficients of current, voltage and peak power

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 600.00 Limit #1: Applicable to any sample

#### 4.5 - MQT 05 - Measurement of Nominal Module Operating Temperature (NMOT)

(id 989) Determination of the NMOT (Nominal Module Operating Temperature) of the module.

NMOT is defined as the equilibrium mean solar cell junction temperature within an open-rack mounted module operating near peak power in the following standard reference environment (SRE):

- Tilt angle: (37 ± 5)°
- Total irradiance: 800 W/m2
- Ambient temperature: 20 °C
- Wind speed: 1 m/s
- Electrical load: A resistive load sized such that the module will operate near its maximum power point at STC or an electronic maximum power point tracker (MPPT).

Accredited [IEC17025]: NO Execution time [day]: 3
Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

#### 4.6.3.1 - MQT 06.1 - Performance measuring at STC Cat. 0

(id 990) Determination of how the electrical performance of the module varies with load at STC (1000 W/m2, 25 °C cell temperature, with the IEC 60904-3 reference solar spectral irradiance distribution)

The values measured are normalized by software using the correction method No.2 of IEC 60891 ed.2 – December 2009.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 600.00 Limit #1: Applicable to any sample

#### 4.6.3.2 - MQT 06.2 - Performance measuring at NMOT

(id 991) Determination of how the electrical performance of the module varies with load at NMOT. The values measured are normalized by software using the correction method No.2 of IEC 60891 ed.2 – December 2009.

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant



# PRICE LIST / Listino prezzi § T - LABORATORY TESTS / Prove di laboratorio

albarubens

QM ref: ALL 7.1.4 Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



#### 4.7 - MQT 07 - Performance at low irradiance

(id 992) Determination of how the electrical performance of the module varies with load at 25 °C and an irradiance of 200 W/m2 (as measured by a suitable reference device), in accordance with IEC 60904-1 using natural sunlight

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 400.00 Limit #1: Applicable to any sample

#### 4.8 - MQT 08 - Outdoor exposure test

(id 993) Assessment of the ability of the module to withstand exposure to outdoor conditions and any synergistic degradation effects which may not be detected by laboratory tests. During the test, the sample is connected to a resistive load sized such that the module will operate near its maximum power point at STC or an electronic maximum power point tracker (MPPT).

Accredited [IEC17025]: NO Execution time [day]: 15
Unit price #1 [EUR]: 500.00 Limit #1: Applicable to any sample

#### 4.9.5.2 - MQT 09.1 - Hot-spot endurance test (WBT)

(id 994) Determination of the ability of the module to withstand hot-spot heating effects, for example solder melting or deterioration of the encapsulation. This defect could be provoked by cracked or mismatched cells, interconnect failures, partial shadowing or soiling (equal to IEC61730 - MST22)

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

#### 4.9.5.3 - MQT 09.2 - Hot-spot endurance test (MLI)

(id 995) Possible subcontract

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

#### 4.10 - MQT 10 - UV preconditioning test

(id 996) The purpose of this test is to precondition the module with ultra-violet (UV) radiation before the thermal cycle/humidity freeze tests to identify those materials and adhesive bonds that are susceptible to UV degradation.

Due to the UV test chamber design, the irradiance at the module test plane does not exceed 250 W/m2 at wavelengths between 280 nm and 400 nm (with no more than 10 % in the wavelength band between 280 nm and 320 nm) and its uniformity is ± 15 % over the test plane.

Accredited [IEC17025]: NO Execution time [day]: 10
Unit price #1 [EUR]: 800.00 Limit #1: Single module up to 1000x2000mm (72 cells)
Unit price #2 [EUR]: 1,800.00 Limit #2: Special sample

#### 4.11 - MQT 11 - Thermal cycling test (200 cycles)

(id 997) The purpose of this test is to determine the ability of the module to withstand thermal mismatch, fatigue and other stresses caused by repeated changes of temperature (200 cycles).

Each module is connected to an appropriate current supply by connecting the positive terminal of the module to the positive terminal of the power supply and the second terminal accordingly. During the thermal cycling test a continuous current flow is set during the heat up cycle to the technology specified current in 4.11.2 at temperature from -40 °C to 80 °C. During cool down, the -40 °C dwell phase and temperatures above 80 °C the continuous current is reduced to no more than 1,0 % of the measured STC peak power current to measure continuity.

The module(s) is (are) subject to cycling between measured module temperatures of  $(-40 \pm 2)$  °C and  $(+85 \pm 2)$  °C, in accordance with the profile in Figure 9 in IEC 61215-2.

The rate of change of temperature between the low and high extremes does not exceed 100 °C/h and the module temperature remains stable at each extreme for a period of at least 10 min. The cycle time does not exceed 6 h unless the module has such a high heat capacity that a longer cycle is required. The number of cycles is as shown in the relevant sequences in Figure 1 of IEC 61215-1:2016.

Accredited [IEC17025]: NO Execution time [day]: 60
Unit price #1 [EUR]: 1,600.00 Limit #1: 200 cycles - single PV up to 1000x2000mm

#### 4.11 - MQT 11 - Thermal cycling test (50 cycles)

(id 998) The purpose of this test is to determine the ability of the module to withstand thermal mismatch, fatigue and other stresses caused by repeated changes of temperature (50 cycles).

Each module is connected to an appropriate current supply by connecting the positive terminal of the module to the positive terminal of the power supply and the second terminal accordingly. During the thermal cycling test a continuous current flow is set during the heat up cycle to the technology specified current in 4.11.2 at temperature from –40 °C to 80 °C. During cool down, the –40 °C dwell phase and temperatures above 80 °C the continuous current is reduced to no more than 1,0 % of the measured STC peak power current to measure continuity.

The module(s) is (are) subject to cycling between measured module temperatures of  $(-40 \pm 2)$  °C and  $(+85 \pm 2)$  °C, in accordance with the profile in Figure 9 in IEC 61215-2.

The rate of change of temperature between the low and high extremes does not exceed 100 °C/h and the module temperature remains stable at each extreme for a period of at least 10 min. The cycle time does not exceed 6 h unless the module has such a high heat capacity that a longer cycle is required. The number of cycles is as shown in the relevant sequences in Figure 1 of IEC 61215-1:2016.

Accredited [IEC17025]: NO Execution time [day]: 15
Unit price #1 [EUR]: 500.00 Limit #1: 50 cycles - single module up to 1000x2000mm (72 cells)



# PRICE LIST / Listino prezzi

## § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

TNDEX



#### 4.12 - MQT 12 - Humidity freeze test

(id 999) The purpose of this test is to determine the ability of the module to withstand the effects of high temperature and humidity followed by sub-zero temperatures. This is not a thermal shock test.

In one full cycle period the is subject to a temperature rising from +25°C till +85°C, with a maximum gradient of 100 °C/h. Once reached 85°C with 85 ± 5 % RH the minimum dwell time is 20 h. Then the temperature decreases with a maximum gradient of 100 °C/h until 0°C and then with a maximum gradient of 200 °C/h until -40°C, with no RH control. Once reached -40°C, the minimum dwell time is 30 min. Then, the temperature rises again until +25°C, where a new cycle begins. The maximum time between two +85°C peaks is 4h.

One cycle is repeated as specified by the standard.

Accredited [IEC17025]: NO Execution time [day]: 10 Limit #1: 10 cycles - Single module up to 1000x200mm (72 cells) Unit price #1 [EUR]: 1,100.00

#### 4.13 - MQT 13 - Damp heat test

(id 1000) The purpose of this test is to determine the ability of the modules/junction boxes to withstand the effects of long-term penetration of humidity.

Accredited [IEC17025]: NO Execution time [day]: 42 Unit price #1 [EUR]: 1,200.00 Limit #1: 1000h - Single module up to 1000x200mm (72 cells) Limit #2: 250h - Single module up to 1000x200mm (72 cells) Unit price #2 [EUR]: 300.00

#### 4.14.2 - MQT 14.1 - Robustness of terminations - jbox

(id 1001) The purpose of the robustness of terminations test is to determine that the terminations, the attachment of the terminations, and the attachment of the cables to the body of the module will withstand stresses that are likely to be applied during normal assembly or handling operations.

A force of 40 N shall be gradually applied in each direction parallel to the mounting surface parallel to the module edges, in steps of 90°

Then, a force of 40 N shall be gradually applied without jerks, in a direction perpendicular to the mounting surface. The pull force should be applied at the centre

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 300.00 Limit #1: Applicable to any sample

#### 4.14.3 - MQT 14.2 - Robustness of terminations - cord

(id 1002) The purpose of the robustness of terminations test is to determine that the terminations, the attachment of the terminations, and the attachment of the cables to the body of the module will withstand stresses that are likely to be applied during normal assembly or handling operations. This test can be omitted if junction box is qualified in accordance to IEC 62790.

During the pull test, the cable is pulled for duration of 1 s, 50 times, without jerks in the direction of the axis with the relevant force as specified in the standard (Table 1).

During the torque test, a torque as specified in Table 2 is applied for 1 min. The twist or torsion inside the cable gland or other cord anchorage shall not exceed 45°. The cable is held in position by the cord anchorage.

Accredited [IEC17025]: NO Execution time [day]: 2 Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 4.15 - MQT 15 - Wet leakage current test

(id 1003) The purpose of this test is to evaluate the insulation of the module under wet operating conditions and verify that moisture from rain, fog, dew or molten snow does not enter the active parts of the module circuitry, where it might cause corrosion, a ground fault, or a safety hazard.

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 150.00 Limit #1: Applicable to any sample

#### 4.16 - MQT 16 - Static mechanical load test

(id 1004) The purpose of this test is to determine the ability of the module to withstand a minimum static load, such as snow or ice.

Additional requirements may apply for certain installations and climates.

The test load is defined as:

Test load =  $ym \times design load$ , where ym is at least > or = 1,5.

The minimum required design load per this standard is 1 600 Pa that results in a minimum test load of 2 400 Pa.

Per each mounting method the manufacturer may specify higher design load(s) for positive (downward) and negative (upward) and also a higher ym for certain applications. Inhomogeneous snow loads are not covered by this test.

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

#### 4.17 - MQT 17 - Hail test

(id 1005) Verification that the sample is capable of withstanding the impact of hailstones in different locations:

- 01 Any corner of the module window, not more than one radius from the module edge
- 02 Any edge of the module, not more than one radius of ice-ball from the module edge.
- 03 Over edges of the circuit (e.g. individual cells).
- 04 Over edges of the circuit (e.g. individual cells).
- 05 Over the circuit near interconnects (i.e. cell interconnects and bus ribbons).
- 06 Over the circuit near interconnects (i.e. cell interconnects and bus ribbons).
- 07 On the module window, not more than half diameter of ice ball from one of the points at which the module is mounted to the supporting structure.
- 08 On the module window, not more than half diameter of ice ball from one of the points at which the module is mounted to the supporting structure.
- 09 On the module window, at points farthest from the points selected above.
- 10 On the module window, at points farthest from the points selected above.

page 138/234



# PRICE LIST / Listino prezzi your value certified § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

#### TNDEX



11 Any points which may prove especially vulnerable to hail impact like over the junction box.

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 600.00 Limit #1: Applicable to pv module Unit price #2 [EUR]: 1,200.00 Limit #2: Applicable to any sample

#### 4.18.1 - MQT 18.1 - Bypass diode thermal test

(id 1006) The test is designed to assess the adequacy of the thermal design and relative long-term reliability of the bypass diodes used to limit the detrimental effects of module hot-spot susceptibility. Its purpose is to determine the diode's temperature characteristic and its maximum diode junction temperature TJ under continuous operation. TJ is obtained from VD at Tamb =75 °C, ID = Isc from a VD versus TJ characteristic obtained according to § 4.18.1.3.

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 400.00 Limit #1: Applicable to any sample

#### 4.18.2 - MQT 18.2 - Bypass diode functionality test

(id 1007) The purpose of this test is to verify that the bypass diode(s) of the test samples remain(s) functional following MQT 09 and MQT 18.1. In case of PV modules without bypass diodes this test can be omitted.

Accredited [IEC17025]: NO Execution time [day]: 2 Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 4.19.5 - MQT 19.1 - Initial stabilization

(id 1008) Exposition to sunlight while open-circuited, before starting any other test sequence. Only irradiance levels > 500 W/m2 contribute to the integrated irradiance.

Accredited [IEC17025]: NO Execution time [day]: 2 Unit price #1 [EUR]: 100.00 Limit #1: Crystalline module

#### 4.19.6 - MQT 19.2 - Final stabilization

(id 1009) ---

Accredited [IEC17025]: NO Execution time [day]: 2 Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 4.6.3.1 - MQT 06.1B - Performance measuring at STC (Flash test)

(id 1010) Flash test with solar simulator for performance measuring at STC

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 150.00 Limit #1: Not relevant

#### - On-field module management

(id 1011) Only for category 3 photovoltaic tests

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 50.00 Limit #1: Not relevant

#### 4.6.3.1 - MQT 06.1 - Performance measuring at STC Cat. III

(id 1566) The purpose of this test is to determine the electrical performance of the module. The values are the result of a differential measurement performed making a comparison between a control module and the module under test.

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1: ask to office Limit #1: Not relevant

#### 4.10 - MQT 10 bis - UV preconditioning test - internal variation for samples that are not PV modules

(id 2127) Test method inspired to the principles of §4.10 - MQT 10 describe in standard IEC 1215-2 but internally developed by the laboratory in order to be adapted also to samples that are not PV modules. This test is not covered by accreditation.

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1: ask to office Limit #1: Not relevant

## IEC 61730-1:2016

(id 287) Photovoltaic (PV) module safety qualification - Part 1: Requirements for construction

#### - Check list-Procedure IEC 61730-1

(id 1285) This part of IEC 61730 describes the fundamental construction requirements for photovoltaic (PV) modules in order to provide safe electrical and mechanical operation during their expected lifetime. This part of IEC 61730 pertains to the particular requirements of construction

Accredited [IEC17025]: NO Execution time [day]: 15 Unit price #1 [EUR]: 1,800.00 Limit #1: Applicable to any sample Unit price #2 [EUR]: 300.00 Limit #2: Update of already existing certificate



# PRICE LIST / Listino prezzi § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



## IEC 61730-2:2016

(id 282) Photovoltaic (PV) module safety qualification - Part 2: Requirements for testing

#### 10.2 - MST 01 - Visual inspection

(id 1147) The purpose of the visual inspection is to detect any visual defects that may cause a risk of reliability loss, including power output. In some instances further testing may be required to finally decide if major visual defects exist or not. The inspection is performed in Illumination conditions of at least 1000 lx.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 50.00 Limit #1: Not relevant

#### 10.3 -MST 02 - Perfomance at STC

(id 1148) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 600.00 Limit #1: Not relevant

#### 10.4 - MST 03 - Maximum power determination

(id 1149) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 150.00 Limit #1: Not relevant

#### 10.5 - MST 04 - Insulation thickness test

(id 1150) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

#### 10.6 - MST 05 - Durability of marking

(id 1151) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

#### 10.7 - MST 06 - Sharp edge test

(id 1152) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

#### 10.8 - MST 07 - Bypass diode functionality test

(id 1153) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 1,400.00 Limit #1: Not relevant

#### 10.9 - MST 11 - Accessibility test

(id 1154) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 150.00 Limit #1: Not relevant

#### 10.10 - MST 12 - Cut susceptibility

(id 1155) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 150.00 Limit #1: Not relevant

#### 10.11 - MST 13 - Continuity test of equipotential bonding

(id 1156) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

#### 10.12 - MST 14 - Impulse voltage test

(id 1157) Possible subcontract (?)

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 500.00 Limit #1: Not relevant

page 140/234



# PRICE LIST / Listino prezzi

## § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

TNDEX

#### 10.13 - MST 16 - Insulation test

(id 1158) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant

#### 10.14 - MST 17 - Wet leakage current test

(id 1159) The purpose of this test is to evaluate the insulation of the module under wet operating conditions and verify that moisture from rain, fog, dew or molten snow does not enter the active parts of the module circuitry, where it might cause corrosion, a ground fault or a safety hazard.

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 150.00 Limit #1: Not relevant

#### 10.15 - MST 21 - Temperature test

(id 1160) This temperature test is designed to determine the maximum reference temperatures for various components and materials used to construct the PV module, in order to verify the suitability of their use.

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 800.00 Limit #1: Not relevant

#### 10.16 - MST 22 - Hot-spot endurance test

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 800.00 Limit #1: Not relevant

#### 10.19 - MST 25 - Bypass diode thermal test

(id 1162) This test is equivalent to MQT 18 in IEC 61215-2. Both MQT 18.1 and MQT 18.2 shall be performed.

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 400.00 Limit #1: Not relevant

#### 10.20 - MST 26 - Reverse current overload test

(id 1163) PV modules contain electrically conductive material contained in an insulating system. Under reverse current fault conditions the electrical conductors and the cells of the PV module are forced to dissipate energy as heat prior to circuit interruption by an over-current protector installed in the system. This test is intended to determine the acceptability of the risk of ignition or fire from this condition.

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 350.00 Limit #1: Not relevant

#### 10.21 - MST 32 - Module breakage test

(id 1164) The purpose of this test is to provide confidence that risk of physical injuries can be minimized if the PV module is broken in its specified installation. For building integrated or overhead applications additional tests may be required according to relevant building codes.

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 800.00 Limit #1: Not relevant

#### 10.22 - MST 33 - Screw connections test

(id 1165) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Limit #1: Not relevant Unit price #1: ask to office

#### 10.23 - MST 34 - Static mechanical load test

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 800.00 Limit #1: Not relevant

#### 10.26 - MST 37 - Materials creep test

(id 1167) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant

#### 10.27 - MST 42 - Robustness of terminations

(id 1168) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 1,300.00 Limit #1: Not relevant



# **PRICE LIST / Listino prezzi**

## § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

TNDEX

#### 10.28 - MST 51 - Thermal cycling test

(id 1169) ---

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 500.00 Limit #1: 50 cycles Unit price #2 [EUR]: 1,600.00 Limit #2: 200 cycles

#### 10.29 - MST 52 - Humidity freeze test

(id 1170) ---

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 1,100.00 Limit #1: Not relevant

#### 10.30 - MST 53 - Damp heat test

Accredited [IEC17025]: NO Execution time [day]: 42 Unit price #1 [EUR]: 1,200.00 Limit #1: 1000h - Single module up to 1000x200mm (72 cells) Unit price #2 [EUR]: 300.00 Limit #2: 200h - Single module up to 1000x200mm (72 cells)

#### 10.31 - MST 54 - UV test

(id 1172) ---

Accredited [IEC17025]: NO Execution time [day]: 0 Limit #1: Single module 1000x2000mm (72 cells) Unit price #1 [EUR]: 800.00 Unit price #2 [EUR]: 1,800.00 Limit #2: Special sample

#### 10.32 - MST 55 - Cold conditioning

(id 1173) ---

Accredited [IEC17025]: NO Execution time [day]: 2 Unit price #1: ask to office Limit #1: Not relevant

#### 10.33 - MST 56 - Dry heat conditioning

(id 1174) ---

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1: ask to office Limit #1: Not relevant

#### A.2 - Module output power

(id 1175) ---

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1: ask to office Limit #1: Not relevant

#### A.3 - Wet insulation test

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1: ask to office Limit #1: Not relevant

#### A.4 - Visual inspection

(id 1177) ---

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1: ask to office Limit #1: Not relevant

#### A.5 - Bypass diodes

(id 1178) ---

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1: ask to office Limit #1: Not relevant

#### A.6 - Continuity test for equipotential bonding

(id 1179) ---

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1: ask to office Limit #1: Not relevant

#### 10.18 - MST 24 - Ignitability test

(id 1336) This test determines the ignitability of PV modules by direct small flame impingement under zero impressed irradiance by external heat sources using vertically oriented test specimens. This test assesses ignitability of outer surfaces. The test method is based on ISO 11925-2

Accredited [IEC17025]: NO Execution time [dav]: 30 Subcontract: Lapi Spa - accredited Laboratory 0086 Limit #1: Module area 4m2 (3x60cells or 2x72 cells) Unit price #1 [EUR]: 900.00

page 142/234



# PRICE LIST / Listino prezzi § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX

#### 10.17 - MST 23 - Fire test

(id 1337) ---

Accredited [IEC17025]: NO Execution time [day]: 30 Subcontract: Lapi Spa - accredited Laboratory 0086 Unit price #1 [EUR]: 2,400.00 Limit #1: Module area 4m2 (3x60cells or 2x72 cells)

#### unref - Wait time 48-96h

(id 1873) ---

Accredited [IEC17025]: NO Execution time [day]: 4
Unit price #1: ask to office Limit #1: Not relevant

#### 10.24 - MST 35 - Peel test

(id 1874) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

### IEC 62208:2023

(id 1986) Empty enclosures for low-voltage switchgear and controlgear assemblies - General requirements

#### 9.14 - Resistance to corrosion

(id 2728) ---

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

### IEC 62262:2002

(id 33) Degrees of protection provided by enclosures for electrical equipment against external mechanical impacts (IK code)

### 6 - Test to verify the protection against mechanical impact

(id 672) ---

#### 6 - OLD? --- Test to verify the protection against mechanical impact on batteries

(id 1034) OLD?

### IEC 62282-2-100:2020

(id 1881) Fuel cell technologies: Fuel cell modules - Safety

#### 5.14.7 - Freeze/thaw cycles tests

(id 2513) This test is only applicable to PEFC modules with a storage or operating temperature below 0°C.

Accredited [IEC17025]: NO Execution time [day]: 5
Unit price #1 [EUR]: 4,400.00 Limit #1: Applicable to any sample

#### 5.3.3 - Gas leakage test - Pressure drop method

(id 2514) This test is not applicable for those gas leakages of the fuel cell modules with:

- operating temperatures higher than the auto-ignition temperature of the combustible gas;
- fuel cells within a gas-tight vessel already proven according to the relevant national regulations.

Where it is impractical to use the full stack, a stack with a reduced, but still representative, number of cells may be used. Leakages shall be calculated based on the ratio of cell numbers.

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample



# § T - LABORATORY TESTS / Prove di laboratorio

**PRICE LIST / Listino prezzi** 

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX

# 5.2 - Shock and vibration test

(id 3029) The fuel cell module shall be subject to the shock and vibration test limits stated in the manufacturer's documentation.

Accredited [IEC17025]: NO Execution time [day]: 10

Unit price #1 [EUR]: 5,000.00 Limit #1: Applicable to any sample

#### 5.3.2 - Gas leakage test - Flow meter method

(id 3030) ---

Accredited [IEC17025]: NO Execution time [day]: 2

Unit price #1 [EUR]: 1,200.00 Limit #1: Applicable to any sample

#### 5.4 - Normal operation test

(id 3031) ---

Accredited [IEC17025]: NO Execution time [day]: 2

Unit price #1 [EUR]: 1,200.00 Limit #1: Applicable to any sample

#### 5.5 - Allowable working pressure test

(id 3032) ---

Accredited [IEC17025]: NO Execution time [day]: 2

Unit price #1 [EUR]: 1,000.00 Limit #1: Applicable to any sample

#### 5.6 - Pressure withstanding test of cooling system

(id 3033) This test shall be performed if the cooling system is not tested during the allowable working pressure test.

Accredited [IEC17025]: NO Execution time [day]: 2

Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

#### 5.7 - Continuous and short-time electrical rating

(id 3034) ---

Accredited [IEC17025]: NO Execution time [day]: 2

Unit price #1 [EUR]: 500.00 Limit #1: Applicable to any sample

#### 5.8 - Overpressure test

(id 3035) ---

Accredited [IEC17025]: NO Execution time [day]: 2

Unit price #1 [EUR]: 2,000.00 Limit #1: Not relevant

#### 5.9 - Dielectric strength test

(id 3036) ---

Accredited [IEC17025]: NO Execution time [day]: 2

Unit price #1 [EUR]: 300.00 Limit #1: Up to 5kV dc

#### 5.10 - Differential pressure test

(id 3037) ---

Accredited [IEC17025]: NO Execution time [day]: 2

Unit price #1 [EUR]: 1,000.00 Limit #1: Applicable to any sample

#### 5.11 - Gas leakage test (repeat)

(id 3038) ---

Accredited [IEC17025]: NO Execution time [day]: 2

Unit price #1 [EUR]: 1,200.00 Limit #1: Applicable to any sample

#### 5.12 - Normal operation (repeat)

(id 3039) ---

Accredited [IEC17025]: NO Execution time [day]: 2

Unit price #1 [EUR]: 1,200.00 Limit #1: Applicable to any sample

#### 5.13 - Flammable concentration test

(id 3040) ---

Accredited [IEC17025]: NO Execution time [day]: 2

Unit price #1 [EUR]: 1,800.00 Limit #1: Applicable to any sample



# PRICE LIST / Listino prezzi § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX

#### 5.14.2 - Fuel starvation test

(id 3041) ---

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 1,400.00 Limit #1: Not relevant

#### 5.14.3 - Oxygen or oxidant starvation test

(id 3042) --

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 1,400.00 Limit #1: Not relevant

#### 5.14.4 - Short-circuit test

(id 3043) ---

Accredited [IEC17025]: NO Execution time [day]: 2

Unit price #1 [EUR]: 1,600.00 Limit #1: Applicable to any sample

#### 5.14.5 - Lack of cooling or impaired cooling test

(id 3044) ---

Accredited [IEC17025]: NO Execution time [day]: 2

Unit price #1 [EUR]: 1,400.00 Limit #1: Applicable to any sample

### 5.14.6 - Crossover monitoring system test

(id 3045) ---

Accredited [IEC17025]: NO Execution time [day]: 2

Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

# IEC 62282-3-100:2019

(id 2100) Fuel cell technologies - Part 3-100: Stationary fuel cell power systems - Safety

#### 5.4.2 - Pneumatic leakage tests

(id 3046) ---

Accredited [IEC17025]: NO Execution time [day]: 2

Unit price #1 [EUR]: 1,200.00 Limit #1: Applicable to any sample

## 5.5.2 - Pneumatic strength tests

(id 3047) ---

Accredited [IEC17025]: NO Execution time [day]: 2

Unit price #1 [EUR]: 1,200.00 Limit #1: Applicable to any sample

#### 5.6 - Normal operation type test

(id 3048) ---

Accredited [IEC17025]: NO Execution time [day]: 2

Unit price #1 [EUR]: 1,200.00 Limit #1: Applicable to any sample

#### 5.7 - Electrical overload test

(id 3049) ---

Accredited [IEC17025]: NO Execution time [day]: 2

Unit price #1 [EUR]: 1,200.00 Limit #1: Applicable to any sample

#### 5.8 - Shutdown parameters

(id 3050) ---

Accredited [IEC17025]: NO Execution time [day]: 2

Unit price #1 [EUR]: 1,800.00 Limit #1: Applicable to any sample

### 5.9 - Burner operating characteristics tests

(id 3051) ---

Accredited [IEC17025]: NO Execution time [day]: 2

Unit price #1 [EUR]: 2,500.00 Limit #1: Not relevant



# **PRICE LIST / Listino prezzi** § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

TNDEX

#### 5.10 - Automatic control of burners and catalytic oxidation reactors

(id 3052) ---

Accredited [IEC17025]: NO Execution time [day]: 2

Unit price #1 [EUR]: 2,500.00 Limit #1: Not relevant

#### 5.11 - Exhaust gas temperature test

Accredited [IEC17025]: NO Execution time [day]: 2

Unit price #1 [EUR]: 1,000.00 Limit #1: Applicable to any sample

#### 5.12 - Surface and component temperatures

Accredited [IEC17025]: NO Execution time [day]: 2

Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

#### 5.13 - Wind tests

(id 3055) ---

Accredited [IEC17025]: NO Execution time [day]: 2

Unit price #1 [EUR]: 8,000.00 Limit #1: Applicable to any sample

#### 5.14 - Rain test

(id 3056) A simulated rain test shall be performed equivalent to a minimum IP rating of 3 (second characteristic numeral 3) as defined in IEC 60529 or the manufacturer's higher declared IP rating. Compliance is defined in 5.14.3.

Accredited [IEC17025]: NO Execution time [day]: 2

Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

#### 5.15 - Emissions

(id 3057) ---

Accredited [IEC17025]: NO Execution time [day]: 2

Unit price #1 [EUR]: 1,800.00 Limit #1: Not relevant

#### 5.16 - Blocked condensate line test

(id 3058) ---

Accredited [IEC17025]: NO Execution time [day]: 2

Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

#### 5.17 - Condensate discharge test

(id 3059) ---

Accredited [IEC17025]: NO Execution time [day]: 2

Unit price #1 [EUR]: 1,000.00 Limit #1: Applicable to any sample

#### 5.18 - Electrical safety tests

(id 3060) ---

Accredited [IEC17025]: NO Execution time [day]: 2

Unit price #1 [EUR]: 500.00 Limit #1: Applicable to any sample

#### 5.19 - EMC test

(id 3061) ---

Execution time [day]: 2 Accredited [IEC17025]: NO Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 5.20 - Venting system leakage test

(id 3062) --

Accredited [IEC17025]: NO Execution time [day]: 2

Limit #1: Applicable to any sample Unit price #1 [EUR]: 4,500.00

#### 5.21 - Leakage tests (repeat)

(id 3063) ---

Accredited [IEC17025]: NO Execution time [day]: 2

Unit price #1 [EUR]: 1,200.00 Limit #1: Applicable to any sample



# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



## IEC 62368-1:2018

(id 1865) Audio/video, information and communication technology equipment - Part 1: Safety requirements

#### 5.4.8 - Humidity conditioning

(id 2481) The sample shall withstand an ordinary or tropical humidity conditioning.

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

#### 5.4.9.1 - Electric strength test (Test procedure for type testing of solid insulation)

(id 2482) The test voltage for the electric strength of basic insulation,

supplementary insulation or reinforced insulation is the highest value determined using one of the three methods proposed by the standard.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 300.00 Limit #1: Up to 5 kV dc

# IEC 62716:2013+COR1:2014

(id 5) Ammonia corrosion testing of photovoltaic (PV) modules

#### 7 - Ammonia resistance test

(id 149) ---

Accredited [IEC17025]: NO Execution time [day]: 25
Unit price #1 [EUR]: 3,200.00 Limit #1: Applicable to any sample

#### 4.2 - Bypass diode functionality test

(id 1098) The purpose of this test is to verify that the bypass diode(s) of the test samples remain(s) functional following the ammonia exposure

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 200.00 Limit #1: Not relevant

# IEC 62984-2:2020

(id 1879) High-temperature secondary batteries - Part 2: Safety requirements and tests

#### 6.3.3 - External fire exposure test

(id 2509) The purpose of this test is to ensure safety in the event that the module or battery is exposed to an external fire.

Accredited [IEC17025]: NO Execution time [day]: 2 Unit price #1 [EUR]: 6,000.00 Limit #1: Applicable to any sample

# IEC 62990-1:2019

(id 1943) Workplace atmospheres - Part 1: Gas detectors - Performance requirements of detectors for toxic gases

### 5.4.2 - Unpowered storage

(id 3112) ---

Accredited [IEC17025]: NO Execution time [day]: 4
Unit price #1 [EUR]: 1,200.00 Limit #1: Applicable to any sample

# 5.4.3.1 - Zero uncertainty, relative expanded uncertainty and calibration curve for Type HM

(id 3113) ---

#### 5.4.3.2 - Calibration curve for Type SM

(id 3114) ---



# § T - LABORATORY TESTS / Prove di laboratorio

**PRICE LIST / Listino prezzi** 

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX

#### 5.4.5.1 - Temperature

(id 3115) ---

Accredited [IEC17025]: NO Execution time [day]: 2

Unit price #1 [EUR]: 2,200.00 Limit #1: Applicable to any sample

#### 5.4.5.2 - Pressure

(id 3116) ---

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 1,200.00 Limit #1: Applicable to any sample

#### 5.4.5.3 - Humidity of test gas

(id 3117) ---

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 2,200.00 Limit #1: Applicable to any sample

#### 5.4.6.1 - Alarm set-point(s)

(id 3118) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 1,200.00 Limit #1: fixed set point, no over range

Unit price #2 [EUR]: 1,600.00 Limit #2: adjustable set point, with over range

#### 5.4.6.2 - Time to alarm or alarm reading

(id 3119) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

#### 5.4.6.4 - Warm-up time

(id 3120) ---

Accredited [IEC17025]: NO Execution time [day]: 2

Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

#### 5.4.6.5 - Time of response

(id 3121) ---

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

#### 5.4.6.6 - Time of recovery

(id 3122) ---

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

### 5.4.6.9 - Gas concentrations above the upper limit of indication

(id 3123) ---

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 1,500.00 Limit #1: Applicable to any sample

#### 5.4.6.10.2 - Extended operation in test gas - Fixed and transportable equipment

(id 3124) ---

Accredited [IEC17025]: NO Execution time [day]: 7

Unit price #1 [EUR]: 1,200.00 Limit #1: Applicable to any sample

#### 5.4.6.11.2 - Orientation tests - Fixed and transportable equipment

(id 3125) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 500.00 Limit #1: Applicable to any sample

#### 5.4.7.1 - Battery capacity for battery-powered equipment

(id 3126) ---

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample



# **PRICE LIST / Listino prezzi**

# § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

TNDEX

#### 5.4.7.2 - Power supply variations (excludes battery-powered equipment)

(id 3127) ---

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

#### 5.4.7.3 - Electromagnetic compatibility

(id 3128) ---

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1: ask to office Limit #1: Not relevant

#### 5.4.8 - Stability

(id 3129) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 4.500.00 Limit #1: Up to 60 days

#### 5.4.9.2 - Low voltage and power interruption

(id 3130) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant

#### 5.4.9.3 - Field wiring faults

(id 3131) ---

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

#### 5.4.9.4 - Digital communication fault

(id 3132) ---

Accredited [IEC17025]: NO Execution time [day]: 2

Unit price #1 [EUR]: 300.00 Limit #1: Applicable to any sample

#### 5.4.9.5 - Measured values below zero

(id 3133) ---

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

#### 5.4.9.6 - Sensor disconnection

(id 3134) ---

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 300.00 Limit #1: Applicable to any sample

#### 5.4.10 - Software controlled equipment

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 3,500.00 Limit #1: Applicable to any sample

#### 5.4.11 - Protection against water - IPX3

(id 3136) ---

Execution time [day]: 1 Accredited [IEC17025]: NO

Unit price #1 [EUR]: 500.00 Limit #1: Very small sample (surface < 1m2)

## IEC TS 60904-1-2:2019

(id 365) Photovoltaic devices Part 1-2: Measurement of current-voltage characteristics of bifacial photovoltaic (PV) devices

### 6.2 - Determination of bifaciality coefficients

(id 1571) Determination of the main I-V characteristics pf the front and rear side at STC.

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

page 149/234



# **PRICE LIST / Listino prezzi**

# § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

TNDEX



#### 6.3.1 - Outdoor power generation gain measurement

(id 1572) Determination of the gain in power generation yielded by the bifaciality of the device under test, as a function of the irradiance on the rear side.

The power generation gain, BiFi, is the linear fit's slope of the Pmax versus GR data series. This linear least squares fit is forced to cross the Pmax axis at PmaxSTC.

PmaxBiFi10= PmaxSTC + BiFi x 100

PmaxBiFi20= PmaxSTC + BiFi x 200

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 650.00

Limit #1: Not relevant

## IEC TS 62804-1:2015

(id 229) Photovoltaic (PV) modules - Test methods for the detection of potential-induced degradation (PID) - Part 1: Crystalline silicon

#### 4.3.2 - Voltage stress damp heat test (stress method a)

(id 1097) Testing and evaluation of the durability of crystalline silicon photovoltaic (PV) modules to the effects of short-term high-voltage stress including potentialinduced degradation (PID).

Accredited [IEC17025]: NO Execution time [day]: 4 Unit price #1 [EUR]: 3,500.00 Limit #1: Applicable to any sample

#### 4.3.3 - Voltage stress test with a conductive electrode (stress method b)

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1: ask to office Limit #1: Not relevant

# IEC/IEEE 60079-30-1:2015

(id 734) Explosive atmospheres - Part 30-1: Electrical resistance trace heating - General and testing requirements

#### 5.1.2 - Dielectric test

(id 2611) ---

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 300.00 Limit #1: Applicable to any sample

#### 5.1.3 - Flectrical insulation resistance test

(id 2612) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 300.00 Limit #1: Applicable to any sample

#### 5.1.4 - Flammability test

(id 2613) A flammability test shall be performed on trace heaters and on trace heaters with integral components.

Accredited [IEC17025]: NO Execution time [day]: 1 Limit #1: Applicable to any sample Unit price #1 [EUR]: 600.00

#### 5.1.5.1 - Room temperature impact test

(id 2614) ---

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 200.00 Limit #1: Applicable to any sample

#### 5.1.5.2 - Minimum temperature impact test

(id 2615) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 300.00 Limit #1: Applicable to any sample

#### 5.1.6 - Deformation test

(id 2616) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 400.00 Limit #1: Not relevant



# § T - LABORATORY TESTS / Prove di laboratorio

PRICE LIST / Listino prezzi

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX

#### 5.1.7 - Cold bend test

(id 2617) ---

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 600.00 Limit #1: Applicable to any sample

#### 5.1.8 - Water resistance test

(id 2618) --

Accredited [IEC17025]: NO Execution time [day]: 14

Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

#### 5.1.9 - Integral components resistance to water test

(id 2619) ---

Accredited [IEC17025]: NO Execution time [day]: 2

Unit price #1 [EUR]: 500.00 Limit #1: Applicable to any sample

#### 5.1.11 - Thermal stability of electrical insulating material

(id 2620) ---

Accredited [IEC17025]: NO Execution time [day]: 28
Unit price #1 [EUR]: 1,000.00 Limit #1: Up to 100°C
Unit price #2 [EUR]: 1,200.00 Limit #2: Up to 250°C
Unit price #3 [EUR]: 1,600.00 Limit #3: Up to 450°C

#### 5.1.14 - Verification of start-up current

(id 2621) The start-up current of the trace heater shall be measured as a function of the start-up temperature, as designated by the manufacturer.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 600.00 Limit #1: Not relevant

#### 5.1.16 - Outdoor exposure test

(id 2622) Only trace heaters and integral components specified for outdoor exposure are subject to this test.

Accredited [IEC17025]: NO Execution time [day]: 42
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

## Internal methods

(id 18) Tests not related to standards, developed by laboratory or under client specification

#### MI 01 - Series resistance measurement - Dark I-V measurement

(id 207) The dark I-V measurement is performed to evaluate both the threshold voltage and the sample's series resistances (R's)

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 120.00 Limit #1: Applicable to any sample

#### MI 02 - Thermal imaging test

(id 265) Infrared (IR) thermography is an interesting technique appropriate to detect hot spots in soldering and cells, bypass diodes and all electrical connections in the PV module. It is a two dimensional non-destructive technique that utilizes the radiation in the infrared range.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 300.00 Limit #1: Applicable to any sample

#### MI 03 - Electroluminescence imaging test

(id 266) Electroluminescence imaging is a non-invasive test which has been established for fast spatially resolved characterization of silicon solar cells.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 300.00 Limit #1: Standard PV module
Unit price #2 [EUR]: 500.00 Limit #2: Big/special PV module

#### MI 04 - Simplified PID reliability test

(id 285) To evaluate the effects on the modules of the stress of system bias voltage. High voltage potential that exists between the active circuit and the grounded module frame or exterior can lead to module degradation – labelled potential-induced degradation

Accredited [IEC17025]: NO Execution time [day]: 30
Unit price #1 [EUR]: 2,000.00 Limit #1: Applicable to any sample



# your value certified § T - LABORATORY TESTS / Prove di laboratorio

**PRICE LIST / Listino prezzi** 

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

TNDEX



#### MI 05 - Determination of the concentration factor for CPV modules

(id 289) Methods described in CEI 82-25 Guideline.

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1: ask to office Limit #1: Applicable to any sample

#### MI 06 - Life test

(id 392) Check the capability of the samples to withstand thermal mismatch, fatigue and other stresses caused by repeated changes of temperature during pressure load changes.

Accredited [IEC17025]: NO Execution time [day]: -1 Unit price #1: ask to office Limit #1: Applicable to any sample

#### MI 07 - Simulated solar radiation on the ground te

(id 393) To determine the ability of the samples to withstand solar radiation and to identify those materials that are susceptible to solar and UV degradation and detect any visual defects.

Accredited [IEC17025]: NO Execution time [day]: 20 Unit price #1 [EUR]: 2,200.00 Limit #1: Applicable to any sample

#### MI 08 - Leakage test

(id 394) Check the tightness of the samples during pressure load

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 300.00 Limit #1: Laboratory temperature, Pmax=2bar Limit #2: Non standard temperature (controlled in climatic chamber: -40 +100°C range) Unit price #2 [EUR]: 800.00 Unit price #3 [EUR]: 1,200.00 Limit #3: Laboratory temperature, high pressure (>2 bar)

#### MI 09 - Gas emission test

(id 395) Determine the actual gas emissions to be expected from the battery system of thesamples and the required volume of the storing/charging sample room to prevent an accumulation of hydrogen, that could cause an explosive atmosphere in the room.

Accredited [IEC17025]: NO Execution time [day]: 7 Unit price #1 [EUR]: 5,500.00 Limit #1: Applicable to any sample

### MI 10 - Load loss of a solar collector

(id 586) ---

Accredited [IEC17025]: NO Execution time [day]: 30 Limit #1: Applicable to any sample Unit price #1: ask to office

#### MI 11 - Dust test at high temperature (70°C)

(id 598) To determine if the envelopes and enclosures of components and groups are able to prevent the penetration of harmful quantities of dust.

Accredited [IEC17025]: NO Execution time [day]: 5 Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

#### MI 12 - Hydrostatic overpressure test

(id 633) Verification of the ability to withstand the pressure performed with the hydrostatic method.

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 400.00 Limit #1: Test to be performed without protections up to 30 bar Unit price #2 [EUR]: 800.00 Limit #2: Test up to 30 bar with oil as medium Unit price #3 [EUR]: 1,200.00 Limit #3: Test to be performed in a hermetic vessel

## MI 13 - Strength test (low temperature)

(id 634) ---

Accredited [IEC17025]: NO Execution time [day]: 30 Unit price #1: ask to office Limit #1: Applicable to any sample

#### MI 14 - Cryogenic Test

(id 721) ---

Accredited [IEC17025]: NO Execution time [day]: 30 Unit price #1: ask to office Limit #1: Applicable to any sample

#### MI 15 - Ageing test for fuel tanks

(id 722) ---

Accredited [IEC17025]: NO Execution time [day]: 30 Unit price #1: ask to office Limit #1: Applicable to any sample



# **PRICE LIST / Listino prezzi**

# your value certified § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

TNDEX



#### MI 16 - Functionality test at Low or High Temperature for solenoid valve

(id 724) The test is intended to verify the ability of samples to work at -40°C or at +100°C ambient temperature with not failures

Accredited [IEC17025]: NO Execution time [day]: 30

Unit price #1 [EUR]: 500.00 Limit #1: Applicable to any sample

#### MI 17 - Short-circuiting test for batteries

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 300.00 Limit #1: Applicable to any sample

## MI 18 - Functioning ageing test

Accredited [IEC17025]: NO Execution time [day]: 40

Unit price #1 [EUR]: 4.000.00 Limit #1: Applicable to any sample

#### MI 19 - Bypass diodes functionality test

(id 844) Test to verify functionality of by-pass diodes, simply checking closed circuit under direct polarization and open circuit under reverse one.

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 200.00 Limit #1: Applicable to any sample

#### MI 20 - CONVERTED in Spec 9.02137/05 - Capitolato FCA 9\_02137-01

(id 847) ANNULLATO, codificato in Spec 9.02137/05 -- Ageing test TUBAZIONI PLASTICHE CON TERMINALI PER CONDUZIONE COMBUSTIBILE

Accredited [IEC17025]: NO Execution time [day]: -1

Unit price #1: ask to office Limit #1: Applicable to any sample

#### MI 21 - Gas diffusion test

(id 875) Capability of enclosures to avoid gas penetration into internal volume, in order to reduce the area classification for the internal equipments.

Accredited [IEC17025]: NO Execution time [day]: 1

Limit #1: Cubic volume 1x1x1 m Unit price #1 [EUR]: 2,500.00

#### MI 22 - CONVERTED IN AI 03

(id 885) See Automotive internal method AI 03

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Applicable to any sample

#### MI 24 - IP 6X - IP 5X on huge size enclosures

(id 903) OBSOLETE

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Applicable to any sample

#### MI 25 - Test in explosive gas mixtures

(id 913) Capability of the equipment to operate in explosive atmosphere without producing ignition. Similar to 60079-0 §26.3 but performed with different gas mixtures (depending on the auto-ignition temperature required) and with hot temperatures.

Accredited [IEC17025]: NO Execution time [day]: 4

Unit price #1 [EUR]: 2,200.00 Limit #1: Test inside small vessel

Unit price #2 [EUR]: 3,500.00 Limit #2: Test in explosion vented chamber

Unit price #3 [EUR]: 12,000.00 Limit #3: Huge size samples

#### MI 26 - Maximum power determination for electrical motors

(id 922) This test aims to measure the maximum power dissipated by an electrical motor.

Accredited [IEC17025]: NO Execution time [day]: 48

Unit price #1 [EUR]: 1,000.00 Limit #1: Applicable to any sample

#### MI 27 - Vibration test

(id 923) Vibration testing on insulated collector for diesel engine

Accredited [IEC17025]: NO Execution time [day]: 2 Subcontract: BPS srl

Unit price #1 [EUR]: 2,500.00 Limit #1: Applicable to any sample

#### MI 28 - Lux measurement

Accredited [IEC17025]: NO Execution time [dav]: -1

Unit price #1 [EUR]: 300.00 Limit #1: Applicable to any sample

page 153/234



# § T - LABORATORY TESTS / Prove di laboratorio

**PRICE LIST / Listino prezzi** 

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX

# .

#### MI 29 - Penetration test

(id 969) Test to verify the behavior of an object as a result of an impact with penetration

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 1,200.00 Limit #1: Explosion risk
Unit price #2 [EUR]: 800.00 Limit #2: Inert sample

#### MI 30 - CONVERTED IN AI 04, then FCA PF.90197 clause 6.3.5

(id 975) ANNULLATO --- See Automotive internal method AI 04, then FCA PF.90197 clause 6.3.5

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1: ask to office Limit #1: ---

#### MI 31 - Ignition tests of flammable substances injected in a hot air stream

(id 1121) This test is performed by spraying the flammable fluid indicated below in hot air flowing in a big pipe simulating a portion of a direct heat exchanger. The test is repeated changing the exhaust flow temperature and speed and the substance flow rate. The direction of the spraying of the flammable fluids is both parallel and orthogonal to the flow through the setup pipe.

The purpose is to verify the conditions of self-ignition for each test condition selected.

If the ignition does not occur, the flammability of the test condition is verified by a countercheck ignition produced by an external spark.

#### MI 32 - Single Measurement

(id 1124) Simple parameters measurement required for evaluation, any dimension

Accredited [IEC17025]: NO
Unit price #1 [EUR]: 150.00

Unit price #2 [EUR]: 300.00

Execution time [day]: 0

Limit #1: Any single measurement

Unit price #2 [EUR]: 300.00

Limit #2: Multiple measurement (up to 12 identical readings on the same sample)

#### MI 33 - Visual inspection

(id 1135) This inspection is performed with illumination conditions of more than 1000 lx

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 50.00 Limit #1: Not relevant
Unit price #2 [EUR]: 100.00 Limit #2: Special condition
Unit price #3 [EUR]: 200.00 Limit #3: Technical examination

### MI 34 - Gas flow catalytic ignition

(id 1201) Verification of the autoignition, due to catalytic effect, of an explosive atmosphere flow through or in contact to activated material.

Accredited [IEC17025]: NO Execution time [day]: 2 Unit price #1 [EUR]: 1,500.00 Limit #1: Filter or catalyzer up to 4"

#### MI 35 - Ignition test of flammable substances in a hot air stream with fixed O2 percentages

(id 1219) The aim of the test is to investigate if flammable fluids injected in a hot air stream with different conditions (Oxygen percentage and temperature) ignite in the most gravous environmental conditions (slow air stream).

If the ignition does not occur, the flammability of the test condition is verified by a countercheck ignition produced by an external spark.

#### MI 36 - Test on interlock Exd plug

(id 1255) This test is performed on a plug-type connector and its auxiliary connector in order to measure the interval of time taken by an operator to disconnect the main connector and the auxiliary connector.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 300.00 Limit #1: All samples

#### MI 37 - Gas sensors functionality test in gas mixture with variable pressures

(id 1260) ---

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,500.00 Limit #1: Not relevant



# PRICE LIST / Listino prezzi

# § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

TNDEX

#### MI 38 - Grounding resistance of operators

(id 1279) The purpose of this test is to verify the antistatic effectiveness of gloves or shoes worn by an operator.

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1 [EUR]: 300.00 Limit #1: Not relevant

#### MI 39 - Pressure resistance

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 300.00 Limit #1: Applicable to any sample

#### MI 23 - Back pressure across flame arrestors

Accredited [IEC17025]: NO Execution time [dav]: 2 Limit #1: Up to 5 m3/min Unit price #1 [EUR]: 600.00 Unit price #2 [EUR]: 1,200.00 Limit #2: From 6 to 25 m3/min Unit price #3 [EUR]: 2,500.00 Limit #3: From 26 to 50 m3/min

#### MI 40 - Functional test for Ex p equipment

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1 [EUR]: 300.00 Limit #1: Not relevant

#### MI 41 - Non-ignition of explosive atmosphere for verification of ATM inertising explosion protection systems

(id 1538) This test aims to verify the operation of the explosion protection system for ATM machines.

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 3,000.00 Limit #1: Not relevant

#### MI 42 - Functional verifications on ticketing system devices according to specification SES8384 Rev.0 §3.9

(id 1573) Check of the basic functionalitys of ticketing system devices.

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 300.00 Limit #1: Not relevant

#### MI 43 - Functionality test for the verification of ATM protection systems based on gas sensors

(id 1576) ---

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 500.00 Limit #1: Not relevant

## MI 44 - Determination of mininum aerosol concentration for C2H2 explosion suppression for ATM protection system

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 2,800.00 Limit #1: Not relevant

#### MI 45 - Gas diffusion test through perimeter walls

(id 1599) ---

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 5,200.00 Limit #1: Not relevant

#### MI 46 - General functional verifications according to customer's requests

(id 1610) ---

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1 [EUR]: 150.00 Limit #1: Not relevant

## MI 47 - Comparative performance measuring at STC - Continuous monitoring

Accredited [IEC17025]: NO Execution time [day]: 3 Unit price #1 [EUR]: 1,500.00 Limit #1: up to 2 modules up to 3 days measurements

### MI 49 - Functional validation of ingition system for Lithium Battery/battery module

Accredited [IEC170251: NO Execution time [dayl: 5 Unit price #1 [EUR]: 4,500.00 Limit #1: Not relevant



# PRICE LIST / Listino prezzi

# § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

TNDEX

#### MI 48 - Construction and validation of dummy battery

(id 1892) ---

Accredited	[IEC17025]: NO	Execution time [day]: 10
Unit price	#1 [EUR]: 1,000.00	Limit #1: <20 dm3
Unit price	#2 [EUR]: 2,000.00	Limit #2: 20-100 dm3
Unit price	#3 [EUR]: 5,000.00	Limit #3: >100 dm3

#### MI 50 - Purging test on highway tunnel (setup preparation and execution of the test)

(id 1907) ---

Accredited [IEC17025]: NO Execution time [day]: 2 Unit price #1 [EUR]: 5,000.00 Limit #1: Not relevant

#### MI 51 - Temperature test on exhaust gases

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1 [EUR]: 300.00 Limit #1: Not relevant

## MI 52 - Test for evaluating exposure effects to gaseous environment of HFC227ea with final IP66 test verification

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 2,800.00 Limit #1: Not relevant

#### MI 53 - Evaluation of characteristics of the explosion of a Lithium battery/battery module

(id 1970) Visual and/or instrumental evaluation of inducted thermal runaway

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 2,000.00 Limit #1: Applicable to any sample

#### MI 54 - Effectiveness of extinguishing agent in controlling Lithium-Ion battery thermal runaway

(id 2006) ---

Accredited [IEC17025]: NO Execution time [day]: 2 Unit price #1 [EUR]: 2,000.00 Limit #1: Small sample Unit price #2 [EUR]: 3,800.00 Limit #2: Medium sample with special condition Unit price #3 [EUR]: 8,000.00 Limit #3: Huge size samples with special condition

#### MI 55 - Functional evaluation of lithium battery estinguishing system

Accredited [IEC17025]: NO Execution time [dav]: 2 Unit price #1 [EUR]: 1,000.00 Limit #1: Applicable to any sample

#### MI 56 - Grayness on pv module

(id 2011) ---

Accredited [IEC17025]: NO Execution time [day]: 2 Unit price #1 [EUR]: 3,200.00 Limit #1: Applicable to any sample

#### MI 58 - Evaluation of thermal and pressure characteristics of the explosion of a cell of a Lithium battery (id 2024) ---

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1: ask to office Limit #1: Not relevant

#### MI 57 - Efficiency of ignition system for Lithium batteries thermal runaway

(id 2025) Attempt to start the thermal runaway of Lithium battery with a specified method

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1: ask to office Limit #1: Short-circuiting Unit price #2 [EUR]: 800.00 Limit #2: Overcharging/Overheating Unit price #3 [EUR]: 2,000.00 Limit #3: Perforation (nail test)

#### MI 59 - Outdoor reproduction of an explosion phenomenon of a cell of a Lithium battery

(id 2026) ---

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1: ask to office Limit #1: Not relevant



# PRICE LIST / Listino prezzi

# § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX

# MI 60 - Video inspection

(id 2027) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 700.00 Limit #1: Applicable to any sample

## MI 61 - Ignition of thermal runaway effect on consecutive cells of a Lithium battery module

(id 2034) ---

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 1,500.00 Limit #1: Applicable to any sample

## MI 62 - Test for non-transmission with detecting flames leakage

(id 2036) ---

Accredited [IEC17025]: NO Execution time [day]: 2

#### MI 63 - Immersion test

(id 2048) Immersion of the sample under test in liquid for a duration and at a depth to be agreed with the customer.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 300.00 Limit #1: Small sample not powered
Unit price #2 [EUR]: 600.00 Limit #2: Small sample powered on

#### MI 64 - Measurement of HF emissions from Lithium batteries

(id 2049) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 2,500.00 Limit #1: Applicable to any sample

#### MI 65 - Non ignition of dust layer on hot surface

(id 2060) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 600.00 Limit #1: one single layer/depth

#### MI 66 - Surface temperature measurement on insulated exhaust pipe model

(id 2125) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 1,200.00 Limit #1: Applicable to any sample

### MI 67 - Measurement of emissions of gaseous by-products

(id 2131) Simulation of an agent discharge event, wanted or not, in a confined environment.

Determination of the emissions it produces, in gaseous or aerosol form, in particular through the search for pollutants.

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 2,000.00 Limit #1: Small setup

#### MI 68 - Validation of safety container for battery transport

(id 2149) The enclosure under test is filled with the specified number of cells and/or modules (typically batteries).

They are ignited to start the thermal runaway, with the production of hot gases.

The scope is the evaluation of the capability to survive to this phenomenon.

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 13,000.00 Limit #1: Up to 40 cells or modules and 50kWh
Unit price #2 [EUR]: 20,000.00 Limit #2: Over 40 cells or modules and 50kWh
Unit price #3 [EUR]: 1,000.00 Limit #3: Up to 2 cells or modules

#### MI 70 - Thermal cycles

(id 2156) Thermal cycles of temperature and/or humidity.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 1,400.00 Limit #1: Not relevant

#### MI 69 - Ageing with given thermal and pressure conditions

(id 2157) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 1,600.00 Limit #1: Applicable to any sample



# § T - LABORATORY TESTS / Prove di laboratorio

PRICE LIST / Listino prezzi

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



#### MI 71 - Obtaining and verifying the maximum battery charge (SOC)

(id 2158) Measurement of the cell/module voltage, compared to the voltage/state-of-charge characteristic. The voltage of each cell is obtained by dividing the whole module voltage by the number of cells.

The cell voltage is compared to the SOC table (annexed to the report).

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 1,200.00 Limit #1: Applicable to any sample

#### MI 72 - Photogrammetry test

(id 2428) Photogrammetry compares the state of the sample before and after a test. It obtains reliable information about physical objects and the environment through the process of recording, measuring and interpreting photographic images and patterns of electromagnetic radiant imagery and other phenomena.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 800.00 Limit #1: Not relevant

#### MI 73 - Verification of sleeves rigidity

(id 2477) This test is intended to verify the rigidity of sleeves depending on temperature and internal pressure.

Accredited [IEC17025]: NO Execution time [day]: 2

Unit price #1 [EUR]: 1,200.00 Limit #1: Applicable to any sample for one temperature

#### MI 74 - Measurement of dangerous gas in a confined environment

(id 2478) This test is intended to measure the maximum level of gas concentration in a confined environment.

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

#### MI 75 - Determination of grading curve

(id 2483) This test is inteded to determine the grading curve of a specific granural material.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 350.00 Limit #1: Applicable to any sample

# MI 76 - Determination of the burning time of substances on hot surfaces

(id 2487) 10 ml of flammable substance are poured onto a known surface of 70cm2 heated to the test temperature. Burning time is measured from the moment the flames develop to the moment they are extinguished.

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 500.00 Limit #1: 1 substance

#### MI 77 - Gas detector calibration (ppm)

(id 2499) Calibration of a gas detector in low concentration (ppm).

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 500.00 Limit #1: Applicable to any sample

#### MI 78 - Gas detector calibration (%vol)

(id 2500) Calibration of a gas detector in high concentration (%vol).

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### MI 79 - Determination of free volume of a complex geometry in an idraulic circuit

(id 2515) This test is intended to determine the free volume of a complex geometry in an idraulic circuit.

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

#### MI 80 - Pyrolysis Test

(id 2591) This test is intended to investigate the syngas emission in case of pyrolysis on plastic material waste.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 600.00 Limit #1: 4 hours sequence

#### MI 81 - Effect of internal explosion

(id 2610) --

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 3,500.00 Limit #1: Test in explosion vented chamber
Unit price #2 [EUR]: 4,200.00 Limit #2: Flame trasmission and pressure measurement



# **PRICE LIST / Listino prezzi**

# § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

TNDEX

#### MI 82 - Leak Test

(id 2668) ---

Accredited [IEC17025]: NO Execution time [day]: 1Unit price #1 [EUR]: 200.00 Limit #1: Test at ambient temperature Unit price #2 [EUR]: 500.00 Limit #2: Test at defined temperature Unit price #3 [EUR]: 800.00 Limit #3: Test with possible danger

#### MI 83 - Mechanical Stress Test at Ambient Temperature

(id 2686) ---

Accredited [IEC17025]: NO Execution time [day]: 2 Unit price #1 [EUR]: 200.00 Limit #1: Up to 24h Unit price #2 [EUR]: 800.00 Limit #2: Up to 120h

#### MI 84 - Dedicated techical evaluation

(id 2750) Dedicated techical evaluation after test with a special setup

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1 [EUR]: 400.00 Limit #1: Applicable to any sample

#### MI 85 - Thermal Cycle NAMI

(id 2752) Thermal cycle developed by Cavalli Srl

Accredited [IEC17025]: NO Execution time [day]: 5 Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### MI 86 - Bubble leak test

(id 2758) ---

Execution time [day]: 0 Accredited [IEC17025]: NO Unit price #1 [EUR]: 200.00 Limit #1: Test at ambient temperature Unit price #2 [EUR]: 500.00 Limit #2: Test at definded temperature

#### MI 87 - Determinazione della temperatura interna di celle al litio

(id 2762) Questo test ha lo scopo di verificare la temperatura interna della cella, per confrontarla con la temperatura in grado di innescare il thermal runaway

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1 [EUR]: 900.00 Limit #1: Applicable to any sample

#### MI 88 - Filter efficiency evaluation

Accredited [IEC17025]: NO Execution time [dav]: 2 Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### MI 89 - Sniffer leak test

(id 2778) This test is aimed at detecting leaks in the test specimen. Generally, it is applied prior to the SHED test and immediately after SHED testing. The method to check for leaks is either by helium leak or by hydrocarbon leak check.

Accredited [IEC17025]: NO Execution time [day]: 2 Unit price #1 [EUR]: 400.00 Limit #1: Applicable to any sample

#### MI 90 - Behaviour of a box in case of runaway of contained batteries

(id 2804) This test is intended to evaluate the behaviour of a box designed for R4 waste in case of internal runaway.

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1 [EUR]: 4,800.00 Limit #1: Applicable to any sample

#### MI 91 - Functional parameter of check valve

(id 2855) ---

Accredited [IEC17025]: NO Execution time [day]: 2 Unit price #1 [EUR]: 1,000.00 Limit #1: Applicable to any sample

### MI 92 - Push on force with pressure

(id 2869) This test is intended to measure the force necessary to connect two couplings with a specific internal pressure.

Accredited [TEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 300.00 Limit #1: Applicable to any sample



# PRICE LIST / Listino prezzi § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX

#### MI 93 - Efficiency of fire suppression on Lithium batteries

(id 2877) ---

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 2,200.00 Limit #1: Small batteries
Unit price #2 [EUR]: 5,000.00 Limit #2: Medium batteries

#### MI 94 - Bloating test at low pressure

(id 2982) ---

#### MI 95 - Functional test at different temperatures

(id 3014) ---

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### MI 96 - Stable combustion test for gas mixture

(id 3025) The purpose of the test is to evaluate the ability of gas mixtures to provide stable combustion.

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

# ISO 10289:1999

(id 1870) Methods for corrosion testing of metallic and other inorganic coatings on metallic substrates — Rating of test specimens and manufactured articles subjected to corrosion tests

#### Unref. - Rating of test specimens and manufactured articles subjected to corrosion tests

(id 2484) This International Standard gives a method of evaluating the condition of decorative and protective metallic and inorganic coated panels or articles which have been exposed to corrosive environments for test or for other purposes. It is applicable to test panels or components exposed to natural atmospheres, in mobile or static conditions, or subjected to accelerated tests.

The result of inspecting a surface, called the performance rating, is recorded as two separate ratings, the protection rating and the appearance rating.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 80.00 Limit #1: Applicable to any sample

# ISO 11997-1:2017

(id 2134) Paints and varnishes - Determination of resistance to cyclic corrosion conditions - Part 1: Wet (salt fog)/dry/humid

#### **Unref. - Cyclical corrosion test**

(id 2959) ---

Accredited [IEC17025]: NO Execution time [day]: 42
Unit price #1 [EUR]: 6,500.00 Limit #1: Applicable to any sample

## ISO 15848-1:2015+A1:2017

(id 1161) Industrial valves - Measurement, test and qualification procedures for fugitive emissions - Part 1: Classification system and qualification procedures for type testing of valves

#### 5.2.4.2 - Mechanical cycles of isolating valves (5 cycles)

(id 1585) Final 5 cycles at RT.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 50.00 Limit #1: Not relevant

#### 5.2.4.10 - Post-test examination

(id 1596) After all the tests have been successfully completed, the test valve shall be disassembled and all sealing components shall be visually examined to record notable wear and any other significant observations for information.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 50.00 Limit #1: Not relevant



# PRICE LIST / Listino prezzi § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



#### A.1.6.2 - Preparation of the valve: cleaning, initial checks and hydrostatic test

(id 1600) Before each test, the valve is cleaned and dried, and the packing tightening checked.

The hydrostatic test shall be performed before testing the valve in high pressure and high temperature conditions.

After the hydrostatic test, the packing shall be dry before any sealing test (when using packing in the stem sealing system). It is recommended that the packing is replaced.

If the tight chamber encloses the entire valve, connection flanges shall be welded to avoid any leaks that come from them. In this case, the measurements correspond to the leaks from stem sealing system and body seals.

#### 5.2.4.4 - Preliminary leak tests at the room temperature for type testing of valves (test 1)

(id 1601) The tests are carried out as follows:

- a) Pressurization of the test valve with the test fluid to the test pressure as specified in a relevant standard.
- b) After the test pressure has been stabilized, measurement of the leakages both from the stem (or shaft) seal and from the body seals, in accordance with Annexes A and B, respectively.

#### 5.2.4.5 - Mechanical cycle test at the room temperature (test 2)

(id 1602) The tests are carried out as follows:

- a) Mechanical cycles at room temperature while the test valve is kept pressurized.
- b) Measurement of the leakage from the stem (or shaft) seal only, in accordance with Annex A.

#### 5.2.4.6 - Static test at the selected test temperature (test 3)

(id 1603) The tests are carried out as follows.

- a) Pressurization of the test valve with the test fluid to the test pressure specified for the selected test temperature.
- b) After the test pressure has been stabilized, adjustment of the valve temperature to the selected test temperature, ensuring that the test pressure does not exceed the level specified.
- c) After the valve temperature has been stabilized with an allowance of ±5 % with a maximum of 15 °C, measurement of the leakage from the stem (or shaft) seal only in accordance with Annex A.

```
Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 500.00 Limit #1: up to 2"
Unit price #2 [EUR]: 800.00 Limit #2: from 2.5" to 10"
Unit price #3 [EUR]: 1,200.00 Limit #3: over 10"
```

#### 5.2.4.7 - Mechanical cycle test at the selected test temperature (test 4)

(id 1604) The tests are carried out as follows.

- a) Mechanical cycles at the selected test temperature while the test valve is kept pressurized.
- b) Measurement of the leakage from the stem (or shaft) seal only in accordance with Annex A.

#### 5.2.4.8 - Intermediate static test at the room temperature (test 5)

(id 1605) The tests are carried out as follows.

- a) Test valve returning to the room temperature, without artificial cooling (or heating).
- b) After the valve temperature has been stabilized, measurement of the leakage from the stem (or shaft) seal only in accordance with Annex A.

```
Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 200.00 Limit #1: up to 2"
Unit price #2 [EUR]: 400.00 Limit #2: from 2.5" to 10"
Unit price #3 [EUR]: 600.00 Limit #3: over 10"
```

#### 5.2.4.9 - Final test at the room temperature (test 6)

(id 1606) The tests are carried out as follows:

- a) Test valve returning to the room temperature, without artificial measures.
- b) After the valve temperature has been stabilized, measurement of the leakage from the stem (or shaft) seal in accordance with Annex A and from body seals in accordance with Annex B.

	page 161/234
	nogo 161/224
Unit price #3 [EUR]. 400.00	Limit #3. Over 10"
Unit price #3 [EUR]: 400.00	Limit #3: over 10"
Unit price #2 [EUR]: 200.00	Limit #2: from 2.5" to 10"
Unit price #1 [EUR]: 100.00	Limit #1: up to 2"
Accredited [IEC17025]: NO	Execution time [day]: 0



# PRICE LIST / Listino prezzi § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

TNDEX



# ISO 16495:2013

(id 35) Packaging - Transport packaging for dangerous goods - Test methods

#### Annex F - Drop test

(id 676) ---

Accredited [IEC17025]: NO

Execution time [day]: 30 Unit price #1: ask to office

Limit #1: Applicable to any sample

#### Annex G - Leakproof test

(id 677) ---

Accredited [IEC17025]: NO Unit price #1: ask to office Execution time [day]: 30

Limit #1: Applicable to any sample

#### Annex H - Hydraulic pressure test

(id 678) ---

Accredited [IEC17025]: NO Unit price #1: ask to office Execution time [day]: 30

Limit #1: Applicable to any sample

#### Annex I - Stacking test

(id 679) ---

Accredited [IEC17025]: NO Unit price #1: ask to office Execution time [day]: 30

Limit #1: Applicable to any sample

#### Annex J - Water spray test

(id 680) ---

Accredited [IEC17025]: NO Unit price #1: ask to office Execution time [day]: 30

Limit #1: Applicable to any sample

#### Annex K - Bottom lift test

(id 681) ---

Accredited [IEC17025]: NO Unit price #1: ask to office Execution time [day]: 30 Limit #1: Applicable to any sample

### Annex L - Top lift test

(id 682) ---

Accredited [IEC17025]: NO Unit price #1: ask to office Execution time [day]: 30 Limit #1: Applicable to any sample

### Annex M - Tear test

(id 683) ---

Accredited [IEC17025]: NO Unit price #1: ask to office Execution time [day]: 30

Limit #1: Applicable to any sample

#### Annex N - Topple test

(id 684) ---

Accredited [IEC17025]: NO Unit price #1: ask to office Execution time [day]: 30

Limit #1: Applicable to any sample

#### Annex O - Righting test

(id 685) ---

Accredited [IEC17025]: NO Execution time [day]: 30

Unit price #1: ask to office Limit #1: Applicable to any sample

#### Annex P - Puncture test

(id 686) ---

Accredited [IEC17025]: NO Execution time [day]: 30

Unit price #1: ask to office Limit #1: Applicable to any sample



# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX

#### Annex Q - Vibration test

(id 687) ---

Accredited [IEC17025]: NO Execution time [day]: 30
Unit price #1: ask to office Limit #1: Applicable to any sample

# ISO 16750-4:2023

(id 2083) Road vehicles — Environmental conditions and testing for electrical and electronic equipment — Part 4: Climatic loads

#### 5.5.3 - Leakage and function test

(id 2934) This test checks the resistance of a system/component to salt mist and salt water on winter streets. The failure mode is an electrical malfunction due to leakage currents caused by the ingress of salt water.

```
Accredited [IEC17025]: NO Execution time [day]: 6
Unit price #1 [EUR]: 1,500.00 Limit #1: Applicable to any sample
```

#### 5.4.2 - Splash water test

(id 2936) The purpose of the test is to simulate the thermal shock induced by cold water splashing over a hot system/component. This happens when driving on wet roads in winter and applies to products in the splash areas of the vehicle. The failure modes are a mechanical cracking of materials or seal failures caused by different temperature expansion coefficients. Additional failure modes not addressed in 5.3.2.1 are loss of tightness and/or intrusion of water into the system/component.

```
Accredited [IEC17025]: NO Execution time [day]: 3
Unit price #1 [EUR]: 1,700.00 Limit #1: Applicable to small sample (one sample at a time)
Unit price #2 [EUR]: 2,500.00 Limit #2: Applicable to medium sample (or more sample contemporary)
```

## ISO 16750-5:2010

(id 1298) Road vehicles - Environmental conditions and testing for electrical and electronic equipment - Part 5: Chemical loads

#### 4 - Chemical load

(id 2033) Components and associated parts that can come into contact with the specified chemical agents shall be resistant to those agents. The components and associated parts shall be tested with all agents they are likely to come into contact with, except for those materials which can be shown by documentary evidence to be immune to the contaminant and which need not be tested. A material is considered to be immune to a contaminant if there is no change in properties sufficient to affect material performances over the time and at the temperature specified. The purpose of the test is to determine whether the device under test (DUT) is unacceptably affected by temporary exposure to contaminating agents.

```
Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample
```

## ISO 20567-1 2017

(id 1335) Paints and varnishes - Determination of stone-chip resistance of coatings - Part 1: Multi-impact testing

#### 8.3 - Projection of grit (stone impact)

(id 2207) ---

```
Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 800.00 Limit #1: Not relevant
```

## ISO 20653:2013

(id 1550) Road vehicles - Degrees of protection (IP-Code) - Protection of electrical equipment against foreign objects, water and access

#### 5 - Dust test for first characteristic numerals IP5K-6K

(id 2012) Test for determining degrees of protection against foreign objects 5K and 6K: dust-protected or dust-tight.

```
Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 600.00 Limit #1: Standard IP chamber (small sample not powered)
Unit price #2 [EUR]: 1,200.00 Limit #2: Medium IP chamber (small sample not powered)
Unit price #3 [EUR]: 2,200.00 Limit #3: Powered on sample
```



# **PRICE LIST / Listino prezzi**

# your value certified § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

TNDEX

#### 6 - Degrees of protection against water IPX6

(id 2013) Strong high-velocity water

Accredited [IEC17	025]: NO	Execution	time [	[day]	: 2				
Unit price #1 [EU	R]: 300.00		Limit	#1:	sample	${\tt movable}$	by	one	operator
Unit price #2 [EU]	R]: 600.00		Limit	#2:	sample	movable	by	two	operator
Unit price #3 [EU]	R]: 1,000.00		Limit	#3:	huge s	ize sampi	le		

#### 6 - Degrees of protection against water IPX9K

(id 2014) Water during high-pressure/steam-jet cleaning

```
Accredited [IEC17025]: NO
                                Execution time [day]: 1
Unit price #1 [EUR]: 800.00
                                           Limit #1: largest dimension < 250 mm
Unit price #2 [EUR]: 1,200.00
                                           Limit #2: largest dimension >= 250 mm
Unit price #3 [EUR]: 1,800.00
                                           Limit #3: huge size sample
```

#### 6 - Degrees of protection against water IPX6K

(id 2045) Strong high-velocity water with increased pressure

```
Accredited [IEC17025]: NO
                                 Execution time [day]: 0
Unit price #1 [EUR]: 300.00
                                           Limit #1: sample movable by one operator
Unit price #2 [EUR]: 600.00
                                           Limit #2: sample movable by two operator
Unit price #3 [EUR]: 1,000.00
                                           Limit #3: huge size samples
```

#### 6 - Degrees of protection against water IPX8

(id 2046) Continuous immersion in water

```
Accredited [IEC17025]: NO
                                 Execution time [day]: 0
Unit price #1 [EUR]: 600.00
                                           Limit #1: Dia 250mm - height 600mm - up to 1 bar (10m)
                                           Limit #2: Dia 800mm - height 400mm - up to 1 bar (10m)
Unit price #2 [EUR]: 1,000.00
Unit price #3 [EUR]: 1,800.00
                                           Limit #3: Dia 2000mm - height 1500mm - up to 60 bar (600m)
```

#### 6 - Degrees of protection against water IPX7

(id 2047) Temporary immersion in water

```
Accredited [IEC17025]: NO
                                 Execution time [day]: 0
Unit price #1 [EUR]: 300.00
                                           Limit #1: sample movable by one operator
Unit price #2 [EUR]: 600.00
                                           Limit #2: sample movable by two operators
Unit price #3 [EUR]: 1,000.00
                                           Limit #3: huge size samples
```

#### 6 - Degrees of protection against water IPX4K

(id 2195) This test aims to verify the protection of electrical equipment within the enclosure against harmful effects resulting from water splashing against the enclosure with increased pressure from any direction.

```
Accredited [IEC17025]: NO
                                  Execution time [dav]: 0
Unit price #1: ask to office
                                            Limit #1: Not relevant
```

## 8.3.3.2 - Dust Intrusion IP6K- Test method 1

(id 2519) This International Standard applies to degrees of protection (IP code) provided by enclosures of the electrical equipment of road vehicles.

```
Accredited [IEC17025]: NO
                                Execution time [day]: 1
Unit price #1 [EUR]: 600.00
                                          Limit #1: Standard IP chamber (small sample not powered)
Unit price #2 [EUR]: 1,200.00
                                           Limit #2: Medium IP chamber (small sample not powered)
Unit price #3 [EUR]: 2,200.00
                                           Limit #3: Powered on sample
```

# ISO 20653:2023

(id 1965) Road vehicles — Degrees of protection (IP code) — Protection of electrical equipment against foreign objects, water and access

#### 5 - Dust test for first characteristic numerals IP5K-6K

(id 2889) Test for determining degrees of protection against foreign objects 5K and 6K: dust-protected or dust-tight.

```
Accredited [IEC17025]: YES
                                 Execution time [day]: 2
                                            Limit #1: Standard IP chamber (small sample not powered)
Unit price #1 [EUR]: 600.00
Unit price #2 [EUR]: 1,200.00
                                            Limit #2: Medium IP chamber (sample not powered)
Unit price #3 [EUR]: 2,200.00
                                            Limit #3: Powered on sample
```

#### 6 - Degrees of protection against water - IPX5

(id 2960) This test is performed to evaluate the protection of the equipment against the access of water in the presence of a water jet hose nozzle of 6.3 mm diameter, and a 12.5 I/min flow rate, spraying water from a 3m distance.

```
Accredited [IEC17025]: YES
                                 Execution time [day]: 1
Unit price #1 [EUR]: 400.00
                                            Limit #1: Very small sample (surface < 1m2 - sample not powered)
Unit price #2 [EUR]: 800.00
                                            Limit #2: Medium sample (surface up to 3m2 - sample not powered)
Unit price #3 [EUR]: 2,000.00
                                            Limit #3: Powered on sample
```

page 164/234



# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX

# - 🔳

#### 6 - Degrees of protection against water IPX6K

(id 2963) Strong high-velocity water with increased pressure.

Accredited [IEC17025]: YES	Execution time [day]: 1
Unit price #1 [EUR]: 400.00	Limit #1: Standard IP chamber (very small sample)
Unit price #2 [EUR]: 800.00	Limit #2: Medium IP chamber (sample up to 90cm height)
Unit price #3 [EUR]: 1,200.00	Limit #3: Huge size samples with movable chamber

#### 6 - Degrees of protection against water IPX9K

(id 2994) Water during high-pressure/steam-jet cleaning

## ISO 2206:1987

(id 38) Packaging - Complete, filled transport packages - Identification of parts when testing

#### - Identification of parts of package

(id 690) ---

```
Accredited [IEC17025]: NO Execution time [day]: 30
Unit price #1: ask to office Limit #1: Applicable to any sample
```

## ISO 2248:1985

(id 37) Packaging - Complete, filled transport packages - Vertical impact test by dropping

#### - Vertical impact test by dropping

(id 689) ---

```
Accredited [IEC17025]: NO Execution time [day]: 30
Unit price #1: ask to office Limit #1: Applicable to any sample
```

# ISO 22609:2004

(id 1520) Clothing for protection against infectious agents - Medical face masks - Test method for resistance against penetration by synthetic blood (fixed volume, horizontally projected)

#### unref - Resistance against penetration by synthetic blood (fixed volume, horizontally projected)

(id 1987) The test method is intended to evaluate the protection of the health care provider's face from exposure to blood and body fluids. It is used to evaluate the resistance of medical face masks to penetration by synthetic blood under high-velocity liquid contact with the medical face mask surface of a fixed volume over a relatively short period of time (0 s to 2.5 s).

Synthetic blood is prepared according instruction of Annex B (ISO 22609:2004) mixing high-performance (HPLC) distilled water (pH 7,0  $\pm$  0,5) with a thickening agent (Acrysol G111/Rhoplex) and a dye containing colorant and surfactant, for reasons of availability in Europe the red dye has been replaced with an orange dye of the same chimical family (Orange G). The formulation is such as to guarantee density and surface tension within the limits required by the standard.

```
Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,600.00 Limit #1: Qualification in package
Unit price #2 [EUR]: 800.00 Limit #2: Single measurement with preconditioning
```

# ISO 23269-2:2011

(id 2136) Ships and marine technology — Breathing apparatus for ships - Part 2: Self-contained breathing apparatus for shipboard firefighters

#### 5.2 - High temperature, high humidity test

(id 2961) ---

```
Accredited [IEC17025]: NO Execution time [day]: 4
Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample
```



# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX

#### 5.3 - Temperature cycling test

(id 2962) ---

Accredited [IEC17025]: NO
Unit price #1 [EUR]: 2,000.00

Execution time [day]: 7

Limit #1: Applicable to any sample

# ISO 2409:2020

(id 1929) Paints and varnishes - Cross-cut test

#### 8 - Cross cutting test

(id 2597) Six parallel cuts are introduced in the coating and another six cuts are introduced perpendicular to the first cuts. Any loose paint particles are removed. The cut area is examined visually and compared to a six-step classification.

Accredited [IEC17025]: NO Unit price #1 [EUR]: 300.00

Execution time [day]: 2

Limit #1: Not relevant

## ISO 2875:2000

(id 36) Packaging - Complete, filled transport packages and unit loads - Water-spray test

#### - Water-spray test

(id 688) ---

Accredited [IEC17025]: NO Unit price #1: ask to office

Execution time [day]: 30

Limit #1: Applicable to any sample

# ISO 4628-2:2016

(id 1931) Paints and varnishes - Evaluation of degradation of coatings - Designation of quantity and size of defects, and of intensity of uniform changes in appearance - Part 2: Assessment of degree of blistering

#### 4 - Assessment of degree of blistering

(id 2598) This part of ISO 4628 specifies a method for assessing the degree of blistering of coatings by comparison with pictorial standards. The pictorial standards provided in this part of ISO 4628 illustrate blisters in the sizes 2, 3, 4, and 5, and so the pictorial standards in the graphics of the pictorial standards.

and each size in the quantities (densities) 2, 3, 4 and 5.

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 100.00 Limit #1: Applicable to any sample

## ISO 6270-1:2017

(id 2163) Paints and varnishes - Determination of resistance to humidity - Part 1: Condensation (single-sided exposure)

#### Unref. - Determination of resistance to humidity - Condensation

(id 2990) A coated test specimen is exposed to continuous condensation, and the effects of the exposure are evaluated by criteria agreed in advance between the interested parties, these criteria usually being of a subjective nature.

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

## ISO 6270-2:2017

(id 1333) Paints and varnishes - Determination of resistance to humidity Condensation (in-cabinet exposure with heated water reservoir)

#### 9.5.3 - Determination of resistance to humidity - Condensation - AHT

(id 1853) Condensation atmosphere with alternating humidity and air temperature

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1: ask to office Limit #1: Not relevant



# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



#### 9.5.4 - Determination of resistance to humidity - Condensation - AT

(id 2890) Condensation atmosphere with alternating air temperature

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 9.5.2 - Determination of resistance to humidity - Condensation - CH

(id 2891) Condensation atmosphere with constant humidity

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

## ISO 6988:1985

(id 1841) Metallic and other non-organic coatings - Sulfur dioxide test with general condensation of moisture

#### 6 - Sulfur dioxide test with general condensation of moisture

(id 2466) Moist air containing sulfur dioxide quickly produces easily visible corrosion of many metals in a form resembling that occurring in industrial environments. It is therefore a test medium well suited to detect pores or other sources of weakness in protective coatings and deficiencies in corrosion resistance associated with unsuitable alloy composition or treatments.

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

## ISO 7628:2010

(id 800) Road vehicles -Thermoplastics tubing for air braking systems

#### 9.3 - Cold impact test

(id 1339) ---

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 300.00 Limit #1: Standard temperature
Unit price #2 [EUR]: 600.00 Limit #2: Low temperature (down to -65°C)

## ISO 7892:1988

(id 2159) Vertical building elements — Impact resistance tests — Impact bodies and general test procedures

#### 4.3 - Urto da corpo duro

(id 2981) ---

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

## ISO 80079-36:2016+COR1:2019

(id 1609) Explosive atmospheres - Part 36: Non-electrical equipment for explosive atmospheres - Basic method and requirements

### 8.2.1 - Determination of the maximum surface temperature - General

(id 1069) ---

Accredited [IEC17025]: YES

Unit price #1 [EUR]: 800.00

Unit price #2 [EUR]: 1,500.00

Unit price #3 [EUR]: 2,500.00

Execution time [day]: 0

Limit #1: small sample movable by one operator

Limit #2: medium sample movable by two operators

Limit #3: assembly or special sample

#### 8.2.2 - Hot Surface Ignition Test

(id 1070) ---



# **PRICE LIST / Listino prezzi**

# § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

TNDEX



#### 8.3.1 - Test for resistance to impact

(id 1071) Impact tests of IEC 60079-0 apply.

The test is made at two separate places on each sample. For equipment with light-transmitting parts (LTP) made of glass, only one of those two impacts is on the alass.

Group II or Group III, portable or transportable luminaires or handlights, are only tested for risk of mechanical danger "high".

Where a part of the equipment (such as a non-metallic overlay of a touch pad) serves multiple functions, such as a light-transmitting area and also as part of the enclosure, the function covering the largest area is used to determine which risk of mechanical danger is applied.

Light-transmitting parts having individual openings from 625 mm2 to 2 500 mm2 are tested without guard.

```
Accredited [IEC17025]: YES
                                  Execution time [day]: 1
Unit price #1 [EUR]: 200.00
                                            Limit #1: Not relevant
```

#### 8.3.2 - Drop test

(id 1072) The drop test of IEC 60079-0 applies. The portable or personal equipment is dropped 4 times onto a horizontal concrete surface.

```
Accredited [IEC17025]: YES
                                 Execution time [day]: 1
Unit price #1 [EUR]: 100.00
                                           Limit #1: Not relevant
```

#### 8.4.4 - Thermal endurance to heat

(id 1073) Evaluation of the thermal endurance to heat of equipment intended for use in explosive gas atmosphere

```
Accredited [IEC17025]: YES
                                  Execution time [day]: 35
Unit price #1 [EUR]: 1,000.00
                                            Limit #1: Service temperature < 75°C
Unit price #2 [EUR]: 1,200.00
                                            Limit #2: 75°C < Service temperature < 100°C
                                            Limit #3: 100°C < Service temperature < 400°C
Unit price #3 [EUR]: 2,400.00
```

#### 8.4.5 - Thermal endurance to cold

(id 1074) Evaluation of the thermal endurance to cold of equipment intended for use in explosive gas atmosphere

```
Accredited [IEC17025]: YES
                                  Execution time [day]: 1
Unit price #1 [EUR]: 400.00
                                            Limit #1: Minimum service temperature > -40°C
                                            Limit #2: Minimum service temperature < -40°C
Unit price #2 [EUR]: 600.00
Unit price #3 [EUR]: 1,200.00
                                            Limit #3: Very big samples
```

#### 8.4.6 - Resistance to chemical substances for Group I equipment

(id 1075) Resistance to chemical agent test of IEC 60079-0 applies. The non-metallic enclosures and non-metallic parts of enclosures are submitted to tests of resistance to the following chemical agents:

- oils and greases (oil IRM 902 according to ASTM D5964);
- hydraulic liquids for mining applications (fire-resistant hydraulic fluid intended for operating at temperatures between -20 °C and +60 °C, comprising an aqueous solution of polymer in 35 % water).

The relevant tests are made on samples of enclosure sealed against the intrusion of test liquids into the interior of the enclosure.

```
Accredited [IEC17025]: YES
                                 Execution time [day]: 2
Unit price #1 [EUR]: 500.00
                                            Limit #1: up to 21
                                            Limit #2: from 21 up to 20 1
Unit price #2 [EUR]: 1,500.00
Unit price #3 [EUR]: 2,200.00
                                           Limit #3: from 211 up to 50 1
```

### 8.4.8 - Surface resistance test of non-conductive parts

(id 1076) This test is performed on non-conductive parts of the equipment relevant for explosion prevention and protection. Surface resistance test of IEC 60079-0 applies.

```
Accredited [IEC17025]: YES
                                  Execution time [day]: 1
Unit price #1 [EUR]: 500.00
                                            Limit #1: Not relevant
```

#### 8.4.9 - Thermal shock test

(id 1077) Thermal shock test of IEC 60079-0 applies.

```
Accredited [IEC17025]: YES
                                  Execution time [day]: 1
Unit price #1 [EUR]: 200.00
                                            Limit #1: Not relevant
```

#### D.4.2.1 - Rubbing with a pure polyamide cloth

(id 1102) Charging test with non-conductive materials

```
Accredited [IEC17025]: NO
                                  Execution time [day]: 0
Unit price #1: ask to office
                                            Limit #1: Not relevant
```

## D.4.2.2 - Rubbing with a cotton cloth

(id 1103) Charging tests with non-conductive materials


Accredited [IEC17025]: NO	Execution time [day]: 0
Unit price #1: ask to office	Limit #1: Not relevant



# QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



(id 1104) Charging tests with non-conductive materials

Accredited [IEC17025]: NO Execution time

albarubens

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1: ask to office Limit #1: Not relevant

#### 8.4.7 - Mechanical resistance tests

(id 1105) This paragraph indicates the conditions to which §8.3.1 and §8.3.2 mechanical tests shall be performed.

Accredited [IEC17025]: YES Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

#### D.4.1 - Conditioning

(id 1116) ---

Accredited [IEC17025]: YES Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

## ISO 80079-37:2016

(id 154) Explosive atmospheres - Part 37: Non-electrical equipment for explosive atmospheres - Non electrical type of protection constructional safety "c", control of ignition source "b", liquid immersion "k"

# 8.3.2 - Increased pressure test on enclosed equipment having a sealed enclosure that contains static or flowing protective liquid

(id 1106) The pressure test shall be performed without the equipment operating. It is not required for the enclosure in the case of open equipment.

#### 8.3.3 - Overpressure test on enclosed equipment having a vented enclosure

(id 1107) The pressure test shall be performed without the equipment operating. It is not required for the enclosure in the case of open equipment.

Accredited [IEC17025]: YES Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

#### 8.1 - Type tests for equipment with Type of Protection constructional safety c

(id 1117) See ISO 80079-36.

Accredited [IEC17025]: YES Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

#### 8.2.1 - Determination of control parameters

(id 1118) ---

#### 8.2.2 - Function and accuracy check of the ignition protection system

(id 1119) Sensors functional check.

#### B.1 - Dry run type test for lubricated sealing arrangements

(id 1120) ---

Accredited [IEC17025]: YES Execution time [day]: 0
Unit price #1 [EUR]: 1,500.00 Limit #1: Small sample

#### B.2 - Type test for determining the maximum engaging time of clutch assembly

(id 1122) ---

Accredited [IEC17025]: YES Execution time [day]: 2
Unit price #1: ask to office Limit #1: Not relevant



# § T - LABORATORY TESTS / Prove di laboratorio

PRICE LIST / Listino prezzi

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



## ISO 8846:1990

(id 1252) Small craft - Electrical devices - Protection against ignition of surrounding flammable gases

#### 4 - External surface temperature test

(id 1732) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 1,000.00 Limit #1: Applicable to any sample

#### 5 - Tests for sealed devices

(id 1733) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 300.00 Limit #1: Applicable to any sample

#### 6 - Tests for non-sealed devices

(id 1734) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 1,000.00 Limit #1: Applicable to any sample

# ISO 9227:2022

(id 1874) Corrosion tests in artificial atmospheres — Salt spray tests

#### 5.2.2 - NSS (Neutral salt spray test)

(id 2498) The NSS test is the test method in which a neutral approximate 5 % sodium chloride solution is atomized under a controlled environment. The NSS test is particularly applicable to:

- metals and their alloys;
- metallic coatings (anodic and cathodic);
- conversion coatings;
- anodic oxide coatings;
- organic coatings on metallic materials.

## ISO 9847:1992

(id 19) Solar energy - Calibration of field pyranometers by comparison to a reference pyranometer

#### 5 - Calibration of PV reference cells or pyranometers

(id 251) The purpose of this method, based on ISO 9847, IEC 60904-4, ASTM E1039-99 and ASTM E1125-10, is the calibration of silicon solar cells and field pyranometers under natural sunlight. This is an outdoor calibration. During the process, the sample is mounted on a tracker, where the sun beams are perpendicular. This leads to the determination of the tracker responsivity. The fixed responsivity is calculated by multiplying the tracker responsivity by the incidence angle modifier coefficient (IAM).

Suggestions to maximize the measurement accuracy while using the reference cell on field after calibration:

- The cell must be mounted on the module plane with an angular error lower than 5°, which corresponds to an error of the coefficient lower than 1%.
- If the plant use a biaxial tracker (which means that the solar radiation is always orthogonal to the module plane), then it is recommended to use the Tracker Responsivity indicated below.
- If the plant is fixed (which means that the solar radiation has got an incident angle typically low), it is recommended to use the Fixed Responsivity indicated below
- If the plant use a mono-axial tracker, it is suggested to use, to a first approximation, the arithmetic mean between Fixed and Tracker Responsivity.

Accredited [IEC17025]: NO Execution time [day]: 3
Unit price #1 [EUR]: 450.00 Limit #1: Applicable to any sample



# PRICE LIST / Listino prezzi

# your value certified § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



# ISO IEC 80079-49:2024

(id 1983) Flame arresters - Performance requirements, test methods and limits for use

#### 6.2, A.2 - Flow measurement (air) - In-line flame arrester

(id 3022) The pressure drop across the flame arrester shall be tested before and after flame transmission tests in accordance with Annex A at a volume flow that is suitable for identifying any alteration (deformation) of the flame arrester, particularly of the flame arrester element.

```
Accredited [IEC17025]: YES

Unit price #1 [EUR]: 500.00

Limit #1: up to 2 inches, standard ambient temperature
Unit price #2 [EUR]: 800.00

Limit #1: up to 2 inches, standard ambient temperature
Unit price #3 [EUR]: 1,200.00

Limit #2: from 3 to 9 inches, standard ambient temperature
Limit #3: from 10 inches, standard ambient temperature
```

#### 7.3.2.2 - Deflagration test - In-line flame arrester (indoor location)

(id 3023) ---

#### 7.3.2.2 - Deflagration test - In-line flame arrester (outdoor location)

(id 3024) ---

```
Accredited [IEC17025]: YES

Unit price #1 [EUR]: 5,000.00

Unit price #2 [EUR]: 6,800.00

Unit price #3 [EUR]: 8,200.00

Execution time [day]: 3

Limit #1: > 5 inches and < 9 inches, standard ambient temperature (max 500kg)

Limit #2: 9 ÷ 10 inches, standard ambient temperature (max 500kg)

Limit #3: > 10 inches and > 500kg, standard ambient temperature
```

#### 7.3.3.2 - Stable detonation without restriction (Type 4)

(id 3065) Detonation flame arresters tested for stable detonations without restriction (see 7.3.3.2) are classified as Type 4.

#### 7.3.2.1 - Deflagration test - End-of-line flame arrester (indoor location)

(id 3094) ---

```
Accredited [IEC17025]: YES
Unit price #1 [EUR]: 2,000.00
Unit price #2 [EUR]: 2,800.00
Unit price #3 [EUR]: 3,600.00

Execution time [day]: 3
Limit #1: < 1 inch, standard ambient temperature
Limit #2: 1 ÷ 2 inches, standard ambient temperature
Limit #3: > 2 inches and up to 5 inches, standard ambient temperature
```

### 6.2, A.3 - Flow measurement (air) - End-of-line flame arrester

(id 3095) The pressure drop across the flame arrester shall be tested before and after flame transmission tests in accordance with Annex A at a volume flow that is suitable for identifying any alteration (deformation) of the flame arrester, particularly of the flame arrester element.

## IVECO 18-0400-A006:1998 IT

(id 1769) CONDUZIONE LIQUIDO DI RAFFREDDAMENTO MOTORE

#### 6.13 - FUNZIONAMENTO SIMULATO

(id 2342) ---

```
Accredited [IEC17025]: NO Execution time [day]: 20
Unit price #1 [EUR]: 2,500.00 Limit #1: Applicable to any sample
```

## IVECO 18-0400-A035:2005

(id 2047) CONDUZIONE ARIA – TUBI DI PLASTICA PER IMPIANTO PNEUMATICO FRENI

#### 35.18 - Prova di fatica a pressione pulsante

(id 2801) ---

Accredited [IEC17025]: NO Execution time [day]: 60
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant



# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX

#### 35.18.4 - Pressione pulsante

(id 2802) ---

Accredited [IEC17025]: NO Execution time [day]: 60

Unit price #1 [EUR]: 5,500.00

Limit #1: Applicable to any sample

# IVECO STD 18-2252:2023

(id 2007) GENERAL CONDITIONS FOR APPLICATIONS OF ELECTRONIC AND ELECTRO-TELEMATIC DEVICES ON INDUSTRIAL VEHICLES

#### 14.4 - Resistence to fuels

(id 2769) The test is intended to determine the capacity of a component or of an assembly to resist chemical aggression of fuels, by atomizing mechanical parts located in engine compartment or similar areas.

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 600.00 Limit #1: Applicable to any sample

#### 14.6 - Resistance to flushing hydrocarbon mixtures

(id 2770) The test is intended to determine the capacity of a component or of an assembly to resist chemical aggression of hydrocarbon mixtures used by repair shops, by spraying cleaning product on mechanical parts located in engine compartment or similar areas.

Accredited [IEC17025]: NO Execution time [day]: 2

Unit price #1 [EUR]: 400.00 Limit #1: Applicable to any sample

## IVECO STD 18-2280:2023

(id 2147) ALTERNATORI

#### 9.3 - Resistenza al fango ed al salmastro

(id 2974) ---

Accredited [IEC17025]: NO Execution time [day]: 2

Unit price #1 [EUR]: 2,800.00 Limit #1: Applicable to any sample

#### 9.5 - Resistenza al lavaggio a vapore ed al lavaggio a pressione

(id 2975) ---

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 1,200.00 Limit #1: largest dimension < 250 mm Unit price #2 [EUR]: 1,600.00 Limit #2: largest dimension >= 250 mm

#### 9.6 - Resistenza all'atmosfera industriale

(id 2976) ---

Accredited [IEC17025]: NO Execution time [day]: 3

Unit price #1 [EUR]: 2,500.00 Limit #1: Applicable to any sample

## IVECO STD 18-3303:2019

(id 1989) SERBATOI COMBUSTIBILE IN MATERIALE PLASTICO

#### 9.2 - Prova di fatica a pressione pulsante-depressione

(id 2741) ---

Accredited [IEC17025]: NO Execution time [day]: 7

Unit price #1 [EUR]: 2,500.00 Limit #1: Applicable to any sample

## 9.3.4 - Cicli ad alta temperatura

(id 2742) ---

Accredited [IEC17025]: NO Execution time [day]: 50
Unit price #1 [EUR]: 6,500.00 Limit #1: Not relevant

#### 9.2.1/9.3.1 - Precondizionamento

(id 2744) ---

Accredited [IEC17025]: NO Execution time [day]: 140

Unit price #1 [EUR]: 6,000.00 Limit #1. Applicable to any sample no fuel change



# **PRICE LIST / Listino prezzi**

# § T - LABORATORY TESTS / Prove di laboratorio

## QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version



#### 6.1 - Tenuta a bassa ed alta pressione

albarubens

(id 2745) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 300.00 Limit #1: Applicable to any sample

#### 9.3.8 - Cicli a bassa temperatura

(id 2746) ---

Accredited [IEC17025]: NO Execution time [day]: 7

Unit price #1 [EUR]: 1,800.00 Limit #1: Applicable to any sample

## IVECO STD 18-3311:2017

(id 1908) SERBATOI PER ADBLUE IN MATERIALE PLASTICO - GAMMA LIGHT -

#### 10.2 - Thermal cycle

(id 2567) Thermal cycle for AdBlue filled tanks.

Accredited [IEC17025]: NO Execution time [day]: 4

Unit price #1 [EUR]: 1,200.00 Limit #1: Applicable to any sample

#### 11 - Resistance to icing-deicing cycle

(id 2568) This test is intended to verify the resistance to icing-deicing cycle of AdBlue filled tanks.

Accredited [IEC17025]: NO Execution time [day]: 12

Unit price #1 [EUR]: 1,800.00 Limit #1: Applicable to any sample - 3 cycles (A)

#### 9 - Pulsating Pressure Test

(id 2584) ---

Execution time [day]: 3 Accredited [IEC17025]: NO

Limit #1: Applicable to any sample Unit price #1 [EUR]: 1,200.00

## 7 - Leak test

(id 2638) Leak test for under pressure tanks immersed in water.

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 150.00 Limit #1: Applicable to any sample

## IVECO STD 18-2102:2008

(id 2148) MOTORI D'AVVIAMENTO

#### - Tenuta all'olio del lato pignone

(id 2977) ---

Accredited [IEC17025]: NO Execution time [day]: 2

Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

# JLR 462 Validation phase

(id 1600) EXTERIOR LIGHTING CORROSION TEST (DVM-0069-EX)

## unref - Exterior lighting corrosion test

(id 2035) The aim of this test is to simulate Kaolin-type dust accelerated corrosion

Accredited [IEC17025]: NO Execution time [day]: 84 Unit price #1 [EUR]: 4,000.00 Limit #1: Not relevant



# PRICE LIST / Listino prezzi

# § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



# Kautex 111120:1998 - rev1 ver3 2012

(id 1760) Minimum performance requirements for elastomer type hose e.g. EPDM used to convey wash fluid in screenwash systems

#### 5.2.1 - Heat Cycle Test

(id 2317) ---

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

#### 5.2.2 - Freeze/Thaw Resistance

(id 2318) ---

Accredited [IEC17025]: NO Execution time [day]: 3

Unit price #1 [EUR]: 1,000.00 Limit #1: Applicable to any sample

#### 5.2.3 - High Humidity Resistance

(id 2319) ---

Accredited [IEC17025]: NO Execution time [day]: 3

Unit price #1 [EUR]: 600.00 Limit #1: Applicable to any sample

#### 5.2.4 - Light Resistance

(id 2320) ---

Accredited [IEC17025]: NO Execution time [day]: 2

Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

#### 5.3.1 - Resistance to Wash Fluids

(id 2321) ---

Accredited [IEC17025]: NO Execution time [day]: 7

Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

#### 5.3.2 - Resistance to Under Bonnet Fluids

(id 2322) ---

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

#### 5.3.3 - Ozone Resistance

(id 2323) ---

Accredited [IEC17025]: NO Execution time [day]: 5
Unit price #1: ask to office Limit #1: Not relevant

### 5.4.1 - Operating Endurance at different temperatures

(id 2324) ---

Accredited [IEC17025]: NO Execution time [day]: 7

Unit price #1 [EUR]: 2,500.00 Limit #1: Applicable to any sample

# **LAH.3P3.827 Version 01**

(id 1810) Component Performance Specification

#### 6.4.4.4 - Changing climate test

(id 2426) Resistance to environmental cycle test as per PV 1200.

Accredited [IEC17025]: NO Execution time [day]: 10
Unit price #1 [EUR]: 3,500.00 Limit #1: Small sample
Unit price #2 [EUR]: 8,800.00 Limit #2: Big sample

## 6.4.4.3 - Thermal Storage Test

(id 2427) The parts have to withstand thermal load that can occur as a result of thermal changes during production, transportation and customer use of the car.

Accredited [IEC17025]: NO Execution time [day]: 5

Unit price #1 [EUR]: 3,000.00 Limit #1: Applicable to any sample



# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



## LP-463PB-22-01

(id 1416) Cycle testing of painted surfaces

## unref - Cycle testing of painted surfaces - Method IV

(id 1881) ---

Accredited [IEC17025]: NO Execution time [day]: 20
Unit price #1 [EUR]: 3,000.00 Limit #1: Not relevant

## LP.7A003:2017

(id 1468) DEF System - Refilling

#### 11.1.1 - Plant Fill - A

(id 2569) Verify the DEF system fills under plant conditions.

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 50.00 Limit #1: Applicable to any sample

## 11.1.2 - Rated -Advertised- Fill Capacity and Quality - B

(id 2570) Verify/determine the DEF system rated fill capacity.

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 400.00 Limit #1: Applicable to any sample

## 11.3 - Cold Fill - F

(id 2571) Verify DEF fill system refills under cold conditions.

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 80.00 Limit #1: Applicable to any sample

#### 11.4 - Frozen Fill - G

(id 2572) Verify tank can accept DEF with 50% rated capacity of frozen DEF.

Accredited [IEC17025]: NO Execution time [day]: 2

Unit price #1 [EUR]: 500.00 Limit #1: Applicable to any sample

## 11.1.3.1 - Round-Up - H

(id 2573) Verify fill quality and ice dome volume is protected.

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 80.00 Limit #1: Applicable to any sample

#### 11.1.3.2 - Trickle Fill - I

(id 2574) Verify ice expansion dome is protected under extreme refilling conditions.

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 80.00 Limit #1: Applicable to any sample

#### 11.1.4 - Filling on Grade - K

(id 2575) Verify the fill quality and capacity in worst case condition.

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 80.00 Limit #1: Applicable to any sample

#### 11.2.1 - Failed Nozzle - C

(id 2576) Verify the DEF system allows for the DEF to be discharged in a manner such that the operator can shut the nozzle off without being sprayed with DEF.

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 80.00 Limit #1: Applicable to any sample

#### 11.2.2 - Market Nozzle Evaluation - D

(id 2577) Verify capability of DEF system with all known market nozzles.

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 80.00 Limit #1: Applicable to any sample



# **PRICE LIST / Listino prezzi** § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

TNDEX

11.2.3 - Short Fill - J

(id 2578) Verify/determine the DEF system fill capacity.

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 80.00 Limit #1: Applicable to any sample

#### 11.2.4 - Nozzle Orientation - E

(id 2579) Verify DEF fill system refills under extreme nozzle orientations.

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 100.00 Limit #1: Applicable to any sample

## LP.7A003:2023

(id 1997) DEF SYSTEM - REFILLING (STELLANTIS HARMONIZED)

#### 12.1.2 - Rated Advertised Fill Capacity and Quality

(id 2764) Verify/determine the DEF system rated fill capacity.

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 400.00 Limit #1: Applicable to any sample

#### 12.1.1 - Plant Fill

(id 2999) Simulate plant conditions and verify fill quality and volume with DEF system buck at assembly plant

Accredited [IEC17025]: NO Execution time [day]: 2 Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 12.1.3 - Failed nozzle

(id 3000) ---

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 12.1.4 - Market Nozzle Evaluation

(id 3001) ---

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 12.1.5 - Nozzle Orientation

(id 3002) ---

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

#### 12.1.6 - Cold Fill

(id 3003) These tests can be performed on vehicle during cold trip

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 12.1.7 - Frozen Fill

(id 3004) ---

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 12.1.8.1 - Overfilling Procedure - Round-Up

(id 3005) ---

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

### 12.1.8.2 - Overfilling Procedure - Trickle Fill

(id 3006) ---

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 200.00 Limit #1: Applicable to any sample with one type of nozzle



# PRICE LIST / Listino prezzi § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



#### 12.1.8.3 - Overfilling Procedure - Excessive Trickle Fill

(id 3007) ---

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 12.1.9 - Short Fill

(id 3008) ---

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 12.1.10 - Filling on grade

(id 3009) ---

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 600.00 Limit #1: Applicable to any sample with one type of nozzle

#### 12.1.11 - Emergency Refueling with vehicle on park grade

(id 3010) This test is requested for xPSA vehicle only.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

## LP.7A005:2017

(id 1437) FCA MINI-SHED Evaporative emission Test procedure

#### 12 - MiniSHED 24h permeation test - tank

(id 2147) The test consists of heating and cooling the fuel tank throughout a specified diurnal temperature cycle while measuring the hydrocarbons that permeate through the surfaces and seams of the fuel tank assembly into the test enclosure.

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 1,500.00 Limit #1: Any tank

#### 9.4 - Helium leak test

(id 2159) This test is aimed to detect leaks in the test specimen. Generally, it is applied prior to the SHED test and immediately after SHED testing. The method to check for leaks is either by helium leak or by hydrocarbon leak check.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 400.00 Limit #1: Applicable to any sample

#### 12/C - MicroSHED permeation test - pipe

(id 2230) Executed before and after preconditioning

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any tube

#### 9.3 - Soak Procedure (5 weeks)

(id 2927) ---

Accredited [IEC17025]: NO Execution time [day]: 35
Unit price #1 [EUR]: 1,200.00 Limit #1: Applicable to any sample

## 9.3/B - Soak Procedure (10 weeks)

(id 2928) ---

Accredited [IEC17025]: NO Execution time [day]: 70 Unit price #1 [EUR]: 2,400.00 Limit #1: Applicable to any sample

#### 12/B - MiniSHED 48h permeation test - tank

(id 3026) The test consists of heating and cooling the fuel tank throughout a specified diurnal temperature cycle while measuring the hydrocarbons that permeate through the surfaces and seams of the fuel tank assembly into the test enclosure.

Accredited [IEC17025]: NO Execution time [day]: 3
Unit price #1 [EUR]: 1,600.00 Limit #1: Any tank

#### 12/D - MiniSHED fast permeation test - tank

(id 3079) This test derives from the standard Minished, with constant temperature and reduced duration. It enables a rapid prequalification, which is useful in comparative R&D activities.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

page 177/234



albarubens

# PRICE LIST / Listino prezzi § T - LABORATORY TESTS / Prove di laboratorio

## QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



## LP.7A006:2024

(id 2202) FUEL SUPPLY SYSTEM - Priming/unpriming checks on test bench

#### 12.1.1 - Minimum volume of fluid necessary to prime the lift pump with tank level [Vp]

(id 3086) ---

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 200.00 Limit #1: Applicable to any sample

#### 12.1.2 - Definition of the residual fuel volume inside the tank at which pump unpriming occurs on the level [Vu]

(id 3087) ---

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 200.00 Limit #1: Applicable to any sample

## 12.1.3 - Definition of the residual fuel volume inside the tank at which pump unpriming occurs on grade "P"

(id 3088) ---

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 200.00 Limit #1: Applicable to any sample

#### 12.1 - Fuel System Priming-Unpriming Checks

(id 3096) ---

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 600.00 Limit #1: Applicable to any sample

## LP.7A008:2018

(id 1861) FUEL SYSTEM - Fuel Tank Volumes and Capacity Determination

## 12.1 - Internal fluid volume as determined by vent port or FLVV position (geometrical)

(id 2489) ---

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 400.00 Limit #1: Applicable to any sample

## 12.3 - Passive vapor dome volume (no connection to exterior)

(id 2490) ---

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 100.00 Limit #1: Applicable to any sample

#### 12.4 - Total vapor dome volume (active + passive)

(id 2491) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 100.00 Limit #1: Applicable to any sample

#### 12.5 - Tank total internal fluid volume

(id 2492) ---

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 100.00 Limit #1: Applicable to any sample

#### 12.6 - Volume of fluid inside filler tube

(id 2493) ---

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 100.00 Limit #1: Applicable to any sample

# 12.8 - Level curve characteristics with simulation of vehicle on level ground $(id\ 2494)$ ---

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample



# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



## LP.7T011 Rev.D:2019

(id 1790) New external document

### **Unref. - ISO For Life Componente**

(id 2380) 10 weeks cycle of climatic chamber and salt mist chamber.

Accredited [IEC17025]: NO Execution time [day]: 70

Unit price #1 [EUR]: 12,000.00 Limit #1: Applicable to any sample

# MBN 10306:2020

(id 2059) Electric and Electronic Components in Motor Vehicles - Environmental Requirements and Tests

#### 8.2 - M-02 Stone chip test

(id 2827) This test simulates the mechanical exposure of the component to grit impact.

The test is intended to verify the resistance of the component to faults such as deformation and cracks.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 7,000.00 Limit #1: Not relevant

#### 8.3 - M-03 Dust test

(id 2828) This test simulates the dust load of the component during vehicle operation.

The test is intended to verify the resistance of the component to electrical and mechanical faults.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 600.00 Limit #1: Standard IP chamber (small sample not powered)
Unit price #2 [EUR]: 1,200.00 Limit #2: Medium IP chamber (small sample not powered)
Unit price #3 [EUR]: 2,200.00 Limit #3: Powered on sample

#### 8.5 - M-05 Mechanical shock

(id 2829) This test simulates the mechanical exposure of components, e.g. when driving over curbs or in accidents.

It is intended to verify the resistance of the component to faults such as cracks or component element displacements.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 8.7 - M-07 Pressure change test coolant circuit

(id 2830) This test simulates the load on the component due to variations in the coolant pressure as well as conditions during the post-heating phase and the filling of the cooling system under vacuum. It shall be exclusively used for components that are connected to a coolant circuit.

It verifies the mechanical strength of component parts exposed to pressure variations in the coolant circuit (e.g. heat sinks of the power module).

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 9.1 - K-01 High-low temperature storage

(id 2831) This test simulates the thermal exposure of the component during storage and transport.

The test is intended to verify the resistance to storage at high or low temperatures, e.g. during the transport of the component (aircraft, ship container).

If the test is carried out at the beginning of a test sequence, it is also intended to adjust all components to the same initial conditions.

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 600.00 Limit #1: Applicable to any sample

#### 9.2 - K-02 Temperature step test

(id 2832) This test simulates the operation of the component at different ambient temperatures.

The test is intended to verify the resistance of the component to malfunctions that may occur within a small interval of the ambient temperature range as well as to verify the start-up behavior of the component over the entire ambient temperature range.

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 3,600.00 Limit #1: Applicable to any sample

#### 9.3 - K-03 Low temperature operation

(id 2833) This test simulates the exposure of the component to low temperatures.

The test is intended to verify the function of the component after a long parking time or operation time at extremely low temperatures.

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample



# PRICE LIST / Listino prezzi your value certified § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

TNDEX



## 9.4 - K-04 Repainting temperature

(id 2834) This test simulates the exposure of the component during repainting.

The test is intended to verify the resistance of the component to faults that occur due to thermal exposure, e.g. crack formation in soldered joints, adhesive joints and welded joints, in bond connections as well as in seals and housings.

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1 [EUR]: 400.00 Limit #1: Applicable to any sample

#### 9.5 - K-05 Thermal shock (component)

(id 2835) This test simulates the thermal exposure of the component to rapid temperature changes during vehicle operation.

The test is intended to verify the resistance of the component to faults that occur due to thermal exposure, e.g. crack formation in soldered joints, adhesive joints and welded joints, in bond connections as well as in seals and housings.

Accredited [IEC17025]: NO Execution time [day]: 30 Unit price #1 [EUR]: 3,800.00 Limit #1: Small sample (up to 300mm) with dedicated technician Unit price #2 [EUR]: 7,200.00 Limit #2: Big or multiple sample with dedicated technician

#### 9.6 - K-06 Salt spray testing, operating, exterior

(id 2836) This test simulates the exposure of the component to air and water containing salt, a situation that may occur in certain areas of the world and under wintry road conditions.

The test is intended to verify the resistance of the component to malfunction when exposed to salt, e.g. due to short circuits and leakage currents caused by the ingress of salt into the component.

Accredited [IEC17025]: NO Execution time [day]: 6 Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 9.9 - K-09 Damp heat, cyclic (with frost)

(id 2837) This test simulates the thermal exposure of a component to cyclic temperature changes (including frost) at high humidity during vehicle operation. The test is intended to verify the resistance of the component when exposed to damp heat.

Accredited [IEC17025]: NO Execution time [day]: 10 Unit price #1 [EUR]: 1,800.00 Limit #1: Applicable to any sample

#### 9.10 - K-10 Water protection - IP X6K

(id 2838) This test simulates the exposure of the component to water.

The test is intended to verify the function of the component, e.g. when exposed to condensation water, rain or spray water.

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1 [EUR]: 400.00 Unit price #2 [EUR]: 1,000.00 Limit #1: very small sample Limit #2: medium/big sample

#### 9.11 - K-11 High-pressure-steam-jet cleaning - IP X9K

(id 2839) This test simulates the exposure of the component to water during vehicle cleaning. The test is intended to verify the function of the component when exposed to high-pressure/steam-jet cleaning.

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1 [EUR]: 1,200.00 Unit price #2 [EUR]: 3,000.00 Limit #1: small sample (automatic spraying) Limit #2: big sample (manual spraying)

#### 9.12 - K-12 Thermal shock with splash water

(id 2840) This test simulates the exposure of the component to splash water as it occurs when driving through puddles. The test is intended to verify the function of the component when exposed to abrupt cooling by means of water.

Accredited [IEC17025]: NO Execution time [dav]: 1 Unit price #1 [EUR]: 4,000.00 Limit #1: Small sample (up to 40cm) Unit price #2 [EUR]: 5,000.00 Limit #2: Medium sample (up to 80cm) Unit price #3 [EUR]: 7,000.00 Limit #3: Big sample (up to 140cm)

## 9.13 - K-13 Thermal shock immersion - IP X7

(id 2841) This test simulates the exposure of the component when immersed in water.

The test is intended to verify the function of the component when subjected to immediate cooling by means of immersion of the heated component.

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1 [EUR]: 3,000.00 Limit #1: Applicable to any sample

#### 9.14 - K-14 Damp heat, steady state

(id 2842) his accelerated test simulates the exposure of the component to damp heat during vehicle service life.

The test is intended to verify the quality and reliability of the component with respect to faults caused by damp heat, e.g. corrosion, migration/dendritic growth, swelling and degradation of plastics, sealing and grouting compounds.

Accredited [IEC17025]: NO Execution time [day]: 2 Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant



# § T - LABORATORY TESTS / Prove di laboratorio

PRICE LIST / Listino prezzi

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



#### 9.15.2 - K-15 b Climate test for components with waterproof housings

(id 2843) This accelerated test simulates the thermal exposure of the component to damp heat taking into account the protective effect of waterproof housings during the vehicle service life.

The test is intended to verify the quality and reliability of the component with respect to faults caused by damp heat, e.g. corrosion, migration/dendritic growth, swelling and degradation of plastics, sealing and embedding compounds.

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 10.1 - C-01 Chemical tests

(id 2844) This test simulates the exposure of the component to different chemicals.

The test is intended to verify the resistance of the component to chemical changes on the housing and impairment of functioning due to chemical reactions.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 11.2 - L-02 High-temperature endurance test

(id 2845) This accelerated test simulates the thermal exposure of the component during the electrical operation over the vehicle service life.

The test is intended to verify the quality and reliability of the component with respect to faults that occur due to thermal exposure such as diffusion, migration and oxidation.

Accredited [IEC17025]: NO Execution time [day]: 27
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 11.3 - L-03 Temperature cycle endurance test

(id 2846) This accelerated test simulates the thermomechanical exposure of the component as a result of temperature changes that occur during vehicle service life.

The test is intended to verify the quality and reliability of the component with respect to faults that occur due to thermomechanical exposure such as aging and cracking in soldered joints, adhesive joints and welded joints, in bond connections as well as in seals or housings.

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 5.4.2 - P-01 Parameter test (function test)

(id 2847) ---

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 5.4.3 - P-02 Parameter test (small)

(id 2848) ---

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 400.00 Limit #1: Applicable to any sample

#### 5.4.4 - P-03 Parameter test (large)

(id 2849) ---

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 2,000.00 Limit #1: Applicable to any sample

#### 5.4.5 - P-04 Physical analysis

(id 2850) ---

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 5.6.2 - Leak test

(id 2851) A leak test is conducted to verify that the component-specific defined leak rate limit of the electronic compartment of a component complies with the specifications.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 300.00 Limit #1: Applicable to any sample

#### Unref. - Parameter test

(id 2852) ---

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant



# PRICE LIST / Listino prezzi § T - LABORATORY TESTS / Prove di laboratorio

# alb**ar**ubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX

## 8.4 - M-04 Vibration test

(id 2857) ---

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### Unref. - Voltage withstand, with preconditioning

(id 2858) --

Accredited [IEC17025]: NO Execution time [day]: 2

Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

#### Unref. - Isolation resistance, with preconditioning

(id 2859) ---

Accredited [IEC17025]: NO Execution time [day]: 2

Unit price #1 [EUR]: 500.00 Limit #1: Applicable to any sample

#### Unref. - Full operation test

(id 2860) ---

Accredited [IEC17025]: NO Execution time [day]: 2

Unit price #1 [EUR]: 400.00 Limit #1: Applicable to any sample

#### Unref. - Equipotential bonding

(id 2861) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 500.00 Limit #1: Applicable to any sample

#### Unref. - Physical analysis

(id 2862) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 100.00 Limit #1: Technical examination

## MBN LV124-2:2013

(id 1784) Electric and Electronic Components in Motor Vehicles up to 3,5t – General Requirements, Test Conditions and Tests Part 2: Environmental Requirements

#### 13.2 - M-02 Stone chip test

(id 2359) This test simulates the mechanical exposure of the component to grit impact. The test is intended to verify the resistance of the component to faults such as deformation and cracks.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,300.00 Limit #1: up to 5 grouped samples and one point
Unit price #2 [EUR]: 7,000.00 Limit #2: up to 5 grouped samples and up to 6 points

#### 13.3 - M-03 Dust test

(id 2360) This test simulates the dust load of the component during vehicle operation. The test is intended to verify the resistance of the component to electrical and mechanical faults

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 2,200.00 Limit #1: Applicable to any sample

#### 14.9 - K-10 Water protection - IP X6K

(id 2361) This test simulates the exposure of the component to water.

The test is intended to verify the function of the component, e.g. when exposed to condensation water, rain or spray water.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 1,000.00 Limit #1: up to 5 grouped samples

#### 14.11 - K-12 Thermal shock with splash water

(id 2362) This test simulates the exposure of the component to splash water as it occurs when driving through puddles. The test is intended to verify the function of the component when exposed to abrupt cooling by means of water.





QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



#### 14.17 - K-18 Corrosion test with flow of mixed gas

(id 2363) This test simulates the influence of corrosive gases on the component, particularly its plug contacts and switches. The test is intended to verify the resistance of the component to faults such as corrosion and component damages.

Accredited [IEC17025]: NO Execution time [day]: 21
Unit price #1 [EUR]: 3,200.00 Limit #1: Applicable to any sample

## Mercedes A2100064099 2020-11-03

(id 1755) Test specification for coolant hoses

#### 5 - Temperature and pressure test on hoses

(id 2264) ---

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

# Metodo Interno per il controllo delle attrezzature Ex m in condizioni iperba

(id 1966) Il seguente metodo interno contiene le varianti dei test contenuti nella norma IEC 60079-18:2014+AMD1:2017, eseguiti in condizioni iperbariche.

#### 8.2.2 - Maximum temperature VAR

(id 2669) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 1,200.00 Limit #1: Applicable to any sample

#### 8.2.3.1 - Thermal endurance to heat VAR

(id 2670) Evaluation of the thermal endurance to heat of non-metallic equipment or parts of equipment under hyperbaric conditions.

#### 8.2.6 - Pressure test for Group I and Group II electrical equipment VAR

(id 2671) This test is required for level of protection "ma" equipment with any individual free spaces between 1 cm3 and 10 cm3 and level of protection "mb" equipment with any individual free spaces between 10 cm3 and 100 cm3.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

#### 8.2.8 - Sealing test for build-in protective devices VAR

(id 2672) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 1,200.00 Limit #1: Applicable to any sample

#### 8.1.1 - Water absorption test VAR

(id 2673) Test on the compound under hyperbaric conditions.

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 600.00 Limit #1: Applicable to any sample

### MIL-STD 810F:2000

(id 1898) DEPARTMENT OF DEFENSE- TEST METHOD STANDARD FOR ENVIRONMENTAL ENGINEERING CONSIDERATIONS AND LABORATORY TESTS

#### 512.4 - Immersion (Procedure I)

(id 2545) The immersion test is performed to determine if materiel can withstand immersion in water and operate as required during or following immersion.

Accredited [IEC17025]: NO Execution time [day]: 2 Unit price #1 [EUR]: 1,200.00 Limit #1: Applicable to any sample



# PRICE LIST / Listino prezzi your value certified § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

TNDEX

#### 509.4 - Salt Fog

(id 2546) ---

Accredited [IEC17025]: NO Execution time [day]: 4

Unit price #1 [EUR]: 1,500.00 Limit #1: Applicable to any sample

#### 506.4 - Rain Test (Procedure I)

(id 2685) Use this method to evaluate materiel likely to be exposed to rain, water spray, or dripping water during storage, transit, or operation. Procedure I is applicable for materiel which will be deployed out-

of-doors and which will be unprotected from rain or blowing rain. The accompanying wind velocity can vary from almost calm to extremely high.

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1 [EUR]: 1,500.00 Limit #1: Applicable to any sample

## MIL-STD-810G CHG-1:2014

(id 167) ENVIRONMENTAL ENGINEERING PROGRAM GUIDELINES

#### 507.6-2.6.3 - Humidity - Procedure II - Aggravated Cycle

(id 1622) The purpose of the aggravated test procedure is to produce representative effects that typically occur when materiel is exposed to elevated temperature-

Accordingly, this procedure does not reproduce naturally occurring or service-induced temperature-humidity scenarios. It may induce problems that are indicative of long-term effects. Test item failures do not necessarily indicate failures in the real environment.

Accredited [IEC17025]: NO Execution time [day]: 10 Unit price #1 [EUR]: 2,500.00 Limit #1: Applicable to any sample

#### 500.6-4.5.2 - Low Pressure: Altitude - Procedure I

(id 2162) Low pressure (altitude) tests are used to determine if materiel can withstand and/or operate in a low pressure environment and/or withstand rapid pressure changes.

Procedure I is appropriate if the materiel is to be transported or stored at high ground elevations or transported by air in its shipping/storage configuration.

Accredited [IEC17025]: NO Execution time [day]: 2 Unit price #1 [EUR]: 2,800.00 Limit #1: Applicable to any sample

### 510.6-4.2 - Sand and dust - Procedure II: Blowing Sand

(id 2163) Procedure II - Blowing Sand is used to investigate the susceptibility of materiel to the effects of blowing sand (150 Pm to 850 Pm).

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1 [EUR]: 2,500.00 Limit #1: Applicable to any sample

#### 506.6-2.3.2 - Rain - Procedure III Drip

(id 2174) Use this Method to evaluate materiel likely to be exposed to rain, water spray, or dripping water during storage, transit, or operation. Procedure III is appropriate when materiel is normally protected from rain but may be exposed to falling water from condensation or leakage from upper surfaces.

Accredited [TEC170251: NO Execution time [day]: 2 Unit price #1 [EUR]: 400.00 Limit #1: both dimensions < 300 mm Unit price #2 [EUR]: 800.00 Limit #2: one dimension <= 300 mm and one dimension >= 300 mm Unit price #3 [EUR]: 1,200.00 Limit #3: both dimensions > 300 mm

#### 501.6 - High temperature - Procedure I

(id 2175) Use high temperature tests to obtain data to help evaluate effects of high temperature conditions on materiel safety, integrity, and performance. Use Procedure I to investigate how high temperatures during storage affect the material (integrity of materials, and safety/performance of the material).

Accredited [IEC17025]: NO Execution time [day]: 7 Unit price #1 [EUR]: 1,000.00 Limit #1: Up to 7 cycles

#### 502.6 - Low temperature - Procedure I

(id 2176) Use low temperature tests to obtain data to help evaluate effects of low temperature conditions on materiel safety, integrity, and performance during storage, operation, and manipulation.

Use Procedure I to investigate how low temperatures during storage affect materiel safety during and after storage, and performance after storage.

Accredited [IEC17025]: NO Execution time [day]: 3 Unit price #1 [EUR]: 600.00 Limit #1: Up to 4 hours Limit #2: Up to 24 hours Unit price #2 [EUR]: 1,000.00 Unit price #3 [EUR]: 1,500.00 Limit #3: Up to 72 hours

#### 516.7-4.6.3 - Shock test - Transport shock - Procedure II

(id 2179) Use this Method to evaluate the physical and functional performance of materiel likely to be exposed to mechanically induced shocks in its lifetime. Procedure II is used to evaluate the response of an item or restraint system to transportation environments that create a repetitive shock load. The procedure uses a classical terminal peak sawtooth, either measured or a synthetic shock waveform, to represent the shock excitation portion of the transportation scenario.

Accredited [IEC17025]: NO Execution time [day]: 2 Subcontract: BPS srl Unit price #1 [EUR]: 1,500.00 Limit #1: Applicable to any sample



# PRICE LIST / Listino prezzi

# § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

TNDEX



#### 514.7 category 20 - Vibration - Annex B AECTP 400

(id 2180) The purpose of this test method is to provide guidance for defining vibration environments material may be exposed to throughout a life cycle and to provide quidance for the conduct of laboratory vibration tests. Vibration tests are performed to:

a. Develop materiel to function in and withstand the vibration exposures of a life cycle including synergistic effects of other environmental factors, materiel duty cycle, and maintenance.

b. Verify that materiel will function in and withstand the vibration exposures of a life cycle.

```
Subcontract: BPS srl
Accredited [IEC17025]: NO
                                  Execution time [day]: 1
Unit price #1 [EUR]: 1,500.00
                                            Limit #1: Applicable to any sample
```

#### 509.6 - 4.5.2 - Salt Fog

(id 2585) The salt fog Method is performed to determine the effectiveness of protective coatings and finishes on materials. It may also be applied to determine the effects of salt deposits on the physical and electrical aspects of materiel.

```
Accredited [IEC17025]: NO
                                  Execution time [day]: 4
Unit price #1 [EUR]: 1,000.00
                                            Limit #1: 48 hours
Unit price #2 [EUR]: 1,400.00
                                            Limit #2: 96 hours
                                            Limit #3: big sample up to 96 hours
Unit price #3 [EUR]: 1,800.00
```

#### 504.2 - Contamination by fluids - Procedure I

(id 2687) Use contamination by fluids tests to determine if materiel (or material samples) is affected by temporary exposure to contaminating fluids (liquids) such as may be encountered and applied during its life cycle, either occasionally, intermittently, or over extended periods.

```
Accredited [IEC17025]: NO
                                  Execution time [day]: 1
Unit price #1 [EUR]: 1,000.00
                                            Limit #1: Occasional contamination
```

#### 510.6 - Sand and dust - Procedure I

(id 2688) Dust (< 150 ?m) procedure. This test is performed to help evaluate the ability of materiel to resist the effects of dust that may obstruct openings, penetrate into cracks, crevices, bearings, and joints, and to evaluate the effectiveness of filters.

```
Accredited [IEC17025]: NO
                                  Execution time [day]: 2
Unit price #1: ask to office
                                            Limit #1: Non eseguibile
```

#### 501.6 - High temperature - Procedure II

(id 2729) Use high temperature tests to obtain data to help evaluate effects of high temperature conditions on materiel safety, integrity, and performance. Use Procedure II to investigate how high ambient temperatures may affect material performance while it is operating.

```
Accredited [IEC17025]: NO
                                 Execution time [day]: 3
Unit price #1 [EUR]: 300.00
                                            Limit #1: Test at static conditions up to 24h
Unit price #2 [EUR]: 500.00
                                            Limit #2: Test at cyclic conditions up to 3 cycles
```

#### 502.6 - Low temperature - Procedure II

```
Accredited [IEC17025]: NO
                                  Execution time [day]: 1
Unit price #1 [EUR]: 400.00
                                            Limit #1: Up to 24h
```

#### 507.6 - Humidity - Procedure I

(id 2747) ---

```
Accredited [IEC17025]: NO
                                  Execution time [day]: 2
Unit price #1 [EUR]: 1,000.00
                                            Limit #1: Not relevant
```

#### 511.6 - 4.5.2 - Explosive atmosphere - Procedure I

(id 2775) This test evaluates the ability of the test item to be operated in a fuel vapor environment without igniting the environment.

```
Accredited [IEC17025]: NO
                                  Execution time [day]: 1
Unit price #1 [EUR]: 1,500.00
                                            Limit #1: Test inside small vessel
```

#### 516.7 - 4.6.2 - Shock - Functional Shock - Procedure I

(id 2776) The intent of this test is to disclose material malfunction that may result from shocks experienced by material during use in the field.

```
Accredited [IEC17025]: NO
                                  Execution time [day]: 1
Unit price #1 [EUR]: 1,000.00
                                            Limit #1: Not relevant
```

#### 503.6 - Temperature Shock

(id 2777) This method includes one test procedure with four variations essentially in the length of the test and the shock itself.

It employs constant temperature at each of the extreme shock conditions.

The four variations are:

- a. Procedure I-A One-way shocks from Constant Extreme Temperature.
- b. Procedure I-B Single Cycle Shock from Constant Extreme Temperature.
- c. Procedure I-C Multi-Cycle Shocks from Constant Extreme Temperature.
- d. Procedure I-D Shocks To or From Controlled Ambient Temperature.

page 185/234



# PRICE LIST / Listino prezzi your value certified § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

TNDEX

Accredited [IEC17025]: NO Execution time [day]: 1

### MIL-STD-810H:2019

Unit price #1 [EUR]: 800.00

(id 1250) Environmental Engineering Considerations and Laboratory Tests

#### 500.6 - Low Pressure: Altitude - Procedure I

(id 2165) Low pressure (altitude) tests are used to determine if materiel can withstand and/or operate in a low pressure environment and/or withstand rapid pressure changes.

Limit #1: One cycle

Procedure I is appropriate if the material is to be transported or stored at high ground elevations or transported by air in its shipping/storage configuration.

Accredited [IEC17025]: NO Execution time [day]: 2 Unit price #1 [EUR]: 2,800.00 Limit #1: Applicable to any sample

#### 507.6 - Humidity - Procedure II - Aggravated Cycle

(id 2166) The purpose of the aggravated test procedure is to produce representative effects that typically occur when materiel is exposed to elevated temperaturehumidity conditions. Accordingly, this procedure does not reproduce naturally occurring or service-induced temperature-humidity scenarios. It may induce problems that are indicative of long-term effects. Test item failures do not necessarily indicate failures in the real environment.

Accredited [IEC17025]: NO Execution time [day]: 11 Unit price #1 [EUR]: 2,800.00 Limit #1: Not relevant

#### 507.6 - Humidity - Procedure II: Aggravated Cycle

(id 2167) The purpose of this test method is to determine the resistance of material to the effects of a warm, humid atmosphere.

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 1,200.00 Limit #1: Applicable to any sample

#### 510.7 - Sand and dust - Procedure II: Blowing Sand

(id 2168) Sand (150 to 850 ?m particle size) procedure. This test is performed to evaluate the ability of materiel to be stored and operated in blowing sand conditions without degrading performance, effectiveness, reliability, and maintainability due to abrasion (erosion) or clogging effects of sharp-edged particles.

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1 [EUR]: 2,500.00 Limit #1: Applicable to any sample

#### 510.7 - Sand and dust - Procedure I Blowing Dust

(id 2872) Dust (< 150 ?m) procedure. This test is performed to evaluate the ability of materiel to resist the effects of dust that may obstruct openings, penetrate into cracks, crevices, bearings, and joints, and to evaluate the effectiveness of filters.

Accredited [IEC17025]: NO Execution time [day]: 2 Unit price #1: ask to office Limit #1: Non eseguibile

#### 506.6 - Rain - Procedure III Drip

(id 2873) Use this Method to evaluate materiel likely to be exposed to dripping water during storage, transit, or operation. Procedure III is appropriate when materiel is normally protected from rain but may be exposed to falling water from condensation or leakage from upper surfaces.

Accredited [IEC17025]: NO Execution time [day]: 2 Unit price #1 [EUR]: 400.00 Limit #1: both dimensions < 300 mm Unit price #2 [EUR]: 800.00 Limit #2: one dimension <= 300 mm and one dimension >= 300 mm Unit price #3 [EUR]: 1,200.00 Limit #3: both dimensions > 300 mm

#### 502.7 - Low Temperature - Procedure I

(id 3098) Use low temperature tests to obtain data to help evaluate effects of low temperature conditions on materiel safety, integrity, and performance during storage, operation, and manipulation.

Use Procedure I to investigate how low temperatures during storage affect materiel safety during and after storage, and performance after storage.

Accredited [IEC17025]: NO Execution time [day]: 3 Unit price #1 [EUR]: 600.00 Limit #1: Up to 4 hours Unit price #2 [EUR]: 1,000.00 Limit #2: Up to 24 hours Unit price #3 [EUR]: 1,500.00 Limit #3: Up to 72 hours

#### 512.6 - Immersion - Procedure I

(id 3099) The immersion test is performed to determine if materiel can withstand immersion or partial immersion in water (e.g., fording), and operate as required during or following immersion.

Procedure I (Immersion) primarily addresses leakage during immersion of encased materiel.

Accredited [IEC17025]: NO Execution time [dav]: 1 Unit price #1 [EUR]: 600.00 Limit #1: Dia 250mm - height 600mm - no pressure Limit #2: Dia 800mm - height 400mm - up to 1 bar (10 meter) Unit price #2 [EUR]: 1,000.00



# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



## MSL.03.01.0035:2013

(id 1811) Fuel Feed Hose Durability Testing

#### unref - Fuel Feed Hose Durability Test

(id 2433) This test procededure is to validate the fuel feed hose material, manufacturing method, connector specification and assembly methods. The fuel feed hose(s) are required to a number of pressure cycles between minimum and maximum operating pressures at an ambient temperature of that expected whilst in service.

Accredited [IEC17025]: NO Execution time [day]: 25
Unit price #1 [EUR]: 5,000.00 Limit #1: Applicable to any sample

## Multimatic ES-2001116-070200

(id 2013) TRANSMISSION OIL COOLING FLUID TRANSFER SYSTEM SPECIFICATIONS

#### 3.D.1 - Impulse Strength

(id 2773) This procedure combines pressure and thermal cycling of cooling system tubing and hose assemblies to provide durability confidence in the system.

Accredited [IEC17025]: NO Execution time [day]: 5
Unit price #1 [EUR]: 5,000.00 Limit #1: Applicable to any sample

### NAVISTAR DR 100159-06 30 October 2019

(id 1648) Electronic module and/or System Environmental Performance

#### 5.3 - Storage temperature test

(id 2097) ---

Accredited [IEC17025]: NO Execution time [day]: 3
Unit price #1 [EUR]: 1,200.00 Limit #1: Applicable to any sample

#### 5.5.1 - Thermal shock test (short)

(id 2098) ---

Accredited [IEC17025]: NO Execution time [day]: 7
Unit price #1 [EUR]: 2,200.00 Limit #1: Applicable to any sample

#### 5.4 - Thermal cycle test

(id 2099) ---

Accredited [IEC17025]: NO Execution time [day]: 30
Unit price #1 [EUR]: 3,800.00 Limit #1: Applicable to any sample

#### 5.6.1 - Humidity test (short)

(id 2100) ---

Accredited [IEC17025]: NO Execution time [day]: 7
Unit price #1 [EUR]: 1,800.00 Limit #1: Applicable to any sample

#### 6.1 - Dust and sand

(id 2101) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 8,000.00 Limit #1: Applicable to any sample

### NFPA 1901:2009

(id 1544) Standard for Automotive Fire Apparatus

#### 24.9 - SCBA or SCUBA Air Cylinder Fill Station

(id 2004) If SCBA and/or SCUBA air cylinders are to be filled from a fire apparatus—mounted air system, the fill station shall meet the requirements of 24.9.1.1 through 24.9.1.6

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 3,000.00 Limit #1: Applicable to any sample



# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



## NFPA 1901:2016

(id 2061) Standard for Automotive Fire Apparatus - This standard defines the requirements for new automotive fire apparatus and trailers designed to be used under emergency conditions to transport personnel and equipment and to support the suppression of fires and mitigation of other hazardous situations.

#### 23.9 - SCBA or SCUBA Air Cylinder Fill Station

(id 2854) If SCBA and/or SCUBA air cylinders are to be filled from a fire apparatus-mounted air system, the fill station shall meet the requirements of 23.9.1.1 through 23.9.1.6.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

## Nissan M0132:1994

(id 1761) Thermal cycle test methods for plastic parts.

#### 7 - Method 3 - Class 2 (former M4-C2)

(id 2327) Thermal cycle test method 3 - class 2 (former M4-C2) for plastic parts

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

## Nissan M0133:2002

(id 1762) Chemical resistance test methods for plastic parts.

#### 8.1 - Method 1 - Immersion

(id 2325) Method 1 - Immersion is applicable both to parts used as containers or equivalent and parts normally exposed to chemicals.

Accredited [IEC17025]: NO Execution time [day]: 7
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 8.3 - Method 4 - Dripping

(id 2326) Method 4 - Dripping is applicable to parts inside engine compartment.

Accredited [IEC17025]: NO Execution time [day]: 4
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

### PF.90004:2021 Rev.D

(id 1877) DIESEL EXHAUST FLUID - TANK ASSEMBLY

#### 6.3.5 - Tank Shell Burst

(id 2502) The DEF tank shell must resist burst during dynamic pressure loading, for the intended useful life of the vehicle. This test is required only for blowmolded tank.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,000.00 Limit #1: Applicable to any sample

#### 6.3.6 - Tank Assembly Burst

(id 2503) The DEF Tank Assembly must resist burst during dynamic pressure loading, for the intended useful life of the vehicle.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,000.00 Limit #1: Applicable to any sample

#### 7.2.2 - Non-Usable Volume in static conditions

(id 2504) The DEF Tank assembly shape shall minimize the unusable volume left in the tank which cannot be drained by DSM in static condition.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample



# PRICE LIST / Listino prezzi your value certified § T - LABORATORY TESTS / Prove di laboratorio

# QM ref: ALL 7.1.4 albarubens

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

TNDEX



#### 7.3 - DEF Tank Volume Characteristics

(id 2505) To verify main tank internal volume characteristics such as total internal volume, volume at shut off and vapor space volume compared to cad data.

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

#### 7.6.1 - DEF Tank Leak Test

(id 2506) The purpose of this test is to verify the static leak integrity of the DEF Tank Assembly.

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1 [EUR]: 160.00 Limit #1: Applicable to any sample

#### 6.4.7.1 - DSM and Quality Sensor functionality test

(id 2507) DSM functional verification on different parameters.

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1 [EUR]: 300.00 Limit #1: Applicable to any sample

#### 5.6 - Heat Resistance with DEF

(id 2629) The DEF tank assembly, including both internal, external components shall be able to withstand the loads and stresses induced by high temperature for extended periods of time during all vehicle conditions and environments for the intended life of the vehicle. The purpose of the heat resistance testing is to validate the DEF tank structure robustness and resistance to heat for extended periods of time within the target life of the vehicle ensuring DEF crystal formation does not negatively affect the integrity of the system.

Accredited [IEC17025]: NO Execution time [day]: 42 Unit price #1 [EUR]: 2,500.00 Limit #1: Applicable to any sample

#### 5.7 - DEF Temperature Soak-Swell test

(id 2630) The DEF tank assembly must resist deformation caused by DEF loading and temperatures extremes that degrade the functional integrity and dimensions at critical clearance locations, during the functional life of the vehicle.

Accredited [IEC17025]: NO Execution time [day]: 15 Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

#### 5.9 - Max Ice Formation

(id 2631) The DEF tank assembly, including both internal and external components, retention system (skid plate, straps), module cover, if any, and DEF filler pipe (if the inlet port is not on the top of the tank), shall be able to withstand the loads and stresses induced by ice formation during all vehicle conditions and environments for the intended life of the vehicle. The purpose of the max ice formation is to verify the DEF tank assembly structure robustness under extreme ice conditions within the target life of the vehicle.

Accredited [IEC17025]: NO Execution time [day]: 48 Unit price #1 [EUR]: 6,000.00 Limit #1: 36 days duration Unit price #2 [EUR]: 7,200.00 Limit #2: 48 days duration

#### 6.3.7 - Tank Drop requirement

(id 2632) The tank drop test is intended to evaluate a DEF tank shells structurally integrity. The test is intended to be used as development tool, during the process development, to ensure that the tank wall thickness and structural integrity is optimized. This test is required only for blowmolded tank.

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1 [EUR]: 400.00 Limit #1: Applicable to any sample

#### 9.3 - Climatic Cycles

(id 2633) The DEF tank assembly, including both internal, external components and retention system shall be able to withstand the loads and stresses induced by thermal cycling during all vehicle conditions and environments for the intended life of the vehicle. The purpose of the climatic cycle testing is to validate the DEF tank assembly durability for thermal cycling within the target life of the vehicle.

Accredited [IEC17025]: NO Execution time [day]: 40 Unit price #1 [EUR]: 7,200.00 Limit #1: Applicable to any sample

#### 9.4 - Pressure - Vacuum Cycling and Deflection

(id 2634) The purpose of the pressure/vacuum cycle test is to measure the ability of the DEF Tank Assembly to withstand the pressure/vacuum events that can occur within the life of a vehicle. The objective of the pressure/vacuum deflection measurements is to ensure the DEF Tank Assembly critical clearances are not compromised during pressure or vacuum events and to determine the amount of permanent deformation that occurs due to pressure and vacuum events.

Accredited [IEC17025]: NO Execution time [day]: 8 Unit price #1 [EUR]: 1,800.00 Limit #1: Applicable to any sample

#### 9.5 - Frozen Slosh Durability

(id 2635) The purpose of the slosh durability test is to prove the robustness of the DEF tank assembly and components when subjected to sloshing events that can occur within the life of a vehicle. This test is intended to test durability of internal components and tank fixation when subjected to slosh effects. The test sample size, amplitude, test duration, and acceleration factors can be adjusted to meet R95C90

reliability. In general, if the test profile/duration represents one useful life (i.e. 15yr, 150K miles), then twenty-four (24) samples would be needed to be tested in order to meet the reliability target. If the supplier elects to accelerate the testing, (through Test-To-Extended-Bogey) then the sample size can be reduced accordingly.

page 189/234



# § T - LABORATORY TESTS / Prove di laboratorio

PRICE LIST / Listino prezzi

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

#### INDEX

Accredited [IEC17025]: NO Execution time [day]: 5
Unit price #1 [EUR]: 1,000.00 Limit #1: Applicable to any sample

#### 7.4 - Measure DEF in tank

(id 2708) To quantify the I/mm and I/signal from DSM on DEF tank.

Accredited [IEC17025]: NO Execution time [day]: 10
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

### PF.90004:2024 Rev.D

(id 2168) DIESEL EXHAUST FLUID - TANK ASSEMBLY (STELLANTIS HARMONIZED)

#### 7.1 - Customer Refueling and Fill Quality

(id 2997) The DEF Tank Assembly as part of DEF System, shall meet the Customer Refueling and Fill Quality and Performance requirements under all commercial refilling conditions and vehicle duty cycles for the intended useful life of the vehicle.

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 7.3 - DEF Tank Volume Characteristics

(id 2998) To verify main tank internal volume characteristics such as total internal volume, volume at shut off and vapor space volume compared to cad data.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

#### 7.4 - Measure DEF in tank (Level ground)

(id 3027) To quantify the I/mm and I/signal from DSM on DEF tank.

Accredited [IEC17025]: NO Execution time [day]: 10
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

### PF.90083:2024 Rev. D

(id 2107) Fuel Tank assembly - Plastic

#### 8.3.1 - Permeation Test (non-NAFTA)

(id 2917) ---

Accredited [IEC17025]: NO Execution time [day]: 140 Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 6.2.15.1 - Soaking (Paragraph B)

(id 2918) ---

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 800.00 Limit #1: Very small sample
Unit price #2 [EUR]: 2,200.00 Limit #2: Medium tank

#### 6.3.9 - Soaking (Paragraph A)

(id 2919) ---

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 2,200.00 Limit #1: Applicable to any sample

#### 9.7.1 - Assembly Shake Vibration Durability - Generic Single Axis Profile

(id 3083) The fuel tank assembly, retention system, and all attached components must sustain all vehicle induced loading and maintain their structural integrity for the intended useful life of the vehicle. The objective of the shake/vibration durability test is to prove the robustness of the fuel tank assembly with respect to the vibration events that can occur within the target useful life of the vehicle.



# PRICE LIST / Listino prezzi § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

#### 7.12.1 - Fuel Measurement Test

(id 3091) ---

Accredited [IEC17025]: NO Execution time [day]: 29

Unit price #1 [EUR]: 1,600.00 Limit #1: Applicable to any sample

#### 7.9.2 - Water Leak Test

(id 3092) The Fuel Tank Assembly shall meet a water leak test, using an ultrasonic leak test method or an alternate method approved by Fuel Systems Engineering whenever helium leak testing is not required or is not appropriate.

For tank assemblies made by STL\_MAKE in production, the cycle time allowed (using water leak test or ultrasonic leak test in water tank) shall be 25-60 seconds, using a pressure of 15-35 kPa.

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 7.8 - Vacuum Leakage

(id 3093) The Fuel Tank Assembly shall meet this vacuum leakage specification for the intended useful life of the vehicle.

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

# Piaggio 7308:2005

(id 2071) Prova alle vibrazioni - Vibration test

#### **Unref. - Vibration test**

(id 2871) To establish the procedure for subjecting parts (or units) to the vibration test, using an electronically controlled vibrating bench, having fixed degrees of severity.

Accredited [IEC17025]: NO Execution time [day]: 2 Subcontract: BPS srl

Unit price #1 [EUR]: 2,300.00 Limit #1: Applicable to any sample

# Piaggio 7316:2009

(id 2070) Prova ai cicli rapidi di temperatura - Thermal Shock Test

#### **Unref. - Thermal Shock Test**

(id 2870) This test is intended to verify the ability of the sample to withstand 10 thermal shock cycles from -20 °C to +80 °C.

Accredited [IEC17025]: NO Execution time [day]: 3

Unit price #1 [EUR]: 1,400.00 Limit #1: Applicable to any sample

# Porsche PPV 4017

(id 1227) Corrosion tests

#### unref - Corrosion test

(id 1639) ---

Accredited [IEC17025]: NO Execution time [day]: 2 Unit price #1 [EUR]: 8,000.00 Limit #1: Up 12 weeks

# Porsche PTL 14100:2020-03

(id 1981) Automotive specification

#### 6.3.4 - Flow ageing

(id 2721) Thermal, chemical and machanical aging (cavitation) of components is achieved by means of coolant flow. The required conditions are binding and must be documented.

Accredited [IEC17025]: NO Execution time [dav]: 45 Unit price #1 [EUR]: 3,000.00

Limit #1: Chamber and fluid temperature <90°C Unit price #2 [EUR]: 4,000.00 Limit #2: Chamber or fluid temperature >90°C Unit price #3 [EUR]: 5,000.00 Limit #3: Chamber and fluid temperature both >90°C

page 191/234



# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



## Porsche PTL 8140 Februar 2017

(id 1699) TechnischeLieferbedingung Interieur Allgemeine Anforderungen an Bauteile und Halbzeugmaterialien Anforderungen und Prüfungen

#### 6/1.1 - Anforderungen und Prüfungen - Feuchtigkeitslagerung

(id 2138) Humidity Storage

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

## Procedura Ferrari DT.M.41 13-03-2009

(id 1203) Tubi flessibili per conduzione liquido di raffreddamento in materiale plastico e/o elastomeri termoplastici

# 7.2 - Resistenza agli shock termici - Resistance to thermal shock (id 1631) ---

### Procedura Ferrari DT.M.42 13-03-2009

(id 1204) Tuli flessibili per conduzione aria in materiale plastico e/o elastomeri termoplastici

#### 7.1 - Resistenza agli shock termici - Resistance to thermal shock

(id 1632) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 1,500.00 Limit #1: 5 cycles
Unit price #2 [EUR]: 2,400.00 Limit #2: 10 cycles
Unit price #3 [EUR]: 3,500.00 Limit #3: 20 cycles

# Proposal 31M/88/NP REV.1

(id 970) Albarubens proposal to TC 31M for 31M/88/NP (Future ISO/IEC 80079-41) - Qualification of insulating materials used to protect hot surfaces - 1st revision

#### x.3 - Thermal endurance to heat

(id 1184) Accelerated ageing test intended to verify the effects of high temperatures on insulating materials used to protect engine hot surfaces. The test is performed in 3 following cycles:

- 1) 168h with an increasing temperature from (Tmax- 350°C) to Tmax in daily steps of 50°C;
- 2) 168h permanently at Tmax;
- 3) 336h at T= Tmax -200°C.

Accredited [IEC17025]: NO Execution time [day]: 28
Unit price #1 [EUR]: 2,400.00 Limit #1: Not relevant

#### x.4 - Thermal endurance to cold

(id 1185) Test intended to verify the effects of low temperatures on insulating materials used to protect engine hot surfaces

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 800.00 Limit #1: Not relevant

#### x.5 - Impact test

(id 1186) Verification of the effects of mechanical impacts on insulating materials used to protect engine hot surfaces. The points of impact are the places considered to be the weakest, e.g., on the external parts which may be exposed to impacts.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 400.00 Limit #1: Not relevant



# § T - LABORATORY TESTS / Prove di laboratorio

PRICE LIST / Listino prezzi

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

TNDEX

# x.6 - Vibration test

(id 1187) Vibration test be performed according to IEC 60068-2-6 Environmental testing - Part 2-6: Tests - Test Fc: Vibration (sinusoidal)

Accredited [IEC17025]: NO Execution time [day]: 0 Subcontract: GESTLabs srl

Unit price #1 [EUR]: 2,500.00 Limit #1: Not relevant

#### x.7 - Maximum surface temperature determination

(id 1188) Determination of the maximum surface temperature of insulating materials used to protect engine hot surfaces by taking into account the presence of possible hot spots

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1 [EUR]: 3,500.00 Limit #1: Not relevant

#### x.8 - Hot-spots ignition effect

(id 1189) Verification of the non ignition of a high temperature explosive mixture around insulating materials used to protect engine hot surfaces, with the hot-spots directly exposed to the explosive atmosphere

Accredited [IEC17025]: NO Execution time [day]: 7 Unit price #1 [EUR]: 5,200.00 Limit #1: Not relevant

#### unref - Special set up for manifold

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 1,200.00 Limit #1: Applicable to any sample

## **PSA B 2261622**

(id 1902) Cooling tests

#### Unref. - PDT test

(id 2555) Cyclic pressure pulsation test for automotive pipes.

Accredited [IEC17025]: NO Execution time [day]: 3 Unit price #1 [EUR]: 4,400.00 Limit #1: Applicable to any sample

## PSA B21 5540:2015 rev.H

(id 1926) STN / NTS

GMP -INTERFACE ADAPTATION GMP/CAISSE / POWERTRAIN -

POWERTRAIN/ BODY INTERFACE ADAPTATION

CIRCUITS CARBURANT INTERVOLUMES / INTERVOLUMES FUEL SYSTEM

CANALISATIONS (ESSENCE-GAZOLE) / PIPES (PETROL-DIESEL)

#### 5.2.3.1 - Liquid fuel pipes ageing test

(id 2589) his test purpose is to validate an assembled fuel pipe with production definition liaisons (male connectors, female connectors, o'rings, pipes, coating, fixings, valves, sensors, regulators etc. and other delivered assembled

components) resistance to fuel and temperature ageing. The connectors and o'rings material (charge type, etc.) is a parameter to vary during tests in order to take into account all the possible configurations.

Accredited [IEC17025]: NO Execution time [day]: 42 Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 5.2.3.2 - Vapor stage pipes ageing test

(id 2590) This test is to validate an assembled fuel pipe fuel and temperature resistance in vapor stage with its connection to the production definition (endings, female connectors, o'rings, pipes, coating, fixings, valves, sensors, regulators,

etc. and other assembled delivered components). The connectors and o'rings material (charge type, etc.) is a parameter to vary during tests in order to take into account all the possible configurations.

Accredited [IEC17025]: NO Execution time [day]: 42 Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 5.5.3.W - Shot blasting test

(id 2677) This characterization is required for the non-protected pipe located under body, but if notified on functional drawing, this characterization could be required for the other pipes.

Accredited [IEC17025]: NO Execution time [dav]: 1 Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant





QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



## PSA B21 7130:2016 Rev.E

(id 1933) Specifications Concerning the Environment of Electrical and Electronic Equipment - Climatic Chemical Character

#### CL17 - Operation in dusty environment

(id 2599) This test is intended to verify the ability of the sample to operate in a dusty environment.

```
Accredited [IEC17025]: NO Execution time [day]: 3
Unit price #1 [EUR]: 1,500.00 Limit #1: 3 days
```

#### CL07c - Ice water thermal shocks - Spraying on 2 sets of electronics

(id 2600) This test is intended to verify che ability of the sample to withstand water thermal shocks.

## PSA B22 6122-A:2020

(id 1823) STN – COOLING CIRCUIT RUBBER AND PLASTIC PIPES

#### 4.7.3.2.2 - Slow Ageing Test - 0075

(id 2455) Pipes shall resist ageing under pressure.

Accredited [IEC17025]: NO	Execution time [day]: 42
Unit price #1 [EUR]: 3,000.00	Limit #1: Chamber and fluid temperature <90°C
Unit price #2 [EUR]: 4,000.00	Limit #2: Chamber or fluid temperature >90°C
Unit price #3 [EUR]: 5,000.00	Limit #3: Chamber and fluid temperature both >90°C

## PSA B22 6142 PEUGEOT-CITROEN

(id 1548) Tube cooling circuit

#### unref - Glicole resistance test

(id 2010) Ageing of the materials due to exposure to glicole

Accredited	[IEC17025]: NO	Execution time [day]: 42	
Unit price	#1 [EUR]: 3,000.00	Limit #1: Chamber and fluid temperature <90°C	
Unit price	#2 [EUR]: 4,000.00	Limit #2: Chamber or fluid temperature >90°C	
Unit price	#3 [EUR]: 5,000.00	Limit #3: Chamber and fluid temperature both >90	°C

# PSA B26-2130:2011

(id 2165) BI-DIRECTIONAL WINDSHIELD WASHER PUMP

#### 11.5 - Resistance to water sprayings

(id 2991) ---

Accredited [IEC17025]: NO	Execution time [day]: 3
Unit price #1 [EUR]: 1,000.00	Limit #1: Not relevant

## PSA B31 5210 rev. C

(id 1903) Thermal cycle on cooling circuit

#### **Unref. - Thermal cycling**

(id 2556) Thermal cycle on cooling circuit.

```
Accredited [IEC17025]: NO Execution time [day]: 20
Unit price #1 [EUR]: 2,200.00 Limit #1: Applicable to any sample
```



# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



## PSA B31 5220:2002

(id 1913) Fuel Lines - Low Pressure Circuit Only - Resistance to Fuels and Temperature

#### 5.1 - TEMPERATURE AGEING WITH FUEL FLOW IN PULSATED PRESSURE

(id 2592) ---

Accredited [IEC17025]: NO Execution time [day]: 42
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 5.2 - TEMPERATURE AGEING WITH FUEL FLOW AT ATMOSPHERIC PRESSURE

(id 2593) ---

Accredited [IEC17025]: NO Execution time [day]: 42
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 5.3 - AGEING IN FUEL IMMERSION

(id 2594) ---

Accredited [IEC17025]: NO Execution time [day]: 21
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

## PSA B31 5250-A:2008

(id 1825) CANALISATIONS CARBURANTS TENUE AU FEU

#### 8.2.1 - Resistance to fire (without pressure)

(id 2456) The purpose of this procedure is to assess the fire behavior of fuel system parts. This test procedure does not apply to fuel tanks. A pipe with or without a connection or any other part of the fuel circuit is filled with fuel and then exposed to flames obtained by the combustion of fuel contained in a container. The time before which a fuel leak appears is measured.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 8.2.2 - Resistance to fire (with pressure)

(id 2457) The purpose of this procedure is to assess the fire behavior of fuel system parts. This test procedure does not apply to fuel tanks. A pipe with or without a connection or any other part of the fuel circuit is filled with fuel and then exposed to flames obtained by the combustion of fuel contained in a container. The time before which a fuel leak appears is measured.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

### PSA B31 5530:2020

(id 1821) Cooling Circuit - Rubber and Plastic Pipes

#### 4.2 - PDT Test: Pressure / Displacement / Temperature

(id 2452) ---

Accredited [IEC17025]: NO Execution time [day]: 6
Unit price #1 [EUR]: 4,800.00 Limit #1: Applicable to any sample

### **PSA BP-054**

(id 2042) New external document

#### Unref. - Extreme high temperature

(id 2781) Extreme high temperature cycle

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample



# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



## PSA D10 5010:2021 - Ind.D

(id 1878) VULCANIZED RUBBERS - THERMOPLASTIC RUBBERS MATERIAL TESTS

#### 9.18 - Fuel Permeation

(id 2508) This test is intended to measure the permeation of a hose filled with test fuel (mixture made according to ISO 1817, Liquid 2: trimethylpentane, toluene, trimethyl-1-pentene, ethanol, methanol, water) and stored in a heating oven for 24h.

Accredited [IEC17025]: NO Execution time [day]: 5
Unit price #1 [EUR]: 600.00 Limit #1: Applicable to any sample

## PSA D47 1847:2003

(id 2064) Resistance to battery acid vapors of molded plastic parts

#### unref - Resistance to battery acid vapour

(id 2856) The purpose of this method is to describe a test making it possible to determine the action of acid vapors from battery on a molded plastic part, in used condition.

The material to be tested is subjected to the vapors of diluted sulfuric acid for a specific period of time, then is observed to assess any alteration in appearance and mechanical strength compared to the initial state.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 700.00 Limit #1: Applicable to any sample

## PV 1200:2004

(id 1798) Vehicle Parts - Testing of Resistance to Environmental Cycle Test (+80/-40) °C

#### 4.1 - Environmental Cycle Test

(id 2419) This Test Specification describes an environmental cycle test (elevated temperature/low temperature cycle) for testing units, e.g. vehicle parts in the engine compartment.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,200.00 Limit #1: 8 cycles

## **QEV 111 AI1YLD:2021**

(id 1872) Technical Component Requirement Specifications - Activated Charcoal Filter

#### 2.1.8.2.4.4.3 - China 6-7

(id 2485) Standard: GB 18352.6-2016 Appendix K (for fuel) - China diurnal test profile (20...35...20 °C)

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 1,500.00 Limit #1: Applicable to any sample

#### 2.1.8.2.4.4.2 - ECE (09-2019)

(id 2486) ECE test fuel as per the regulation: ECE / Trans/ WP.29 / GRPE / 2018/4, Annex 2 - EU diurnal test profile (20...35...20 °C)

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 1,500.00 Limit #1: Applicable to any sample

# Renault 34 - 04 - 816 / - - N (2018)

(id 1542) FUEL SYSTEM PARTS CYCLED PRESSURE ENDURANCE TEST

#### 4.1. - Preliminary aging

(id 1996) ---

Accredited [IEC17025]: NO Execution time [day]: 9
Unit price #1 [EUR]: 1,200.00 Limit #1: Applicable to any sample

#### 4.4.1 - Conditioning

(id 1997) The test cycles shall be performed at a relative pressure equal to 3 (± 1/6) Ps (where Ps=relative service pressure).

Accredited [IEC17025]: NO Execution time [day]: 3 Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

page 196/234



# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX

#### 4.4.2 - Preliminary test

(id 1998) ---

Accredited [IEC17025]: NO Execution time [day]: 3

Unit price #1 [EUR]: 1,000.00 Limit #1: Applicable to any sample

#### 4.4.3 - Endurance test

(id 1999) The relative pressure applied shall vary between the vapor pressure (P.V.) and the service pressure indicated into the design plan.

Accredited [IEC17025]: NO Execution time [day]: 4

Unit price #1 [EUR]: 1,500.00 Limit #1: Applicable to any sample

#### 4.5 - Test until relative non-burst limit pressure

(id 2000) ---

Accredited [IEC17025]: NO Execution time [day]: 5

Unit price #1 [EUR]: 800.00 Limit #1: Applicable to any sample

#### 4.6 - Determination of the burst pressure

(id 2001) ---

Accredited [IEC17025]: NO Execution time [day]: 3

Unit price #1 [EUR]: 800.00 Limit #1: Not relevant

#### 4.4.4 - Final test

(id 2002) ---

Accredited [IEC17025]: NO Execution time [day]: 3

Unit price #1 [EUR]: 600.00 Limit #1: Applicable to any sample

# Renault 34 -04 -831 /-- K (2018)

(id 1543) ACCELERATED AGEING FOR FUEL CIRCUIT PARTS - TANK AND ASSEMBLED PARTS - FUEL LINES

#### 5.2 - Ageing of fuel lines (liquid and vapor)

(id 2003) ---

Accredited [IEC17025]: NO Execution time [day]: 75

Unit price #1 [EUR]: 4,500.00 Limit #1: 40 cycles, 45h each, 1800 h

Unit price #2 [EUR]: 3,200.00 Limit #2: duration 1000 h

#### 5.2 - Ageing of fuel lines - modified for H2

(id 2761) ---

Accredited [IEC17025]: NO Execution time [day]: 75

Unit price #1 [EUR]: 6,000.00 Limit #1: 40 cycles, 45h each, 1800 h

## Renault 39-06-008:2019 Rev.H

(id 1751) HOSES FOR COOLANT CIRCUITS

#### 8.4.1 - Resistance of the internal connections to the cycled pressure

(id 2242) Validate that the internal connections resists to the conditions of the cycled pressure.

Accredited [IEC17025]: NO Execution time [day]: 10
Unit price #1 [EUR]: 3,200.00 Limit #1: Applicable to any sample

#### 8.4.2 - Resistance of the internal connections to temperature variations

(id 2243) Evaluate the resistance of the internal connections to the thermic conditions and to the cycled pressure

Accredited [IEC17025]: NO Execution time [day]: 7

Unit price #1 [EUR]: 3,800.00 Limit #1: Applicable to any sample



# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



# RENAULT NISSAN DESIGN SPECIFICATION RNDS-A-20020\_2.0 (2019-3-29

(id 1345) PLASTIC FUEL LINES SYSTEM (Part/module generic specifications, containing RNDS-A-20020 v1.0)

#### B - Aging test

(id 1858) ---

Accredited [IEC17025]: NO Execution time [day]: 75
Unit price #1 [EUR]: 2,000.00 Limit #1: Large quantity of small assembly samples
Unit price #2 [EUR]: 5,000.00 Limit #2: Applicable to any sample

#### C - Fuel resistance test

(id 1859) ---

#### E - Resistance to alternating pressure

(id 1860) ---

#### H - Start and Stop pressure test

(id 1861) ---

Accredited [IEC17025]: NO Execution time [day]: 15
Unit price #1 [EUR]: 4,500.00 Limit #1: up to 2600000 cycles

#### I - Start and Stop cold oscillating test

(id 1862) ---

Accredited [IEC17025]: NO Execution time [day]: 6
Unit price #1 [EUR]: 1,200.00 Limit #1: up to 40 000 cycles
Unit price #2 [EUR]: 2,200.00 Limit #2: up to 250000 cycles
Unit price #3 [EUR]: 4,500.00 Limit #3: up to 2600000 cycles

#### D - Thermal cycle test

(id 2224) Thermal cycle test between high and low temperature

Accredited [IEC17025]: NO Execution time [day]: 13
Unit price #1 [EUR]: 2,500.00 Limit #1: Applicable to any sample

# RENAULT PROCEDURE D'ESSAI 34 - 04 - 898 / - - C (2004)

(id 1414) CANALISATIONS POUR CIRCUITS DE CARBURANT TENUE AUX DEBATTEMENTS A FROID

#### unref - Tenue aux débattements à froid - Cold oscillating test

(id 2058) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 1,500.00 Limit #1: up to 40 000 cycles
Unit price #2 [EUR]: 6,000.00 Limit #2: up to 2600000

# RENAULT PRODUCT SPECIFICATION 34 - 04 - 892 / - - H (2015)

(id 1202) RENAULT PRODUCT SPECIFICATION - FUEL SYSTEM LINES

#### 6.2 - Resistance to fuel and temperature

(id 1929) The test is performed according to Test Procedure 34-04-831 (ageing), resistance to temperature, 40 cycles. This test is also used to check for any deterioration of the physical conditions indicted in other paragraphs. The test shall be conducted using a fuel that is representative in terms of rates of easter, alcohol or diester used for those destinations.

Accredited [IEC17025]: NO Execution time [day]: 3
Unit price #1 [EUR]: 5,000.00 Limit #1: Not relevant



# PRICE LIST / Listino prezzi

# § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

TNDEX

(id 1930) The test is performed according to the test method D47 1698 with fluid Np, the specimens are vacuum-dried at 70 °C until a constant weight is obtained

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1: ask to office Limit #1: Not relevant

6.3.2 - Resistance to peroxidised petrol (plastic lines)

#### 6.5 - Resistance to oscillations

(id 1931) The test is performed according to Test Procedure 34-04-898, on a new line representative of volumeproduction for 40 000 cycles

Accredited [IEC17025]: NO Execution time [day]: 0 Limit #1: up to 40000 cycles Unit price #1 [EUR]: 1,500.00

#### 6.6 - Resistance to coupling vibrations

(id 1932) The test is performed at 23 °C on each type of coupling at the end of the lines under the installation conditions described in 34-04-892 specification. The test is performed with fuel inside the line without any pressure .

Accredited [IEC17025]: NO Execution time [day]: 12 Unit price #1 [EUR]: 12,000.00 Limit #1: Not relevant

#### 6.8 - Occasional contact with vehicle fluids - Ambient temperature (23°C)

(id 1933) The test is performed according to the test method D47 1924, all the couplings at each end of thepipe(s) must be in contact with the fluids. The couplings at each end are blanked off.

Accredited [IEC17025]: NO Execution time [day]: 7 Limit #1: up to 24 h Unit price #1 [EUR]: 500.00 Unit price #2 [EUR]: 800.00 Limit #2: up to 48 h Unit price #3 [EUR]: 1,500.00 Limit #3: up to 7 days

#### 6.10 - Resistance to alternating pressure (fatigue)

(id 1934) The test is performed according to Test Procedure 34-04-816 on previously shaped lines or on atypical line as defined in Test Procedure 34-04-898. The relations between working pressure, proofpressure and non-bursting pressure are defined in Standard ISO 7751.

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1: ask to office Limit #1: Not relevant

#### 6.17 - Resistance to fire

(id 1935) The test is performed according to Test Procedure 34-04-875

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1 [EUR]: 2,000.00 Limit #1: Not relevant

#### 6.8 - Occasional contact with vehicle fluids - High temperature

(id 1939) The test is performed according to the test method D47 1924, all the couplings at each end of the pipe(s) must be in contact with the fluids. The couplings at each end are blanked off.

Accredited [IEC17025]: NO Execution time [day]: 7 Unit price #1 [EUR]: 800.00 Limit #1: up to 24h Unit price #2 [EUR]: 1,200.00 Limit #2: up to 48h Unit price #3 [EUR]: 2,000.00 Limit #3: up to 7 days

# Renault product specification 36 - 00 - 802 / - - P

(id 1157) Renault product specification 36 - 00 - 802 / - - P

#### CH/01 - Solid tightness - Dust test

(id 1579) ---

Accredited [IEC17025]: NO Execution time [day]: 2 Limit #1: Not relevant Unit price #1 [EUR]: 1,800.00

#### CH/04 - Liquid tightness - High-pressure test (Kärcher Test)

(id 1580) ---

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1 [EUR]: 2,200.00 Limit #1: Not relevant



# PRICE LIST / Listino prezzi

# § T - LABORATORY TESTS / Prove di laboratorio

albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

TNDEX



# RNDS-A-20020 v3.0 (2020-12-16)

(id 1845) Plastic Fuel Lines System

#### B - Ageing test

(id 2469) This procedure defines an accelerated ageing of the parts including fluctuation of temperature under constant pressure and fuel circulation.

```
Execution time [day]: 75
Accredited [IEC17025]: NO
                                            Limit #1: Large quantity of small assembly samples
Unit price #1 [EUR]: 2,500.00
Unit price #2 [EUR]: 5,000.00
                                            Limit #2: Applicable to any sample
```

#### C - Fuel resistance test

(id 2470) This procedure defines a test evaluating the compatibility of the fuel line material regarding different fuel composition for compatibility testing.

```
Accredited [IEC17025]: NO
                                 Execution time [day]: 10
Unit price #1 [EUR]: 1,800.00
                                            Limit #1: Static method (240h)
Unit price #2 [EUR]: 2,500.00
                                            Limit #2: Cyclic method (720h)
```

#### D - Thermal cycle test

(id 2471) Thermal cycle test between high and low temperature shall be carried out.

```
Accredited [IEC17025]: NO
                                  Execution time [day]: 12
Unit price #1 [EUR]: 2,500.00
                                            Limit #1: Applicable to any sample
```

# RNDS-A-20020 4.0

(id 2065) Plastic Fuel Lines System

#### B - Ageing test

(id 2864) This procedure defines an accelerated ageing of the parts including fluctuation of temperature under constant pressure and fuel circulation.

```
Accredited [IEC17025]: NO
                                 Execution time [day]: 75
Unit price #1 [EUR]: 3,800.00
                                            Limit #1: Large quantity of small assembly samples
Unit price #2 [EUR]: 5,000.00
                                            Limit #2: Applicable to any sample
```

#### C - Fuel resistance test

(id 2865) This procedure defines a test evaluating the compatibility of the fuel line material regarding different fuel composition for compatibility testing.

```
Accredited [IEC17025]: NO
                                  Execution time [day]: 30
Unit price #1 [EUR]: 1,800.00
                                            Limit #1: Static method (240h)
Unit price #2 [EUR]: 2,500.00
                                            Limit #2: Cyclic method (720h)
```

#### D - Thermal cycle test

(id 2866) Thermal cycle test between high and low temperature shall be carried out.

```
Accredited [IEC17025]: NO
                                 Execution time [day]: 12
Unit price #1 [EUR]: 2,500.00
                                            Limit #1: Applicable to any sample
```

## RTCA/DO-160G:2010

(id 668) Environmental Conditions and Test Procedures for Airborne Equipment

#### 12.4 - Dust

(id 1522) This test determines the resistance of the equipment to the effects of blowing dust where carried by air movement at moderate speeds.

```
Accredited [IEC17025]: NO
                                  Execution time [day]: 0
Unit price #1 [EUR]: 3.000.00
                                            Limit #1: Applicable to any sample
```

#### 12.5 - Sand

(id 1523) This test determines the resistance of the equipment to the effects of blowing sand where carried by air movement at moderate speeds.

```
Accredited [IEC17025]: NO
                                 Execution time [day]: 0
Unit price #1: ask to office
                                            Limit #1: Not relevant
```

#### 10.3.2 - Drip proof test

(id 1524) This test determines whether the equipment can withstand the effects of liquid water falling on the equipment

```
Accredited [IEC17025]: NO
                                 Execution time [day]: 0
Unit price #1 [EUR]: 600.00
                                           Limit #1: not relevant
```

page 200/234



# § T - LABORATORY TESTS / Prove di laboratorio

PRICE LIST / Listino prezzi

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



#### 11.4.1 - Spray test

(id 1525) This test determines whether the materials used in the construction of the equipment can withstand the deleterious effects of contaminants.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 2,200.00 Limit #1: Applicable to any sample

#### 9.6.2 - Non ignition test

(id 2076) This test specifies requirements and procedures for aircraft equipment that may come into contact with flammable fluids and vapors such as those specified herein. It also refers to normal and fault conditions that could occur in areas that are or may be subjected to flammable fluids and vapors during flight operations.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,500.00 Limit #1: Small sample

#### 10.3.4 - Continuous stream proof test

(id 2077) This test determines whether the equipment can withstand the effects of liquid water being sprayed on the equipment.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

#### 14.xxx - Salt Fog

(id 2078) IN FIERI

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

#### 10.3.3 - Spray Proof Test

(id 2628) ---

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 9.6.1 - Containment Test

(id 3015) The purpose of this procedure is to intentionally fill the case or enclosure under test with a volatile gas mixture then ignite it.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 9.6.3 - Components or Surface Temperature Test

(id 3016) ---

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 26.6.2 - Flammability Test - Vertical Bunsen Burner Test for Cabin and Cargo Compartment Materials

(id 3017) ---

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 26.6.3 - Flammability Test - Horizontal Bunsen Burner Test for Miscellaneous Materials

(id 3018) ---

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 26.6.4 - Flammability Test - 60-Degree Bunsen Burner Test for Electric Wire

(id 3019) ---

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

## SAE J 1737 : 2019

(id 2166) Test Procedure to Determine the Hydrocarbon Losses from Fuel Tubes, Hoses, Fittings, and Fuel Line Assemblies by Recirculation

#### 7.2 - Preconditioning Test Specimen

(id 2995) ---

Accredited [IEC17025]: YES Execution time [day]: 42
Unit price #1 [EUR]: 6,800.00 Limit #1: 1000h preconditioning



# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX

#### 7.10 - Permeation measurement

(id 2996) ---

Accredited [IEC17025]: NO Execution time [day]: 5

Unit price #1 [EUR]: 4,300.00 Limit #1: Applicable to any sample

#### 7.10 FID - Permeation measurement using FID

(id 3111) ---

Accredited [IEC17025]: NO Execution time [day]: 5

Unit price #1 [EUR]: 4,500.00 Limit #1: Not relevant

## SAE J1455:2006

(id 1963) Recommended Environmental Practices for Electronic Equipment Design in Heavy-Duty Vehicle Applications

#### 4.7 - Dust and Sand

(id 2666) Dust creates a harsh environment for chassis, under-hood and exterior-mounted devices, and can be a long-term problem in interior locations, such as under the dash and seats. Sand, primarily windblown, is an important environmental consideration for components mounted in the chassis, exterior and under-hood areas.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 2,200.00 Limit #1: Applicable to any sample

#### 4.8 - Gravel Bombardment

(id 2667) Bombardment by gravel is significant for chassis, lower engine and exterior-mounted electronic components. Gravel from unimproved roads, gravel from gravel haulers and highway salt dispersing equipment are just some of the sources of gravel damage.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,200.00 Limit #1: Applicable to any sample

## **SAE J1645 REV. OCT2011**

(id 1626) Surface vehicle recommended practice (R) Fuel Systems and Components - Electrostatic Charge Mitigation

#### 5.2.3 - Resistance tests for components (not installed in the vehicle)

(id 2051) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 500.00 Limit #1: Not relevant

#### A.4.4.2 - Ignition Energy

(id 2748) Spark discharges can ignite flammable air fuel mixtures if the discharge has usfficient energy.

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 1,800.00 Limit #1: Applicable to any sample with preconditioning

#### **SAE J2044 AUG2009**

(id 1440) Surface vehicle recommended practice - (R) Quick Connect coupling specification for liquid fuel and vapor/Emissions Systems

#### unref - Life cycle test: internal method - Variation of §7.5

(id 2059) ---

Accredited [IEC17025]: NO Execution time [day]: 15
Unit price #1 [EUR]: 4,000.00 Limit #1: Not relevant

## **SAE J2045 NOV2012**

(id 1439) Surface vehicle recommended practice - (R) Fuel Systems and Components - Electrostatic Charge Mitigation

#### 4.1 - Leak Tightness

(id 2440) A device capable of applying the recommended internal pressure specified for both liquid fuel and fuel vapor line assemblies. Test is intended to be performed on liquid fuel/fuel vapor assemblies that duplicate the design intended for vehicle application, including applicable end fittings and/or connections.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

page 202/234



albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



## **SAE J2260 REV. NOV2004**

(id 1268) Surface vehicle standard - Nonmetallic Fuel System Tubing with One or More Layers

#### 7.6.5.1.1 - Recirculation method for "C" Test Fluid

(id 2978) ---

Accredited [IEC17025]: NO Execution time [day]: 42
Unit price #1 [EUR]: 3,800.00 Limit #1: Applicable to any sample

#### 7.6.5.2.2 - Recirculation method for CE-10 Test Fluid

(id 2979) ---

Accredited [IEC17025]: NO Execution time [day]: 209
Unit price #1 [EUR]: 15,000.00 Limit #1: 5000h

#### 7.7 - Methanol Resistance

(id 2980) ---

Accredited [IEC17025]: NO Execution time [day]: 42
Unit price #1 [EUR]: 3,800.00 Limit #1: Applicable to any sample

### **SAE J2334 REV. DEC2003**

(id 1328) Surface vehicle standard - Laboratory Cyclic Corrosion Test

#### unref - Cycling corrosion test

(id 2094) The SAE J2334 test procedure should be used when determining corrosion performance for a particular coating system, substrate, process, or design.

Accredited [IEC17025]: NO Execution time [day]: 150
Unit price #1 [EUR]: 12,000.00 Limit #1: Manual Operation - small sample only
Unit price #2 [EUR]: 15,000.00 Limit #2: Automatic Operation
Unit price #3 [EUR]: 8,000.00 Limit #3: Automatic Operation - very small sample only

### **SAE J2578:2014**

(id 1990) Recommended Practice for General Fuel Cell Vehicle Safety

#### Appendix F - Test method for drive-away protection

(id 2749) The selection of a force of 668N was based on the maximum allowable force for break-away at hydrogen dispensers. Such break-away devices are required to separate the fuel nozzle from the dispenser in the event of a drive-away. The selection of 668N is also consistent with a force that an adult can apply.

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 1,500.00 Limit #1: Applicable to any sample

### **SAE J2599:2003**

(id 1956) Fuel Filler Pipe Assembly Design Practice to Meet Low Evaporative Emission Requirements

#### 13 - Durability Test

(id 2645) Filler pipe assemblies must continue to function for the life of the vehicle and they will be subjected to evaporative emission scrutiny through vehicle In-Use testing after the vehicle has seen service for up to 15 years or 241 350 km. Metallic corrosion, polymeric material life testing, and mechanical durability testing are therefore in order

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

## **SAE J2600:2015**

(id 1759) Compressed Hydrogen Surface Vehicle Fueling Connection Devices

#### 4.2.4.5 - Materials of Construction

(id 2294) Materials of Construction - Materials used in the construction of nozzles, receptacles and dust caps shall be non-sparking or spark-reducing.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 100.00 Limit #1: Not relevant

page 203/234



# PRICE LIST / Listino prezzi

# § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

TNDEX

# 4.2.6.2 - Materials suitable and compatible with Hydrogen

(id 2295) Hydrogen - Nozzles and receptacles shall be manufactured of materials suitable and compatible for use with compressed hydrogen at the pressure and the temperature ranges to which it will be subjected.

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 100.00 Limit #1: Not relevant

#### 4.3.3 - Flow Requirements (Cv)

(id 2296) The minimum Cv value for the receptacle shall be 0.35 or as listed in the manufacturer's literature. Determination may be by test or analysis. Measuring method is to be documented with the Cv value. Refer to ANSI/ISA-75.02.01 or the EN 60534 series of documents for guidance.

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 100.00 Limit #1: Not relevant

#### 5.2 - User-Machine Interface Test

(id 2297) The objective of this test is to verify that the connection and disconnection forces and torques for only the nozzles in the unpressurized and pressurized states.

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 300.00 Limit #1: Not relevant

#### 5.3 - Valve Operating Handle Test

(id 2298) The objective of this test is to verify that nozzles with valve operating handles can withstand a maximum force without damage. This test is applicable to test nozzles whose design includes a valve operating handle.

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 300.00 Limit #1: Not relevant

#### 5.5 - Hydrogen Resistance Test

(id 2299) Verify sealing material resistance to damage from hydrogen diffusion and subsequent depressurization. This test is applicable to sealing materials in test nozzles or test receptacles.

Accredited [IEC17025]: NO Execution time [day]: 14 Unit price #1 [EUR]: 2,000.00 Limit #1: Not relevant

#### 5.6 - Room Temperature Leak Test

(id 2300) Verify leakage rate of nozzle, receptacle, connector, and receptacle check valve at room temperature. This test is applicable to test nozzles and test receptacles

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 300.00 Limit #1: Not relevant

#### 5.7 - Low and High Temperature Leak Test

(id 2301) Verify leakage rate and operation of nozzle, receptacle, and connector at low and high temperatures as stipulated in section 4.2.3.

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1 [EUR]: 800.00 Limit #1: Not relevant

#### 5.8 - Hydrostatic (or Ultimate) Strength Test

(id 2302) Verify that the nozzle, receptacle, and connector will not fail when subjected to pressures resulting in stresses near their Ultimate Strength. Because the devices are subjected to stresses near their Ultimate Strength, do not use the test samples for testing or any other pressurecontaining use after this test. This test is applicable to test nozzles and test receptacles.

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 5.9 - Drop Test

(id 2303) Verify that a nozzle can withstand a drop of 2 m under -40 °C conditions. This test is applicable to test nozzles.

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 5.10 - Receptacle Vibration Resistance Test

(id 2304) Verify receptacle and protective cap resistance to vibration.

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant



# PRICE LIST / Listino prezzi

# § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX

## 5.11 - Abnormal Loads Test

(id 2305) Verify that the nozzle can withstand abnormal loads in service, both axial and moments about the end fitting. This test is applicable to test nozzles.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 500.00 Limit #1: Not relevant

#### 5.12 - Durability Cycling Test

(id 2306) Verify that the nozzle and receptacle can withstand anticipated life-time cycling (100 000 and 15 000 cycles, respectively). This test is applicable to test nozzles and test receptacles.

Accredited [IEC17025]: NO Execution time [day]: 35
Unit price #1 [EUR]: 6,500.00 Limit #1: Not relevant

#### 5.13 - Corrosion Resistance Test

(id 2307) Verify nozzle and receptacle resistance to corrosion. This test is applicable to test nozzles and test receptacles.

Accredited [IEC17025]: NO Execution time [day]: 4
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 5.14 - Deformation of Field Connected/Assembled Parts Test

(id 2308) Verify that field connected/assembled parts can withstand a specified installation over-torque. This test is applicable to the nozzle with a representative hose attached following the installation/maintenance manual procedures. This test is also applicable to the receptacle with a representative fuel system attached following the installation/maintenance manual procedures.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 500.00 Limit #1: Not relevant

#### 5.15 - Contamination Test

(id 2309) Verify that the nozzle and receptacle can withstand contamination. This test is applicable to test nozzles and test receptacles.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 800.00 Limit #1: Not relevant

#### 5.16 - Thermal Cycle Test

(id 2310) Verify that the nozzle and receptacle can withstand thermal cycling. This test is applicable to test nozzles and test receptacles.

Accredited [IEC17025]: NO Execution time [day]: 25
Unit price #1 [EUR]: 10,000.00 Limit #1: Not relevant

#### 5.17 - Pre-Cooled Hydrogen Exposure Test

(id 2311) Verify that the H35 and H70 hardware can withstand exposure to pre-cooled hydrogen during fuelling.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 5.18 - Misconnected Nozzle Test

(id 2312) The objective of this test is to verify that a misconnected Type C nozzle shall not flow gas. This misconnection is presumed to be due to foreign material collecting around a receptacle installed in a vehicle, preventing proper insertion of the receptacle into the nozzle. This test is applicable to only Type C nozzles.

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 5.19 - H70 Receptacle/Low Pressure Nozzle Compatibility Test

(id 2313) Verify that a lower operating pressure nozzle design is tolerant and compatible with the H70 seal at a hydrogen mass flow rate of 60 g/s.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 5.20 - Rocking

(id 2314) Verify the receptacle/nozzle connection shall not be loosened or damaged when subjected to the following test.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 800.00 Limit #1: Not relevant



# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX

#### 5.4 - Oxygen Aging Test

(id 2315) Verify sealing material resistance to aging. This test is applicable to sealing materials in nozzles and/or receptacles being tested.

Accredited [IEC17025]: NO Execution time [day]: 4
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

## **SAE J30:2022**

(id 2170) SURFACE VEHICLE STANDARD - Fuel and Oil Hoses

#### 6.19 - Reservoir Method for Fuel Hose Permeation

(id 3012) ---

Accredited [IEC17025]: NO Execution time [day]: 14
Unit price #1 [EUR]: 4,300.00 Limit #1: Applicable to any sample

## **SAE J400:2002**

(id 1954) Test for Chip Resistance of Surface Coatings

#### 4.4.3 - Gravel Bombardment - Method C

(id 2646) This test is intended to evaluate the resistance of surface coating to chipping by gravel impact. The test is designed to reproduce the effect of gravel or other media striking exposed paint or coated surfaces of an automobile at ambient temperature.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,200.00 Limit #1: Only one impact area
Unit price #2 [EUR]: 1,500.00 Limit #2: Up to 5 impact areas

# SCANIA specification STD4471:2016

(id 966) Plastic pipes - Air brake system

#### 4.2 - Impact

(id 1486) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 200.00 Limit #1: Not relevant

#### 4.3 - Impact after accelerated heat ageing

(id 1487) ---

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 500.00 Limit #1: Not relevant

## SSCM00FUQR00075/750044-A-00-GD-0012

(id 1420) Smart subsea control module - SSCM qualification procedure

#### 15 - Minimum and maximum design temperature test

(id 1888) This test is to ensure the operation of the SSCM at a test temperature equal to or less than the minimum design temperature classification and at a test temperature equal to or greater than the maximum design temperature classification.

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1: ask to office Limit #1: Not relevant

## ST/SG/AC.10/11/Rev.7

(id 1896) United Nations: Manual of Tests and Criteria (Former: Recommendations on the Transport of Dangerous Goods)

#### 38.3 - Lithium metal and lithium ion batteries

(id 2533) This section presents the procedures to be followed for the classification of lithium metal and lithium ion cells and batteries (see UN Nos. 3090, 3091, 3480 and 3481, and the applicable special provisions of Chapter 3.3 of the Model Regulations).



# PRICE LIST / Listino prezzi § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX

Accredited [IEC17025]: NO
Unit price #1 [EUR]: 1,000.00

•

Limit #1: Not relevant

Execution time [day]: 2

# Stellantis draft FTIV test sequence

(id 1741) Proposal for the evaluation and qualification of FTIV valves

#### 1 - FTIV valve elegibility

(id 2217) The scope of the test is to exclude the FTIV valve that can put in corking the FLVV/ROV valve

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 200.00 Limit #1: Applicable to any sample

#### 2 - FTIV valve depressurization time

(id 2218) The scope of this test is to determine tank pressure depressurization time

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 500.00 Limit #1: Applicable to any sample

#### 3 - FTIV valve electric consumption

(id 2219) The scope of this test is to determine FTIV electric consumption (Wh) during refueling operation

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 200.00 Limit #1: Applicable to any sample

#### 4 - FTIV valve audible noise

(id 2220) Verify the noise during actuation in cold and hot conditions

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 500.00 Limit #1: Applicable to any sample

#### 5 - FTIV valve acceleration

(id 2221) Acceleration of the valve body along 3-axis

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 300.00 Limit #1: Applicable to any sample

## **Stellantis PF.90315:2022**

(id 2082) PLASTIC PIPES FOR ENGINE COOLANT LINES

#### 7.2 - Burst Test

(id 2874) ---

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 7.4 - High Temperature Burst Test

(id 2875) ---

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 9.3.2 - Pressure Cycling Test - Standard Duty Test Specification

(id 2876) ---

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

# Sumiriko proprietary test methods

(id 1745) New external document

#### unref - Fuel tank Conditioning with E22 fuel

(id 1533) The purpose of this test is to verify the compatibility of the fuel tank and component materials with applicable fuels.

Accredited [IEC17025]: NO Execution time [day]: 43
Unit price #1 [EUR]: 4,800.00 Limit #1: Not relevant

page 207/234



# § T - LABORATORY TESTS / Prove di laboratorio

PRICE LIST / Listino prezzi

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX

#### unref - Sumiriko internal test: Fuel conditioning test

(id 1578) ---

Accredited [IEC17025]: NO Execution time [day]: 24 Unit price #1 [EUR]: 1,800.00 Limit #1: Not relevant

#### unref - Sumiriko internal test: resistance to thermal cycles

(id 1611) --

Accredited [IEC17025]: NO	Execution time [day]: 0
Unit price #1 [EUR]: 1,200.00	Limit #1: 5 cycles
Unit price #2 [EUR]: 2,000.00	Limit #2: 10 cycles
Unit price #3 [EUR]: 3,200.00	Limit #3: 20 cycles

#### unref - SCR: thermal cycle durability test - Engine bay section

(id 1871) The heated SCR supply line assembly shall be able to withstand thermal stresses during all vehicle conditions and environment for the intended life of the vehicle.

Cycle engine bay line from -30°C to +160°C, temperature ramp shall be around 5°C/min, dwell at 160°C for 0.5 hours, dwell at -30°C for 0.5 hours with the heated SCR supply line assembly completely filled with SCR in vehicle position.

SCR supply line shall be maintained with internal pressure of 8.5 bar for test temperature greater than +10°C. During temperature decrease phase, while chamber temperature is crossing +10°C, SCR supply line internal pressure shall be set at ambient pressure.

SCR supply line internal pressure below +10°C has not to be imposed and will change according closed end condition and chamber temperature.

SCR Heaters shall be on and functioning only when chamber temperature is growing up from -30°C up to +80°C.

Heaters shall be deactivated during all the rest of the test.

Repeat for 1000 cycles.

Accredited [IEC17025]: NO Execution time [day]: 98
Unit price #1 [EUR]: 15,000.00 Limit #1: Applicable to any sample

#### 1.3 Rev.2 - SCR: thermal cycle durability test Underfloor and wheelhouse sections

(id 1872) The heated SCR supply line assembly shall be able to withstand thermal stresses during all vehicle conditions and environment for the intended life of the vehicle (thermal cycles between 80°C and -30°C).

SCR supply line shall be maintained filled with AdBlue in static conditions at an internal pressure of 8.5 bar for test temperatures greater than +10°C. During temperature decrease phase, while chamber temperature is crossing +10°C, SCR supply lines shall be partially emptied and internal pressure setted at ambient pressure.

SCR subbly line shall be refilled of AdBlue when chamber temperature is growing up from -30°C while chamber temperature is crossing +10°C.

SCR Heaters shall be on and functioning only when chamber temperature is growing up from -30°C.

Heaters shall be deactivated during all the rest of the test.

Accredited [IEC17025]: NO Execution time [day]: 84
Unit price #1 [EUR]: 12,000.00 Limit #1: Applicable to any sample

#### unref - Sumiriko internal specification named 80-80

(id 2203) Quick evaluation of protection by corrosion.

Accredited [IEC17025]: NO Execution time [day]: 6
Unit price #1 [EUR]: 1,500.00 Limit #1: Not relevant

#### unref - Sumiriko internal test: Fuel exposure preconditioning (SAEJ2260)

(id 2206) ---

#### Unref. - IMA Heat duration performance in cold storage 5.9C 160°C

(id 2458) The purpose of this test is to verify the performance of pipes during a 3000h test using AdBlue as a medium.

Accredited [IEC17025]: NO Execution time [day]: 125
Unit price #1 [EUR]: 36,000.00 Limit #1: Applicable to any sample

#### Unref. - Sumiriko internal test: Static ageing

(id 2606) 168 hours static ageing on fuel pipes at 130°C with gasoline as fluid.

Accredited [IEC17025]: NO Execution time [day]: 7
Unit price #1 [EUR]: 1,000.00 Limit #1: Applicable to any sample

#### Unref. - SRK - Pressure test

(id 2825) SRK pressure test on cap and capless systems.

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 600.00 Limit #1: Up to 20 liters of fuel at static temperature
Unit price #2 [EUR]: 700.00 Limit #2: Up to 50 liters of fuel at static temperature
Unit price #3 [EUR]: 1,000.00 Limit #3: Up to 100 liters of fuel or thermal cycle

page 208/234



# PRICE LIST / Listino prezzi

# § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



## TL 226:2018

(id 1795) Paintwork on Materials of Vehicle Interior Equipment

#### Table 3 - 2.1 - Cross-cut test

(id 2382) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 150.00 Limit #1: Not relevant

#### Table 3 - 2.2 - Cross-cut test (St. Andrew's cross)

(id 2383) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 150.00 Limit #1: Not relevant

#### Table 3 - 3 - Scratch resistance

(id 2384) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 100.00 Limit #1: Not relevant

#### Table 3 - 4.1 - Heat resistance in the heat cabinet (recirculated air with closed throttle valve)

(id 2385) Thermal stability and environmental cycle resistance

#### Table 3 - 4.2 - Environmental cycle test

(id 2386) This test, as per PV 1200, describes an environmental cycle test (elevated temperature/low temperature cycle) for testing units, e.g. vehicle parts in the engine compartment.

Accredited [IEC17025]: NO Execution time [day]: 4
Unit price #1 [EUR]: 1,500.00 Limit #1: Not relevant

#### Table 3 - 5.1 - Condensation atmosphere with constant humidity

(id 2387) Weather resistance and aging resistance

Accredited [IEC17025]: NO Execution time [day]: 10
Unit price #1 [EUR]: 1,200.00 Limit #1: Not relevant

#### Table 3 - 5.2 - Lightfastness

(id 2388) Weather resistance and aging resistance - Testing as per PV 1303

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### Table 3 - 5.3 - Hydrolysis aging

(id 2389) Weather resistance and aging resistance

Accredited [IEC17025]: NO Execution time [day]: 3
Unit price #1 [EUR]: 600.00 Limit #1: Not relevant

#### Table 3 - 5.4 - Sunlight simulation

(id 2390) Weather resistance and aging resistance

Accredited [IEC17025]: NO Execution time [day]: 10 Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### Table 3 - 6.1.1 - 2 000 double strokes, dry

(id 2391) Resistance to chemicals and behavior on exposure to abrasion

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 800.00 Limit #1: Not relevant

#### Table 3 - 6.1.2 - 100 double strokes, dry

(id 2392) Resistance to chemicals and behavior on exposure to abrasion

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 200.00 Limit #1: Not relevant



# PRICE LIST / Listino prezzi

# § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX

## Table 3 - 6.1.3 - 100 double strokes, wet (distilled water)

(id 2393) Resistance to chemicals and behavior on exposure to abrasion

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 200.00 Limit #1: Not relevant

#### Table 3 - 6.1.4 - 10 double strokes with tenside solution

(id 2394) Resistance to chemicals and behavior on exposure to abrasion

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 100.00 Limit #1: Not relevant

#### Table 3 - 6.1.5 - 10 double strokes with ammoniacal alcoholic cleaning solution

(id 2395) Resistance to chemicals and behavior on exposure to abrasion

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 100.00 Limit #1: Not relevant

#### Table 3 - 6.1.6 - 10 double strokes with cleaner's naphtha

(id 2396) Resistance to chemicals and behavior on exposure to abrasion

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 100.00 Limit #1: Not relevant

#### Table 3 - 6.1.7 - 10 double strokes with methylated spirit

(id 2397) Resistance to chemicals and behavior on exposure to abrasion

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 100.00 Limit #1: Not relevant

#### Table 3 - 6.1.8 - 10 double strokes with synthetic perspiration solution A

(id 2398) Resistance to chemicals and behavior on exposure to abrasion

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 100.00 Limit #1: Not relevant

#### Table 3 - 6.1.9 - 10 double strokes with synthetic perspiration solution B

(id 2399) Resistance to chemicals and behavior on exposure to abrasion

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 100.00 Limit #1: Not relevant

#### Table 3 - 6.2 - Droplet test

(id 2400) Resistance to chemicals and behavior on exposure to abrasion

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 400.00 Limit #1: Not relevant

#### Table 3 - 6.3 - Cream resistance

(id 2401) ---

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

## TL 52231:2016

(id 1802) ABS and PC Polymer Blends - Materials Requirements

#### Table 1 - 2 - Flexural strength

(id 2402) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### Table 1 - 3 - Notched impact strength

(id 2403) ---

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 200.00 Limit #1: Not relevant



# PRICE LIST / Listino prezzi § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX

#### Table 1 - 4 - Notched impact strength

(id 2404) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 200.00 Limit #1: Not relevant

Table 1 - 5 - Impact strength

(id 2405) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 200.00 Limit #1: Not relevant

#### Table 1 - 6 - Dimensional stability under heat

(id 2406) ---

Accredited [IEC17025]: NO Execution time [day]: 2

Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### Table 1 - 7 - High-temperature behavior

(id 2407) ---

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 400.00 Limit #1: Not relevant

#### Table 1 - 8 - Low-temperature behavior

(id 2408) ---

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 400.00 Limit #1: Not relevant

#### Table 1 - 9 - Ball drop test

(id 2409) ---

Accredited [IEC17025]: NO Execution time [day]: 1

Unit price #1 [EUR]: 200.00 Limit #1: Applicable to any sample but withour preconditioning
Unit price #2 [EUR]: 600.00 Limit #2: Applicable to any sample with 24h preconditioning

## TL 52439: 2008-12

(id 980) Polyamide Pipe - Material Requirements

#### 6.4 - Equivalent stress

(id 1519) The equivalent stress is calculated according to the following formula:

 $?V = PB \times dm/20 \text{ s [N/mm2]}, \text{ where:}$ 

-PB = burst pressure, in bar

-dm = d1 - s, average pipe diameter

- d1 = outer diameter, measured

-s = pipe wall thickness

-?V = equivalent stress in radial direction

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 400.00 Limit #1: Not relevant

#### 6.5 - Low-temperature behavior

(id 1520) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 200.00 Limit #1: Not relevant

#### 6.7 - Aging resistance

(id 1521) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1 [EUR]: 700.00 Limit #1: Not relevant

page 211/234



# § T - LABORATORY TESTS / Prove di laboratorio

PRICE LIST / Listino prezzi

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



## TL 52680:2016

(id 1801) Components Made of Long-Fiber-Reinforced or Continuous-Filament-Reinforced Plastic Composites (FRP) - Materials and Component Requirements

#### 5.1.1 - Salt spray test

(id 2410) ---

Accredited [IEC17025	]: NO	Execution	time [	day]	: 4
Unit price #1 [EUR]:	3,200.00		Limit	#1:	96h
Unit price #2 [EUR]:	5,500.00		Limit	#2:	720h

#### 5.1.2 - Corrosion cycle test

(id 2411) ---

Accredited [IEC17025]: NO	Execution time [day]: 2
Unit price #1 [EUR]: 1,000.00	Limit #1: 10 cycles

#### 6.4 - 3-point bending

(id 2412) ---

Accredited [IEC17025]: NO	Execution time [day]: 0
Unit price #1 [EUR]: 1,000.00	Limit #1: Not relevant

#### 6.5 - Impact strength

(id 2413) ---

Accredited [IEC17025]: NO	Execution time [day]: 0
Unit price #1 [EUR]: 200.00	Limit #1: Not relevant

#### Table 5 - 7 - Penetration test

(id 2414) ---

Accredited [IEC17025]: NO	Execution time [day]: 0
Unit price #1 [EUR]: 100.00	Limit #1: Not relevant

#### 6.6 - High-temperature behavior

(id 2415) ---

Accredited [IEC17025]: NO	Execution time [day]: 1
Unit price #1 [EUR]: 400.00	Limit #1: Not relevant

#### 6.7 - Low-temperature behavior

(id 2416) ---

Accredited	[IEC17025]: NO	Execution time [day]: 1
Unit price	#1 [EUR]: 400.00	Limit #1: Not relevant

#### Table 6 - Environmental cycle test

(id 2417) This test, as per PV 1200, describes an environmental cycle test (elevated temperature/low temperature cycle) for testing units, e.g. vehicle parts in the engine compartment.

```
Accredited [IEC17025]: NO Execution time [day]: 4
Unit price #1 [EUR]: 1,500.00 Limit #1: 8 cycles
```

#### 5.8 - Evaluation of FRP surfaces

(id 2418) The surfaces must be evaluated before and after aging as per PV 1200, Audi parts, 20 cycles.

```
Accredited [IEC17025]: NO Execution time [day]: 4
Unit price #1 [EUR]: 1,800.00 Limit #1: 8 cycles
```

#### 5.9 - Evaluation of the paintwork

(id 2420) ---

```
Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant
```

#### 6.8 - Direct visual inspection without aids

(id 2421) Component surfaces made of an FRP are evaluated as per DIN EN 13018 by means of direct visual inspection (visual inspection without interrupted beam path between the eye and the inspection surface, but with aids).

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 100.00 Limit #1: Not relevant



# PRICE LIST / Listino prezzi § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX

#### Table A.1 - Ball drop test

(id 2422) This test determines the material behavior in a ball drop test at constant temperature (e.g., susceptibility to fracture and cracking at low temperatures) for components such as sheets, films, or trims made of organic materials.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 200.00 Limit #1: Applicable to any sample but withour preconditioning
Unit price #2 [EUR]: 600.00 Limit #2: Applicable to any sample with 24h preconditioning

#### Table A.1 - Scratch resistance

(id 2423) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 100.00 Limit #1: Not relevant

#### Table A.1 - Xenon arc light aging

(id 2424) ---

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 6.3 - Glass transition temperature

(id 2429) --

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

## ${\bf 6.3}$ - Glass transition temperature after the second heat-up cycle

(id 2430) ---

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

#### 6.3.1 - Storage modulus

(id 2431) ---

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

### TL 82253:2022

(id 1827) Fuel Lines - Functional requirements

#### 6.5.2 - Aging with fuel

(id 2461) ---

Accredited [IEC17025]: NO Execution time [day]: 7
Unit price #1 [EUR]: 2,200.00 Limit #1: Single pipe
Unit price #2 [EUR]: 3,000.00 Limit #2: Assembly pipes

### TL 82421:2021

(id 1826) Quick Couplings, Quick Coupling System - Functional Requirements

#### 7.10.3 - Preloading by aging in test fuel

(id 2460) ---

Accredited [IEC17025]: NO Execution time [day]: 4
Unit price #1 [EUR]: 800.00 Limit #1: 48h aging duration
Unit price #2 [EUR]: 1,500.00 Limit #2: 100h aging duration

# Toyota TSM1640G:2002 Rev.2

(id 1165) TSM1640G Toyota Engineering Standard

#### 20 - Fuel resistance test

(id 1598) ---

Accredited [IEC17025]: NO Execution time [day]: 10 Unit price #1 [EUR]: 800.00 Limit #1: Not relevant

page 213/234



# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



# Toyota TSM1640G:2020 Rev.5

(id 2000) Standard test method for material characteristic of plastic tube with connector for fuel line

#### 25 - Rotation Durability Test

(id 2766) ---

Accredited [IEC17025]: NO Execution time [day]: 12

Unit price #1 [EUR]: 4,000.00 Limit #1: Applicable to any sample

#### 26 - Flexing Durability test (Main Test)

(id 2853) ---

Accredited [IEC17025]: NO Execution time [day]: 5

Unit price #1 [EUR]: 1,800.00 Limit #1: up to 50000 cycles
Unit price #2 [EUR]: 2,200.00 Limit #2: up to 200000 cycles
Unit price #3 [EUR]: 4,500.00 Limit #3: up to 2000000 cycles

# TS026\_2\_24/01/2018

(id 994) Sumiriko Technical Specification: Battery Venting Line - Assembly Plastic Pipe - Blowby line

#### 7 - Fitting pull-off

(id 1529) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant

#### 8 - Burst test

(id 1530) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant

#### 9 - Cold impact test

(id 1531) ---

Accredited [IEC17025]: NO Execution time [day]: 0

Unit price #1: ask to office Limit #1: Not relevant

#### 6 - Leak test

(id 1532) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

## UL 1204:2018

(id 1876) Standard for safety - Parts Cleaners

#### 28.2 a) - Fluid Leakage Test (a)

(id 2495) A parts cleaner with an integral solvent container shall not exhibit leakage of the solvent after being tested. The product is to be conditioned and filled with liquid as indicated and observed for leakage.

Accredited [IEC17025]: NO Execution time [day]: 61
Unit price #1 [EUR]: 4,500.00 Limit #1: Applicable to any sample

#### 28.2 b) - Fluid Leakage Test (b)

(id 2496) A parts cleaner with an integral solvent container shall not exhibit leakage of the solvent after being tested. The product is to be conditioned and filled with liquid as indicated and observed for leakage.

Accredited [IEC17025]: NO Execution time [day]: 61
Unit price #1 [EUR]: 3,500.00 Limit #1: Applicable to any sample

#### 28.2 c) - Fluid Leakage Test (c)

(id 2497) A parts cleaner with an integral solvent container shall not exhibit leakage of the solvent after being tested. The product is to be conditioned and filled with liquid as indicated and observed for leakage.

Accredited [IEC17025]: NO Execution time [day]: 21

Unit price #1 [EUR]: 6,000.00 Limit #1: Applicable to any sample



# PRICE LIST / Listino prezzi

# § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

TNDEX



## UL 125:2020

(id 1955) Safety Flow Control Valves for Anhhydrous Ammonia and LP-Gas

#### 23 - Deformation Test

(id 2640) Joints in a valve shall not leak, nor shall there be evidence of loosening of joints, distortion, or other damage resulting from the stress imposed on pipethreaded sections when tested in accordance with these requirements.

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 200.00 Limit #1: Applicable to any sample

#### 24 - External Leakage Test

(id 2641) A valve in the "as-received" condition or after being subjected to the Deformation Test, shall be free from leakage through stem or body seals or other joints, and shall not show evidence of porosity in castings when tested at any aerostatic pressure between 0 and 1.5 times the rated service pressure of the valve.

Execution time [day]: 0 Accredited [IEC17025]: NO Limit #1: up to 1" Unit price #1 [EUR]: 200.00 Unit price #2 [EUR]: 250.00 Limit #2: greater than 1" up to 3"

#### 26.2 - Seat Leakage Test for primary shutoff valves

(id 2642) The following tests for seat leakage are to be conducted on samples which have previously been subjected to the Deformation test, par. 23 and External leakage test, par. 24.

Accredited [IEC17025]: NO Execution time [day]: 0 Unit price #1 [EUR]: 200.00 Limit #1: up to 1" Unit price #2 [EUR]: 250.00 Limit #2: greater than 1" up to 3"

#### 27 - Endurance Test

(id 2643) A shutoff valve shall be capable of complying with the applicable leakage requirements of 24.1 and 26.2, or 24.1 and 26.3, after being subjected to the number of cycles of opening and closing as specified in Table 27.1, for each type of valve and under the test conditions described in 27.2 - 27.7.

Accredited [IEC17025]: NO Execution time [day]: 2 Unit price #1 [EUR]: 300.00 Limit #1: up to 1" Unit price #2 [EUR]: 500.00 Limit #2: greater than 1" up to 3"

## UL 1973:2010

(id 2066) Batteries for Use in Stationary and Motive Auxiliary Power Applications

#### 8.3 - Crush test

(id 2863) ---

Accredited [IEC17025]: NO Execution time [day]: 2 Unit price #1 [EUR]: 800.00 Limit #1: Not relevant

## UL 1973:2018

(id 1434) Standard for Batteries for Use in Stationary, Vehicle Auxiliary Power and Light Electric Rail (LER) Applications

#### 38 - External fire exposure test

(id 1905) ---

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1 [EUR]: 8,000.00 Limit #1: Not relevant

### UL 1973:2022

(id 2089) Batteries for Use in Stationary and Motive Auxiliary Power Applications

#### - Resistance to Moisture Test

(id 2879) The purpose of this test is to determine that the battery system can safely withstand exposure to moisture anticipated in the end use.

Accredited [IEC17025]: NO Execution time [dav]: 2 Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant



# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



### UN 38.3 - 6th Ed.

(id 2161) UN Manual of Tests and Criteria, Sixth Revised Edition Sub-Section 38.3 - Lithium metal and lithium ion batteries

#### 38.3.4.2 - Test T2 - Thermal Rest

(id 2988) This test assesses cell and battery seal integrity and internal electrical connections. The test is conducted using rapid and extreme temperature changes.

Accredited [IEC17025]: NO Execution time [day]: 11
Unit price #1 [EUR]: 5,000.00 Limit #1: Up to 144h / 6 days
Unit price #2 [EUR]: 8,000.00 Limit #2: Up to 264h / 11 days

#### 38.3.4.1 - Test T.1 - Altitude simulation

(id 3064) This test simulates air transport under low-pressure conditions.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 500.00 Limit #1: small sample up to 400mm diameter, 800mm lenght
Unit price #2 [EUR]: 1,000.00 Limit #2: medium sample up to 680mm diameter, 2000mm lenght
Unit price #3 [EUR]: 2,200.00 Limit #3: big sample up to 3000mm diamiter

# UN Regulation No 154 E/ECE/TRANS/505/Rev.3/Add.153/Rev.3/Amend.1

(id 2195) Uniform provisions concerning the approval of light duty passenger and commercial vehicles with regards to criteria emissions, emissions of carbon dioxide and fuel consumption and/or the measurement of electric energy consumption and electric range (WLTP)

Addendum 153 – UN Regulation No. 154

Revision 3 - Amendment 1

Supplement 1 to the 03 series of amendments – Date of entry into force: 15 June 2024

ECE/TRANS/WP.29/2023/128

#### C3 - 5.1.1 - Ageing through exposure to temperature cycling

(id 2674) ---

Accredited [IEC17025]: NO Execution time [day]: 8
Unit price #1 [EUR]: 2,200.00 Limit #1: Applicable to any sample

#### C3 - 5.1.2 - Ageing through exposure to vibration

(id 2675) ---

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,500.00 Limit #1: Not relevant

### C3 - 5.1.3 - Ageing through exposure to fuel vapour and determining BWC300

(id 2676) Ageing shall consist of repeatedly loading with fuel vapour and purging with laboratory air.

Accredited [IEC17025]: NO Execution time [day]: 5
Unit price #1 [EUR]: 700.00 Limit #1: Applicable to any sample

#### C3 - 5.1.3 - Ageing through exposure to fuel vapour and determining GWC

(id 3081) Ageing shall consist of repeatedly loading with fuel vapour and purging with laboratory air.

Accredited [IEC17025]: NO Execution time [day]: 25
Unit price #1 [EUR]: 2,800.00 Limit #1: Applicable to any sample

# **UN Transport of dangerous goods**

(id 1167) ST/SG/AC.10/11/Rev.6

Recommendations on the transport of dangerous goods. Manual of Tests and Criteria. United Nations Publication, 2015

#### 16.6.1 - Test 6 (c): External fire (bonfire) test

(id 1607) This is a test performed on packages of an explosive substance or explosive articles, or unpackaged explosive articles, to determine whether there is a mass explosion or hazard from dangerous projections, radiant heat and/or violent burning when involved in a fire.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 2,200.00 Limit #1: Up to 50 liters



albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



# UNECE E/ECE/324/Rev.1/Add.43/Rev.3-E/ECE/TRANS/505/Rev.1/Add.43/Rev.

(id 1289) Agreement Concerning the Adoption of Uniform Technical Prescriptions for Wheeled Vehicles, Equipment and Parts which can be Fitted and/or be Used on Wheeled Vehicles and the Conditions for Reciprocal Recognition of Approvals Granted on the Basis of these Prescriptions\* - Regulation No. 44

Uniform provisions concerning the approval of restraining devices for child occupants of power-driven vehicles ("Child Restraint Systems").

### 8.1.1/Ann.4 - Corrosion test

(id 1812) ---

```
Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant
```

# **UNI EN ISO 16852:2017**

(id 442) Fermafiamma - Requisiti prestazionali, metodi di prova e limiti di utilizzo

### 7.3.2.1 - Deflagration test - End of line flame arrester

(id 1389) The end-of-line flame arrester under test shall be mountend on an enclosure and enclosed in a plastic bag.

The enclosure and the plastic bag shall be filled with an explosive test mixture as specified in § 6.8.2 ISO 16852:2016.

After disconnecting the mixture supply the test mixture shall be ignited separately at three positions inside the external bag: one at about 30cm from the flame arrester, one at about 130cm, and one at about 230cm.

Two tests for each position resulting in a total of six tests.

Test set up as for figure 1 of standard ISO 16852:2016.

```
Accredited [IEC17025]: YES

Unit price #1 [EUR]: 2,400.00

Unit price #2 [EUR]: 3,600.00

Unit price #3 [EUR]: 5,000.00

Execution time [day]: 3

Limit #1: up to 2 inches, standard ambient temperature (indoor location)

Limit #2: from 3 to 9 inches, standard ambient temperature (indoor location)

Limit #3: from 10 inches, standard ambient temperature (outdoor location, up to 500kg)
```

### 7.3.2.2 - Deflagration test - In line flame arrester

(id 1390) ---

```
Accredited [IEC17025]: YES
Unit price #1 [EUR]: 2,400.00
Unit price #2 [EUR]: 3,600.00
Unit price #3 [EUR]: 5,000.00
Limit #3: from10 inches, standard ambient temperature (indoor location)
Limit #3: from10 inches, standard ambient temperature (outdoor location, up to 500kg)
```

### 6.7, A.3 - Flow measurement (air)

(id 1391) Test to be performed before and after flame transmission tests

### 7.3.2.3 - Deflagration test - Pre volume flame arrester

(id 1865) An enclosure with a volume/shape as similar as possible to the original configuration simulates the pre-volume of the equipment under test.

Pre-volume applications using end-of-line types shall be mounted on the enclosure and enclosed in a plastic bag; pre-volume applications using in-line types shall

be connected to the actual pipe work or equipment on the protected side, or to pipe work simulating the actual length, diameter and volume. The enclosure and the plastic bag or pipe shall be filled with an explosive test mixture as specified in § 6.8.2 ISO 16852:2016.

After disconnecting the mixture supply the test mixture shall be ignited separately at three positions inside the enclosure: one as close as possible to the flame arrester, one at the most likely position of an ignition source and one as far away from the flame arrester as possible.

Two tests for each position resulting in a total of six tests.

If the enclosure has more than one outlet, all flame arresters shall be used and tested simultaneously.

```
Accredited [IEC17025]: YES Execution time [day]: 2
Unit price #1 [EUR]: 1,000.00 Limit #1: Small pre volume (up to 5 liters)
Unit price #2 [EUR]: 2,600.00 Limit #2: Medium pre volume (up to 15 liters)
Unit price #3 [EUR]: 5,000.00 Limit #3: Medium pre volume (over 15 liter)
```

### 7.3.3.2 - Detonation test: stable detonation without restriction

(id 2052) This test applies to detonation flame arresters tested for stable detonations without restriction, classified as Type 4.

Accredited [IEC17025]: YES	Execution time [day]: 3
Unit price #1 [EUR]: 2,800.00	Limit #1: up to 2 inches, standard ambient temperature (indoor location)
Unit price #2 [EUR]: 4,800.00	Limit #2: from 3 to 9 inches, standard ambient temperature (outdoor location)
Unit price #3 [EUR]: 5,600.00	Limit #3: from 10 inches, standard ambient temperature (outdoor location, up to 500kg)



# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX

# .

#### 7.3.5 - Endurance burning test

(id 2053) ---

### 6.7, A.2 - Flow measurement (air)

(id 2054) Test to be performed before and after flame transmission tests

#### 6.6 - Leakage test

(id 2092) ---

Accredited [IEC17025]: YES Execution time [day]: 0
Unit price #1 [EUR]: 500.00 Limit #1: Applicable to any sample

#### 6.5 - Pressure test

(id 2093) ---

Accredited [IEC17025]: YES Execution time [day]: 0
Unit price #1 [EUR]: 500.00 Limit #1: Applicable to any sample

### 7.3.3.2 a) - Deflagration tests after stable detonation without restriction (7.3.3.2)

(id 2197) 5 deflagration tests shall be carried out with a test set-up shall be in accordance with Figure 2, in addition to Stable detonation without restriction §7.3.3.2.

Accredited [IEC17025]: YES Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

## **UNI EN ISO 2812-2:2007**

(id 1842) Paints and varnishes — Determination of resistance to liquids — Part 2: Water immersion method

### 6.3 - Water Immersion

(id 2468) A production coated part shall resist immersed in water at 180°F for 10 days.

Accredited [IEC17025]: NO Execution time [day]: 10
Unit price #1 [EUR]: 1,000.00 Limit #1: Applicable to any sample

# UNI ISO 10263-3:2008

(id 1772) Earth-moving machinery - Operator enclosure environment - Part 3: Operator enclosure pressurization test method

### 6 - Pressurization test

(id 2347) This part of ISO 10263 specifies a test method which provides uniform measurement of the pressurization inside an operator enclosure of an earthmoving machine fitted with a device to pressurize the enclosure.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 1,200.00 Limit #1: Applicable to any sample on field

# UNI/PdR 90.2:2020

(id 1581) Maschere di comunità - Parte 2: Metodi di prova (Prassi di riferimento)

### APPENDICE A - Misurazione efficienza di rimozione particellare su mascherine di comunità semirigide

(id 2028) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

### APPENDICE B - Misurazione della resistenza al moto

(id 2029) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant



# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



### APPENDICE C - Rimozione della carica elettrostatica con vapori di alcool isopropilico

(id 2031) ---

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1: ask to office Limit #1: Not relevant

# **VDA AK**

(id 1914)

#### 1.4.2 - Salt Mist Test

(id 2580) Cyclic salt mist corrosion test specific for automotive.

Accredited [IEC17025]: NO Execution time [day]: 7
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

# VDA233-102

(id 1774) Mixed salt-mist and damp-heat test, specific for automotive.

### - Cyclic corrosion testing of materials and components in automotive construction

(id 1203) Mixed salt-mist and damp-heat test, specific for automotive.

Accredited [IEC17025]: NO Execution time [day]: 42
Unit price #1 [EUR]: 12,000.00 Limit #1: Big chamber
Unit price #2 [EUR]: 7,200.00 Limit #2: Small chamber

# **VDI 2440 November 2000**

(id 1695) Emission control - Mineral oil refineries - Guideline

### Annex - Leakage rate of moving parts

(id 2136) This test method is intended to determine the leakage rate at pressures, temperatures and spindle/shaft movements corresponding to the operating conditions of shut-off and control fittings

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 5,000.00 Limit #1: Not relevant

### Annex - Leakage rate of static parts

(id 2148) This test method is intended to determine the leakage rate at pressures, temperatures corresponding to the operating conditions of flanges or type of fittings without any moving part.

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 4,000.00 Limit #1: Not relevant

# Volkswagen Group standard PV 3905: 2015-04

(id 983) Organic Materials - Ball Drop Test

### 4 - Ball drop test

(id 2425) This standard describes the test for determining the material behavior in a ball drop test at constant temperature (e.g., susceptibility to fracture and cracking at low temperatures) for components such as sheets, films, or trims made of organic materials.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 200.00 Limit #1: Applicable to any sample but withour preconditioning
Unit price #2 [EUR]: 600.00 Limit #2: Applicable to any sample with 24h preconditioning

# Volkswagen LV 124 VW 80000: 2017-10

(id 1681) Electric and Electronic Components in Motor Vehicles up to 3,5 t - General Component Requirements, Test Conditions and Tests

### 11.6 - K-06 Salt spray test with operation, exterior

(id 2128) This test simulates the load on the component in a saline atmosphere with saline water, as it may occur in certain areas of the world or in winter road conditions.

It is meant to safeguard the component against any malfunctions when exposed to a salt load, e.g., from short circuits and leakage currents due to the ingress of salt into the component.

page 219/234



your value certified

# PRICE LIST / Listino prezzi

# § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

#### TNDEX



This test is to be applied to Components mounted in or on doors, hoods, and tailgates/trunk lids.

The test is carried out as per DIN EN 60068-2-11 with the parameters specified in Volkswagen LV 124 VW 80000: 2017-10.

The operating mode of the DUT is:

- During the spraying phase: 1 h "Off-grid parkingmin" and 1 h
- "OperationMax," intermittently
- During the resting phase: "Off-grid parkingMin".

```
Accredited [IEC17025]: NO
                                  Execution time [day]: 6
Unit price #1 [EUR]: 3,600.00
                                            Limit #1: Applicable to any sample
```

#### 10.3 - M-03 Dust test

(id 2129) This test simulates the dust load of the component during vehicle operation.

It is meant to verify the resistance of the component to electrical and mechanical error/flaw patterns.

The test is carried out as per ISO 20653 with the parameters specified in Volkswagen LV 124 VW 80000: 2017-10 and the degree of protection defined in the

The operating mode of the DUT is:

- For electrical/electronic components: "Operationmin"
- For mechatronic components (e.g., for components with rotating parts, controls): "Operationmax" and "Operationmin" intermittently as per Figure 25.

```
Accredited [IEC17025]: NO
                                 Execution time [day]: 1
Unit price #1 [EUR]: 600.00
                                            Limit #1: Standard IP chamber (small sample not powered)
                                            Limit #2: Medium IP chamber (small sample not powered)
Unit price #2 [EUR]: 1,200.00
Unit price #3 [EUR]: 2,200.00
                                            Limit #3: Powered on sample
```

### 11.9 - K-09 Damp heat, cyclic (with frost)

(id 2130) This test simulates the thermal loading (including frost) of the component by cyclic temperature changes with high humidity during vehicle operation. It is meant to verify the resistance of the components to damp heat.

The test is carried out as per DIN EN 60068-2-38 with the parameters specified in Volkswagen LV 124 VW 80000: 2017-10.

```
Accredited [IEC17025]: NO
                                  Execution time [day]: 11
Unit price #1 [EUR]: 2,000.00
                                            Limit #1: Not relevant
```

### 11.2 - K-02 Incremental temperature test

(id 2184) This test simulates the operation of the component at different ambient temperatures.

```
Accredited [IEC17025]: NO
                                 Execution time [day]: 2
Unit price #1 [EUR]: 3,600.00
                                            Limit #1: Applicable to any sample
```

## undef - Parameter test (small) for lighting

(id 2185) These checks are aimed at checking the correct functioning of the light after each test.

```
Accredited [IEC17025]: NO
                                 Execution time [day]: 0
Unit price #1 [EUR]: 300.00
                                            Limit #1: Applicable to any sample
```

### 11.5 - K-05 Temperature shock

(id 2188) This test simulates the thermal load of a component imposed through shock-type temperature changes during vehicle operation. The test serves to verify the resistance of the component to faults that occur due to thermal load, e.g. cracking in oldered joints, adhesive joints and welded joints, in bond connections as well as in seals or housings.

```
Accredited [IEC17025]: NO
                                 Execution time [day]: 0
Unit price #1 [EUR]: 3,800.00
                                            Limit #1: Small sample (up to 300mm) with dedicated technician
Unit price #2 [EUR]: 7,200.00
                                            Limit #2: Big or multiple sample with dedicated technician
```

### 11.12 - K-12 Temperature shock with splash water

(id 2194) This test simulates the load of the component when exposed to splash water as it occurs when driving through puddles. The test serves to verify functioning of the component when subjected to shock-type cooling by means of water.

```
Accredited [IEC17025]: NO
                                  Execution time [day]: 14
Unit price #1 [EUR]: 4,000.00
                                            Limit #1: Small sample
                                                                    (up to 40cm)
                                            Limit #2: Medium sample (up to 80cm)
Unit price #2 [EUR]: 5,000.00
Unit price #3 [EUR]: 7,000.00
                                            Limit #3: Big sample (up to 140cm)
```

### 11.10 - K-10 Water protection - IP X4K

(id 2196) This test simulates the load on the component when exposed to water. It is meant to verify the function of the component, e.g., when exposed to condensed water, rain, or splash water. This test is carried out as per ISO 20653, in order to verify the protection against harmful effects resulting from water splashing against the sample under test with increased pressure from any direction.

Accredited [IEC17025]: NO	Execution time [day]: 0
Unit price #1 [EUR]: 600.00	Limit #1: Applicable to any sample



# PRICE LIST / Listino prezzi your value certified § T - LABORATORY TESTS / Prove di laboratorio

# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

TNDEX



#### 10.2 - M-02 Stone impact test

(id 2352) This test simulates the mechanical load on the component due to stone impact. It is meant to verify the resistance of the component to flaw patterns, e.g., deformation or cracks.

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1 [EUR]: 7,000.00 Limit #1: up to 5 grouped samples

### 11.10 - K-10 Water protection - IP X6K

(id 2353) This test simulates the load on the component when exposed to water. It is meant to verify the function of the component, e.g., when exposed to condensed water, rain, or splash water.

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1 [EUR]: 400.00 Limit #1: Applicable to any sample Unit price #2 [EUR]: 1,000.00 Limit #2: up to 5 pipes contemporary

### 11.18 - K-18 Harmful gas test

(id 2354) This test simulates the influence of harmful gases on the component, primarily on its plug contacts and switches. It is meant to verify the resistance of the component to flaw patterns, such as corrosion and component damage

Accredited [IEC17025]: NO Execution time [day]: 21 Unit price #1 [EUR]: 3,200.00 Limit #1: Applicable to any sample

### 11.11 - K-11 High-pressure cleaning/pressure washing (IP X9K)

(id 2566) This test simulates the load on the component when subjected to water during vehicle cleaning. It is meant to verify the function of the component when exposed to high-pressure cleaning/pressure washing.

Accredited [IEC17025]: NO Execution time [day]: 1 Unit price #1 [EUR]: 1,200.00 Limit #1: small sample (automatic spraying) Unit price #2 [EUR]: 3,000.00 Limit #2: big sample (manual spraying)

# Volkswagen Technical Supply Specification TL 244:2017-06

(id 953) Zinc/Nickel Alloy Coatings Surface Protection Requirements

### 3.11.2 - Corrosion properties: test method as per PV 1209

(id 1480) ---

Accredited [IEC17025]: NO Execution time [day]: 2 Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

# Volkswagen Test Specification PV 1209: 2016-02

(id 954) Add-On Parts/Hang-On Parts with a Zinc or Zinc Alloy Coating and Aluminum - Environmental Corrosion Cycle Test

## 5.1 - Leak tightness test for aluminum add-on parts/hang-on parts that form cavities

(id 1478) ---

Accredited [IEC17025]: NO Execution time [day]: 2 Unit price #1 [EUR]: 120.00 Limit #1: Not relevant

## unref - Environmental corrosion cycle test as per Test Specification PV 1209

(id 1479) This test method is used to evaluate the corrosion resistance of components subjected to high corrosive loads. It applies particularly to components with a zinc-nickel coating with passivation and sealing, and to heat exchangers (e.g., condensers, radiators, and charge air coolers), lines, and other components made of aluminum alloys.

Accredited [IEC17025]: NO Execution time [dav]: 0 Unit price #1 [EUR]: 1,500.00 Limit #1: 1 week Unit price #2 [EUR]: 5,500.00 Limit #2: 4 week

# Volkswagen TL 934

(id 1783) Wiper Systems - Functional and Materials Requirements

### 6.2.6.2 - Wet operation endurance test

(id 2357) The test simulates the load on the whole wiper system over the service life.

Accredited [IEC17025]: NO Execution time [day]: 23 Unit price #1 [EUR]: 4,500.00 Limit #1: Applicable to any sample



# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



### 6.2.6.3 - Wipe angle constriction

(id 2358) The test simulates the load on the whole wiper system in winter operation (wipe angle constriction due to snow and ice deposits in the reversal points) over the service life.

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 1,000.00 Limit #1: Applicable to any sample

### 5.5 a - Corrosion test on wiper motor with linkage ASSY/rear wiper motor

(id 2453) This test serves to monitor and evaluate corrosion properties or corrosion protection solutions for these parts when exposed to static load. This test simulates the component load, e.g., in salty air or water containing de-icing salt, that can occur under wintry road conditions.

Accredited [IEC17025]: NO Execution time [day]: 21
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

### 5.5 c - Corrosion test on wiper lever ASSY/rear wiper lever

(id 2454) This test serves to monitor and evaluate corrosion properties or corrosion protection solutions for these parts when exposed to static load. This test simulates the component load, e.g., in salty air or water containing de-icing salt, that can occur under wintry road conditions.

Accredited [IEC17025]: NO Execution time [day]: 94
Unit price #1 [EUR]: 1,000.00 Limit #1: Not relevant

### 6.2.7.2 - High Temperature

(id 2517) This test simulates the operation of the component at higher ambient temperatures.

Accredited [IEC17025]: NO Execution time [day]: 4
Unit price #1 [EUR]: 1,500.00 Limit #1: Applicable to any sample

### 6.2.7.1 - Low Temperature

(id 2518) This test simulates the operation of the component at a low ambient temperature.

Accredited [IEC17025]: NO Execution time [day]: 2
Unit price #1 [EUR]: 1,800.00 Limit #1: Applicable to any sample

## 6.2.7.3 - Cold Impact Tests

(id 2532) The aim of this test is to check the mechanical strength of the wiper lever ASSY at low temperatures.

Accredited [IEC17025]: NO Execution time [day]: 1
Unit price #1 [EUR]: 400.00 Limit #1: Applicable to any sample

## Volto STD 423-0036:2008

(id 2219) Impact strength, falling-weight method - Organic materials

### Unref. - Impact strength

(id 3103) The purpose of this test method is to describe the operating conditions for evaluating the impact strength of plastics- or rubber-based materials by means of a falling weight. It applies to rubber parts, plastics parts, rubber- or plastics-coated textiles, floor coverings and linings, sound-absorbing materials, etc.

Accredited [IEC17025]: NO Execution time [day]: 0
Unit price #1 [EUR]: 200.00 Limit #1: Applicable to any sample

# Volvo Group 0001-14-8879

(id 1923) Frigoriferi TR

### 2.2.3.5 - Climate Cycling

(id 2586) The component must be able to survive and maintain its functionality under varying temperatures.

Accredited [IEC17025]: NO Execution time [day]: 14
Unit price #1 [EUR]: 4,200.00 Limit #1: Applicable to any sample

### 2.2.4.3 - Moisture Resistance - Ageing

(id 2587) The objective of this test is to ensure that humidity does not have a critical effect on the components in terms of deformation, changes in appearance, loss of functionality and mechanical properties.

Accredited [IEC17025]: NO Execution time [day]: 14
Unit price #1 [EUR]: 1,200.00 Limit #1: Applicable to any sample



# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version



# Volvo STD 20743840 index 09

(id 2218) Volvo specification

### 5.5.1 - Heat aging test

(id 3102) ---

Accredited [IEC17025]: NO Unit price #1 [EUR]: 3,600.00

Execution time [day]: 14

Limit #1: Applicable to any sample

### 5.5.2 - Cold temp test

(id 3106) ---

Accredited [IEC17025]: NO Unit price #1 [EUR]: 1,200.00 Execution time [day]: 3

Limit #1: Applicable to any sample

5.9.2 - Cold Temperature

(id 3109) ---

Accredited [IEC17025]: NO

Unit price #1 [EUR]: 1,500.00

Execution time [day]: 3

Limit #1: Applicable to any sample





QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



# § P

# **TESTS PACKAGES / Pacchetti di prove**

PRICES FOR PACKED TEST SEQUENCES, DISCOUNTED BECAUSE PERFORMED ALL TOGETHER

Prezzi a pacchetto per sequenze di prove, scontate perché eseguite tutte assieme

### Ex nR Ex t enclosure (IEC 60079-0:2017, EN60079-15:2010)

(id 132) Test sequence for standard Ex nR enclosure

Unit price #1: ask to office

Limit #1: ---

Included tests: all the ones required, ask for quotation including detailed list

# Adaptation to IEC 61215-2:2016 of PV modules already tested according to IEC 61215:2005

(id 99) Test procedure applied to PV modules for the confirmation of the materials

Price #1 [euro]: 4,800.00

T.imit #1: ---

Included tests: all the ones required, ask for quotation including detailed list

# Additional tests for J-box qualification according to EN 50548:2011/A2:2014 requirements

(id 105) ---

Price #1 [euro]: 4,500.00

Limit #1: ---

Included tests: all the ones required, ask for quotation including detailed list

# Determination of safety characteristics of Lithium batteries/battery modules

(id 143) ---

Unit price #1: ask to office

Limit #1: ---

Included tests: all the ones required, ask for quotation including detailed list

### Determination of safety characteristics of Lithium cells

(id 142) This test sequence is composed of internal test methods for the determination of safety characteristics of Lithium cells.

After a first phase of searching for the most suitable method to ignite the cell, an explosion due to the runaway thermal effect activation is produced, in order to analyse the thermal and pressure characteristics of the possible explosion, together with the volume of gases produced. This operation is conducted at first in normal air mixture, then in inert gas mixture (typically with N2), simulating the case in which the cell is placed in an enclosed case without air.

The purpose of these measurments is to evaluate the effect of the explosion on the battery container cnsidering the mechanical resistance of the enclosure and providing a basis for the calculation of a vent.

The explosion is then reproduced into an open air vessel in order to allow a video shooting of the phenomenon in order to show the esplosion extent and characteristics.

Unit price #1: ask to office

Limit #1: ---

Included tests: all the ones required, ask for quotation including detailed list

# Electrostatic hand-held spraying equipment for ignitable coating powder

(id 101) ---

Price #1 [euro]: 7,000.00

Limit #1: ---

Included tests: all the ones required, ask for quotation including detailed list

# End-of-line deflagration flame arrester without pre-volume, BC=c (no burn time), tested according t

Unit price #1: ask to office

Limit #1: ---


Included tests: all the ones required, ask for quotation including detailed list

page 224/234



albarubens

# PRICE LIST / Listino prezzi § P - TESTS PACKAGES / Pacchetti di prove



Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



# Equipment protection by pressurized room "pb" or "pc"

(id 103) ---

Unit price #1: ask to office Limit #1: ---

Included tests: all the ones required, ask for quotation including detailed list

## Ex d and Ex t flameproof enclosure, IP65 or 66 or 67, transparent windows

(id 65) Test sequence for standard Exd enclosure, both for ATEX and IECEx qualification.

Valid for groups I and II and standard IP.

Price #1 [euro]: 5,200.00 Limit #1: V<=40 [dm3]

Price #2 [euro]: 7,000.00 Limit #2: 40 < V < 200 [dm3] or L>1000mm

Price #3 [euro]: 12,000.00 Limit #3: V>=200 [dm3] or L>1500mm

# Included tests: all the ones required, ask for quotation including detailed list Ex d and Ex t rotating machine, IP65 or 66 or 67

(id 71) Test sequence for standard Exd rotating machine, both for ATEX and IECEx qualification. Valid for groups I and II and standard IP.

Price #1 [euro]: 5,200.00 Limit #1: Shaft height < 112. Price #2 [euro]: 8,000.00 Limit #2: Shaft height > 112.

Price #3 [euro]: 12,000.00 Limit #3: Extended Tamb= - 40°C + 70°C

Included tests: all the ones required, ask for quotation including detailed list

# Ex d and Ext flameproof enclosure, IP65 or 66 or 67, transparent windows (IEC 60079-0:2017)

(id 129) Test sequence for standard Exd enclosure (includes IEC 60079-0:2017)

Price #2 [euro]: 7,000.00 Limit #2: 40 < V < 200 [dm3] or L>1000mm Price #3 [euro]: 12,000.00 Limit #3: V>=200 [dm3] or L>1500mm

Included tests: all the ones required, ask for quotation including detailed list

# Ex d and Ext rotating machine, IP65 or 66 or 67, (IEC 60079-0:2017)

(id 130) Test sequence for standard Exd rotating machine, for Groups I and II and standard IP.

Price #1 [euro]: 5,200.00 Limit #1: Shaft height <112

Included tests: all the ones required, ask for quotation including detailed list

### Ex d enclosure Battery mounting system

(id 82) Additional tests for qualification of battery mounting system for inside Ex d enclosure

Price #1 [euro]: 2,500.00 Limit #1: ---

Included tests: all the ones required, ask for quotation including detailed list

# Ex de cable glands family with different sizes - not integral, threaded and not threaded (Group I only

(id 88) Qualification of cable glands. For sample preparation before tensile and mechanical strength tests and for the degree of protection IP, see the relevant clauses of Annex A (A.3) of IEC 60079-0 and Annex C (C.3) of 60079-1.

Price #1 [euro]: 2,200.00 Limit #1: Temperature -40°C +80°C Price #2 [euro]: 2,800.00 Limit #2: Temperature -60°C +180°C

Included tests: all the ones required, ask for quotation including detailed list

### Ex e and Ex to low voltage rotating electrical machines (Gas and dust)

(id 76) Test sequence for standard low voltage rotanting electrical machine (60079-7+Annex A). Valid for groups I and II and standard IP.

Price #1 [euro]: 3,800.00 Limit #1: < 2 liters
Price #2 [euro]: 6,000.00 Limit #2: > 2 liters

Included tests: all the ones required, ask for quotation including detailed list

## Ex e connection and junction boxes (metallic or non metallic for gas and dust)

(id 73) ---

Price #1 [euro]: 4,500.00 Limit #1: standard equipments

Price #2 [euro]: 6,000.00 Limit #2: big enclosures or temperature more than -40°C +100°C

Price #3 [euro]: 8,500.00 Limit #3: special equipments

Included tests: all the ones required, ask for quotation including detailed list

page 225/234



# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



# Ex e connection and junction boxes (metallic or non metallic for gas and dust, IEC 60079-0:2017)

(id 131) ---

Price #1 [euro]: 4,500.00 Limit #1: standard equipment
Price #2 [euro]: 6,000.00 Limit #2: big enclosures or temperature more than -40°C +100°C
Price #3 [euro]: 8,500.00 Limit #3: special equipments

Included tests: all the ones required, ask for quotation including detailed list

# Ex e qualification of battery pack >25Ah

(id 69) ---

Price #1 [euro]: 2,400.00 Limit #1: ---

Included tests: all the ones required, ask for quotation including detailed list

# Ex e resistance heating device intended for immersion

(id 75) Ex e resistance heating device or unit intended for immersion

Price #1 [euro]: 4,500.00 Limit #1: Service temperature up to 6 points
Price #2 [euro]: 6,500.00 Limit #2: Service temperature from 7 to 30 points
Price #3 [euro]: 8,000.00 Limit #3: Service temperature from 31 to 60 points

Included tests: all the ones required, ask for quotation including detailed list

# Ex eb/ec lead-acid battery elements qualification

(id 83) Qualification of battery elements only, for certfication as 'U' component

Price #1 [euro]: 4,000.00 Limit #1: Capacity <=100Ah
Price #2 [euro]: 5,500.00 Limit #2: Capacity 101-150Ah

Included tests: all the ones required, ask for quotation including detailed list

## Ex h non-electrical equipment for explosive atmospheres

(id 121) ---

Unit price #1: ask to office Limit #1: ---

Included tests: all the ones required, ask for quotation including detailed list

# Ex ia Ex ib Ex ic equipment with metallic enclosure (with IP/Annex F)

(id 59) ---

Price #1 [euro]: 4,500.00 Limit #1: not relevant

Included tests: all the ones required, ask for quotation including detailed list

# Ex ia Ex ib Ex ic equipment with plastic enclosure (with IP/Annex F)

(id 60) ---

Price #1 [euro]: 7,200.00 Limit #1: not relevant

Included tests: all the ones required, ask for quotation including detailed list

# Ex ia Ex ib equipment (without IP/Annex F/STA)

(id 93) This package includes all the tests usually required for the ATEX qualification of electronic equipment, qualified as "intrinsically safe" under the standard IEC/EN60079-0 + EN60079-11.<br/>br/>It doen't include thermal endurance and IP, because it's intended for real intrinsic safe, not asking for the application of Annex F of the -11 standard.<br/>br/>The ignition and the surface resistence tests are required only in case of plastic enclosures.

Price #1 [euro]: 2,200.00 Limit #1: ---

Included tests: all the ones required, ask for quotation including detailed list

### Ex ia/Ex ib batteries qualification under IEC 60079-11:2023

(id 72) Test sequence for primary/secondary batteries to be used inside intrinsic safe apparatus

Price #1 [euro]: 1,400.00 Limit #1: 4 Ah
Price #2 [euro]: 1,700.00 Limit #2: 25 Ah
Price #3 [euro]: 2,000.00 Limit #3: > 25 Ah

Included tests: all the ones required, ask for quotation including detailed list

page 226/234



# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



# Ex ic equipment with plastic enclosure

(id 168) ---

Price #1 [euro]: 7,000.00

imit #1: ---

Included tests: all the ones required, ask for quotation including detailed list

## Ex LPG filling systems

(id 102) LPG dispensing nozzles for automotive use: all the tests required by the specific harmonized technical standard for the assessment of conformity to the ATEX Directive

Price #1 [euro]: 8,500.00

Limit #1: ---

Included tests: all the ones required, ask for quotation including detailed list

## Ex nA enclosure (Group II only)

(id 100) Test sequence for

Ex nA enclosure (plastic or metallic)

Price #1 [euro]: 4,000.00

Limit #1: ---

Included tests: all the ones required, ask for quotation including detailed list

## Ex nA lamps

(id 109) ---

Price #1 [euro]: 4,300.00

Limit #1: ---

Included tests: all the ones required, ask for quotation including detailed list

### Ex nR Ex t enclosure

(id 74) Test sequence for standard Ex nR enclosure

Price #1 [euro]: 4,000.00

Limit #1: ---

Included tests: all the ones required, ask for quotation including detailed list

## Ex o - Equipment protection by liquid immersion (metallic enclosure)

(id 107) Levels of protection 'ob' and 'oc' intended for use in explosive gas atmosphere

Price #1 [euro]: 3,200.00

Limit #1: ---

Included tests: all the ones required, ask for quotation including detailed list

# Ex o - Equipment protection by liquid immersion (plastic enclosure)

(id 116) Levels of protection 'ob' and 'oc' intended for use in explosive gas atmosphere

Unit price #1: ask to office

Limit #1: ---

Included tests: all the ones required, ask for quotation including detailed list

### Ex q equipment

(id 115) ---

Price #1 [euro]: 5,000.00 Price #2 [euro]: 8,000.00 Price #3 [euro]: 14,000.00 Limit #1: ---

Limit #2: not relevant

Limit #3: not relevant

# Ex t (dust) qualification (EN 60079-0:2012)

(id 80) Test sequence for standard Ex t qualification (Group III only)

Price #1 [euro]: 5,000.00

Limit #1: ---

Included tests: all the ones required, ask for quotation including detailed list

Included tests: all the ones required, ask for quotation including detailed list

page 227/234





QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



# Ex t (dust) qualification (IEC 60079-0:2017+IEC 60079-31:2013)

(id 136) Test sequence for standard Ex t qualification (Group III)

Price #1 [euro]: 5,000.00 Limit #1: --

Included tests: all the ones required, ask for quotation including detailed list

# Extension to different gas of detectors for flammable gases according to IEC 60079-29-1

(id 151) ---

Price #1 [euro]: 1,800.00

Limit #1: ---

Included tests: all the ones required, ask for quotation including detailed list

### Fire test for quarter-turn valves and valves equipped with non metallic seats

(id 112) ---

Price #1 [euro]: 2,800.00 Limit #1: Up to 3"
Price #2 [euro]: 4,500.00 Limit #2: from 4" to 10"
Price #3 [euro]: 7,500.00 Limit #3: from 12" to 20"

Included tests: all the ones required, ask for quotation including detailed list

# Homologation of plastic fuel tank under Commission Delegated Regulation (EU) No 44/2014, Annex

Price #1 [euro]: 15,800.00 Limit #1: ---

Included tests: all the ones required, ask for quotation including detailed list

# In line detonation flame arresters tested for stable detonations without restriction, classified as Type

Unit price #1: ask to office Limit #1: ---

Included tests: all the ones required, ask for quotation including detailed list

### LV 124 VW 80000:2013-06 for SCR pipes Parallel Tests, Sequence A3

(id 126) Electric and electronic components in motor vehicles up to 3.5 t

Unit price #1: ask to office Limit #1: --

Included tests: all the ones required, ask for quotation including detailed list

### LV 124 VW 80000:2013-06 for SCR pipes Test sequence A2

(id 125) Electric and electronic components in motor vehicles up to 3.5 t

Price #1 [euro]: 82,000.00 Limit #1: ---

Included tests: all the ones required, ask for quotation including detailed list

## MINISHED permeation test (non-NAFTA, plastic tanks, 20 weeks) - PF.90083 Rev.D 2024 §8.3.1

(id 163) The following tests are for non-NAFTA market and are based on LP.7A005:2017 as indicated in the main specification PF.90083 Rev.D 2024, paragraph 8.3.1.

Price #1 [euro]: 18,000.00 Limit #1: ---

Included tests: all the ones required, ask for quotation including detailed list

## Non-static Ex p enclosure

(id 70) ---

Price #1 [euro]: 3,400.00 Limit #1: < 500 [dm3]
Price #2 [euro]: 6,400.00 Limit #2: 500 < 3000 [dm3]
Price #3 [euro]: 8,000.00 Limit #3: > 3000 [dm3]



# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



## Percentage power loss due to desert storm

(id 110) ---

Price #1 [euro]: 4,000.00 Limit #1: Severity 2h

Included tests: all the ones required, ask for quotation including detailed list

# Performance requirements of detectors for flammable gases according to IEC 60079-29-1

Unit price #1: ask to office

Limit #1: ---

Included tests: all the ones required, ask for quotation including detailed list

### Prequalification of filtering material for face masks

(id 141) <h4>GENERAL</h4>

This sequence of tests is intended to verify if a material is suitable for the manufacturing of face masks.
/p>This sequence of tests is intended to get quick info at low cost, in order to help manufacturers in selection of proper materials for the prototyping of the face masks, focusing the filtering efficiency and the breathability only.
/p>
<h4>TESTS</h4>

Absolute minimum is n. 4 pieces 150x150mm. Larger clothes are recommended.
Larger clothes are recommended.
Findividually packed, clearing marking the specific material type code on the package.
Larger clothes are recommended.

Price #1 [euro]: 1,200.00 Limit #1: Medical type mask
Price #2 [euro]: 1,500.00 Limit #2: Pre-formed type mask

Included tests: all the ones required, ask for quotation including detailed list

# PV - AMMONIA resistance (IEC 62716:2013 + IEC 61215-2:2016 + IEC 61730-2:2004) - crystalline PV

(id 90) Test sequence useful to determine the resistance of crystalline or thin film PV modules to ammonia (NH3). The tests are combined to provide means to evaluate possible faults caused in PV modules when operating under wet atmospheres having high concentration of dissolved ammonia (NH3).

Price #1 [euro]: 5,000.00 Limit #1: Size: 2000x1000 max

Weight: 100kg max; Power: 600W max; Voltage: 500V; Current: 30A
Price #2 [euro]: 13,000.00
Limit #2: Particular samples (example CIGS) and/or special cond:

Limit #2: Particular samples (example CIGS) and/or special conditions like: weight, size,

voltage

Included tests: all the ones required, ask for quotation including detailed list

# PV - DESERT STORM resistance (IEC60068-2-68:1996 + IEC 61215-2:2016 + IEC 61730-2:2004)

(id 89) ---

Price #3 [euro]: 15,000.00

Price #1 [euro]: 5,500.00 Limit #1: Severity 2 hours

Max: 2000x1000mm-100kg-600W-500V-30A- <25%rh-20m/s wind - 5g/m³ quartz sand Price #2 [euro]: 7.000.00 Limit #2: Severity 4 hours

Size: 2000x1200x600 max
Weight: 100kg max
Power: 600W max

Voltage: 500V Current: 30A Temperature 40°C

Relative humidity <25% Wind speed  $20\sim25\,\text{m/s}$  Sand Concentration 5 g /  $\text{m}^3$ 

Sand type: quartz

Limit #3: Particular samples (example CIGS) and/or special conditions like: weight, size,

Relative humidity <25% - Wind speed  $20\sim25\text{m/s}$  - Sand Concentration 5 g/m³ - Sand type: quart



# your value certified

# PRICE LIST / Listino prezzi

# § P - TESTS PACKAGES / Pacchetti di prove



QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

TNDEX



### PV - Fast track confirmation (IEC 61215:2016 + IEC 61730-2:2004)

(id 97) Test procedure applied to PV modules for the confirmation of the materials

Price #1 [euro]: 12,000.00 Limit #1: Size: 2800x2400x1400 max

Weight: 100kg max Power: 600W max Voltage: 500V Current: 30A

Included tests: all the ones required, ask for quotation including detailed list

## PV - Fire Safe test (IEC61730-2:2004)

(id 96) The two tests required for fire safe

Price #1 [euro]: 3,200.00

Included tests: all the ones required, ask for quotation including detailed list

# PV - Fire Safe tests according to IEC 61703-2:2016

(id 106) The two tests required for fire safe

Price #1 [euro]: 3,200.00

Limit #1: ---

Included tests: all the ones required, ask for quotation including detailed list

## PV - SALT MIST resistance (EN 61701:2012 + IEC 61215-2:2016 + IEC 61730-2:2004) - Severity 1

(id 91) Test sequence useful to determine the resistance of crystalline or thin-film PV modules to corrosion from salt mist containing CI- (NaCI, MgCl2, etc..).<br/>
The tests are combined to provide means to evaluate possible faults caused in PV modules when operating under wet atmospheres having high concentration of dissolved salt (NaCl).<br/>
The severity grade is 1.

Price #1 [euro]: 6,000.00

Limit #1: Severity 1 Size: 2000x1000 mm max Weight: 100kg max Power: 600W max Voltage: 500V Current: 30A

Included tests: all the ones required, ask for quotation including detailed list

# PV - SALT MIST resistance (EN 61701:2012+ IEC 61215-2:2016 + IEC 61730-2:2004) - Severity 3 or 4

(id 134) Test sequence useful to determine the resistance of crystalline or thin film PV modules to corrosion from salt mist. The tests are combined to provide means to evaluate possible faults caused in PV modules when operating under wet atmospheres having high concentration of dissolved salt (NaCl), Severity 3 or 4.

Price #1 [euro]: 4,000.00

Limit #1: Severity 3 Size: 2000x1000 mm max Weight: 100kg max Power: 600W max Voltage: 500V Current: 30A Limit #2: Severity 4 Size: 2000x1000 mm max

Price #2 [eurol: 5.000.00 Weight: 100kg max Power: 600W max Voltage: 500V

Included tests: all the ones required, ask for quotation including detailed list

# PV - SALT MIST resistance (EN 61701:2012+ IEC 61215-2:2016 + IEC 61730-2:2004) - Severity 5 or 6

(id 135) Test sequence useful to determine the resistance of crystalline or thin film PV modules to corrosion from salt mist. The tests are combined to provide means to evaluate possible faults caused in PV modules when operating under wet atmospheres having high concentration of dissolved salt (NaCI). Severity 5 or 6.

Price #1 [euro]: 6,000.00 Limit #1: Severity 5

Size: 2000x1000 mm max Weight: 100kg max Power: 600W max Voltage: 500V Current: 30A Limit #2: Severity 6

Price #2 [euro]: 8,000.00

Size: 2000x1000 mm max Weight: 100kg max Power: 600W max Voltage: 500V

Current: 30A

page 230/234

Included tests: all the ones required, ask for quotation including detailed list

Document automatically generated by the Albarubens WebApp - rev. on 30 Nov 2025 (22:05)





QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



## PV crystalline modules AMMONIA resistance test (IEC 62716:2013)

(id 18) Test sequence useful to determine the resistance of crystalline PV modules to ammonia (NH3). The tests are combined to provide means to evaluate possible faults caused in crystalline PV modules when operating under wet atmospheres having high concentration of dissolved ammonia (NH3).

Price #1 [euro]: 5,000.00 Limit #1: Size: 2000x1200x600 max

Weight: 100kg max; Power: 600W max; Voltage: 500V; Current: 30A

Included tests: all the ones required, ask for quotation including detailed list

## PV modules complete qualification (IEC 61215-2:2016 + IEC 61730-2:2004)

(id 85) Standard IEC 61215 "Crystalline silicon terrestrial photovoltaic (PV) modules - Design qualification and type approval" plus Standard IEC 61730-2 "Photovoltaic (PV) modules safety qualification-Part 2: Requirements for testing"

Price #1 [euro]: 22,000.00 Limit #1: Size: 2800x2400x1400 max

Weight: 100kg max Power: 600W max Voltage: 500V Current: 30A

Included tests: all the ones required, ask for quotation including detailed list

## PV PID (Potential Induced Degradation) Reliability Test

(id 127) Evaluation of potential induced degradation (PID) of PV modules with a voltage stress test in damp heat chamber (test method A).

Price #1 [euro]: 5,000.00 Limit #1: Test with both polarity

Max: 2000x1000mm - 100kg - 600W - 500V - 30A Price #2 [euro]: 3,500.00 Limit #2: Test su una polarità

Limit #2: Test su una polarità Size: 2800x2400x1400 max

Weight: 100kg max

Power: 600W max Voltage: 500V Current: 30A

Included tests: all the ones required, ask for quotation including detailed list

## PV Simplified retesting for frameless qualification (IEC 61215-2:2016 + IEC 61730-2)

(id 86) Point k) of DSH 0647A (Retest Guidelines for IEC 61215 and 61646) and DSH 770 (Retest Guidelines for IEC 61730 -1 -2):

Qualification of a frameless module after the design has received certification as a framed module when the frame is not part of the package seal and the superstrate is tempered glass.

Price #1 [euro]: 2,000.00 Limit #1: Size: 2800x2400x1400 max

Weight: 100kg max Power: 600W max Voltage: 500V Current: 30A

Included tests: all the ones required, ask for quotation including detailed list

### PV thin-film modules AMMONIA resistance test (IEC 62716:2013)

(id 19) Test sequence useful to determine the resistance of thin-film PV modules to ammonia (NH3). The tests are combined to provide means to evaluate possible faults caused in thin-film PV modules when operating under wet atmospheres having high concentration of dissolved ammonia (NH3). Severity 1. Severity (3) to (6) under request.

The duration of the test depends on the final light soaking.

Price #1 [euro]: 13,000.00 Limit #1: Size: 2000x1200x600 max

Weight: 100kg max Power: 600W max Voltage: 500V Current: 30A

Included tests: all the ones required, ask for quotation including detailed list

# Qualifica semplificata maschere facciali

Price #1 [euro]: 7,000.00 Limit #1: All types



# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



# Qualification of ball valves for gas installations for buildings, according to EN 331:2015

(id 162) This test sequence is intended to qualify performances and reliability of manually operated ball valves and closed bottom taper plug valves for gas installations for buildings. The sequence is suitable for the small size (up to 1"), the medium size (from 1" to 3") and the big size (beyond 3"). The reference standard is EN 331:2015.

N.B. In the sequence the test "Resistance to high temperatures", Annex A, is excluded.

```
Price #1 [euro]: 3,800.00 Limit #1: Size up to 1"

Price #2 [euro]: 5,000.00 Limit #2: Size from 1" to 3"

Price #3 [euro]: 6,600.00 Limit #3: Size beyond 3"
```

Included tests: all the ones required, ask for quotation including detailed list

# Qualification of end of line flame arrester through deflagration, according to ISO 16852:2016

(id 159) ---

```
Price #1 [euro]: 4,000.00 Limit #1: up to 2 inches, standard ambient temperature (indoor location)

Price #2 [euro]: 5,500.00 Limit #2: from 3 to 9 inches, standard ambient temperature (indoor location)

Price #3 [euro]: 7,500.00 Limit #3: from 10 inches, standard ambient temperature (outdoor location, up to 500kg)
```

Included tests: all the ones required, ask for quotation including detailed list

### Qualification of Ex eb electrical motors

(id 156) This tests sequence has the aim to check electrical motors for Group I M2. The sequence is composed by general tests (EN 60079-0) and specific tests (EN 60079-7 and 60034-5).

```
Price #1 [euro]: 5,000.00 Limit #1: ---
Price #2 [euro]: 8,500.00 Limit #2: not relevant
Price #3 [euro]: 12,500.00 Limit #3: not relevant
```

Included tests: all the ones required, ask for quotation including detailed list

### Qualification of Ex ec electrical motors

(id 149) This tests sequence has the aim to check electrical motors placed in zone 2. The sequence is composed by general tests (EN 60079-0) and specific tests (EN 60079-7 and 60034-5).

```
Price #1 [euro]: 5,000.00 Limit #1: ---
```

Included tests: all the ones required, ask for quotation including detailed list

## Qualification of Ex fixed gas sensor Group II under EN IEC 60079-29-1

(id 155) This test sequence is intended to qualify performances and reliability of an equipment that provides display indication, alarm functions, output contacts or alarm signal outputs or any combination when operated with gas detection transmitter(s). Exception made for fixed installations for which it is possible to skip the vibration test (5.4.12), the sample supplied shall be subjected to each test included in this sequence, which is prescribed by EN IEC 60079-29-1. The test sequence will be applied only on n.1 sample. As our laboratory does not carry out tests for 5.4.21 Electromagnetic compatibility and 5.4.23 Software validation, the tests sequence will be completed by the customer in order to achive full compliance to EN IEC 60079-29-1. These tests shall be excluded from our offer.

```
Price #1 [euro]: 18,600.00 Limit #1: ---
```

Included tests: all the ones required, ask for quotation including detailed list

## Qualification of Ex ia / Ex ib optical isolators (photocouplers), under IEC/EN 60079-11

(id 154) This test sequence is intended to qualify the optical isolators (photocouplers) for use as safety device inside the intrinsic safety barriers, under the standard IEC/EN 60079-11.

```
Price #1 [euro]: 3,500.00 Limit #1: All the types
```

Included tests: all the ones required, ask for quotation including detailed list

### Qualification of fixed SM gas sensors according to IEC 62990-1:2019

(id 170) This test sequence is intended to qualify fixed Safety Monitoring gas sensors according to IEC 62990-1:2019 with one gas only.

Price #1 [euro]: 20,600.00 Limit #1: ---





QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



# Qualification of flame arrester through detonation, according to EN ISO 16852:2016

(id 161) ---

Price #1 [euro]: 4,500.00 Limit #1: up to 2 inches, standard ambient temperature (indoor location)

Price #2 [euro]: 6,800.00 Limit #2: from 3 to 9 inches, standard ambient temperature

Price #3 [euro]: 8,000.00 Limit #3: from 10 inches, standard ambient temperature (outdoor location, up to 500kg)

Included tests: all the ones required, ask for quotation including detailed list

## Qualification of hydrogen nozzles under SAE J2600

(

Price #1 [euro]: 52,000.00

Limit #1: ---

Included tests: all the ones required, ask for quotation including detailed list

# Qualification of in line flame arrester through deflagration, according to ISO 16852:2016

(id 160) ---

Price #1 [euro]: 4,000.00 Limit #1: up to 2 inches, standard ambient temperature (indoor location)

Price #2 [euro]: 5,500.00 Limit #2: from 3 to 9 inches, standard ambient temperature (indoor location)

Price #3 [euro]: 7,500.00 Limit #3: from 10 inches, standard ambient temperature (outdoor location, up to 500kg)

Included tests: all the ones required, ask for quotation including detailed list

## Qualification of insulating materials used to protect hot surfaces

(id 98) This internal test sequence is intended to qualify the protection systems of hot surfaces realized using insulating materials, instead of cooling systems.

Price #1 [euro]: 12,500.00 Limit #1: ---

Included tests: all the ones required, ask for quotation including detailed list

# Qualification of metal fuel tank under EPA 40 CFR 1051.515, Short Test without DF Determination

(id 164) ---

Price #1 [euro]: 8,800.00

Limit #1: ---

Limit #1: ---

Included tests: all the ones required, ask for quotation including detailed list

### Qualification of optical isolators

(id 128) Qualification of optical isolators under IEC 60079-11

Price #1 [euro]: 4,800.00

Included tests: all the ones required, ask for quotation including detailed list

# Qualification of plastic fuel tank under Commission Delegated Regulation (EU) No 44/2014, Annex I

(id 165) This test sequence is intended to qualify plastic fuel tanks for the EU market. All the tests contained in this sequence are under Commission Delegated Regulation (EU) No 44/2014, Annex IX. Overturn Test and Hydraulic Test are excluded from this package. Permeability Test is performed according to Commission Delegated Regulation (EU) No 134/2014, Annex V, Appendix I.

Price #1 [euro]: 9,600.00 Limit #1: ---

Included tests: all the ones required, ask for quotation including detailed list

# Qualification of R9 hoses according to SAE J30 and EPA 40 CFR 1051.501

(id 169) This test sequence is intended to qualify R9 hoses according to SAE J30 combined with the prescriptions given by EPA 40 CFR 1051.501. The purpose of the test sequence is limited to the verifications of environmental parameters as requested by the United States Environmental Protection Agency.

Price #1 [euro]: 5,000.00 Limit #1: --

Included tests: all the ones required, ask for quotation including detailed list

### Qualification of Rectangular connectors according with EN 175301-803:2006

(id 133) Flat contacts, 0, 8 mm

thickness, locking screw not detachable

Unit price #1: ask to office Limit #1: ---

Included tests: all the ones required, ask for quotation including detailed list

page 233/234



# albarubens

QM ref: ALL 7.1.4

Records status checked on 30 Nov 2025 (22:05) Restricted: NO Copy uncontrolled when printed: check the updated version

INDEX



# Qualification of valves under ANSI Z21.15:2021

(id 158) This test sequence is intended to qualify performances and reliability of valves with a particular focus on leakages. The sequence is suitable both for the small size (up to 1") and the big size (from 1" to 3"). The reference standard is ANSI Z21.15:2021.

Price #1 [euro]: 3,700.00 Limit #1: Size up to 1"
Price #2 [euro]: 5,200.00 Limit #2: Size from 1" to 3"

Included tests: all the ones required, ask for quotation including detailed list

# RTA compliance for LED street lightning fittings

(id 108) ---

Price #1 [euro]: 1,000.00 Limit #1: ---

Included tests: all the ones required, ask for quotation including detailed list

# Safety requirements for engines for use in flammable gas atmospheres (Group II)

(id 78) Test on engine simulator with 1 flame arrester, fully supplied by the manufacturer

Price #1 [euro]: 7,000.00 Limit #1: No turbo, <18 liters
Price #2 [euro]: 10,000.00 Limit #2: Turbo, <18 liters
Price #3 [euro]: 15,000.00 Limit #3: Turbo, >18 liters

Included tests: all the ones required, ask for quotation including detailed list

# Static detonation flame arrester without pre-volume, BC=c (no burn time), tested according to ISO 1

(id 147) ---

Price #1 [euro]: 5,000.00 Limit #1: ---

Included tests: all the ones required, ask for quotation including detailed list

# Static Ex p enclosure

(id 81) Test sequence for static Ex p enclosure

Price #1 [euro]: 1,800.00 Limit #1: < 500 [dm3]

Included tests: all the ones required, ask for quotation including detailed list

# Test sequence to qualify refrigerating components under IEC 60335-2-24, according to IEC 60079-1

(id 152) This test sequence is designed to qualify refrigerating appliances that uses R290. Being this an inflammable gas, the reference standard is IEC 60079-15:2017 under ATEX scheme. The test sequence follows the non-incendive method from chapter 7. All the components under test, in their normal usage, are placed inside the machine, thus outdoor exposure tests are to be considered excluded.

NOTE: Thermal endurance mentioned in table 1 of IEC 60079-15 is considered applicable. This pre-conditoning test refers to chapter 7.2 of IEC 60079-0 that, in turn, refers to chapters 26.8 and 26.9 of the same standard. This is considered to be applicable because electrical contacts can oxidize due to heat and humidity and this can cause ignition.

Price #1 [euro]: 3,200.00 Limit #1: ---

Included tests: all the ones required, ask for quotation including detailed list

# Type testing of valves for fugitive emissions (isolating valves, Class BH, CO1, 205 cycles)

(id 122) This sequence of tests has the aim to verifying the fugitive emission class of valves according to ISO 15848-1:2015. It is made to qualify tested valves in endurance class "CO1". The minimum required number of mechanical cycles for shut-off valves must be 205 cycles (full stroke) with two thermal cycles (for a total of 50 cycles at RT, 50 cycles at test temperature, 50 cycles at RT, 50 cycles at test temperature and 5 cycles at RT).

Price #1 [euro]: 3,200.00 Limit #1: Up to 3"
Price #2 [euro]: 5,000.00 Limit #2: from 4" to 10"
Price #3 [euro]: 8,500.00 Limit #3: from 12" to 20"

Included tests: all the ones required, ask for quotation including detailed list

# Verifications of fuel tank volumes and capacity under paragraph 7.12 of PF.90258 for metallic tanks

(id 157) ---

Price #1 [euro]: 1,200.00 Limit #1: ---

Included tests: all the ones required, ask for quotation including detailed list

End of document - Certificate on the cover

Fine del documento - Certificato sulla copertina

page 234/234